

STN	Kvapalné ropné výrobky. Stanovenie obsahu sodíka, draslíka, vápnika, fosforu, medi a zinku v motorovej naftě. Metóda optickej emisnej spektrometrie s indukčne viazanou plazmou (ICP-OES).	STN EN 16476 65 6528
------------	---	--

Liquid petroleum products - Determination of Sodium, Potassium, Calcium, Phosphorus, Copper and Zinc contents in diesel fuel - Method via Inductively Coupled Plasma Optical Emission Spectrometry (ICP OES)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 10/14

Obsahuje: EN 16476:2014

119676

Úrad pre normalizáciu, metrológiu a skúšobníctvo SR, odbor SÚTN, 2014
Podľa zákona č. 264/1999 Z. z. v znení neskorších predpisov sa môžu slovenské technické normy rozmnožovať a rozširovať iba so súhlasom Úradu pre normalizáciu, metrológiu a skúšobníctvo SR.

EUROPEAN STANDARD

EN 16476

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2014

ICS 75.160.20

English Version

Liquid petroleum products - Determination of Sodium,
Potassium, Calcium, Phosphorus, Copper and Zinc contents in
diesel fuel - Method via Inductively Coupled Plasma Optical
Emission Spectrometry (ICP OES)

Produits pétroliers liquides - Détermination des
concentrations en sodium, potassium, calcium, phosphore,
cuivre et zinc dans le gazole - Méthode par spectrométrie
d'émission atomique par plasma à couplage inductif (ICP
OES)

Flüssige Mineralölzeugnisse - Bestimmung des Gehalts
an Natrium, Kalium, Calcium, Phosphor, Kupfer, Zink in
Dieselkraftstoffen - Direkte Bestimmung durch optische
Emissionsspektrometrie mit induktiv gekoppeltem Plasma
(ICP OES)

This European Standard was approved by CEN on 20 March 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
Foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Principle.....	5
4 Reagents.....	5
5 Apparatus	6
6 Sampling.....	8
7 Preparation of calibration solutions	8
7.1 General.....	8
7.2 Internal Standard solution	8
7.3 Calibration solutions	8
7.4 Calibration check solution.....	9
8 Calibration	9
8.1 General.....	9
8.2 Calibration of the ICP OES spectrometer.....	9
8.3 Procedure A.....	10
8.4 Procedure B.....	10
8.5 Check of calibration	10
9 Sample analysis.....	11
9.1 Sample preparation	11
9.2 Sample measurement.....	11
10 Calculation.....	12
11 Expression of result	12
12 Precision.....	12
12.1 General.....	12
12.2 Repeatability.....	12
12.3 Reproducibility.....	12
13 Test report	13
Bibliography.....	14

Foreword

This document (EN 16476:2014) has been prepared by Technical Committee CEN/TC 19 “Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin”, the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2014 and conflicting national standards shall be withdrawn at the latest by November 2014.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

This document specifies a method for determining concentration of the six most essential ash forming elements. The number of six was chosen in order to limit the complexity of the test and to be able to determine a decent method precision. The six elements were specified by the vehicle manufacturers. The test provides additional information to the more common ash concentration determination methods which do aim towards filter problems. Metals have more impact on the durability of modern DPF exhaust gas after-treatment systems in diesel vehicles than ash in general.

1 Scope

This European Standard specifies an inductively coupled plasma optical emission spectrometry (ICP OES) method for the determination of sodium, potassium, calcium, phosphorus, copper and zinc concentrations of diesel fuels, including those containing up to 30 % (V/V) fatty acid methyl ester (FAME), in the range detailed in Table 1. These six elements are considered as the most essential ash forming elements.

Table 1 — Application ranges for ash forming elements

Element	Range mg/kg
Sodium	1,2 to 2,5
Potassium	0,9 to 2,5
Calcium	0,3 to 2,5
Phosphorus	0,9 to 2,5
Copper	0,2 to 2,5
Zinc	0,2 to 2,5

NOTE For the purposes of this European Standard, the term “% (V/V)” is used to represent the volume fraction, φ .

WARNING — The use of this European Standard may involve hazardous materials, operations and equipment. This European Standard does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this European Standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 1042, *Laboratory glassware - One-mark volumetric flasks (ISO 1042)*

EN ISO 3170, *Petroleum liquids - Manual sampling (ISO 3170)*

EN ISO 3171, *Petroleum liquids - Automatic pipeline sampling (ISO 3171)*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN