

<b>STN</b>	<b>Zdravotnícke elektrické prístroje. Časť 2-34: Osobitné požiadavky na základnú bezpečnosť a nevyhnutné prevádzkové vlastnosti invazívnych monitorov krvného tlaku.</b>	<b>STN EN 60601-2-34</b>  36 4800
------------	--	---

Táto norma obsahuje anglickú verziu európskej normy.  
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 11/14

Obsahuje: EN 60601-2-34:2014, IEC 60601-2-34:2011

Oznámením tejto normy sa od 27.06.2017 ruší  
STN EN 60601-2-34 (36 4800) z júla 2002

**119717**

---

Úrad pre normalizáciu, metrológiu a skúšobníctvo SR, odbor SÚTN, 2014  
Podľa zákona č. 264/1999 Z. z. v znení neskorších predpisov sa môžu slovenské technické normy  
rozmnžovať a rozširovať iba so súhlasom Úradu pre normalizáciu, metrológiu a skúšobníctvo SR.

English Version

**Medical electrical equipment - Part 2-34: Particular requirements  
for the basic safety and essential performance of invasive blood  
pressure monitoring equipment  
(IEC 60601-2-34:2011)**

Appareils électromédicaux - Partie 2-34: Exigences  
particulières pour la sécurité de base et les performances  
essentiels des appareils de surveillance de la pression  
sanguine prélevée directement  
(CEI 60601-2-34:2011)

Medizinische elektrische Geräte - Teil 2-34: Besondere  
Festlegungen für die Sicherheit einschließlich der  
wesentlichen Leistungsmerkmale von invasiven Blutdruck-  
Überwachungsgeräten  
(IEC 60601-2-34:2011)

This European Standard was approved by CENELEC on 2011-05-16. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

## Foreword

The text of document 62D/906/FDIS, future edition 3 of IEC 60601-2-34, prepared by SC 62D, "Electromedical equipment", of IEC TC 62, "Electrical equipment in medical practice", was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60601-2-34:2014.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2014-12-27
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2017-06-27

This document supersedes EN 60601-2-34:2000.

EN 60601-2-34:2014 was revised to align structurally with EN 60601-1:2006.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this standard;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;

- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive(s) see informative Annex ZZ, which is an integral part of this document.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

### **Endorsement notice**

The text of the International Standard IEC 60601-2-34:2011 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following note has to be added for the standard indicated:

IEC 80601-2-30      NOTE Harmonized as EN 80601-2-30.

## Annex ZA (normative)

### Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

*Annex ZA of the general standard applies, except as follows:*

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<i>Replacement:</i>				
IEC 60601-1-2 (mod)	2007	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests	EN 60601-1-2 + corr. March	2007 2010
IEC 60601-1-8	2006	Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance - Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	EN 60601-1-8 + corr. March	2007 2010
<i>Addition:</i>				
IEC 60601-2-2	2009	Medical electrical equipment - Part 2-2: Particular requirements for basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories	EN 60601-2-2 + A11	2009 2011
IEC 60601-2-27	-	Medical electrical equipment - Part 2-27: Particular requirements for the basic safety and essential performance of electrocardiographic monitoring equipment	EN 60601-2-27 <sup>1)</sup>	-
IEC 60601-2-49	-	Medical electrical equipment - Part 2-49: Particular requirements for the basic safety and essential performance of multifunction patient monitoring equipment	EN 60601-2-49 <sup>1)</sup>	-
ISO 15223-1	2007	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements	EN ISO 15223-1	2012

---

<sup>1)</sup> At draft stage.

## **Annex ZZ** (informative)

### **Coverage of Essential Requirements of EC Directives**

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and within its scope the standard covers all relevant essential requirements as given in Annex I of the EC Directive 93/42/EEC except as follows:

- Essential Requirement 6a
- Essential Requirement 7.4
- Essential Requirement 7.5 paragraph 2 & 3
- Essential Requirement 13.6 (q)

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

**WARNING** - Other requirements and other EC Directives may be applicable to the products falling within the scope of this standard.



# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

---

**Medical electrical equipment –  
Part 2-34: Particular requirements for the basic safety and essential performance  
of invasive blood pressure monitoring equipment**

**Appareils électromédicaux –  
Partie 2-34: Exigences particulières pour la sécurité de base et les performances  
essentielle des appareils de surveillance de la pression sanguine prélevée  
directement**





## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2011 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland  
Email: [inmail@iec.ch](mailto:inmail@iec.ch)  
Web: [www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

- Catalogue of IEC publications: [www.iec.ch/searchpub](http://www.iec.ch/searchpub)

The IEC on-line Catalogue enables you to search by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, withdrawn and replaced publications.

- IEC Just Published: [www.iec.ch/online\\_news/justpub](http://www.iec.ch/online_news/justpub)

Stay up to date on all new IEC publications. Just Published details twice a month all new publications released. Available on-line and also by email.

- Electropedia: [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary of electronic and electrical terms containing more than 20 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary online.

- Customer Service Centre: [www.iec.ch/webstore/custserv](http://www.iec.ch/webstore/custserv)

If you wish to give us your feedback on this publication or need further assistance, please visit the Customer Service Centre FAQ or contact us:

Email: [csc@iec.ch](mailto:csc@iec.ch)  
Tel.: +41 22 919 02 11  
Fax: +41 22 919 03 00

### A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

- Catalogue des publications de la CEI: [www.iec.ch/searchpub/cur\\_fut-f.htm](http://www.iec.ch/searchpub/cur_fut-f.htm)

Le Catalogue en-ligne de la CEI vous permet d'effectuer des recherches en utilisant différents critères (numéro de référence, texte, comité d'études,...). Il donne aussi des informations sur les projets et les publications retirées ou remplacées.

- Just Published CEI: [www.iec.ch/online\\_news/justpub](http://www.iec.ch/online_news/justpub)

Restez informé sur les nouvelles publications de la CEI. Just Published détaille deux fois par mois les nouvelles publications parues. Disponible en-ligne et aussi par email.

- Electropedia: [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International en ligne.

- Service Clients: [www.iec.ch/webstore/custserv/custserv\\_entry-f.htm](http://www.iec.ch/webstore/custserv/custserv_entry-f.htm)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions, visitez le FAQ du Service clients ou contactez-nous:

Email: [csc@iec.ch](mailto:csc@iec.ch)  
Tél.: +41 22 919 02 11  
Fax: +41 22 919 03 00





IEC 60601-2-34

Edition 3.0 2011-05

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

---

**Medical electrical equipment –  
Part 2-34: Particular requirements for the basic safety and essential performance  
of invasive blood pressure monitoring equipment**

**Appareils électromédicaux –  
Partie 2-34: Exigences particulières pour la sécurité de base et les performances  
essentielle des appareils de surveillance de la pression sanguine prélevée  
directement**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

PRICE CODE **XB**  
CODE PRIX

---

ICS 11.040.55

ISBN 978-2-88912-495-4

## CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
201.1 Scope, object and related standards .....	7
201.2 Normative references .....	8
201.3 Terms and definitions.....	9
201.4 General requirements.....	10
201.5 General requirements for testing of ME EQUIPMENT.....	10
201.6 Classification of ME EQUIPMENT and ME SYSTEMS .....	11
201.7 ME EQUIPMENT identification, marking and documents.....	11
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	13
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	20
201.10 Protection against unwanted and excessive radiation HAZARDS.....	21
201.11 Protection against excessive temperatures and other HAZARDS.....	21
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	23
201.13 HAZARDOUS SITUATIONS and fault conditions.....	29
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS) .....	29
201.15 Construction of ME EQUIPMENT .....	29
201.16 ME SYSTEMS .....	30
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS .....	30
202 Electromagnetic compatibility – Requirements and tests .....	30
208 General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems.....	36
Annexes .....	45
Annex AA (informative) Particular guidance and rationale .....	46
Annex BB (informative) Alarm diagrams 208/IEC 60601-1-8:2006.....	59
Bibliography.....	61
Index of defined terms used in this particular standard.....	62
Figure 201.101 – Dynamic test for limitation of energy from different parts – Recovery test .....	15
Figure 201.102 – Diaphragm leak test.....	16
Figure 201.103 – Measuring circuit for PATIENT LEAKAGE CURRENT via the PATIENT CONNECTION(S) of an F-TYPE APPLIED PART to earth caused by an external voltage on the PATIENT CONNECTION(S) .....	17
Figure 201.104 – Measuring circuit for the PATIENT LEAKAGE CURRENT from the APPLIED PART to earth of CLASS I ME EQUIPMENT caused by an external voltage on a metal ACCESSIBLE PART that is not PROTECTIVELY EARTHED.....	18
Figure 201.105 – Measuring circuit for the PATIENT LEAKAGE CURRENT from the APPLIED PART to earth of INTERNALLY POWERED ME EQUIPMENT caused by an external voltage on a metal ACCESSIBLE PART that is not PROTECTIVELY EARTHED.....	19
Figure 201.106– Clarification of leakage current tests.....	20
Figure 201.107 – Over-pressure test.....	21
Figure 201.108 – Test for accuracy of pressure measurements .....	25

Figure 201.109 – Test sequence for sensitivity, repeatability, non-linearity, drift and hysteresis .....	26
Figure 201.110 – Pressure measurement system for accuracy of systolic and diastolic pressure .....	27
Figure 201.111 – Frequency response of ME EQUIPMENT and TRANSDUCER.....	28
Figure 202.101 – Test layout for conducted and radiated emission and radiated immunity test .....	31
Figure 202.102 – Test circuit for high-frequency surgery interference measurement, when the isolation of the APPLIED PART is in the monitor .....	34
Figure 202.103 – Test circuit for high-frequency surgery protection, when the isolation of the APPLIED PART is in the TRANSDUCER.....	35
Figure 202.104 – Test set-up for high-frequency surgery protection.....	36
Figure 208.101 – Test for delay times of ALARM SIGNALS indicating PHYSIOLOGICAL ALARM CONDITIONS .....	41
Figure 208.102 – Test for delay times of ALARM SIGNALS indicating PHYSIOLOGICAL ALARM CONDITIONS .....	42
Figure AA.1 – Pressure TRANSDUCER error band .....	53
Figure BB.101 – NON-LATCHING ALARM SIGNALS without ALARM RESET .....	59
Figure BB.102 – NON-LATCHING ALARM SIGNALS with ALARM RESET .....	59
Figure BB.103 – LATCHING ALARM SIGNALS with ALARM RESET .....	60
Figure BB.104 – Two ALARM CONDITIONS with ALARM RESET.....	60
Table 201.101 – ESSENTIAL PERFORMANCE requirements .....	10
Table 208.101 – ALARM CONDITION priorities .....	37
Table 208.102 – Characteristics of the BURST of auditory ALARM SIGNALS.....	38

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEDICAL ELECTRICAL EQUIPMENT –****Part 2-34: Particular requirements for the basic safety and essential performance of invasive blood pressure monitoring equipment**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-34 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition of IEC 60601-2-34 published in 2001 and constitutes a technical revision. This edition was revised to align structurally with the 2005 edition of IEC 60601-1.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/906/FDIS	62D/923/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

## INTRODUCTION

This particular standard concerns the BASIC SAFETY and ESSENTIAL PERFORMANCE of INVASIVE BLOOD PRESSURE MONITORING EQUIPMENT. It amends and supplements IEC 60601-1 (third edition, 2005): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereinafter referred to as the general standard.

The aim of this third edition is to bring this particular standard up to date with reference to the third edition of the general standard through reformatting and technical changes.

The requirements of this particular standard take priority over those of the general standard.

A “General guidance and rationale” for the more important requirements of this particular standard is included in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, Annex AA does not form part of the requirements of this standard.

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-34: Particular requirements for the basic safety and essential performance of invasive blood pressure monitoring equipment

#### 201.1 Scope, object and related standards

Clause 1 of the general standard<sup>1</sup> applies, except as follows:

##### 201.1.1 \*Scope

*Replacement:*

This particular standard applies to BASIC SAFETY and ESSENTIAL PERFORMANCE of INVASIVE BLOOD PRESSURE MONITORING EQUIPMENT as defined in 201.3.63, hereinafter referred to as ME EQUIPMENT.

This particular standard does not apply to catheter tubing, catheter needles, Luer locks, taps and tap tables that connect to the DOME.

This particular standard does not apply to NON-INVASIVE BLOOD PRESSURE MONITORING EQUIPMENT.

##### 201.1.2 Object

*Replacement:*

The object of this particular standard is to establish BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for INVASIVE BLOOD PRESSURE MONITORING EQUIPMENT, as defined in 201.3.63.

##### 201.1.3 Collateral standards

*Addition:*

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2:2007 and IEC 60601-1-8:2006 apply as modified in Clauses 202 and 208 respectively. IEC 60601-1-3 does not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

##### 201.1.4 Particular standards

*Replacement:*

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

---

<sup>1</sup> The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

## **201.2 Normative references**

Clause 2 of the general standard applies, except as follows:

### *Replacement:*

IEC 60601-1-2:2007, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests*

IEC 60601-1-8:2006, *Medical electrical equipment – Part 1-8: General requirements for basic safety and essential performance – Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*

### *Addition:*



IEC 60601-2-2:2009, *Medical electrical equipment – Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories*

IEC 60601-2-27, *Medical electrical equipment – Part 2-27: Particular requirements for the basic safety and essential performance of electrocardiographic monitoring equipment*<sup>2</sup>

IEC 60601-2-49, *Medical electrical equipment – Part 2-49: Particular requirements for the basic safety and essential performance of multifunction patient monitoring equipment*<sup>3</sup>

ISO 15223-1:2007 *Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements*

NOTE Informative references are listed in the bibliography beginning on page 61.

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**