

STN	Kardiovaskulárne implantáty a umelé orgány. Výmenníky pre krv-plyn (oxygenátory) (ISO 7199: 2009).	STN EN ISO 7199 85 6250
------------	---	---

Cardiovascular implants and artificial organs - Blood-gas exchangers (oxygenators) (ISO 7199:2009 + Amd 1:2012)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 01/15

Obsahuje: EN ISO 7199:2014, ISO 7199:2009, ISO 7199:2009/Amd 1:2012

Oznámením tejto normy sa ruší
STN EN 12022 (85 6250) z januára 2001

119974

English Version

Cardiovascular implants and artificial organs - Blood-gas exchangers (oxygenators) (ISO 7199:2009 + Amd 1:2012)

Implants cardiovasculaires et organes artificiels -
Échangeurs gaz/sang extracorporels (oxygénateurs) (ISO
7199:2009 + Amd 1:2012)

Kardiovaskuläre Implantate und künstliche Organe -
Blutgasaustauscher (Oxygenatoren) (ISO 7199:2009 + Amd
1:2012)

This European Standard was approved by CEN on 17 July 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....**3**

Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices**4**

Foreword

The text of ISO 7199:2009 + Amd 1:2012 has been prepared by Technical Committee ISO/TC 150 "Implants for surgery" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 7199:2014 by Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2015, and conflicting national standards shall be withdrawn at the latest by February 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 12022:1999.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 7199:2009 + Amd 1:2012 has been approved by CEN as EN ISO 7199:2014 without any modification.

Annex ZA (informative)

Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to Essential Requirements of the New Approach Directive 93/42/EEC on medical devices.

Once this standard is cited in the Official Journal of the European Union under that Directive and has been implemented as a national standard in at least one Member State, compliance with the clauses of this standard given in table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

Table ZA.1 — Correspondence between this European Standard and Directive 93/42/EEC on medical devices

Clause(s)/sub-clause(s) of this EN	Essential Requirements (ERs) of Directive 93/42/EEC	Qualifying remarks/Notes
4.1.1; 4.1.2; 5.2; 6.2.1; 6.2.2; 6.3	7.2	Some redundancies, but all these EN sections address performance and sterility issues.
4.3.3; 5.2; 5.4.3	7.3	
4.2.1; 4.2.2; 4.2.4	7.5	
4.1.1; 6.2.1; 6.2.2; 6.3	8.1	
6.2.1; 6.2.2; 6.3	8.3	
6.2.1; 6.2.2; 6.3	8.4	
4.1.1; 5.2.1; 6.2.1; 6.2.2; 6.2.3	8.5	
6.2.1	8.6	
6.1; 6.2.1	8.7	
4.2.4	9.1	
6.1; 6.2; 6.3	13	

WARNING — Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.

**Cardiovascular implants and artificial
organs — Blood-gas exchangers
(oxygenators)**

*Implants cardiovasculaires et organes artificiels — Échangeurs
gaz/sang extracorporels (oxygénateurs)*



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2009

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Requirements	3
4.1 Biological characteristics	3
4.2 Physical characteristics	3
4.3 Performance characteristics	3
5 Tests and measurements to determine compliance with this International Standard	4
5.1 General	4
5.2 Biological characteristics	4
5.3 Physical characteristics	5
5.4 Performance characteristics	5
6 Information supplied by the manufacturer	8
6.1 Information to be given on the oxygenator	8
6.2 Information to be given on the packaging	8
6.3 Information to be given in the accompanying documents	9
6.4 Information to be given in the accompanying documents in a prominent form	10
7 Packaging	10
Bibliography	11

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 7199 was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*.

This second edition cancels and replaces the first edition (ISO 7199:1996), which has been technically revised.

Introduction

This International Standard is intended to ensure that devices designed to affect the exchange of gases in support of, or as a substitution for, the normal respiratory function of the lungs have been adequately tested for both their safety and function, and that extracorporeal device characteristics are appropriately disclosed when labeling the device.

This International Standard therefore contains procedures to be used for evaluation of extracorporeal blood-gas exchangers (oxygenators). Type test procedures for determination of the gas transfer, blood cell damage and heat exchanger performance are described, although limits for these characteristics are not specified. Ready identification of the performance characteristics should, however, assist the user in the selection of an oxygenator that will suit the needs of the patient.

This International Standard also includes minimum reporting requirements, which will allow the user to compare performance characteristics of oxygenators of different designs in a standard way.

This International Standard makes reference to other International Standards in which methods for determination of characteristics common to medical devices can be found.

No provisions have been made for quantification of microbubble generation or for non-formed elements of bovine blood because there currently is no consensus regarding satisfactorily reproducible test methods.

Requirements for animal and clinical studies have not been included in this International Standard. Such studies may be parts of a manufacturer's quality system.

This International Standard contains only those requirements that are specific to oxygenators. Non-specific requirements are covered by references to other International Standards listed in the normative references section. Since non-toxicity is anticipated to be the subject of a future horizontal/level 1 standard, this International Standard does not cover non-toxicity.

Cardiovascular implants and artificial organs — Blood-gas exchangers (oxygenators)

1 Scope

This International Standard specifies requirements for sterile, single-use, extracorporeal blood-gas exchangers (oxygenators) intended for supply of oxygen to, and removal of carbon dioxide from, the blood of humans.

This International Standard also applies to heat exchangers that are integral parts of oxygenators and to external equipment unique to the use of the device.

This International Standard does not apply to:

- implanted oxygenators;
- liquid oxygenators;
- extracorporeal circuits (blood tubing);
- separate heat exchangers;
- separate ancillary devices.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing*

ISO 10993-7, *Biological evaluation of medical devices — Part 7: Ethylene oxide sterilization residuals*

ISO 10993-11, *Biological evaluation of medical devices — Part 11: Tests for systemic toxicity*

ISO 11135:1994, *Medical devices — Validation and routine control of ethylene oxide sterilization*

ISO 11137-1, *Sterilization of health care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*

ISO 11607-2, *Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes*

ISO 17665-1, *Sterilization of health care products — Moist heat — Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN