

|            |   |  |
|------------|---|--|
| <b>STN</b> | <b>Očné implantáty. Vnútročné šošovky. Časť 2:<br/>Optické vlastnosti a skúšobné metódy (ISO<br/>11979-2:2014).</b> | <b>STN<br/>EN ISO 11979-2</b><br><br>19 5300 |
|------------|---|--|

Ophthalmic implants - Intraocular lenses - Part 2: Optical properties and test methods (ISO 11979-2:2014)

Táto norma obsahuje anglickú verziu európskej normy.  
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 01/15

Obsahuje: EN ISO 11979-2:2014, ISO 11979-2:2014

Oznámením tejto normy sa ruší  
STN EN ISO 11979-2 (19 5300) z júna 2001

## 119999

---

Úrad pre normalizáciu, metrológiu a skúšobníctvo SR, odbor SÚTN, 2015  
Podľa zákona č. 264/1999 Z. z. v znení neskorších predpisov sa môžu slovenské technické normy  
rozmnožovať a rozširovať iba so súhlasom Úradu pre normalizáciu, metrológiu a skúšobníctvo SR.

English Version

## Ophthalmic implants - Intraocular lenses - Part 2: Optical properties and test methods (ISO 11979-2:2014)

Implants ophtalmiques - Lentilles intraoculaires - Partie 2:  
Propriétés optiques et méthodes d'essai (ISO 11979-  
2:2014)

Ophthalmische Implantate - Intraokularlinsen - Teil 2:  
Optische Eigenschaften und Prüfverfahren (ISO 11979-  
2:2014)

This European Standard was approved by CEN on 25 July 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

**Contents**

Page

**Foreword.....3**

## **Foreword**

This document (EN ISO 11979-2:2014) has been prepared by Technical Committee ISO/TC 172 "Optics and photonics" in collaboration with Technical Committee CEN/TC 170 "Ophthalmic optics" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2015, and conflicting national standards shall be withdrawn at the latest by February 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11979-2:1999.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

### **Endorsement notice**

The text of ISO 11979-2:2014 has been approved by CEN as EN ISO 11979-2:2014 without any modification.

---

---

**Ophthalmic implants — Intraocular  
lenses —**

Part 2:  
**Optical properties and test methods**

*Implants ophtalmiques — Lentilles intraoculaires —  
Partie 2: Propriétés optiques et méthodes d'essai*





**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2014

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

|   |           |
|---|-----------|
| <b>Foreword</b> .....   | <b>iv</b> |
| <b>Introduction</b> .....   | <b>v</b>  |
| <b>1 Scope</b> .....  | <b>1</b>  |
| <b>2 Normative references</b> .....                                   | <b>1</b>  |
| <b>3 Terms and definitions</b> .....                                  | <b>1</b>  |
| <b>4 Requirements</b> .....   | <b>1</b>  |
| 4.1 General.....  | 1         |
| 4.2 Dioptric power.....   | 2         |
| 4.3 Determination of imaging quality.....                             | 3         |
| 4.4 Spectral transmittance.....                                       | 5         |
| <b>Annex A (normative) Measurement of dioptric power</b> .....        | <b>6</b>  |
| <b>Annex B (normative) Measurement of resolution efficiency</b> ..... | <b>14</b> |
| <b>Annex C (normative) Measurement of MTF</b> .....                   | <b>17</b> |
| <b>Bibliography</b> .....   | <b>22</b> |

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 11979-2 was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

This second edition cancels and replaces the first edition (ISO 11979-2:1999), which has been technically revised. It also incorporates the Technical Corrigendum ISO 11979-2:1999/Cor.1:2003.

ISO 11979 consists of the following parts, under the general title *Ophthalmic implants — Intraocular lenses*:

- *Part 1: Vocabulary*
- *Part 2: Optical properties and test methods*
- *Part 3: Mechanical properties and test methods*
- *Part 4: Labelling and information*
- *Part 5: Biocompatibility*
- *Part 6: Shelf-life and transport stability testing*
- *Part 7: Clinical investigations*
- *Part 8: Fundamental requirements*
- *Part 9: Multifocal intraocular lenses*
- *Part 10: Phakic intraocular lenses*



## Introduction

This part of ISO 11979 initially addressed monofocal IOLs and now has been revised to include the requirements and test methods for spherical monofocal, aspheric monofocal, toric, multifocal, and accommodative IOLs. This part of ISO 11979 contains several test methods for which associated requirements are given and one test method for which no requirement is formulated. The former are directly connected to the optical functions of intraocular lenses. The latter, the test for spectral transmittance, has been provided for information about UV transmission and in specific situations, e.g. when using laser light sources for diagnosis and treatment.

For the original spherical monofocal IOLs, extensive interlaboratory testing was carried out before setting the limits specified. During this testing some basic problems were encountered as described in Reference [1]. The accuracy in the determination of dioptric power has an error that is not negligible in relation to the half dioptre steps in which intraocular lenses are commonly labelled. The dioptric power tolerances take this fact into account. Hence the limits set may lead to some overlap into the next labelled power, especially for high dioptre lenses. Reference [1] gives further discussion on this subject.

The majority of lenses hitherto implanted were qualified using the method described in [Annex B](#) or [Annex C](#) (model eye 1). The method in [Annex B](#) is limited in its applicability, however. The limits for the more general method in [Annex C](#) have been set in terms of MTF in a model eye, following two approaches. The first is by correlation to the method and limit in [Annex B](#). Further discussion can be found in Reference [2]. The second is set as a percentage of what is calculated as theoretical maximum for the design, with the rationale that a minimum level of manufacturing accuracy be guaranteed. For common PMMA lenses, these two limits correspond well with each other. For lenses made of materials with lower refractive index, or with certain shape factors, or for extreme power lenses in general, the latter limit is lower than the former. However, such lenses are already in use, indicating clinical acceptance. The question of which is the absolute lowest limit that is compatible with good vision arises. No definite answer can be found, but following clinical data presented to the working group, an absolute lower limit has been set for the calculation method.

# Ophthalmic implants — Intraocular lenses —

## Part 2: Optical properties and test methods

### 1 Scope

This part of ISO 11979 specifies requirements and test methods for certain optical properties of intraocular lenses (IOLs) with any of spherical, aspheric, monofocal, toric, multifocal, and/or accommodative optics. The generic descriptor 'IOL' used throughout this document also includes phakic intraocular lenses (PIOL).

### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 6328:2000, *Photography — Photographic materials — Determination of ISO resolving power*

ISO 9334, *Optics and photonics — Optical transfer function — Definitions and mathematical relationships*

ISO 9335, *Optics and photonics — Optical transfer function — Principles and procedures of measurement*

ISO 11979-1, *Ophthalmic implants — Intraocular lenses — Part 1: Vocabulary*

ISO 11979-3, *Ophthalmic implants — Intraocular lenses — Part 3: Mechanical properties and test methods*

ISO 11979-4, *Ophthalmic implants — Intraocular lenses — Part 4: Labelling and information*

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**