

STN	Zdravotnícke elektrické prístroje. Časť 1-2: Všeobecné požiadavky na základnú bezpečnosť a nevyhnutné prevádzkové vlastnosti. Pridružená norma: Elektromagnetické rušenia. Požiadavky a skúšky.	STN EN 60601-1-2
		36 4800

Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 02/16

Obsahuje: EN 60601-1-2:2015, IEC 60601-1-2:2014

Oznámením tejto normy sa od 31.12.2018 ruší
STN EN 60601-1-2 (36 4800) z apríla 2008

122365

EUROPEAN STANDARD

EN 60601-1-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 2015

ICS 11.040.01; 33.100.10; 33.100.20

Supersedes EN 60601-1-2:2007

English Version

**Medical electrical equipment - Part 1-2: General requirements for
basic safety and essential performance - Collateral Standard:
Electromagnetic disturbances - Requirements and tests
(IEC 60601-1-2:2014)**

Appareils électromédicaux - Partie 1-2: Exigences
générales pour la sécurité de base et les performances
essentielles - Norme collatérale: Perturbations
électromagnétiques - Exigences et essais
(IEC 60601-1-2:2014)

Medizinische elektrische Geräte - Teil 1-2: Allgemeine
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale - Ergänzungsnorm:
Elektromagnetische Störgrößen - Anforderungen und
Prüfungen
(IEC 60601-1-2:2014)

This European Standard was approved by CENELEC on 2014-04-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

European foreword

The text of document 62A/916/FDIS, future edition 4 of IEC 60601-1-2, prepared by SC 62A, "Common aspects of electrical equipment used in medical practice", of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60601-1-2:2015.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2016-03-18
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2018-12-31

This document supersedes EN 60601-1-2:2007.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 60601-1-2:2014 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60601-1-2:2007	NOTE	Harmonized as EN 60601-1-2:2007 (not modified)
IEC 60601-2-27:2011	NOTE	Harmonized as EN 60601-2-27:2006 (not modified)
IEC 60601-2-44:2009	NOTE	Harmonized as EN 60601-2-44:2009 (not modified)
IEC 61000-3-11:2000	NOTE	Harmonized as EN 61000-3-11:2000 (not modified)
IEC 61000-3-12:2011	NOTE	Harmonized as EN 61000-3-12:2011 (not modified)
IEC 61000-3-12:2011	NOTE	Harmonized as EN 61000-3-12:2011 (not modified)
IEC 60601-6-1:2005	NOTE	Harmonized as EN 60601-6-1:2007 (not modified)
IEC 60601-6-2:2005	NOTE	Harmonized as EN 60601-6-2:2005 (not modified)
IEC 61496-1:2008	NOTE	Harmonized as EN 61496-1:2008 (not modified)
CISPR 16-1-1:2010	NOTE	Harmonized as EN 55016-1-1:2010 (not modified)
CISPR 16-2-3:2010	NOTE	Harmonized as EN 55016-2-3:2010 (not modified)
CISPR 24:2010	NOTE	Harmonized as EN 55024:2010 (not modified)
CISPR 25:2008	NOTE	Harmonized as EN 55025:2008 (not modified)
ISO 17025:2005	NOTE	Harmonized as EN ISO/IEC 17025:2005 (not modified)

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies. However, for any use of this standard "within the meaning of Annex ZZ", the user must always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

When the IEC or ISO standard is referred to in the IEC text standard, this must be understood as a normative reference to the parallel EN standard, as outlined below, including the foreword and the Annexes ZZ.

NOTE 1 The way in which referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

NOTE 2 When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD and IEC/ISO</u>	<u>Year</u>
IEC 60417	Data base	Graphical symbols for use on equipment available from http://www.graphical-symbols.info/equipment	IEC 60417	2004
IEC 60601-1	2005	Medical electrical equipment Part 1: General requirements for basic safety and essential performance	EN 60601-1	2006
A1	2012		A1	2013
IEC 60601-1-8	2006	Medical electrical equipment Part 1-8: General requirements for basic safety and essential performance - Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	EN 60601-1-8 + corr. March	2007 2010
A1	2013		A1	2013
IEC 60601-1-11	2010	Medical electrical equipment Part 1-11: General requirements for basic safety and essential performance - Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	EN 60601-1-11	2010
IEC 60601-1-12	2014	Medical electrical equipment Part 1-12: General requirements for basic safety and essential performance - Collateral standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment		
IEC 60601-2-2	2010	Medical electrical equipment Part 2-2: Particular requirements for the basic safety and essential performance of high		

EN 60601-1-2:2015

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD and IEC/ISO</u>	<u>Year</u>
IEC 60601-2-3	2012	frequency surgical equipment and high frequency surgical accessories Medical electrical equipment – Part 2-3: Particular requirements for the basic safety and essential performance of short-wave therapy equipment		
IEC 61000-3-2	2005	Electromagnetic compatibility (EMC) - Part 3-2: Limits - Limits for harmonic current emissions (equipment input current ≤ 16 A per phase)	EN 61000-3-2	2006
A1	2008		+A1	2009
A2	2009		+A2	2009
IEC 61000-3-3	2013	Electromagnetic compatibility (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection	EN 61000-3-3	2013
IEC 61000-4-2	2008	Electromagnetic compatibility (EMC) - Part 4-2: Testing and measuring techniques - Electrostatic discharge immunity test	EN 61000-4-2	2009
IEC 61000-4-3	2006	Electromagnetic compatibility (EMC) - Part 4-3: Testing and measurement techniques - Radiated, radio-frequency, electromagnetic field immunity test	EN 61000-4-3	2006
A1	2007		+A1	2008
A2	2010		+IS1 +A2	2009 2010
IEC 61000-4-4	2012	Electromagnetic compatibility (EMC) - Part 4-4: Testing and measurement techniques - Electrical fast transient/burst immunity test	EN 61000-4-4	2012
IEC 61000-4-5	2005	Electromagnetic compatibility (EMC) - Part 4-5: Testing and measurement techniques - Surge immunity test	EN 61000-4-5	2006
IEC 61000-4-6	2013	Electromagnetic compatibility (EMC) - Part 4-6: Testing and measurement techniques - Immunity to conducted disturbances, induced by radio-frequency fields		
IEC 61000-4-8	2009	Electromagnetic compatibility (EMC) - Part 4-8: Testing and measurement techniques - Power frequency magnetic field immunity test	EN 61000-4-8	2010
IEC 61000-4-11	2004	Electromagnetic compatibility (EMC) - Part 4-11: Testing and measurement techniques - Voltage dips, short interruptions and voltage variations immunity tests	EN 61000-4-11	2004
CISPR 11	2009	Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics - Limits and methods of measurement	EN 55011 (mod)	2009
A1	2010			
CISPR 14-1	2005	Electromagnetic compatibility - Requirements for household appliances, electric tools and similar apparatus - Part 1: Emission	EN 55014-1 +A1 +A2	2006 2009 2011
CISPR 16-1-2	2003	Specification for radio disturbance and immunity measuring apparatus and methods - Part 1-2: Radio disturbance and immunity	EN 55016-1-2	2004

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD and IEC/ISO</u>	<u>Year</u>
		measuring apparatus - Ancillary equipment - Conducted disturbances		
A1	2004		+A1	2005
A2	2006		+A2	2006
CISPR 32	2012	Electromagnetic compatibility of multimedia equipment – Emission requirements	EN 55032	2012
ISO 7137	1995	Aircraft – Environmental conditions and test procedures for airborne equipment		
ISO 7637-2	2011	Road vehicles – Electrical disturbances from conduction and coupling – Part 2: Electrical transient conduction along supply lines only		
ISO 14971	2007	Medical devices – Application of risk management to medical devices	EN ISO 14971	2012

EN 60601-1-2:2015

Application of Annexes of the EN 60601 series

The Annex ZZ of EN 60601-1:2006+A1:2013 applies.

Annex ZZ (informative)

Coverage of Essential Requirements of EU Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association to provide a means of conforming to the Essential Requirements given in Annex I of EU Directive 93/42/EEC.

General Guidance:

Once this standard will be cited in the Official Journal of the European Union under that Directive, compliance with the clauses of this standard given in Table ZZ.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements (ERs) of that Directive and associated EFTA regulations.

NOTE 1 The standard's scope is limited to the specific uses, environments, contexts, objective situations specifically indicated. It cannot provide for presumption of conformity in other conditions. Some clauses or subclauses may be not applicable due to the specific type of equipment under consideration.

NOTE 2 Only prescriptions contained in the normative parts of the text are relevant to the presumption of conformity of this standard. Informative parts may, however, support users to interpret such prescriptions correctly.

NOTE 3 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 93/42/EEC. This means that risks have to be reduced "as far as possible", "to a minimum", "to the lowest possible level", "minimized" or "removed", according to the wording of the corresponding essential requirement which must be interpreted and applied in such a way as to take account of technology and practice existing at the time of design and of technical and economical considerations compatible with a high level of protection of health and safety.

NOTE 4 The manufacturer's policy for determining acceptable risk must be in compliance with essential requirements 1, 2, 5, 6, 7, 8, 9, 11 and 12 of the directive.

NOTE 5 For all parts of this standard that a) refer in their clauses to specific national legislation possibly exempting manufacturers from the thorough application of relevant provisions of this standard or b) link the completion of a relevant process/prescription to any discretionary choice/power of manufacturers, the user of the standard should check that such clauses are in compliance with Directive 93/42/EEC.

NOTE 6 This Annex ZZ is based on Normative References according to Annex ZA, replacing the references in the core text.

NOTE 7 According to the scope of this standard the coverage in Table ZZ.1 only applies to protection of ME equipment and ME systems against electromagnetic disturbances. This European Standard lists in Table ZZ.1 only the essential requirements covered.

WARNING: Other requirements and other EU Directives and Regulations may be applicable to the product(s) falling within the scope of this standard.

Table ZZ.1: Relationship between Essential Requirements of Directive 93/42/EEC, and Clauses and Subclauses of this standard

No.	Essential Requirements	Coverage EN 60601-1-2
I.	GENERAL REQUIREMENTS	
1.	General Guidance notes 1-7 shall be observed	
1	<p>The devices must be designed and manufactured in such a way that, when used under the conditions and for the purposes intended, they will not compromise the clinical condition or the safety of patients, or the safety and health of users or, where applicable, other persons, provided that any risks which may be associated with their intended use constitute acceptable risks when weighed against the benefits to the patient and are compatible with a high level of protection of health and safety.</p> <p>This shall include:</p>	<p>If the manufacturer follows this standard in his design and manufacturing process, this European Standard gives a valuable set of technical requirements to assist in fulfilling this ER with regard to EMC-aspects of the clinical condition or the safety of patients, or the safety and health of users or, where applicable, other persons (refer to clauses 4 to 9 of this collateral standard) without covering the risk benefit balancing.</p>
	<ul style="list-style-type: none"> - reducing, as far as possible, the risk of use error due to the ergonomic features of the device and the environment in which the device is intended to be used (design for patient safety), and 	<p>Covered in respect to 7.1 Protection of radio services and other equipment, 7.2 Protection of the public mains network, and 8.9 Immunity test levels</p>
2.	General Guidance notes 1-7 shall be observed	
2	<p>The solutions adopted by the manufacturer for the design and construction of the devices must conform to safety principles, taking account of the generally acknowledged state of the art.</p> <p>In selecting the most appropriate solutions, the manufacturer must apply the following principles in the following order:</p>	<p>1st paragraph: Covered under the condition that 2nd paragraph (including the following 3 bullets) is taken into account.</p> <p>2nd paragraph (including the following 3 bullets): Not covered.</p>
	<ul style="list-style-type: none"> - eliminate or reduce risks as far as possible (inherently safe design and construction), 	
	<ul style="list-style-type: none"> - where appropriate take adequate protection measures including alarms if necessary, in relation to risks that cannot be eliminated, 	
	<ul style="list-style-type: none"> - inform users of the residual risks due to any shortcomings of the protection measures adopted. 	
II.	REQUIREMENTS FOR DESIGN AND CONSTRUCTION	
	General Guidance notes 1-7 shall be observed	
9.2	Devices must be designed and manufactured in such a way as to remove or minimize as far as is	

No.	Essential Requirements	Coverage EN 60601-1-2
	possible:	
	<ul style="list-style-type: none"> - the risk of injury, in connection with their physical features, including the volume/pressure ratio, dimensional and where appropriate ergonomic features; 	
	<ul style="list-style-type: none"> - risks connected with reasonably foreseeable environmental conditions, such as magnetic fields, external electrical influences, electrostatic discharge, pressure, temperature or variations in pressure and acceleration; 	Covered in respect to electromagnetic disturbances, see 8.9 Immunity test levels.
	<ul style="list-style-type: none"> - the risks of reciprocal interference with other devices normally used in the investigations or for the treatment given; 	Covered in respect to electromagnetic disturbances, see 7.1 Protection of radio services and other equipment, 7.2 Protection of the public mains network.
	<ul style="list-style-type: none"> - risks arising where maintenance or calibration are not possible (as with implants), from ageing of materials used or loss of accuracy of any measuring or control mechanism. 	
11	Protection against radiation	General Guidance note 1-7 shall be observed
11.1	General	
11.1.1	Devices shall be designed and manufactured in such a way that exposure of patients, users and other persons to radiation shall be reduced as far as possible compatible with the intended purpose, whilst not restricting the application of appropriate specified levels for therapeutic and diagnostic purposes.	Covered in respect to electromagnetic disturbances, see 7.1 Protection of radio services and other equipment.
12	Requirements for medical devices connected to or equipped with an energy source	General Guidance notes 1-7 shall be observed
12.5	Devices must be designed and manufactured in such a way as to minimize the risks of creating electromagnetic fields which could impair the operation of other devices or equipment in the usual environment.	Covered in respect to electromagnetic disturbances, see 7.1 Protection of radio services and other equipment.
13	Information supplied by the manufacturer	General Guidance notes 1-7 shall be observed
13.5	Wherever reasonable and practicable, the devices and detachable components must be identified, where appropriate in terms of batches, to allow all appropriate action to detect any potential risk posed by the devices and detachable components.	Covered in respect to aspects related to accessories, components and subassemblies contained in 5.2.1.1 d) and e)



INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 1-2: General requirements for basic safety and essential performance –
Collateral Standard: Electromagnetic disturbances – Requirements and tests**

**Appareils électromédicaux –
Partie 1-2: Exigences générales pour la sécurité de base et les performances
essentielles – Norme collatérale: Perturbations électromagnétiques – Exigences
et essais**



THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2014 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
 3, rue de Varembe
 CH-1211 Geneva 20
 Switzerland

Tel.: +41 22 919 02 11
 Fax: +41 22 919 03 00
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

IEC Catalogue - webstore.iec.ch/catalogue

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

IEC publications search - www.iec.ch/searchpub

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing more than 30 000 terms and definitions in English and French, with equivalent terms in 14 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

More than 55 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: csc@iec.ch.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Catalogue IEC - webstore.iec.ch/catalogue

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

Recherche de publications IEC - www.iec.ch/searchpub

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient plus de 30 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 14 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

Plus de 55 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: csc@iec.ch.



INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 1-2: General requirements for basic safety and essential performance –
Collateral Standard: Electromagnetic disturbances – Requirements and tests**

**Appareils électromédicaux –
Partie 1-2: Exigences générales pour la sécurité de base et les performances
essentiels – Norme collatérale: Perturbations électromagnétiques – Exigences
et essais**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE **XD**
CODE PRIX

ICS 11.040.01, 33.100.10, 33.100.20

ISBN 978-2-8322-1413-8

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

CONTENTS	2
FOREWORD	6
INTRODUCTION	9
1 Scope, object and related standards	11
1.1 * Scope	11
1.2 Object	11
1.3 Related standards	11
1.3.1 IEC 60601-1	11
1.3.2 Particular standards	11
2 Normative references	11
3 Terms and definitions	13
4 General requirements	17
4.1 RISK MANAGEMENT PROCESS for ME EQUIPMENT and ME SYSTEMS	17
4.2 * Non-ME EQUIPMENT used in an ME SYSTEM	17
4.3 General test conditions	17
4.3.1 * Configurations	17
4.3.2 Artificial hand	18
4.3.3 * Power input voltages and frequencies	18
5 ME EQUIPMENT and ME SYSTEMS identification, marking and documents	20
5.1 Additional requirements for marking on the outside of ME EQUIPMENT and ME SYSTEMS that are specified for use only in a shielded location SPECIAL ENVIRONMENT	20
5.2 ACCOMPANYING DOCUMENTS	20
5.2.1 Instructions for use	20
5.2.2 Technical description	21
6 Documentation of the tests	23
6.1 General	23
6.2 Test plan	23
6.3 Test report	23
7 ELECTROMAGNETIC EMISSIONS requirements for ME EQUIPMENT and ME SYSTEMS	23
7.1 Protection of radio services and other equipment	23
7.1.1 * General	23
7.1.2 Operating modes	23
7.1.3 Multimedia equipment	24
7.1.4 * Subsystems	24
7.1.5 ME EQUIPMENT and ME SYSTEMS specified for use only in a shielded location SPECIAL ENVIRONMENT	24
7.1.6 ME EQUIPMENT and ME SYSTEMS that include radio equipment	24
7.1.7 * ME EQUIPMENT whose main functions are performed by motors and switching or regulating devices	25
7.1.8 ME EQUIPMENT and ME SYSTEMS containing X-ray generators	25
7.1.9 PATIENT physiological simulation	25
7.1.10 Artificial hand	25
7.1.11 PATIENT-coupled cables	25
7.1.12 PERMANENTLY INSTALLED LARGE ME EQUIPMENT and LARGE ME SYSTEMS	25
7.2 Protection of the PUBLIC MAINS NETWORK	26

	7.2.1	* Harmonic distortion	26
	7.2.2	* Voltage fluctuations and flicker	26
	7.3	EMISSIONS requirements summary	26
8		Electromagnetic IMMUNITY requirements for ME EQUIPMENT and ME SYSTEMS	27
	8.1	* General	27
	8.2	PATIENT physiological simulation	30
	8.3	Termination of PATIENT-COUPLED parts	30
	8.4	HAND-HELD ME EQUIPMENT and parts intended to be HAND-HELD	30
	8.5	* Subsystems	31
	8.6	PERMANENTLY INSTALLED LARGE ME EQUIPMENT and LARGE ME SYSTEMS	31
	8.7	* Operating modes	31
	8.8	* Non-ME EQUIPMENT	32
	8.9	* IMMUNITY TEST LEVELS	32
	8.10	* IMMUNITY to proximity fields from RF wireless communications equipment	39
9		* Test report	41
		Annex A (informative) General guidance and rationale	43
	A.1	Safety and performance	43
	A.2	Testing of normally non-observable functions	43
	A.3	Rationale for particular clauses and subclauses	43
		Annex B (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	57
	B.1	Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts	57
	B.2	ACCOMPANYING DOCUMENTS, instructions for use	57
	B.3	ACCOMPANYING DOCUMENTS, technical description	57
		Annex C (informative) Guidance in classification according to CISPR 11	59
	C.1	General	59
	C.2	Separation into groups	59
	C.3	Division into classes	60
		Annex D (informative) Guidance in the application of IEC 60601-1-2 to particular standards	61
	D.1	General	61
	D.2	Recommended modifications	61
		D.2.1 Testing requirements	61
		D.2.2 ACCOMPANYING DOCUMENTS	61
	D.3	Cautions	61
		Annex E (informative) Determination of IMMUNITY TEST LEVELS for SPECIAL ENVIRONMENTS	63
	E.1	General	63
	E.2	Summary of method for E.1 a)	66
	E.3	Summary of method for E.1 b), c) and d)	66
	E.4	Determination of EM DISTURBANCE level reduction	66
	E.5	Assessment of EM DISTURBANCE sources	66
	E.6	Reasonably foreseeable maximum EM DISTURBANCE levels	67
	E.7	Determination of IMMUNITY TEST LEVELS	67
	E.8	RF radiators in SPECIAL ENVIRONMENTS	67
	E.9	Examples of mitigations and special conditions	68
		Annex F (informative) RISK MANAGEMENT for BASIC SAFETY and ESSENTIAL PERFORMANCE with regard to ELECTROMAGNETIC DISTURBANCES	69

F.1	General.....	69
F.2	General requirements for RISK MANAGEMENT	70
F.3	RISK ANALYSIS.....	71
F.4	RISK EVALUATION.....	74
F.5	RISK CONTROL.....	75
	F.5.1 RISK CONTROL option analysis	75
	F.5.2 Implementation of RISK CONTROL measure(s).....	75
	F.5.3 RESIDUAL RISK EVALUATION.....	75
	F.5.4 RISK/benefit analysis.....	76
	F.5.5 RISKS arising from RISK CONTROL measures	76
	F.5.6 Completeness of RISK CONTROL	76
F.6	Evaluation of overall RESIDUAL RISK acceptability	76
F.7	RISK MANAGEMENT report	76
F.8	Production and post-production information	77
Annex G (informative)	Guidance: Test plan.....	78
	G.1 Test plan contents	78
Annex H (informative)	PATIENT-coupled cables EMISSIONS	80
	H.1 * Protection of other equipment from PATIENT cable conducted EMISSIONS	80
	H.2 Test method.....	80
	H.3 Rationale	80
Annex I (informative)	Identification of IMMUNITY pass/fail criteria	82
	I.1 General.....	82
	I.2 IMMUNITY pass/fail criteria principles.....	82
	I.2.1 General	82
	I.2.2 IMMUNITY pass/fail criteria for non-ME EQUIPMENT used in an ME SYSTEM	82
	I.2.3 IMMUNITY pass/fail criteria determination	82
	I.3 IMMUNITY pass/fail criteria examples.....	83
	I.3.1 General examples.....	83
	I.3.2 Example of IMMUNITY pass/fail criteria for a radiological table system.....	84
	Bibliography.....	86
	Index of defined terms used in this collateral standard	89
	Figure 1 – RC element of the artificial hand	18
	Figure 2 – PORTS of ME EQUIPMENT and ME SYSTEMS	27
	Figure 3 – Examples of environments of INTENDED USE.....	33
	Figure A.1 – Examples of PORTS (from IEC 61000-6-1:2005).....	47
	Figure A.2 – IEC 61000-4-2 Figure A.1 – Maximum values of electrostatic voltages to which OPERATORS can be charged while in contact with the materials mentioned in A.2	54
	Figure E.1 – Test plan development flow when SPECIAL ENVIRONMENTS are known	64
	Figure E.2 – Sub-process for determination of IMMUNITY TEST LEVELS for SPECIAL ENVIRONMENTS.....	65
	Figure F.1 – Function of this collateral standard in the RISK MANAGEMENT PROCESS.....	69
	Figure F.2 – Examples of multiple VERIFICATION methods for improving confidence in RISK levels	70
	Figure H.1 – Setup for PATIENT-COUPLED cables conducted EMISSIONS test for ME EQUIPMENT and ME SYSTEMS that conform to IEC 60601-2-27	81

Table 1 – Power input voltages and frequencies during the tests (1 of 2)	19
Table 2 – EMISSION limits per environment	26
Table 3 – Procedure for continuing to test ME EQUIPMENT or ME SYSTEMS that are damaged by an IMMUNITY test signal	28
Table 4 – * ENCLOSURE PORT	34
Table 5 – * Input a.c. power PORT (1 of 2)	35
Table 6 – Input d.c. power PORT	37
Table 7 – * PATIENT coupling PORT	38
Table 8 – Signal input/output parts PORT	39
Table 9 – Test specifications for ENCLOSURE PORT IMMUNITY to RF wireless communications equipment.....	40
Table 10 – * Minimum test report contents (1 of 2)	41
Table A.1 – IEC/TR 61000-2-5 information considered in specifying IMMUNITY TEST LEVELS for each IMMUNITY TEST	49
Table B.1 – Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts	57
Table B.2 – ACCOMPANYING DOCUMENTS, instructions for use	57
Table B.3 – ACCOMPANYING DOCUMENTS, technical description	58
Table E.1 – Examples of specific mitigations / environmental conditions	68
Table F.1 – Examples of EM phenomena that should be considered in a RISK ANALYSIS	72
Table G.1 – Recommended minimum test plan contents (1 of 2)	78
Table H.1 – PATIENT-COUPLED conducted EMISSIONS recommended limit.....	80
Table I.1 – Example of IMMUNITY pass criteria for a radiological table system	85

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-1-2 has been prepared by IEC subcommittee 62A: Common aspects of electrical equipment used in medical practice of IEC technical committee 62: Electrical equipment in medical practice.

This fourth edition cancels and replaces the third edition of IEC 60601-1-2, published in 2007, and constitutes a technical revision.

This fourth edition constitutes a collateral standard to IEC 60601-1: *Medical electrical equipment – Part 1: General requirements for safety and essential performance* hereafter referred to as the general standard.

The most significant changes with respect to the previous edition include the following modifications:

- specification of IMMUNITY TEST LEVELS according to the environments of INTENDED USE, categorized according to locations that are harmonized with IEC 60601-1-11: the professional healthcare facility environment, the HOME HEALTHCARE ENVIRONMENT and SPECIAL ENVIRONMENTS;
- specification of tests and test levels to improve the safety of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS when PORTABLE RF communications equipment is used closer to the MEDICAL ELECTRICAL EQUIPMENT than was recommended based on the IMMUNITY TEST LEVELS that were specified in the third edition;
- specification of IMMUNITY tests and IMMUNITY TEST LEVELS according to the PORTS of the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM;
- specification of IMMUNITY TEST LEVELS based on the reasonably foreseeable maximum level of ELECTROMAGNETIC DISTURBANCES in the environments of INTENDED USE, resulting in some IMMUNITY TEST LEVELS that are higher than in the previous edition; and
- better harmonization with the RISK concepts of BASIC SAFETY and ESSENTIAL PERFORMANCE, including deletion of the defined term “life-supporting”;

and the following additions:

- guidance for determination of IMMUNITY TEST LEVELS for SPECIAL ENVIRONMENTS;
- guidance for adjustment of IMMUNITY TEST LEVELS when special considerations of mitigations or INTENDED USE are applicable;
- guidance on RISK MANAGEMENT for BASIC SAFETY and ESSENTIAL PERFORMANCE with regard to ELECTROMAGNETIC DISTURBANCES; and
- guidance on identification of IMMUNITY pass/fail criteria.

The text of this collateral standard is based on the following documents:

FDIS	Report on voting
62A/916/FDIS	62A/924/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. ALARM SYSTEMS).

In this collateral standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this collateral standard, the term

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 1 includes 1.1, 1.2, etc.);

- “subclause” means a numbered subdivision of a clause (e.g. 1.1, 1.2 and 1.3.1 are all subclauses of Clause 1).

References to clauses within this collateral standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this collateral standard are by number only.

In this collateral standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this collateral standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this collateral standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this collateral standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this collateral standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex A.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip them for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for mandatory implementation nationally not earlier than 3 years from the date of publication.

IMPORTANT – The “colour inside” logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.

INTRODUCTION

The need for establishing specific standards for BASIC SAFETY and ESSENTIAL PERFORMANCE with regard to ELECTROMAGNETIC DISTURBANCES for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS is well recognized.

The requirements and tests specified by this collateral standard are generally applicable to MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS as defined in 3.63 and 3.64 in the general standard. For certain types of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, these requirements might need to be modified by the special requirements of a particular standard. Writers of particular standards are encouraged to refer to Annex D for guidance in the application of this collateral standard.

MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS are expected to provide their BASIC SAFETY and ESSENTIAL PERFORMANCE without interfering with other equipment and systems in the ELECTROMAGNETIC ENVIRONMENTS in which they are intended by their MANUFACTURER to be used. The application of ELECTROMAGNETIC EMISSION standards is essential for the protection of:

- safety services;
- other MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS;
- non-ME EQUIPMENT (e.g. computers);
- telecommunications (e.g. radio/TV, telephone, radio-navigation).

Of even more importance, the application of ELECTROMAGNETIC IMMUNITY standards is essential to ensure safety of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS. To ensure safety, MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS are expected to provide their BASIC SAFETY and ESSENTIAL PERFORMANCE in the ELECTROMAGNETIC ENVIRONMENTS of INTENDED USE throughout their EXPECTED SERVICE LIFE.

This collateral standard specifies IMMUNITY TEST LEVELS for safety for ME EQUIPMENT and ME SYSTEMS intended by their MANUFACTURER for use in the professional healthcare facility environment or the HOME HEALTHCARE ENVIRONMENT. It recognizes that RF wireless communications equipment can no longer be prohibited from most PATIENT ENVIRONMENTS because in many cases it has become essential to the efficient provision of healthcare. This collateral standard also recognizes that, for certain SPECIAL ENVIRONMENTS, higher or lower IMMUNITY TEST LEVELS than those specified for the professional healthcare facility environment and the HOME HEALTHCARE ENVIRONMENT might be appropriate. This collateral standard provides guidance in determining appropriate IMMUNITY TEST LEVELS for SPECIAL ENVIRONMENTS.

The IMMUNITY TEST LEVELS specified for BASIC SAFETY and ESSENTIAL PERFORMANCE are based on the reasonably foreseeable maximum of the ELECTROMAGNETIC DISTURBANCE phenomena in the applicable environments of INTENDED USE.

Not all ELECTROMAGNETIC DISTURBANCE phenomena are covered by this collateral standard, as it is not practical to do so. MANUFACTURERS of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS need to address this during their RISK ASSESSMENT and evaluate if other ELECTROMAGNETIC DISTURBANCE phenomena could make their product unsafe. This evaluation should be based on the environments of INTENDED USE and the reasonably foreseeable maximum levels of ELECTROMAGNETIC DISTURBANCES expected throughout the EXPECTED SERVICE LIFE.

This collateral standard recognizes that the MANUFACTURER has the responsibility to design and perform VERIFICATION of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS to meet the requirements of this collateral standard and to disclose information to the RESPONSIBLE ORGANIZATION or OPERATOR so that the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM will remain safe throughout its EXPECTED SERVICE LIFE.

This collateral standard provides guidance in incorporating considerations regarding ELECTROMAGNETIC DISTURBANCES into the RISK MANAGEMENT PROCESS.

This collateral standard is based on existing IEC standards prepared by subcommittee 62A, technical committee 77 (ELECTROMAGNETIC COMPATIBILITY between electrical equipment including networks), ISO (International standards organization), and CISPR (International special committee on radio interference).

MEDICAL ELECTRICAL EQUIPMENT –

Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests

1 Scope, object and related standards

1.1 * Scope

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, hereafter referred to as ME EQUIPMENT and ME SYSTEMS.

This collateral standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of ME EQUIPMENT and ME SYSTEMS in the presence of ELECTROMAGNETIC DISTURBANCES and to ELECTROMAGNETIC DISTURBANCES emitted by ME EQUIPMENT and ME SYSTEMS.

BASIC SAFETY with regard to ELECTROMAGNETIC DISTURBANCES is applicable to all ME EQUIPMENT and ME SYSTEMS.

1.2 Object

The object of this collateral standard is to specify general requirements and tests for BASIC SAFETY and ESSENTIAL PERFORMANCE with regard to ELECTROMAGNETIC DISTURBANCES and for ELECTROMAGNETIC EMISSIONS of ME EQUIPMENT and ME SYSTEMS. They are in addition to the requirements of the general standard and serve as the basis for particular standards.

1.3 Related standards

1.3.1 IEC 60601-1

For ME EQUIPMENT and ME SYSTEMS, this collateral standard complements IEC 60601-1.

When referring to IEC 60601-1 or to this collateral standard, either individually or in combination, the following conventions are used:

- "the general standard" designates IEC 60601-1 alone (IEC 60601-1:2005+A1:2012);
- "this collateral standard" designates IEC 60601-1-2 alone;
- "this standard" designates the combination of the general standard and this collateral standard.

1.3.2 Particular standards

A requirement in a particular standard takes priority over the corresponding requirement in this collateral standard.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

IEC 60601-1:2005¹⁾, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
Amendment 1:2012

IEC 60601-1-8:2006²⁾, *Medical electrical equipment – Part 1-8: General requirements for basic safety and essential performance – Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*
Amendment 1:2012

IEC 60601-1-11:2010, *Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*

IEC 60601-1-12__³⁾ *Medical electrical equipment – Part 1-12: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment*

IEC 60601-2-2:2009, *Medical electrical equipment – Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories*

IEC 60601-2-3:2012, *Medical electrical equipment – Part 2-3: Particular requirements for the basic safety and essential performance of short-wave therapy equipment*

IEC 61000-3-2:2005⁴⁾, *Electromagnetic compatibility (EMC) – Part 3-2: Limits – Limits for harmonic current emissions (equipment input current ≤ 16 A per phase)*
Amendment 1:2008
Amendment 2:2009

IEC 61000-3-3:2013, *Electromagnetic compatibility (EMC) – Part 3-3: Limits – Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection*

IEC 61000-4-2:2008, *Electromagnetic compatibility (EMC) – Part 4-2: Testing and measurement techniques – Electrostatic discharge immunity test*

IEC 61000-4-3:2006⁵⁾, *Electromagnetic compatibility (EMC) – Part 4-3: Testing and measurement techniques – Radiated, radio-frequency, electromagnetic field immunity test*
Amendment 1:2007
Amendment 2:2010

IEC 61000-4-4:2012, *Electromagnetic compatibility (EMC) – Part 4-4: Testing and measurement techniques – Electrical fast transient/burst immunity test*

1) There exists a consolidated edition 3.1, including IEC 60601-1:2005 and its Amendment 1:2012.

2) There exists a consolidated edition 2.1, including IEC 60601-1-8:2006 and its Amendment 1:2012.

3) To be published.

4) There exists a consolidated edition 3.2, including IEC 61000-3-2:2005 and its Amendment 1:2008 and Amendment 2:2009.

5) There exists a consolidated edition 3.2, including IEC 61000-4-3:2006 and its Amendment 1:2007 and Amendment 2:2010.

IEC 61000-4-5:2005, *Electromagnetic compatibility (EMC) – Part 4-5: Testing and measurement techniques – Surge immunity test*

IEC 61000-4-6:2013, *Electromagnetic compatibility (EMC) – Part 4-6: Testing and measurement techniques – Immunity to conducted disturbances, induced by radio-frequency fields*

IEC 61000-4-8:2009, *Electromagnetic compatibility (EMC) – Part 4-8: Testing and measurement techniques – Power frequency magnetic field immunity test*

IEC 61000-4-11:2004, *Electromagnetic compatibility (EMC) – Part 4-11: Testing and measuring techniques – Voltage dips, short interruptions and voltage variations immunity tests*

CISPR 11:2009⁶⁾, *Industrial, scientific and medical equipment – Radio-frequency disturbance characteristics – Limits and methods of measurement*
Amendment 1:2010

CISPR 14-1:2005, *Electromagnetic compatibility – Requirements for household appliances, electric tools and similar apparatus – Part 1: Emission*

CISPR 16-1-2:2003⁷⁾, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-2: Radio disturbance and immunity measuring apparatus – Ancillary equipment – Conducted disturbances*
Amendment 1:2004
Amendment 2:2006

CISPR 32:2012, *Electromagnetic compatibility of multimedia equipment – Emission requirements*

ISO 7137:1995, *Aircraft – Environmental conditions and test procedures for airborne equipment*

ISO 7637-2:2011, *Road vehicles – Electrical disturbances from conduction and coupling – Part 2: Electrical transient conduction along supply lines only*

ISO 14971:2007, *Medical devices – Application of risk management to medical devices*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN

6) There exists a consolidated edition 5.1, including CISPR 11:2009 and its Amendment 1:2010.

7) There exists a consolidated edition 1.2, including CISPR 16-1-2:2003 and its Amendment 1:2004 and Amendment 2:2006.