

STN	Implantáty pre chirurgiu. Keramické materiály na báze yttriom stabilizovaného tetragonálneho zirkónia (Y - TZP) (ISO 13356: 2015).	STN EN ISO 13356 85 2927
------------	---	--

Implants for surgery - Ceramic materials based on yttria-stabilized tetragonal zirconia (Y-TZP) (ISO 13356:2015)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 02/16

Obsahuje: EN ISO 13356:2015, ISO 13356:2015

Oznámením tejto normy sa ruší
STN EN ISO 13356 (85 2927) z augusta 2013

122437

Úrad pre normalizáciu, metrológiu a skúšobníctvo SR, 2016
Podľa zákona č. 264/1999 Z. z. v znení neskorších predpisov sa môžu slovenské technické normy
rozmnžovať a rozširovať iba so súhlasom Úradu pre normalizáciu, metrológiu a skúšobníctvo SR.

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 13356

September 2015

ICS 11.040.40

Supersedes EN ISO 13356:2013

English Version

Implants for surgery - Ceramic materials based on yttria-stabilized tetragonal zirconia (Y-TZP) (ISO 13356:2015)

Implants chirurgicaux - Produits céramiques à base de zircone tétragonal stabilisée à l'yttrium (Y-TZP) (ISO 13356:2015)

Chirurgische Implantate - Keramische Werkstoffe aus yttriumstabilisiertem tetragonalem Zirkonium (Y-TZP) (ISO 13356:2015)

This European Standard was approved by CEN on 22 August 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

European foreword.....	3
------------------------	---

European foreword

This document (EN ISO 13356:2015) has been prepared by Technical Committee ISO/TC 106 “Dentistry” in collaboration with Technical Committee CEN/TC 55 “Dentistry” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2016, and conflicting national standards shall be withdrawn at the latest by March 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 13356:2013.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 13356:2015 has been approved by CEN as EN ISO 13356:2015 without any modification.

Implants for surgery — Ceramic materials based on yttria-stabilized tetragonal zirconia (Y-TZP)

Implants chirurgicaux — Produits céramiques à base de zircone tétragonal stabilisée à l'yttrium (Y-TZP)



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Physical and chemical properties	2
3.1 Test category	2
3.1.1 General	2
3.1.2 Category 1	2
3.1.3 Category 2	2
4 Test methods	4
4.1 General	4
4.2 Bulk density	4
4.3 Chemical composition	4
4.4 Microstructure	4
4.4.1 Principle	4
4.4.2 Test report	5
4.4.3 Amount of monoclinic phase	5
4.5 Biaxial flexural strength	6
4.5.1 Principle	6
4.5.2 Apparatus	6
4.5.3 Preparation of test specimens	6
4.5.4 Procedure	7
4.5.5 Calculation of results	8
4.5.6 Test report	8
4.6 Four-point bending strength	8
4.7 Weibull modulus	9
4.8 Young's modulus	9
4.9 Hardness	9
4.10 Cyclic fatigue	9
4.10.1 Principle	9
4.10.2 Apparatus	9
4.10.3 Sample size and preparation of test specimens	9
4.10.4 Procedure and sample requirement	9
4.10.5 Test Report	10
4.11 Radioactivity	10
4.11.1 Principle	10
4.11.2 Apparatus	10
4.11.3 Sample preparation	11
4.11.4 Isotope identification - Energy calibration	11
4.11.5 Quantitative analysis	11
4.11.6 Expression of results	12
4.11.7 Test report	12
4.12 Accelerated aging test	12
4.12.1 General	12
4.12.2 Procedure	12
4.12.3 Evaluation of accelerated aging outcome	12
Bibliography	13

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary Information](#)

The committee responsible for this document is ISO/TC 150, *Implants for surgery*, Subcommittee SC 1, *Materials*.

This third edition cancels and replaces the second edition (ISO 13356:2008), which has been technically revised.

Introduction

No known surgical implant material has ever been found to cause absolutely no adverse reactions in the human body. However, long-term clinical experience regarding the use of the material referred to in this International Standard has shown that an acceptable level of biological response can be expected if the material will be used in appropriate applications.

Implants for surgery — Ceramic materials based on yttria-stabilized tetragonal zirconia (Y-TZP)

1 Scope

This International Standard specifies the requirements and corresponding test methods for a biocompatible and biostable ceramic bone-substitute material based on yttria-stabilized tetragonal zirconia (yttria tetragonal zirconia polycrystal, Y-TZP) for use as a material for surgical implants.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3310-1, *Test sieves — Technical requirements and testing — Part 1: Test sieves of metal wire cloth*

ISO 3611, *Geometrical product specifications (GPS) — Dimensional measuring equipment: Micrometers for external measurements — Design and metrological characteristics*

ISO 7500-1, *Metallic materials — Verification of static uniaxial testing machines — Part 1: Tension/compression testing machines — Verification and calibration of the force-measuring system*

ISO 13383-1, *Fine ceramics (advanced ceramics, advanced technical ceramics) — Microstructural characterization — Part 1: Determination of grain size and size distribution*

ISO 14704, *Fine ceramics (advanced ceramics, advanced technical ceramics) — Test method for flexural strength of monolithic ceramics at room temperature*

ISO 14705, *Fine ceramics (advanced ceramics, advanced technical ceramics) — Test method for hardness of monolithic ceramics at room temperature*

ISO 17561, *Fine ceramics (advanced ceramics, advanced technical ceramics) - Test method for elastic moduli of monolithic ceramics at room temperature by sonic resonance*

ISO 18754, *Fine ceramics (advanced ceramics, advanced technical ceramics) — Determination of density and apparent porosity*

ISO 20501, *Fine ceramics (advanced ceramics, advanced technical ceramics) — Weibull statistics for strength data*

ISO 22214, *Fine ceramics (advanced ceramics, advanced technical ceramics) — Test method for cyclic bending fatigue of monolithic ceramics at room temperature*

EN 623-2, *Advanced technical ceramics — Monolithic ceramics — General and textural properties — Part 2: Determination of density and porosity*

EN 843-2, *Advanced technical ceramics — Mechanical properties of monolithic ceramics at room temperature — Part 2: Determination of Young's modulus, shear modulus and Poisson's ratio*

EN 843-4, *Advanced technical ceramics — Mechanical properties of monolithic ceramics at room temperature — Part 4: Vickers, Knoop and Rockwell superficial hardness*

EN 843-5, *Advanced technical ceramics — Mechanical properties of monolithic ceramics at room temperature — Part 5: Statistical analysis*

ASTM C1161, Standard Test Method for Flexural Strength of Advanced Ceramics at Ambient Temperature

ASTM C1198, Standard Test Method for Dynamic Young's Modulus, Shear Modulus, and Poisson's Ratio for Advanced Ceramics by Sonic Resonance

ASTM C1239, Standard Practice for Reporting Uniaxial Strength Data and Estimating Weibull Distribution Parameters for Advanced Ceramics

ASTM C1259, Standard Test method for Dynamic Young's Modulus, Shear Modulus, and Poisson's Ratio for Advanced Ceramics by Impulse Excitation of Vibration

ASTM C1327, Standard Test Method for Vickers Indentation Hardness of Advanced Ceramics

ASTM C1331, Standard Test Method for Measuring Ultrasonic Velocity in Advanced Ceramics with Broadband Pulse-Echo Cross-Correlation Method

ASTM E112, Standard Test Method for Determining Average Grain Size

koniec náhľadu – text ďalej pokračuje v platenej verzii STN