

STN	Kaly, upravené bioodpady a zemina. Stanovenie ortuti v mineralizátoch lúčavky kráľovskej a kyseliny dusičnej. Časť 1: Metóda absorpčnej atómovej spektrometrie so studenými parami (CVAAS).	STN EN 16175-1 83 8434
------------	--	--

Sludge, treated biowaste and soil - Determination of mercury - Part 1: Cold-vapour atomic absorption spectrometry (CV-AAS)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 03/17

Obsahuje: EN 16175-1:2016

Oznámením tejto normy sa ruší
STN P CEN/TS 16175-1 (83 8434) z augusta 2013

124258

Úrad pre normalizáciu, metrológiu a skúšobníctvo Slovenskej republiky, 2017
Podľa zákona č. 264/1999 Z. z. o technických požiadavkách na výrobky a o posudzovaní zhody a o zmene a doplnení niektorých zákonov v znení neskorších predpisov sa slovenská technická norma a časti slovenskej technickej normy môžu rozmnožovať alebo rozširovať len so súhlasom slovenského národného normalizačného orgánu.

EUROPEAN STANDARD

EN 16175-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2016

ICS 13.030.01; 13.080.10

Supersedes CEN/TS 16175-1:2013

English Version

Sludge, treated biowaste and soil - Determination of mercury - Part 1: Cold-vapour atomic absorption spectrometry (CV-AAS)

Boues, bio-déchets traités et sols - Détermination du mercure - Partie 1: Spectrométrie d'absorption atomique de vapeur froide (SAA-VP)

Schlamm, behandelter Bioabfall und Boden - Bestimmung von Quecksilber - Teil 1: Kaltdampf-Atomabsorptionsspektrometrie (CV-AAS)

This European Standard was approved by CEN on 19 March 2016.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
European foreword	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Principle	5
4 Interferences	5
5 Reagents	6
6 Apparatus	7
6.1 Usual laboratory apparatus.....	7
6.2 Atomic absorption spectrometer (AAS).....	7
6.3 Automated sample introduction system	8
6.4 Cold-vapour generator, batch system or an automated flow injection analysis system (FIA)	8
7 Procedure	8
7.1 Test solution	8
7.2 Test blank solution.....	8
7.3 Preparation of the calibration solutions	8
7.4 Calibration	8
7.5 Measurement of test sample	9
8 Calculation and expression of results	9
8.1 Calculation	9
8.2 Expression of results.....	9
9 Performance data	9
10 Test report	10
Annex A (informative) Repeatability and reproducibility data	11
A.1 Materials used in the interlaboratory comparison study	11
A.2 Interlaboratory comparison results	11
Bibliography	13

European foreword

This document (EN 16175-1:2016) has been prepared by Technical Committee CEN/TC 444 “Test methods for environmental characterization of solid matrices”, the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2017, and conflicting national standards shall be withdrawn at the latest by April 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN/TS 16175-1:2013.

The preparation of the previous edition of this analytical method by CEN is based on a mandate by the European Commission (Mandate M/330), which assigned the development of standards on sampling and analytical methods for hygienic and biological parameters as well as inorganic and organic determinants, aiming to make these standards applicable to sludge, treated biowaste and soil as far as this is technically feasible.

EN 16175, *Sludge, treated biowaste and soil — Determination of mercury* comprises the following parts:

- *Part 1: Cold-vapour atomic absorption spectrometry (CV-AAS);*
- *Part 2: Cold-vapour atomic fluorescence spectrometry (CV-AFS).*

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

This European Standard is applicable and validated for several types of matrices as indicated in Table 1 (see Annex A for the results of validation).

Table 1 — Matrices for which this European Standard is applicable and validated

Matrix	Materials used for validation
Sludge	Municipal sludge
Biowaste	Compost
Soil	Soil

WARNING — Persons using this European Standard should be familiar with usual laboratory practice. This European Standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this European Standard be carried out by suitably trained staff.

1 Scope

This European Standard specifies a method for the determination of mercury in *aqua regia* or nitric acid digests of sludge, treated biowaste and soil, obtained according to EN 16173 or EN 16174 using cold-vapour atomic absorption spectrometry (CV-AAS). The lower working range limit is 0,03 mg/kg (dry matter basis).

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15934, *Sludge, treated biowaste, soil and waste — Calculation of dry matter fraction after determination of dry residue or water content*

EN 16173, *Sludge, treated biowaste and soil — Digestion of nitric acid soluble fractions of elements*

EN 16174, *Sludge, treated biowaste and soil — Digestion of aqua regia soluble fractions of elements*

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696)*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN