

<b>STN</b>	<b>Ochrana ovzdušia. Stacionárne zdroje emisií. Meranie hmotnostnej koncentrácie oxidu uhľnatého. Štandardná referenčná metóda: nedisperzná infračervená spektrometria.</b>	<b>STN EN 15058</b>  83 4740
------------	---	--

Stationary source emissions - Determination of the mass concentration of carbon monoxide - Standard reference method: non-dispersive infrared spectrometry

Táto norma obsahuje anglickú verziu európskej normy.  
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 05/17

Obsahuje: EN 15058:2017

Oznámením tejto normy sa ruší  
STN EN 15058 (83 4740) z marca 2007

**124733**

Úrad pre normalizáciu, metrológiu a skúšobníctvo Slovenskej republiky, 2017  
Podľa zákona č. 264/1999 Z. z. o technických požiadavkách na výrobky a o posudzovaní zhody a o zmene a doplnení niektorých zákonov v znení neskorších predpisov sa slovenská technická norma a časti slovenskej technickej normy môžu rozmnožovať alebo rozširovať len so súhlasom slovenského národného normalizačného orgánu.



EUROPEAN STANDARD

**EN 15058**

NORME EUROPÉENNE

EUROPÄISCHE NORM

January 2017

ICS 13.040.40

Supersedes EN 15058:2006

English Version

## Stationary source emissions - Determination of the mass concentration of carbon monoxide - Standard reference method: non-dispersive infrared spectrometry

Émissions de sources fixes - Détermination de la concentration massique de monoxyde de carbone - Méthode de référence normalisée : spectrométrie infrarouge non dispersive

Emissionen aus stationären Quellen - Bestimmung der Massenkonzentration von Kohlenmonoxid - Standardreferenzverfahren: Nicht-dispersive Infrarotspektrometrie

This European Standard was approved by CEN on 26 September 2016.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

<b>Contents</b>	<b>Page</b>
European foreword.....	5
<b>1 Scope</b> .....	<b>6</b>
<b>2 Normative references</b> .....	<b>6</b>
<b>3 Terms and definitions</b> .....	<b>6</b>
<b>4 Symbols and abbreviations</b> .....	<b>13</b>
4.1 Symbols.....	13
4.2 Abbreviated terms .....	14
<b>5 Principle</b> .....	<b>14</b>
5.1 General.....	14
5.2 Measuring principle .....	14
<b>6 Description of the measuring system</b> .....	<b>15</b>
6.1 General.....	15
6.2 Sampling and sample gas conditioning system.....	16
6.2.1 Sampling probe .....	16
6.2.2 Filter.....	16
6.2.3 Sample gas line.....	16
6.2.4 Sample gas conditioning system.....	16
6.2.5 Sample gas pump.....	17
6.2.6 Secondary filter.....	17
6.2.7 Flow controller and flow meter .....	17
6.3 Analyser equipment.....	17
6.3.1 General.....	17
6.3.2 Pressure and temperature effects .....	18
6.3.3 Sampling pump for the analyser.....	18
6.3.4 Interferences due to infrared absorbing gases .....	18
<b>7 Performance characteristics of the SRM</b> .....	<b>18</b>
<b>8 Suitability of the measuring system for the measurement task</b> .....	<b>20</b>
<b>9 Field operation</b> .....	<b>20</b>
9.1 Measurement planning.....	20
9.2 Sampling strategy.....	21
9.2.1 General.....	21
9.2.2 Measurement section and measurement plane.....	21
9.2.3 Minimum number and location of measurement points.....	21
9.2.4 Measurement ports and working platform .....	21
9.3 Choice of the measuring system .....	21
9.4 Setting of the measuring system on site .....	22
9.4.1 General.....	22
9.4.2 Preliminary zero and span check, and adjustments .....	22
9.4.3 Zero and span checks after measurement .....	23

<b>10</b>	<b>Ongoing quality control</b> .....	<b>23</b>
<b>10.1</b>	<b>Introduction</b> .....	<b>23</b>
<b>10.2</b>	<b>Frequency of checks</b> .....	<b>23</b>
<b>11</b>	<b>Expression of results</b> .....	<b>24</b>
<b>12</b>	<b>Equivalence of an alternative method</b> .....	<b>25</b>
<b>13</b>	<b>Measurement report</b> .....	<b>25</b>
	<b>Annex A (informative) Validation of the method in the field</b> .....	<b>26</b>
<b>A.1</b>	<b>General</b> .....	<b>26</b>
<b>A.2</b>	<b>Characteristics of installations</b> .....	<b>26</b>
<b>A.3</b>	<b>Repeatability and reproducibility in the field</b> .....	<b>27</b>
<b>A.3.1</b>	<b>General</b> .....	<b>27</b>
<b>A.3.2</b>	<b>Repeatability</b> .....	<b>28</b>
<b>A.3.3</b>	<b>Reproducibility</b> .....	<b>29</b>
	<b>Annex B (informative) Schematics of non-dispersive infrared spectrometer</b> .....	<b>31</b>
	<b>Annex C (informative) Calculation of the uncertainty associated with a concentration expressed on dry gas and at an oxygen reference concentration</b> .....	<b>33</b>
<b>C.1</b>	<b>Uncertainty associated with a concentration expressed on dry gas</b> .....	<b>33</b>
<b>C.2</b>	<b>Uncertainty associated with a concentration expressed at a oxygen reference concentration</b> .....	<b>35</b>
	<b>Annex D (informative) Example of assessment of compliance of non-dispersive infrared method for CO with requirements on emission measurements</b> .....	<b>37</b>
<b>D.1</b>	<b>General</b> .....	<b>37</b>
<b>D.2</b>	<b>Elements required for the uncertainty determinations</b> .....	<b>37</b>
<b>D.2.1</b>	<b>Model equation</b> .....	<b>37</b>
<b>D.2.2</b>	<b>Combined uncertainty</b> .....	<b>38</b>
<b>D.2.3</b>	<b>Expanded uncertainty</b> .....	<b>38</b>
<b>D.2.4</b>	<b>Determination of uncertainty contributions in case of rectangular distributions</b> .....	<b>39</b>
<b>D.2.5</b>	<b>Determination of uncertainty contributions by use of sensitivity coefficients</b> .....	<b>39</b>
<b>D.3</b>	<b>Example of an uncertainty calculation</b> .....	<b>40</b>
<b>D.3.1</b>	<b>Site specific conditions</b> .....	<b>40</b>
<b>D.3.2</b>	<b>Performance characteristics</b> .....	<b>41</b>
<b>D.3.3</b>	<b>Determination of the uncertainty contributions</b> .....	<b>42</b>
<b>D.3.4</b>	<b>Result of uncertainty calculation</b> .....	<b>45</b>
<b>D.3.4.1</b>	<b>Standard uncertainties</b> .....	<b>45</b>
<b>D.3.4.2</b>	<b>Combined uncertainty</b> .....	<b>46</b>
<b>D.3.4.3</b>	<b>Expanded uncertainty</b> .....	<b>46</b>
<b>D.3.4.4</b>	<b>Evaluation of the compliance with the required measurement quality</b> .....	<b>46</b>
	<b>Annex E (informative) Example of correction of data from drift effect</b> .....	<b>47</b>

<b>Annex F (informative) Significant technical changes.....</b>	<b>49</b>
<b>Bibliography .....</b>	<b>50</b>

## **European foreword**

This document (EN 15058:2017) has been prepared by Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

This document supersedes EN 15058:2006.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2017, and conflicting national standards shall be withdrawn at the latest by July 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

Annex F provides details of significant technical changes between this document and the previous edition.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## 1 Scope

This European Standard specifies the standard reference method (SRM) based on the infrared (IR) absorption principle. It includes the sampling and the gas conditioning system, and allows the determination of the carbon monoxide CO in flue gases emitted to the atmosphere from ducts and stacks.

This European Standard specifies the characteristics to be determined and the performance criteria to be fulfilled by portable automated measuring systems (P-AMS) using the IR measurement method. It applies for periodic monitoring and for the calibration or control of automated measuring systems (AMS) permanently installed on a stack, for regulatory or other purposes.

This European Standard specifies criteria for demonstration of equivalence of an alternative method (AM) to the SRM by application of EN 14793:2017.

This European Standard has been validated during field tests on waste incineration, co-incineration and large combustion plants and on a recognized test bench. It has been validated for CO concentrations with sampling periods of 30 min in the range of 0 mg/m<sup>3</sup> to 400 mg/m<sup>3</sup> for large combustion plants and 0 mg/m<sup>3</sup> to 740 mg/m<sup>3</sup> for waste and co-incineration. Directive 2010/75/EU lays down emission values which are expressed in mg/m<sup>3</sup>, on dry basis at a specified value of oxygen and at standard conditions (273 K and 101,3 kPa).

NOTE The characteristics of installations, the conditions during field tests and the values of repeatability and reproducibility in the field are given in Annex A.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 14793:2017, *Stationary source emission – Demonstration of equivalence of an alternative method with a reference method*

EN 15259:2007, *Air quality - Measurement of stationary source emissions - Requirements for measurement sections and sites and for the measurement objective, plan and report*

EN 15267-4:2017, *Air quality — Certification of automated measuring systems — Part 4: Performance criteria and test procedures for automated measuring systems for periodic measurements of emissions from stationary sources*

EN ISO 14956:2002, *Air quality - Evaluation of the suitability of a measurement procedure by comparison with a required measurement uncertainty (ISO 14956:2002)*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**