

STN	Zdravotnícke elektrické prístroje Osobitné požiadavky na základnú bezpečnosť a nevyhnutné prevádzkové vlastnosti zdravotníckych lôžok pre deti	STN EN 50637 36 4800
------------	---	--

Medical electrical equipment - Particular requirements for the basic safety and essential performance of medical beds for children

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 04/18

Obsahuje: EN 50637:2017

126630

EUROPEAN STANDARD

EN 50637

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2017

ICS 11.140

English Version

Medical electrical equipment - Particular requirements for the basic safety and essential performance of medical beds for children

Appareils électromédicaux - Exigences particulières de sécurité de base et de performances essentielles des lits médicaux pour enfants

Medizinische elektrische Geräte - Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von medizinischen Betten für Kinder

This European Standard was approved by CENELEC on 2017-08-30. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

EN 50637:2017 (E)**Contents**

Page

European foreword	5
Introduction	6
201.1 Scope, object and related standards	7
201.1.1 * Scope	7
201.1.2 Object	7
201.1.3 Collateral standards	7
201.1.4 Particular standards	8
201.2 Normative references	8
201.3 Terms and definitions	9
201.4 General requirements	14
201.4.2 *Risk management process for ME EQUIPMENT and ME SYSTEM	14
201.5 General requirements for testing of ME EQUIPMENT	14
201.5.101 * Entrapment test TOOLS	14
201.5.102 Loading pad	17
201.5.103 Impactor	18
201.5.104 Side impactor	19
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	19
201.6.2 *Protection against electrical shock	19
201.7 ME EQUIPMENT identification, marking and documents	19
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	28
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	28
201.9.1 *Mechanical HAZARDS of ME EQUIPMENT	29
201.9.2 HAZARDS associated with moving parts	36
201.9.3 HAZARDS associated with surfaces, corners and edges	40
201.9.8 HAZARDS associated with support systems	45
201.9.101 *Protection against inadvertent PATIENT falls	54
201.10 Protection against unwanted and excessive radiation HAZARDS	56
201.11 Protection against excessive temperatures and other HAZARDS	56
201.11.7 Biocompatibility of MEDICAL BED	59
201.11.8 *Interruption of the power supply/SUPPLY MAINS to ME EQUIPMENT	59
201.12 Accuracy of controls and instruments and protection against hazardous outputs	59
201.12.2 USABILITY	59
201.13 HAZARDOUS SITUATIONS and fault conditions	59
201.14 Programmable electrical medical systems (PEMS)	60
201.15 Construction of ME EQUIPMENT	60
201.15.3 Mechanical strength	60
201.15.4 ME EQUIPMENT components and general assembly	61
201.16 ME SYSTEMS	63
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	63
Annex AA (normative) Particular guidance and rationale	64
AA.1 General guidance	64
AA.1.1 Rationale for particular clauses and subclauses	64
Annex BB (normative) Design requirements and recommendations for MEDICAL BED	77
BB.1 General	77
BB.2 Human factors (ergonomic)	77
BB.2.1 General	77
BB.2.2 Range of adjustment of the height of the mattress support platform	77
BB.2.3 Dimensions for handles and pedals	78
BB.2.4 Operating forces for handles and pedals	78
BB.3 Functionality	78
BB.3.1 General	78
BB.3.2 Combination of MEDICAL BED and MOBILE hoists	78

BB.3.3	Combination of MEDICAL BED and mattress.....	79
BB.3.4	Range of adjustment of moveable sections of the MATTRESS SUPPORT PLATFORM.....	79
Annex CC	(informative) Particular guidance for assessing risk of entrapment in v-shaped openings.....	81
Bibliography		87
FIGURE 201.101	– MEDICAL BED, GENERAL ARRANGEMENT (EXAMPLE, SCHEMATIC PRESENTATION ONLY).....	13
FIGURE 201.102A	– CONE TOOL	15
FIGURE 201.102B	– FINGER PROBE FOR MESH.....	15
FIGURE 201.102C	– CYLINDER TOOL	15
FIGURE 201.102D	– BALL CHAIN	16
FIGURE 201.102E	– BALL CHAIN LOOP AND SPHERICAL MASS.....	16
FIGURE 201.102	– ENTRAPMENT TEST TOOLS.....	16
FIGURE 201.103	– LOADING PAD.....	17
FIGURE 201.104	– IMPACTOR	18
FIGURE 201.105	– SIDE IMPACTOR.....	19
FIGURE 201.106	– GRAPHICAL SYMBOL FOR MAXIMUM PATIENT WEIGHT AND SAFE WORKING LOAD	20
FIGURE 201.107	– MARKING FOR COMPATIBLE MATTRESSES SPECIFIED BY THE MANUFACTURER	21
FIGURE 201.108	– MARKING FOR DETACHABLE SIDE RAILS SPECIFIED BY THE MANUFACTURER	21
FIGURE 201.109	DESCRIPTION OF ALLOWED LENGTH OF PATIENT.....	22
FIGURE 201.110	EXAMPLE OF MARKING ON THE MEDICAL BED FOR POSITIONING PATIENT IN MEDICAL BED	23
FIGURE 201.111	– MEDICAL BED FUNCTION CONTROLS AND/OR ACTUATORS: GUIDELINES FOR CREATING GRAPHICAL SYMBOLS.....	25
FIGURE 201.112	MARKING ON THE MEDICAL BED OF STORAGE LOCATION FOR PENDANT CONTROL	26
FIGURE 201.113	– EXAMPLE OF MEDICAL BED WITH SEGMENTED OR SPLIT SIDE RAIL	30
FIGURE 201.114	– EXAMPLE OF MEDICAL BED WITH SINGLE PIECE SIDE RAIL AND PROTECTION SIDE RAIL	31
FIGURE 201.115	– DIMENSION OF HANDLE FOR LIFTING POLE.	35
FIGURE 201.116	– ALLOWABLE SPACING FOR FINGERS IN AREAS OF NORMAL REACH AROUND THE PERIMETER OF THE MATTRESS SUPPORT PLATFORM	37
FIGURE 201.117	– EXAMPLE USING BARRIERS FOR CLEARANCE MEASUREMENT AROUND THE PERIMETER OF THE MATTRESS SUPPORT PLATFORM TO MITIGATE PATIENT-FINGER ENTRAPMENT	38
FIGURE 201.118A	– FOOT AND TOE CLEARANCE AREA BETWEEN MOVING PARTS AND THE FLOOR.....	38
FIGURE 201.118B	– TOE CLEARANCE AREA BETWEEN MOVING PARTS AND THE FLOOR.....	39
FIGURE 201.118	– CLEARANCE AREAS	39
FIGURE 201.119	– REQUIRED MINIMUM RADII OF EDGES AND CORNERS	40
FIGURE 201.120	– RETENTION OF LOOP AND MASS.....	41
FIGURE 201.121	– LATERAL STABILITY TEST ALONG THE SIDE OF THE MEDICAL BED.....	42
FIGURE 201.122	– LONGITUDINAL STABILITY TEST WITH REMOVABLE FOOT BOARD	43
FIGURE 201.123	– LONGITUDINAL STABILITY TEST WITH FIXED HEAD/FOOT BOARDS.....	43
FIGURE 201.124	– DISTRIBUTION OF SAFE WORKING LOAD FOR TESTS	47
FIGURE 201.125	– POSITION OF LOADING PAD AND IMPACTOR	49
FIGURE 201.126	– IMPACT TO SLATS AND SOLID ELEMENTS OF MEDICAL BEDS.....	51
FIGURE 201.127	– APPLICATION OF FORCES FOR TEST OF SIDE RAIL.....	53
FIGURE 201.128	– HEIGHT OF SIDE RAIL / PROTECTION SIDE RAIL	55
FIGURE 201.129A	– ANGLE γ BETWEEN THE BACK SECTION AND THE LEG SECTION OF THE MATTRESS SUPPORT PLATFORM	61
FIGURE 201.129B	– ANGLE γ BETWEEN THE BACK SECTION AND THE UPPER LEG SECTION OF THE MATTRESS SUPPORT PLATFORM	61
FIGURE 201.129C	– ANGLE γ BETWEEN THE ANGLED BACK SECTION AND UPPER LEG SECTION OF THE MATTRESS SUPPORT PLATFORM	62
FIGURE 201.129D	– ANGLE γ BETWEEN THE ANGLED BACK SECTION AND THE LEG/UPPER LEG SECTION OF THE MATTRESS SUPPORT PLATFORM	62
FIGURE 201.129	– CONFIGURATIONS OF THE MATTRESS SUPPORT PLATFORM	62
FIGURE AA.1	– RESULTANT FORCES WITHOUT MATTRESS.....	68
FIGURE AA.2	– RESULTANT FORCES WITH MATTRESS.....	68

EN 50637:2017 (E)

FIGURE AA.3 – EXAMPLE OF 40 MM GAP MEASUREMENT OF B	68
FIGURE AA.4 – ANGLE MEASUREMENT EXAMPLE OF B	68
FIGURE AA.5 – PLACEMENT OF MEASUREMENT TOOL FOR MEASUREMENT OF D_x	69
FIGURE AA.6 – EXAMPLE OF AREA D_x MEASUREMENT THAT PASSES	70
FIGURE AA.7 – EXAMPLE OF AREA D_x MEASUREMENT THAT FAILS	70
FIGURE AA.8 – EXAMPLE OF AREA D_x MEASUREMENT THAT FAILS (ON LIMIT)	70
FIGURE AA.9 – EXAMPLE OF POTENTIAL PATIENT ENTRAPMENT IN AREA A WITHIN THE SIDE RAIL.....	71
FIGURE AA.10 – EXAMPLE OF POTENTIAL PATIENT ENTRAPMENT IN AREA A BELOW THE SIDE RAIL	71
FIGURE AA.11 – EXAMPLE OF POTENTIAL PATIENT ENTRAPMENT IN AREA B.....	71
FIGURE AA.12 – EXAMPLE OF POTENTIAL PATIENT ENTRAPMENT IN AREA C BETWEEN SPLIT SIDE RAIL	71
FIGURE AA.13 – EXAMPLE OF POTENTIAL PATIENT ENTRAPMENT IN AREA C BETWEEN SIDE RAIL AND HEAD BOARD ...	71
FIGURE AA.14 – EXAMPLE OF POTENTIAL PATIENT ENTRAPMENT IN AREA D_x	71
FIGURE AA.15 – EXAMPLE OF POTENTIAL PATIENT ENTRAPMENT IN AREA A BELOW A SINGLE PIECE SIDE RAIL	72
FIGURE BB.1 – SCHEMATIC PRESENTATION OF UNDER MEDICAL BED CLEARANCE	79
FIGURE BB.2 – RECOMMENDATIONS AND REQUIREMENTS REGARDING ANGLES FOR DIFFERENT SECTIONS OF THE MATTRESS SUPPORT PLATFORM.....	80
FIGURE CC.1 – WEDGE TOOL	82
FIGURE CC.2 – V-SHAPED OPENING IN RELATION TO B	83
FIGURE CC.3 – PASS/FAIL IN RELATION TO AREA B	83
FIGURE CC.4 – POSITIONING OF WEDGE TOOL	84
FIGURE CC.5 – PASS/FAIL IN RELATION TO AREA C BETWEEN HEAD BOARD AND FOOT BOARD	85
FIGURE CC.6 – PASS/FAIL IN RELATION TO AREA C BETWEEN SPLIT SIDE RAILS	86
*TABLE 201.101 – PROTECTION AGAINST PATIENT ENTRAPMENT.....	32
TABLE 201.102 – PROTECTION AGAINST INADVERTENT PATIENT FALLS AND CLIMBING OUT OF MEDICAL BED	56
TABLE 24 – ALLOWABLE MAXIMUM TEMPERATURES FOR SKIN CONTACT WITH MEDICAL BED APPLIED PARTS.....	57
TABLE AA.1 – PROTECTION AGAINST PATIENT ENTRAPMENT IN NON-MOVING PARTS	67
TABLE AA.2 – HEIGHT OF PROTECTIVE BARRIER ACCORDING TO AGE.....	75

European foreword

This document (*EN 50637:2017*) has been prepared by CLC/TC 62 "*Electrical equipment in medical practice*".

The following dates are fixed:

- latest date by which this document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2018-08-30
- latest date by which the national standards conflicting with this document have to be withdrawn (dow) 2020-08-30

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN and CENELEC by the European Commission.

EN 50637:2017 (E)**Introduction**

EN 60601-2-52 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL BEDS for ADULTS, hence not covering requirement for beds for CHILDREN and ADULTS with atypical anatomy. This particular standard is based on the Mandate M/467 Medical beds issued by the European Commission with the following background information:

It appears, from a first analysis undertaken by Competent Authorities, that the current set of standards is not adapted to the needs of CHILDREN or ADULTS with an atypical anatomy. EN 60601-2-52 does not foresee a maximum distance for the bars that is small enough to prevent accidents.

According to the Competent Authorities' representatives, a part of the safety problem is due to the fact that medical beds for ADULTS are not appropriately labelled as being designed only for ADULTS with a normal anatomy. Users are therefore not always aware of the risk of medical beds for young PATIENTS or for ADULTS with an atypical anatomy. Hospital administrations do not always see a need to buy medical beds which are appropriate for CHILDREN or for ADULTS with an atypical anatomy. Therefore, clear labelling of the targeted PATIENT groups for medical beds complying with EN 60601-2-52 could reduce the risk of inappropriate use of this kind of medical beds for CHILDREN or for ADULTS with an atypical anatomy. Labeling is taken care of by Amendment 1 of EN 60601-2-52.

Competent Authorities' representatives also stated that there is a need for the development of requirements for MEDICAL BEDS and COTS for CHILDREN and ADULTS with an atypical anatomy.

In order to prevent EN 60601-2-52 from being extraordinary complex to use, CLC/TC 62 decided to develop this particular standard rather than further amending EN 60601-2-52 in relation to use for CHILDREN and ADULTS with an atypical anatomy.

This standard is based on input from the following standards and reports:

- EN 60601-2-52, *Medical electrical equipment — Part 2-52: Particular requirements for basic safety and essential performance of medical beds (IEC 60601-2-52)*¹⁾;
- EN 716-1, *Furniture — Children's cots and folding cots for domestic use — Part 1: Safety requirements*;
- EN 716-2, *Furniture — Children's cots and folding cots for domestic use — Part 2: Test methods*;
- EN 1130-2, *Furniture — Cribs and cradles for domestic use — Part 2: Test methods*;
- EN 747-1, *Furniture — Bunk beds and high beds — Part 1: Safety, strength and durability requirements*;
- EN 747-2, *Furniture — Bunk beds and high beds — Part 2: Test methods*;
- CEN/TR 13387 (all parts), *Child use and care articles — General safety guidelines*;
- DIN 32623, *Hospital children's cots made from metal and plastic — Safety requirements and testing*;
- Nordic Requirements specification for Adjustable beds for disabled children.

1) This document is currently impacted by the amendment EN 60601-2-52:2010/A1:2015 and the corrigendum EN 60601-2-52:2010/AC:2011.

201.1 Scope, object and related standards

Clause 1 of EN 60601-1:2006, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*, applies, except as follows:

201.1.1 * Scope

Replacement:

This Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL BEDS, hereafter referred to as MEDICAL BEDS as defined in 201.3.218, intended for CHILDREN as defined in 201.3.207, and ADULTS with atypical anatomy (ADULTS ranging outside the definition for ADULTS in 201.3.201).

This standard applies to medical beds with nonadjustable and electrical / mechanical adjustable functions.

This Standard applies to MEDICAL BEDS with an internal length of up to 180 cm suitable to a body length of 155 cm.

NOTE 1 The limitation of 180 cm is in order to minimize the foreseeable misuse, of a parent sharing the bed with the child or that the bed will be used by an ADULT.

If a manufacturer wishes to make a bed that can be used by both a child and an ADULT, e.g. length of 180 cm or more, then it will fulfil both EN 60601-2-52 and this particular standard.

This Standard does not apply to MEDICAL BEDS intended for ADULTS as defined in 201.3.201 (covered by EN 60601-2-52).

This Standard does not apply to :

- incubators covered by EN 60601-2-19 ;
- beds for children, covered by EN 716-1 and EN 716-2 ;
- cribs and cradles covered by EN 1130 (all parts) ;
- bunk beds and high beds, covered by EN 747-1 and 747-2.

If a clause or subclause is specifically intended to be applicable to a MEDICAL BED only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to MEDICAL BEDS and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of MEDICAL BED or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of EN 60601-1:2006.

NOTE 2 See also 4.2 of EN 60601-1:2006.

NOTE 3 Body length is measured from crown to sole.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements and test methods for MEDICAL BEDS as defined in 201.3.218 intended for CHILDREN as defined in 201.3.207.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of EN 60601-1:2006 standard and its associated amendments and corrigenda and Clause 2 of this particular standard.

EN 60601-1-3 and EN 60601-1-10 do not apply. All other published collateral standards in the EN 60601-1 series apply as published.

EN 50637:2017 (E)**201.1.4 Particular standards***Replacement:*

In the EN 60601 series, particular standards may modify, replace or delete requirements contained in EN 60601-1:2006 and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of this particular standard takes priority over EN 60601-1:2006.

The numbering of clauses and subclauses of this particular standard corresponds to that of EN 60601-1:2006 with the prefix "201" (e.g. 201.1 in this standard addresses the content of Clause 1 of EN 60601-1:2006) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the EN 60601-1-2:2015 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the EN 60601-1-3:2008 collateral standard, etc.). The changes to the text of EN 60601-1:2006 are specified by the use of the following words:

"Replacement" means that the clause or subclause of EN 60601-1:2006 or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of EN 60601-1:2006 or applicable collateral standard.

"Amendment" means that the clause or subclause of EN 60601-1:2006 or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of EN 60601-1:2006 are numbered starting from 201.101. However, due to the fact that definitions in EN 60601-1:2006 are numbered 3.1 through 3.147, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for EN 60601-1-2, 203 for EN 60601-1-3, etc.

The term "this standard" is used to make reference to EN 60601-1:2006, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of EN 60601-1:2006 or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of EN 60601-1:2006 or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the Bibliography.

Clause 2 of EN 60601-1:2006, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*, applies except as follows:

Addition:

EN 71-3, *Safety of toys — Part 3: Migration of certain elements*

EN 716-2, *Furniture — Children's cots and folding cots for domestic use — Part 2: Test methods*

EN 1021-1, *Furniture — Assessment of the ignitability of upholstered furniture — Part 1: Ignition source smouldering cigarette*

EN 1021-2, *Furniture — Assessment of the ignitability of upholstered furniture — Part 2: Ignition source match flame equivalent*

EN 13501-1, *Fire classification of construction products and building elements — Part 1: Classification using data from reaction to fire tests*

EN 50525-2-21, *Electric cables — Low voltage energy cables of rated voltages up to and including 450/750 V (Uo/U) — Part 2-21: Cables for general applications — Flexible cables with crosslinked elastomeric insulation*

EN 60068-2-31, *Environmental testing — Part 2-31: Tests — Test Ec: Rough handling shocks, primarily for equipment-type specimens (IEC 60068-2-31)*

Replacement of the reference to ISO 10993 (all parts) (reference to the corresponding EN ISO series):

EN ISO 10993 (all parts), *Biological evaluation of medical devices (ISO 10993, all parts)*

Replacement of the references to IEC 60227-1:1993 and IEC 60245-1:2003 (not to be dated anymore):

IEC 60227-1, *Polyvinyl chloride insulated cables of rated voltages up to and including 450/750 V — Part 1: General requirements*

IEC 60245-1, *Rubber insulated cables — Rated voltages up to and including 450/750 V — Part 1: General requirements*

Replacement of the references to EN 60601-1-2 and EN 60601-1-3 (now to be dated):

EN 60601-1-2:2015, *Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral Standard: Electromagnetic disturbances — Requirements and tests (IEC 60601-1-2:2014)*

EN 60601-1-3:2008, *Medical electrical equipment — Part 1-3: General requirements for basic safety and essential performance — Collateral Standard: Radiation protection in diagnostic X-ray equipment (IEC 60601-1-3:2008)*

Deletion:

ISO 9614-1, *Acoustics — Determination of sound power levels of noise sources using sound intensity — Part 1: Measurement at discrete points*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN