Stomatológia Laserové zváranie a plniace materiály (ISO 28319: 2018) STN EN ISO 28319 85 6357

Dentistry - Laser welding and filler materials (ISO 28319:2018)

Táto norma obsahuje anglickú verziu európskej normy. This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 10/18

Obsahuje: EN ISO 28319:2018, ISO 28319:2018

Oznámením tejto normy sa ruší STN EN ISO 28319 (85 6357) z novembra 2010

EUROPEAN STANDARD

EN ISO 28319

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2018

ICS 11.060.10

Supersedes EN ISO 28319:2010

English Version

Dentistry - Laser welding and filler materials (ISO 28319:2018)

Médecine bucco-dentaire - Soudage par laser et matériaux d'apport (ISO 28319:2018)

Zahnheilkunde - Laserschweißen (ISO 28319:2018)

This European Standard was approved by CEN on 15 March 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 28319:2018 (E)

Contents	Page
European foreword	3

European foreword

This document (EN ISO 28319:2018) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2018, and conflicting national standards shall be withdrawn at the latest by November 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 28319:2010.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 28319:2018 has been approved by CEN as EN ISO 28319:2018 without any modification.

INTERNATIONAL STANDARD

ISO 28319

Second edition 2018-04

Dentistry — Laser welding and filler materials

Médecine bucco-dentaire — Soudage par laser et matériaux d'apport



ISO 28319:2018(E)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Fax: +41 22 749 09 47 Email: copyright@iso.org Website: www.iso.org

Published in Switzerland

Cor	Contents Pa		
Fore	word		iv
1	Scop	De	1
2	Nori	mative references	1
3			
		ns and definitions	
4	Req i 4.1	uirements Chemical composition	
	4.1	4.1.1 Metallic materials to be joined	
		4.1.2 Filler material	
		4.1.3 Hazardous elements in filler material	
	4.2	Biocompatibility	
	4.3 4.4	Mechanical strength of laser welded joint (tensile strength)	
	4.4	4.4.1 Static immersion test	
		4.4.2 Appearance after corrosion exposure	
	4.5	Laser welding process	
5	Sam	pling	3
6	Preparation of specimens		3
	6.1	General	
	6.2	Specimens for tensile testing	
		6.2.1 General	
	6.3	6.2.2 Procedure Specimens for corrosion testing	
	0.5	6.3.1 General	
		6.3.2 Seam geometry	
7	Mea	surement and test methods	6
-	7.1	Visual inspection	
	7.2	Chemical composition	
	7.3	Tensile testing	
		7.3.1 General	
		7.3.2 Evaluation of tensile testing results 7.3.3 Calculation of tensile strength	
	7.4	Corrosion resistance by static immersion test	
		7.4.1 Reagents	7
		7.4.2 Apparatus	
		7.4.3 Test solution	
		7.4.5 Analysis	
		7.4.6 Microscopic inspection	
		7.4.7 Report	
8	Inst	ruction for use	8
9	Mar	king and labelling	8
	9.1	Marking	
	9.2	Labelling	
10		report	
Anne	ex A (in	nformative) Laser welding process	10
Ribli	iogranl	hv	13

ISO 28319:2018(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 2, *Prosthodontic materials*.

This second edition cancels and replaces the first edition (ISO 28319:2010), which has been technically revised.

The main changes compared to the previous edition are as follows:

- a) reference to the corrosion standard ISO 10271:2011, for corrosion test methods and measurements has been added;
- b) a corrosion limit for the static corrosion test has been specified;
- c) Annex A has been revised in order to describe the laser welding process.

Dentistry — Laser welding and filler materials

1 Scope

This document specifies requirements and test methods for laser welding and the filler materials thereto used in the dental laboratory for welding of metallic restorations and appliances.

For filler materials used in laser welding, this document also specifies the information given in the instructions for use, marking and labelling.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, Dentistry — Vocabulary

ISO 6344-1, Coated abrasives — Grain size analysis — Part 1: Grain size distribution test

ISO 10271:2011, Dentistry — Corrosion test methods for metallic materials

ISO 15223-1:2016, Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements

ISO 22674:2016, Dentistry — Metallic materials for fixed and removable restorations and appliances

koniec náhľadu – text ďalej pokračuje v platenej verzii STN