

STN	Sterilné drenážne katétre a príslušenstvo na jednorazové použitie (ISO 20697: 2018)	STN EN ISO 20697 85 5828
------------	--	--

Sterile drainage catheters and accessory devices for single use (ISO 20697:2018)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 12/18

Obsahuje: EN ISO 20697:2018, ISO 20697:2018

Oznámením tejto normy sa ruší
STN EN 1617 (85 5828) z augusta 2001

127495

EUROPEAN STANDARD

EN ISO 20697

NORME EUROPÉENNE

EUROPÄISCHE NORM

July 2018

ICS 11.040.25

Supersedes EN 1617:1997

English Version

Sterile drainage catheters and accessory devices for single use (ISO 20697:2018)

Sondes et dispositifs auxiliaires stériles de drainage
non réutilisables (ISO 20697:2018)

Sterile Drainagekatheter und Zubehör zur einmaligen
Verwendung (ISO 20697:2018)

This European Standard was approved by CEN on 3 May 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 20697:2018 (E)

Contents	Page
European foreword.....	3

European foreword

This document (EN ISO 20697:2018) has been prepared by Technical Committee ISO/TC 84 " Devices for administration of medicinal products and catheters " in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2019, and conflicting national standards shall be withdrawn at the latest by January 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 1617:1997.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 20697:2018 has been approved by CEN as EN ISO 20697:2018 without any modification.

INTERNATIONAL STANDARD

ISO
20697

First edition
2018-06

Corrected version
2018-09

Sterile drainage catheters and accessory devices for single use

Sondes et dispositifs auxiliaires stériles de drainage non réutilisables



Reference number
ISO 20697:2018(E)

© ISO 2018

ISO 20697:2018(E)**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Intended performance	3
5 General requirements	3
5.1 Risk management	3
5.2 Biocompatibility	4
5.3 Detectability	4
5.4 Surface finish	4
5.5 Size designation	4
5.5.1 General	4
5.5.2 Outer diameter	4
5.5.3 Effective length	4
5.5.4 Nominal balloon inflation volume	5
5.6 Connector	5
5.7 MRI compatibility	6
5.8 Sterilization	6
6 Specific requirements	6
6.1 Kink stability	6
6.2 Corrosion resistance	6
6.3 Resistance to deformation	6
6.4 Peak tensile force	6
6.4.1 Connections	6
6.4.2 Drainage catheters and other accessory devices	7
6.5 Impact resistance	7
6.6 Flow rate	7
6.7 Retention strength	7
6.8 Balloon safety	7
6.9 Drainage catheter inflation lumen integrity and volume maintenance	8
6.9.1 General	8
6.9.2 Compliant balloon	8
6.9.3 Non-compliant balloon	8
6.10 Inflated balloon resistance to traction	8
6.11 Freedom from leakage during aspiration or vacuum	8
7 Information supplied by the manufacturer	9
7.1 General	9
7.2 Marking on the device and/or packaging	9
7.3 Instructions for use	9
Annex A (informative) Test method for determining kink stability	11
Annex B (normative) Test method for corrosion resistance	13
Annex C (normative) Test method for resistance to deformation by suction	14
Annex D (normative) Test method for determining peak tensile force of connections	15
Annex E (normative) Test method for determining peak tensile force of drainage catheter	16
Annex F (normative) Test method for impact resistance of collection device	18
Annex G (normative) Test method for determination of flow rate through drainage catheter	20
Annex H (informative) Test method for retention strength	22

ISO 20697:2018(E)

Annex I (normative) Test method for determining balloon safety	24
Annex J (normative) Test method for determining inflation lumen leakage and/or function and/or balloon deflation (drainage catheter with compliant balloon)	27
Annex K (normative) Test method for determining balloon size and deflation reliability (drainage catheter with non-compliant balloon)	29
Annex L (normative) Test method for determining inflated balloon resistance to traction	31
Annex M (normative) Test method for resistance to leakage during aspiration or vacuum	35
Bibliography	37

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 84, *Devices for administration of medicinal products and catheters*.

This document is based on EN 1617, *Sterile drainage catheters for single use*.

This corrected version of ISO 20697:2018 incorporates the following corrections:

- correction of measurements and units in [6.9.3](#), [6.10](#), [H.2.2 a](#)) and [L.2.1](#);
- deletion of EN 980 from Bibliography;
- minor editorial changes.

ISO 20697:2018(E)**Introduction**

Guidance on transition periods for implementing the requirements of this document is given in ISO/TR 19244.

Sterile drainage catheters and accessory devices for single use

1 Scope

This document specifies requirements for sterile, single use drainage catheters, wound and fluid accumulation drainage systems, surgical drainage catheters and their components, where the catheter is placed in a body cavity or wound, surgically or percutaneously, for drainage of fluid or air to the exterior.

The drainage catheter is left to drain naturally or connected to a suction source for faster tissue granulation.

This document is not applicable to:

- a) suction catheters;
- b) tracheal catheters;
- c) urethral catheters;
NOTE See ISO 20696.
- d) ureteral stents, biliary stents, and other stents;
NOTE See ISO 14630 and ASTM F1828-97 for stents requirements.
- e) drainage catheters placed in digestive tracts percutaneously with gastrostomy technique;
- f) neuraxial catheters used for removal of cerebrospinal fluid;
NOTE See ISO 20698.
- g) enteral catheters used for removal of solutions or substances from the gastrointestinal tract;
NOTE See ISO 20695.
- h) coatings.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 14155, *Clinical investigation of medical devices for human subjects — Good clinical practice*

ISO 80369-1, *Small bore connectors for liquids and gases in healthcare applications — Part 1: General requirements*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN