

STN	Ultrazvukové domové plynomery	STN EN 14236 25 7866
------------	--------------------------------------	--

Ultrasonic domestic gas meters

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 04/19

Obsahuje: EN 14236:2018

Oznámením tejto normy sa ruší
STN EN 14236 (25 7866) z augusta 2007

128312

EUROPEAN STANDARD**EN 14236****NORME EUROPÉENNE****EUROPÄISCHE NORM**

October 2018

ICS 91.140.40

Supersedes EN 14236:2007

English Version

Ultrasonic domestic gas meters

Compteurs de gaz domestiques à ultrasons

Ultraschall-Haushaltsgaszähler

This European Standard was approved by CEN on 20 November 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents

Page

European foreword.....	5
1 Scope.....	6
2 Normative references.....	6
3 Terms, definitions and symbols.....	8
3.1 Terms and definitions.....	8
3.2 Symbols.....	11
4 Normal operating conditions.....	13
4.1 Flow range.....	13
4.2 Maximum working pressure.....	13
4.3 Temperature range.....	13
4.4 Range of gases.....	14
4.5 Orientation.....	14
5 Metrological performance.....	14
5.1 General.....	14
5.2 Mode comparison.....	15
5.3 Permissible errors.....	15
5.4 Gas — air relationship.....	16
5.5 Pressure absorption.....	17
5.6 Metrological stability.....	18
5.7 Immunity to contaminants in gas stream.....	18
5.8 Installation effects.....	20
5.9 Zero flow.....	20
5.10 Reverse flow.....	20
5.11 Low flow registration.....	21
5.12 High flow registration.....	21
5.13 Pulsed (unsteady) flow.....	21
5.14 Temperature sensitivity.....	22
6 Construction and materials.....	23
6.1 General.....	23
6.2 Robustness of meter case.....	23
6.3 Corrosion protection.....	32
6.4 Casework decorative finish.....	38
6.5 Ageing of non-metallic casework.....	38
6.6 Ageing of external surfaces of the meter, including index windows and adhesion of the index window.....	39
6.7 Protection against solar radiation.....	39
6.8 Resistance to external humidity.....	40
6.9 Flame retardance of external surfaces.....	41
6.10 Resistance to storage temperature range.....	41
6.11 Resistance to the effects of toluene/iso-octane vapour.....	41
6.12 Resistance to water vapour.....	44
6.13 Ageing.....	45
7 Optional features.....	46

7.1	Pressure measuring point.....	46
7.2	Resistance to high ambient temperature	46
7.3	Meters with temperature conversion	48
7.4	Ancillary devices (if fitted)	48
7.5	Use in hazardous zones.....	49
8	Index	49
8.1	Recording and storage.....	49
8.2	Display.....	49
8.3	Segmental display	49
8.4	Non-volatile memory	50
8.5	Display reset	51
9	Marking.....	51
9.1	All meters	51
9.2	Two-pipe meters	52
9.3	Durability and legibility of marking.....	52
9.4	Accompanying information	53
10	Software.....	54
10.1	Requirements	54
10.2	Test.....	54
11	Communications.....	54
11.1	General.....	54
11.2	Character transmission.....	54
11.3	Communications protocol.....	55
11.4	Data	55
11.5	Test-mode	56
11.6	Data optical port.....	58
11.7	Galvanic port (optional)	58
11.8	Diagnostics	58
12	Battery.....	59
12.1	General.....	59
12.2	Voltage interruptions	59
12.3	Minimum operating voltage	60
12.4	Battery life	60
13	Immunity to electromagnetic disturbances	60
13.1	General.....	60
13.2	Electrostatic discharge.....	60
13.3	Radio frequency electromagnetic field	61
13.4	Electromagnetic induction (power frequency).....	61
13.5	Electromagnetic induction (pulsed field).....	62
13.6	Radio interference suppression	62
14	Ultrasonic (acoustic) noise interference	62
14.1	Requirements	62
14.2	Test.....	63
15	Meters supplied for testing.....	63
Annex A	(informative) Test gases	66
A.1	General.....	66
A.2	Test gas properties.....	66

EN 14236:2018 (E)

Annex B (normative) Production requirements for gas meters	67
B.1 Specification	67
B.2 Technical requirements.....	67
B.3 Certificates of conformity.....	67
Annex C (normative) Meters with gas temperature conversion devices.....	69
C.1 Scope	69
C.2 Metrological performance	69
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 2014/32/EU Measuring Instruments Directive aimed to be covered.....	73
Bibliography.....	79

European foreword

This document (EN 14236:2018) has been prepared by Technical Committee CEN/TC 237 “Gas meters”, the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2019 and conflicting national standards shall be withdrawn at the latest by April 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 14236:2007.

Changes from previous editions include:

- conformity with the MID 2014/32/EU regarding declared errors of the same sign.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive 2014/32/EU.

For relationship with EU Directive 2014/32/EU, see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

EN 14236:2018 (E)

1 Scope

This European Standard specifies requirements and tests for the construction, performance and safety of class 1,0 and class 1,5 battery powered ultrasonic gas meters (hereinafter referred to as meters), having co-axial single pipe, or two pipe connections, used to measure volumes of distributed fuel gases of the second and/or third family, as given in EN 437, at maximum working pressures not exceeding 0,5 bar¹⁾ and maximum actual flow rates of up to 10 m³/h over a minimum ambient temperature range of -10 °C to +40 °C, and minimum gas temperature span of 40 K, for domestic applications. This European Standard applies to meters where the measuring element and the register(s) are enclosed in the same case.

This European Standard applies to meters with and without built-in temperature conversion, that are installed in locations with vibration and shocks of low significance and in

- closed locations (indoor or outdoor with protection as specified by the manufacturer) with condensing or with non-condensing humidity

or, if specified by the manufacturer,

- open locations (outdoor without any covering) with condensing humidity or with non-condensing humidity

and in locations with electromagnetic disturbances.

Unless otherwise stated, all pressures given in this European Standard are gauge pressures.

When more than one meter type is submitted for testing, then each meter type is required to be tested against this European Standard.

Clauses 1 to 15 and Annex C are for design and type testing only.

NOTE See Annex A for production requirements.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 55032, *Electromagnetic compatibility of multimedia equipment - Emission Requirements (CISPR 32)*

EN 60068-2-5, *Environmental testing - Part 2: Tests - Test Sa: Simulated solar radiation at ground level (IEC 60068-2-5)*

EN 60068-2-30, *Environmental testing - Part 2-30: Tests - Test Db: Damp heat, cyclic (12 h + 12 h cycle) (IEC 60068-2-30)*

EN 60079-0, *Explosive atmospheres - Part 0: Equipment - General requirements (IEC 60079-0)*

EN 60079-10, *(all parts), Explosive atmospheres - Part 10: Classification of areas (IEC 60079-10, all parts)*

1) 1 bar = 1 000 mbar = 10⁵ Pa.

EN 60079-11, *Explosive atmospheres - Part 11: Equipment protection by intrinsic safety "i" (IEC 60079-11)*

EN 60079-15, *Explosive atmospheres - Part 15: Equipment protection by type of protection "n" (IEC 60079-15)*

EN 60086-1, *Primary batteries - Part 1: General (IEC 60086-1)*

EN 60086-4, *Primary batteries - Part 4: Safety of lithium batteries (IEC 60086-4)*

EN 60529, *Degrees of protection provided by enclosures (IP Code) (IEC 60529)*

EN 60695-11-5, *Fire hazard testing - Part 11-5: Test flames - Needle-flame test method - Apparatus, confirmatory test arrangement and guidance (IEC 60695-11-5)*

EN 60695-11-10, *Fire hazard testing - Part 11-10: Test flames - 50 W horizontal and vertical flame test methods (IEC 60695-11-10)*

EN 60730-1:2000, *Automatic electrical controls for household and similar use - Part 1: General requirements (IEC 60730-1:1999)*

EN 61000-4-2, *Electromagnetic compatibility (EMC) - Part 4-2: Testing and measurement techniques - Electrostatic discharge immunity test (IEC 61000-4-2)*

EN 61000-4-3, *Electromagnetic compatibility (EMC) - Part 4-3: Testing and measurement techniques - Radiated, radio-frequency, electromagnetic field immunity test (IEC 61000-4-3)*

EN 61000-4-8, *Electromagnetic compatibility (EMC) - Part 4-8: Testing and measurement techniques - Power frequency magnetic field immunity test (IEC 61000-4-8)*

EN 61000-4-9, *Electromagnetic compatibility (EMC) - Part 4-9: Testing and measurement techniques — Impulse magnetic field immunity test (IEC 61000-4-9)*

EN 61000-6-1, *Electromagnetic compatibility (EMC) - Part 6-1: Generic standards - Immunity for residential, commercial and light-industrial environments (IEC 61000-6-1)*

EN 61000-6-2, *Electromagnetic compatibility (EMC) - Part 6-2: Generic standards - Immunity for industrial environments (IEC 61000-6-2)*

EN 62056-21:2002, *Electricity metering - Data exchange for meter reading, tariff and load control - Part 21: Direct local data exchange (IEC 62056-21:2002)*

EN ISO 1518 (all parts), *Paints and varnishes - Determination of scratch resistance (ISO 1518)*

EN ISO 2409, *Paints and varnishes - Cross-cut test (ISO 2409)*

EN ISO 2812-1:2017, *Paints and varnishes - Determination of resistance to liquids - Part 1: Immersion in liquids other than water (ISO 2812-1:2017)*

EN ISO 4628-2:2016, *Paints and varnishes - Evaluation of degradation of coatings - Designation of quantity and size of defects, and of intensity of uniform changes in appearance - Part 2: Assessment of degree of blistering (ISO 4628-2:2016)*

EN 14236:2018 (E)

EN ISO 4628-3:2016, *Paints and varnishes - Evaluation of degradation of coatings - Designation of quantity and size of defects, and of intensity of uniform changes in appearance - Part 3: Assessment of degree of rusting (ISO 4628-3:2016)*

EN ISO 4892-3, *Plastics - Methods of exposure to laboratory light sources - Part 3: Fluorescent UV lamps (ISO 4892-3)*

EN ISO 6270-1, *Paints and varnishes - Determination of resistance to humidity - Part 1: Condensation (single-sided exposure) (ISO 6270-1)*

EN ISO 6272-1, *Paints and varnishes - Rapid-deformation (impact resistance) tests - Part 1: Falling-weight test, large-area indenter (ISO 6272-1)*

EN ISO 9227, *Corrosion tests in artificial atmospheres - Salt spray tests (ISO 9227)*

ISO 834-1, *Fire-resistance tests - Elements of building construction - Part 1: General requirements*

ISO 7724-3, *Paints and varnishes - Colorimetry - Part 3: Calculation of colour differences*

ASTM D471, *Standard Test Method for Rubber Property - Effect of Liquids*

ASTM D1003, *Standard Test Method for Haze and Luminous Transmittance of Transparent Plastics*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN