

Molekulárno-diagnostické vyšetrenia in vitro Špecifikácie postupov pred vyšetrením zmrazeného tkaniva Časť 1: Izolovaná RNA (ISO 20184-1: 2018)

STN EN ISO 20184-1

85 6573

Molecular in vitro diagnostic examinations - Specifications for pre-examination processes for frozen tissue - Part 1: Isolated RNA (ISO 20184-1:2018)

Táto norma obsahuje anglickú verziu európskej normy. This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 05/19

Obsahuje: EN ISO 20184-1:2018, ISO 20184-1:2018

Oznámením tejto normy sa ruší STN P CEN/TS 16826-1 (85 6573) z januára 2016

128343

EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

EN ISO 20184-1

December 2018

ICS 11.100.10

Supersedes CEN/TS 16826-1:2015

English Version

Molecular in vitro diagnostic examinations - Specifications for pre-examination processes for frozen tissue - Part 1: Isolated RNA (ISO 20184-1:2018)

Analyses de diagnostic moléculaire in vitro -Spécifications relatives aux processus préanalytiques pour les tissus congelés - Partie 1: ARN extrait (ISO 20184-1:2018) Molekularanalytische in-vitro-diagnostische Verfahren - Spezifikationen für präanalytische Prozesse für schockgefrorene Gewebeproben - Teil 1: Isolierte RNS (ISO 20184-1:2018)

This European Standard was approved by CEN on 30 November 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 20184-1:2018 (E)

Contents	Page
_	
European foreword	3

European foreword

This document (EN ISO 20184-1:2018) has been prepared by Technical Committee ISO/TC 212 "Clinical laboratory testing and in vitro diagnostic test systems" in collaboration with Technical Committee CEN/TC 140 "In vitro diagnostic medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2019, and conflicting national standards shall be withdrawn at the latest by June 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN/TS 16826-1:2015.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 20184-1:2018 has been approved by CEN as EN ISO 20184-1:2018 without any modification.

INTERNATIONAL STANDARD

ISO 20184-1

First edition 2018-11

Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for frozen tissue —

Part 1: **Isolated RNA**

Analyses de diagnostic moléculaire in vitro — Spécifications relatives aux processus préanalytiques pour les tissus congelés —

Partie 1: ARN extrait



ISO 20184-1:2018(E)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Fax: +41 22 749 09 47 Email: copyright@iso.org Website: www.iso.org

Website: www.iso.org
Published in Switzerland

ISO 20184-1:2018(E)

Con	itents		Page
Foreword			
Intro	duction		v
1	Scope		1
2	-	ative references	
3		s and definitions	
4	Genei	al considerations	5
5	Outsi 5.1	Specimen collection 5.1.1 General 5.1.2 Information about the specimen donor/patient 5.1.3 Information about the specimen 5.1.4 Specimen processing Fresh tissue transport requirements 5.2.1 General 5.2.2 Preparations for the transport 5.2.3 During transport	5 6 6 6 7
6	6.1	the laboratory	
	6.2 6.3 6.4 6.5	Evaluation of the pathology of the specimen and selection of the sample(s) Freezing of the specimen or sample(s) Storage requirements Isolation of RNA 6.5.1 General	
		6.5.2 Requirements and recommendations6.5.3 Using commercial kits6.5.4 Using the laboratories own protocols	12
	6.6	Quantity and quality assessment of isolated RNA	12
	6.7	Storage of isolated RNA	12 13
Anne		ormative) Impact of pre-examination variables on RNA profiles obtained from I liver tissue samples collected during and after routine surgery	14
Bibli	ography	7	18

ISO 20184-1:2018(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*.

A list of all parts in the ISO 20184 can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Molecular in vitro diagnostics, including molecular pathology, has enabled a significant progress in medicine. Further progress is expected with new technologies analysing nucleic acids, proteins, and metabolites in human tissues and body fluids. However, the profiles and/or integrity of these molecules can change drastically during specimen collection, transport, storage, and processing thus making the outcome from diagnostics or research unreliable or even impossible because the subsequent examination assay will not determine the situation in the patient but an artificial profile generated during the pre-examination process.

Therefore, a standardization of the entire process from specimen collection to the RNA examination is needed. Studies have been undertaken to determine the important influencing factors. This document draws upon such work to codify and standardize the steps for frozen tissue with regard to RNA examination in what is referred to as the pre-examination phase.

RNA profiles in tissues can change drastically before, during and after collection (due to e.g. gene induction or gene down regulation). RNA species can change differently in different donor's patients' tissues.

Therefore, it is essential to take special measures to minimize the described RNA profile changes and modifications within the tissue for subsequent examination.

In this document, the following verbal forms are used:

- "shall" indicates a requirement;
- "should" indicates a recommendation;
- "may" indicates a permission;
- "can" indicates a possibility or a capability.

Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for frozen tissue —

Part 1: **Isolated RNA**

1 Scope

This document gives guidelines on the handling, documentation, storage and processing of frozen tissue specimens intended for RNA examination during the pre-examination phase before a molecular assay is performed.

This document is applicable to any molecular in vitro diagnostic examination performed by medical laboratories and molecular pathology laboratories that evaluate RNA extracted from frozen tissue. It is also intended to be used by laboratory customers, in vitro diagnostics developers and manufacturers, biobanks, institutions and commercial organisations performing biomedical research, and regulatory authorities.

Tissues that have undergone chemical stabilization pre-treatment before freezing are not covered in this document.

NOTE International, national or regional regulations or requirements can also apply to specific topics covered in this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 15189:2012, Medical laboratories — Requirements for quality and competence

koniec náhľadu – text ďalej pokračuje v platenej verzii STN