

<b>STN</b>	<b>Monitorovanie koncentrácií aktivity rádioaktívnych látok vo vzduchu na pracovisku jadrových zariadení (ISO 16639: 2017)</b>	<b>STN EN ISO 16639</b>  40 1009
------------	--	--

Surveillance of the activity concentrations of airborne radioactive substances in the workplace of nuclear facilities (ISO 16639:2017)

Táto norma obsahuje anglickú verziu európskej normy.  
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 11/19

Obsahuje: EN ISO 16639:2019, ISO 16639:2017

**129884**

EUROPEAN STANDARD

EN ISO 16639

NORME EUROPÉENNE

EUROPÄISCHE NORM

June 2019

ICS 13.280

English Version

## Surveillance of the activity concentrations of airborne radioactive substances in the workplace of nuclear facilities (ISO 16639:2017)

Surveillance de l'activité volumique des substances radioactives dans l'air des lieux de travail des installations nucléaires (ISO 16639:2017)

Überwachung der Aktivitätskonzentrationen von luftgetragenen radioaktiven Substanzen an Arbeitsplätzen kerntechnischer Einrichtungen (ISO 16639:2017)

This European Standard was approved by CEN on 8 March 2019.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

**EN ISO 16639:2019 (E)**

<b>Contents</b>	<b>Page</b>
<b>European foreword.....</b>	<b>3</b>

## **European foreword**

The text of ISO 16639:2017 has been prepared by Technical Committee ISO/TC 85 "Nuclear energy, nuclear technologies, and radiological protection" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 16639:2019 by Technical Committee CEN/TC 430 "Nuclear energy, nuclear technologies, and radiological protection" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2019, and conflicting national standards shall be withdrawn at the latest by December 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## **Endorsement notice**

The text of ISO 16639:2017 has been approved by CEN as EN ISO 16639:2019 without any modification.

**INTERNATIONAL  
STANDARD**

**ISO  
16639**

First edition  
2017-01

---

---

**Surveillance of the activity  
concentrations of airborne radioactive  
substances in the workplace of  
nuclear facilities**

*Surveillance de l'activité volumique des substances radioactives dans  
l'air des lieux de travail des installations nucléaires*



Reference number  
ISO 16639:2017(E)

© ISO 2017

**ISO 16639:2017(E)****COPYRIGHT PROTECTED DOCUMENT**

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
copyright@iso.org  
www.iso.org

# Contents

	Page
<b>Foreword</b> .....	<b>v</b>
<b>Introduction</b> .....	<b>vi</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>2</b>
<b>3 Terms and definitions</b> .....	<b>2</b>
<b>4 Symbols</b> .....	<b>5</b>
<b>5 Developing the surveillance program</b> .....	<b>5</b>
5.1 Reasons for conducting a surveillance programme.....	5
5.1.1 General.....	5
5.1.2 Sampling when respiratory protective equipment is used.....	6
5.1.3 Sampling to establish air contamination areas.....	6
5.1.4 Air sampling as a basis for determining worker intakes.....	6
5.1.5 Air monitoring for early warning of elevated air concentrations.....	6
5.2 Graded approach to sampling.....	7
5.3 Frequency of sampling.....	8
5.3.1 General.....	8
5.3.2 Grab vs. continuous sampling.....	8
5.3.3 Continuous monitoring of activity concentrations.....	8
5.3.4 Prompt analysis of certain samples.....	9
5.4 Substitutes for air sampling.....	9
<b>6 Location of samplers and monitors</b> .....	<b>9</b>
6.1 General.....	9
6.2 Types of air flow studies.....	9
6.2.1 General.....	9
6.2.2 Qualitative airflow studies.....	9
6.2.3 Quantitative airflow studies.....	10
6.3 Location of samplers for estimating committed effective dose.....	10
6.4 Location of samplers for evaluating effectiveness of containment.....	11
6.5 Location of samplers for posting of air contamination areas.....	11
6.6 Location of portable samplers.....	12
6.7 Location of CAM for continuous monitoring of the activity concentration.....	12
<b>7 Collection of samples</b> .....	<b>12</b>
7.1 General.....	12
7.2 Sampling of aerosol particles.....	12
7.3 Gas Sampling.....	13
<b>8 Evaluation of sampling results</b> .....	<b>14</b>
8.1 Determining the average activity concentration.....	14
8.2 Uncertainty.....	14
8.3 Techniques for correcting for radon progeny interference.....	15
8.4 Evaluating changes in activity concentration over time.....	15
8.5 Review of sampling results.....	15
<b>9 Evaluating the effectiveness of the sampling program</b> .....	<b>16</b>
9.1 General.....	16
9.2 Dose-based assessment of the adequacy of the sampling program.....	16
<b>10 Quality assurance and quality control</b> .....	<b>17</b>
10.1 General.....	17
10.2 Sample identification, handling, and storage.....	17
10.3 Sampling and monitoring equipment.....	17
10.3.1 General.....	17
10.3.2 Performance of measuring instruments.....	18

**ISO 16639:2017(E)**

10.3.3	Air in-leakage testing.....	18
10.4	Documentation and record keeping.....	18
<b>Annex A</b>	<b>(informative) Examples for the determination of uncertainty, decision threshold and detection limit according to ISO 11929.....</b>	<b>20</b>
<b>Annex B</b>	<b>(informative) Correcting for the interference of radon progeny.....</b>	<b>27</b>
<b>Annex C</b>	<b>(informative) Normalized concentration and exposure.....</b>	<b>29</b>
<b>Annex D</b>	<b>(informative) Example applications of evaluating sampling program sensitivity from the viewpoint of potential missed exposure.....</b>	<b>30</b>
<b>Bibliography</b>	.....	<b>32</b>



## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

The committee responsible for this document is ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*.

## ISO 16639:2017(E)

### Introduction

Sampling of airborne radionuclides and monitoring of activity concentration in workplaces are critically important for maintaining worker safety at facilities where dispersible radioactive substances are used. Specifically, air sampling and monitoring are critical for evaluation of containment integrity, evaluation of effectiveness of contamination control programs and work practices, providing measurements for qualitative dose assessment, providing a general assessment of the level of the airborne hazard in a room, and for providing workers an immediate warning when the activity concentration exceeds safe levels.

This document sets forth guidelines and performance criteria for sampling airborne radioactive substances and monitoring activity concentration in the workplace of nuclear facilities. Emphasis is on health protection for workers in indoor environments. This document provides best practices and performance-based criteria for the use of sampling devices and systems, including delayed radioactivity measurement samplers and continuous air monitors. Specifically, this document covers air sampling program objectives, design of sampling and monitoring programs to meet program objectives, methods for air sampling and monitoring in the workplace, and quality assurance to ensure system performance toward protecting workers against unnecessary inhalation exposures. Taken together, these activities constitute the sampling or surveillance program.

The primary purpose of the surveillance of airborne activity concentrations in the workplace is to evaluate and mitigate inhalation hazards to workers in facilities where these may become airborne. Results often provide the basis for development and evaluation of control procedures and may indicate if engineering controls or operational changes are necessary.

The surveillance can consist of two general techniques. The first is retrospective sampling, in which constituents of the air are sampled, the collection medium is removed and taken to a radiation detector system and analysed for radioactive substances, and the activity concentration results made available at a later time. In this context, the measured activity concentrations are evaluated retrospectively. The second approach is real-time monitoring, in which activity concentrations are continuously monitored so that workers can be warned that a significant release of airborne activity may have occurred. In implementing an effective sampling program, it is important to achieve a proper balance between the two general approaches of the program. The specific balance depends on the hazard level of the work and the characteristics of each facility.

When designing a surveillance program, the optimization of worker protection minimizes internal and external exposures while balancing social, technical, economic, practical, and public policy considerations that are associated with the use of the radioactive substance.

A comprehensive surveillance program should also consider that the monitoring program is only one element of a comprehensive radiation protection program. Therefore, individuals involved with the monitoring program should interact with personnel working in the other elements of the radiation protection program, such as contamination control and internal dosimetry.

# Surveillance of the activity concentrations of airborne radioactive substances in the workplace of nuclear facilities

## 1 Scope

This document provides guidelines and performance criteria for sampling airborne radioactive substances in the workplace. Emphasis is on health protection of workers in the indoor environment.

This document provides best practices and performance-based criteria for the use of air sampling devices and systems, including retrospective samplers and continuous air monitors. Specifically, this document covers air sampling program objectives, design of air sampling and monitoring programs to meet program objectives, methods for air sampling and monitoring in the workplace, and quality assurance to ensure system performance toward protecting workers against unnecessary inhalation exposures.

The primary purpose of the surveillance of airborne activity concentrations in the workplace is to evaluate and mitigate inhalation hazards to workers in facilities where these can become airborne. A comprehensive surveillance program can be used to

- determine the effectiveness of administrative and engineering controls for confinement,
- measure activity concentrations of radioactive substances,
- alert workers to high activity concentrations in the air,
- aid in estimating worker intakes when bioassay methods are unavailable,
- determine signage or posting requirements for radiation protection, and
- determine appropriate protective equipment and measures.

Air sampling techniques consist of two general approaches. The first approach is retrospective sampling, in which the air is sampled, the collection medium is removed and taken to a radiation detector system and analysed for radioactive substance, and the concentration results made available at a later time. In this context, the measured air concentrations are evaluated retrospectively. The second approach is continuous real-time air monitoring so that workers can be warned that a significant release of airborne radioactivity may have just occurred. In implementing an effective air sampling program, it is important to achieve a balance between the two general approaches. The specific balance depends on hazard level of the work and the characteristics of each facility.

A special component of the second approach which can apply, if properly implemented, is the preparation of continuous air monitoring instrumentation and protocols. This enables radiation protection monitoring of personnel that have been trained and fitted with personal protective equipment (PPE) that permit pre-planned, defined, extended stay time in elevated concentrations of airborne radioactive substances. Such approaches can occur either as part of a planned re-entry of a contaminated area following an accidental loss of containment for accident assessment and recovery, or part of a project which involves systematic or routine access to radioactive substances (e.g. preparing process material containing easily aerosolized components), or handling objects such as poorly characterized waste materials that may contain radioactive contaminants that could be aerosolized when handled during repackaging. In this special case, the role of continuous air monitoring is to provide an alert to health physics personnel that the air concentrations of concern have exceeded a threshold such that the planned level of protection afforded by PPE has been or could be exceeded. This level would typically be many 10's or 100's of times higher than the derived air concentration (DAC) established for unprotected workers. The monitoring alarm or alert would therefore be designed not to be confused with the normal

**ISO 16639:2017(E)**

monitoring alarm, and the action taken in response would be similarly targeted at the specific site and personnel involved.

The air sampling strategy should be designed to minimize internal exposures and balanced with social, technical, economic, practical, and public policy considerations that are associated with the use of the radioactive substance.

A comprehensive air sampling strategy should also consider that the air sampling program is only one element of a broader radiation protection program. Therefore, individuals involved with the air sampling program should interact with personnel working in other elements of the radiation protection program, such as contamination control and internal dosimetry.

This document does not address outdoor air sampling, effluent monitoring, or radon measurements.

## **2 Normative references**

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11929, *Determination of the characteristic limits (decision threshold, detection limit and limits of the confidence interval) for measurements of ionizing radiation — Fundamentals and application*

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**