

<b>TNI</b>	<b>Prístroje na ochranu pred žiarením Určovanie neistoty pri meraní</b>	<b>TNI CLC IEC/TR 62461</b>  35 6606
------------	---	--

Radiation protection instrumentation - Determination of uncertainty in measurement

Táto technická normalizačná informácia obsahuje anglickú verziu CLC IEC/TR 62461:2019, IEC/TR 62461:2015.  
This Technical standard information includes the English version of CLC IEC/TR 62461:2019,  
IEC/TR 62461:2015.

Táto technická normalizačná informácia bola oznámená vo Vestníku ÚNMS SR č. 12/19

**130035**

TECHNICAL REPORT

**CLC IEC/TR 62461**

RAPPORT TECHNIQUE

TECHNISCHER BERICHT

May 2019

ICS 13.280

English Version

**Radiation protection instrumentation - Determination of  
uncertainty in measurement  
(IEC/TR 62461:2015)**

Instrumentation pour la radioprotection - Détermination de  
l'incertitude de mesure  
(IEC/TR 62461:2015)

Strahlenschutz-Messgeräte - Bestimmung der Unsicherheit  
beim Messen  
(IEC/TR 62461:2015)

This Technical Report was approved by CENELEC on 2019-05-20.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

**CLC IEC/TR 62461:2019 (E)****European foreword**

This document (CLC IEC/TR 62461:2019) consists of the text of IEC/TR 62461:2015 prepared by SC 45B "Radiation protection instrumentation" of IEC/TC 45 "Nuclear instrumentation".

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

**Endorsement notice**

The text of the International Standard IEC/TR 62461:2015 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 61526:2010	NOTE	Harmonized as EN 61526:2013 (modified)
IEC 60846-1:2009	NOTE	Harmonized as EN 60846-1:2014 (modified)
IEC 62387:2012	NOTE	Harmonized as EN 62387:2016 (modified)
IEC 61005:2003	NOTE	Harmonized as EN 61005:2004 (modified)
IEC 61577-2:2014	NOTE	Harmonized as EN 61577-2:2017 (modified)
IEC 61577-3:2011	NOTE	Harmonized as EN 61577-3:2014 (modified)
IEC 60325:2002	NOTE	Harmonized as EN 60325:2004 (modified)

## Annex ZA (normative)

### Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: [www.cenelec.eu](http://www.cenelec.eu).

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60050	series	International Electrotechnical - Vocabulary General Index		-
ISO/IEC Guide 98-3	2008	Uncertainty of measurement - Part 3: - Guide to the expression of uncertainty in measurement (GUM:1995)		-
ISO/IEC Guide 98-3 Supplement 1	2008	Uncertainty of measurement – Part 3: - Guide to the expression of uncertainty in measurement (GUM:1995) - Supplement 1: Propagation of distributions using a Monte Carlo method		-



IEC TR 62461

Edition 2.0 2015-01

# TECHNICAL REPORT



---

**Radiation protection instrumentation – Determination of uncertainty in measurement**



**THIS PUBLICATION IS COPYRIGHT PROTECTED****Copyright © 2015 IEC, Geneva, Switzerland**

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
Fax: +41 22 919 03 00  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

**About the IEC**

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

**About IEC publications**

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

**IEC Catalogue - [webstore.iec.ch/catalogue](http://webstore.iec.ch/catalogue)**

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

**IEC publications search - [www.iec.ch/searchpub](http://www.iec.ch/searchpub)**

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

**IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)**

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

**Electropedia - [www.electropedia.org](http://www.electropedia.org)**

The world's leading online dictionary of electronic and electrical terms containing more than 30 000 terms and definitions in English and French, with equivalent terms in 15 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

**IEC Glossary - [std.iec.ch/glossary](http://std.iec.ch/glossary)**

More than 60 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

**IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)**

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [csc@iec.ch](mailto:csc@iec.ch).



IEC TR 62461

Edition 2.0 2015-01

# TECHNICAL REPORT



---

## Radiation protection instrumentation – Determination of uncertainty in measurement

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

---

ICS 13.280

ISBN 978-2-8322-2216-4

**Warning! Make sure that you obtained this publication from an authorized distributor.**

## CONTENTS

FOREWORD .....	5
INTRODUCTION .....	7
1 Scope .....	8
2 Normative references .....	8
3 Terms and definitions .....	9
4 List of symbols .....	12
5 The GUM and the GUM S1 concept .....	14
5.1 General concept of uncertainty determination .....	14
5.1.1 Overview in four steps .....	14
5.1.2 Summary of the analytical method for steps 3 and 4 .....	15
5.1.3 Summary of the Monte Carlo method for steps 3 and 4 .....	15
5.1.4 Which method to use: Analytical or Monte Carlo? .....	16
5.2 Example of a model function .....	16
5.3 Collection of data and existing knowledge for the example .....	18
5.3.1 General .....	18
5.3.2 Calibration factor for the example .....	19
5.3.3 Zero reading for the example .....	20
5.3.4 Reading for the example .....	21
5.3.5 Relative response or correction factor for the example .....	21
5.3.6 Comparison of probability density distributions for input quantities .....	23
5.4 Calculation of the result of a measurement and its standard uncertainty (uncertainty budget) .....	25
5.4.1 General .....	25
5.4.2 Analytical method .....	25
5.4.3 Monte Carlo method .....	26
5.4.4 Uncertainty budgets .....	26
5.5 Statement of the measurement result and its expanded uncertainty .....	27
5.5.1 General .....	27
5.5.2 Analytical method .....	28
5.5.3 Monte Carlo method .....	28
5.5.4 Representation of the output distribution function in a simple form (Monte Carlo method) .....	31
6 Results below the decision threshold of the measuring device .....	31
7 Overview of the annexes .....	32
Annex A (informative) Example of an uncertainty analysis for a measurement with an electronic ambient dose equivalent rate meter according to IEC 60846-1:2009 .....	33
A.1 General .....	33
A.2 Model function .....	33
A.3 Calculation of the complete result of the measurement (measured value, probability density distribution, associated standard uncertainty, and the coverage interval) .....	34
A.3.1 General .....	34
A.3.2 Low level of consideration of measuring conditions .....	35
A.3.3 High level of consideration of measuring conditions .....	37
Annex B (informative) Example of an uncertainty analysis for a measurement with a passive integrating dosimetry system according to IEC 62387:2012 .....	40



B.1	General.....	40
B.2	Model function .....	40
B.3	Calculation of the complete result of the measurement (measured value, probability density distribution, associated standard uncertainty, and the coverage interval) .....	41
B.3.1	General .....	41
B.3.2	Low level of consideration of workplace conditions .....	41
B.3.3	High level of consideration of workplace conditions .....	43
Annex C (informative) Example of an uncertainty analysis for a measurement with an electronic direct reading neutron ambient dose equivalent meter according to IEC 61005:2003 .....		46
C.1	General.....	46
C.2	Model function .....	46
C.3	Calculation of the complete result of the measurement (measured value, probability density distribution, associated standard uncertainty, and the coverage interval) .....	47
C.3.1	General .....	47
C.3.2	Analytical method .....	47
C.3.3	Monte Carlo method .....	48
C.3.4	Comparison of the result of the analytical and the Monte Carlo method .....	49
Annex D (informative) Example of an uncertainty analysis for a calibration of radon activity monitor according to the IEC 61577 series .....		51
D.1	General.....	51
D.2	Model function .....	51
D.3	Calculation of the complete result of the measurement (measured value, probability density distribution, associated standard uncertainty, and the coverage interval) .....	51
Annex E (informative) Example of an uncertainty analysis for a measurement of surface emission rate with a contamination meter according to IEC 60325:2002 .....		54
E.1	General.....	54
E.2	Model function .....	54
E.3	Calculation of the complete result of the measurement (measured value, probability density distribution, associated standard uncertainty, and the coverage interval) .....	54
E.3.1	General .....	54
E.3.2	Effects of distance .....	55
E.3.3	Contamination non-uniformity .....	55
E.3.4	Surface absorption .....	56
E.3.5	Other influence quantities .....	56
E.3.6	Uncertainty budget .....	56
Bibliography.....		59
Figure 1 – Triangular probability density distribution of possible values $n$ for the calibration factor $N$ .....		20
Figure 2 – Rectangular probability density distribution of possible values $g_0$ for the zero reading $G_0$ .....		21
Figure 3 – Gaussian probability density distribution of possible values $g$ for the reading $G$ .....		21
Figure 4 – Comparison of different probability density distributions of possible values: rectangular (broken line), triangular (dotted line) and Gaussian (solid line) distribution .....		24
Figure 5 – Distribution function $Q$ of the measured value .....		29

Figure 6 – Probability density distribution (PDF) of the measured value .....	30
Figure C.1 – Results of the analytical (red dashed lines) and the Monte Carlo method (grey histogram and blue dotted and solid lines) for $\dot{H}^*(10)$ .....	50
Figure D.1 – Result of the analytical (red dashed lines) and the Monte Carlo method (grey histogram and blue dotted lines) for $K_T$ .....	53
Table 1 – Symbols (and abbreviated terms) used in the main text (excluding annexes).....	12
Table 2 – Standard uncertainty and method to compute the probability density distributions shown in Figure 4.....	24
Table 3 – Example of an uncertainty budget for a measurement with an electronic dosimeter using the model function $M = NK(G - G_0)$ and low level of consideration of the workplace conditions, see 5.3.5.2 .....	27
Table 4 – Example of an uncertainty budget for a measurement with an electronic dosimeter using the model function $M = NK(G - G_0)$ and high level of consideration of the workplace conditions, see 5.3.5.3 .....	27
Table A.1 – Example of an uncertainty budget for a dose rate measurement according to IEC 60846-1:2009 with an instrument having a logarithmic scale and low level of consideration of the measuring conditions, see text for details .....	36
Table A.2 – Example of an uncertainty budget for a dose rate measurement according to IEC 60846-1:2009 with an instrument having a logarithmic scale and high level of consideration of the measuring conditions, see text for details .....	38
Table B.1 – Example of an uncertainty budget for a photon dose measurement with a passive dosimetry system according to IEC 62387-1:2007 and low level of consideration of the workplace conditions, see text for details .....	42
Table B.2 – Example of an uncertainty budget for a photon dose measurement with a passive dosimetry system according to IEC 62387-1:2007 and high level of consideration of the measuring conditions, see text for details .....	44
Table C.1 – Example of an uncertainty budget for a neutron dose measurement according to IEC 61005:2003 using the analytical method.....	48
Table C.2 – Example of an uncertainty budget for a neutron dose rate measurement according to IEC 61005:2003 using the Monte Carlo method .....	49
Table C.3 – Results of the analytical and the Monte Carlo method .....	50
Table D.1 – List of quantities used in formula (D.1).....	51
Table D.2 – List of data available for the input quantities of formula (D.1).....	52
Table D.3 – Example of an uncertainty budget for the calibration of a radon monitor according to IEC 61577, see text for details .....	52
Table E.1 – Example of an uncertainty budget for a surface emission rate measurement according to IEC 60325:2002, see text for details .....	57
Table E.2 – Example of an uncertainty budget for a surface emission rate measurement according to IEC 60325:2002 for the determination of the uncertainty at a measured value of zero.....	58

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

---

### **RADIATION PROTECTION INSTRUMENTATION – DETERMINATION OF UNCERTAINTY IN MEASUREMENT**

#### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a technical report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC 62461, which is a technical report, has been prepared by subcommittee 45B: Radiation protection instrumentation, of IEC technical committee 45: Nuclear instrumentation.

This second edition of IEC TR 62461 cancels and replaces the first edition, published in 2006, and constitutes a technical revision. The main changes with respect to the previous edition are as follows:

- add to the analytical method for the determination of uncertainty the Monte Carlo method for the determination of uncertainty according to supplement 1 of the Guide to the Expression of uncertainty in measurement (GUM S1), and
- add a very simple method to judge whether a measured result is significantly different from zero or not based on ISO 11929.

The text of this technical report is based on the following documents:

Enquiry draft	Report on voting
45B/783/DTR	45B/813/RVD

Full information on the voting for the approval of this technical report can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

**IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.**

## INTRODUCTION

The ISO/IEC Guide 98-3:2008, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)* as well as its Supplement 1:2008, *Propagation of distributions using a Monte Carlo method (GUM S1)*, are general guides to assess the uncertainty in measurement. This Technical Report lays emphasis on their application in the area of radiation protection and serves as a practical introduction to the GUM and its supplement 1 (GUM S1).

The process of determining the uncertainty delivers not only a numerical value of the uncertainty; in addition it produces the best estimate of the quantity to be measured which may differ from the indication of the instrument. Thus, it can also improve the result of the measurement by using information beyond the indicated value of the instrument, e.g. the energy dependence of the instrument.

## RADIATION PROTECTION INSTRUMENTATION – DETERMINATION OF UNCERTAINTY IN MEASUREMENT

### 1 Scope

This Technical Report gives guidelines for the application of the uncertainty analysis according to ISO/IEC Guide 98-3:2008 (GUM describing an analytical method for the uncertainty determination) and its Supplement 1:2008 (GUM S1 describing a Monte Carlo method for the uncertainty determination) for measurements covered by standards of IEC Subcommittee 45B. It does not include the uncertainty associated with the concept of the measuring quantity, e. g., the difference between  $H_p(10)$  on the ISO water slab phantom and on the person.

This Technical Report explains the principles of the ISO/IEC Guide 98-3:2008 (GUM), its Supplement 1:2008 (GUM S1) and the special considerations necessary for radiation protection at an example taken from individual dosimetry of external radiation. In the informative annexes, several examples are given for the application on instruments, for which SC 45B has developed standards.

This Technical Report is supposed to assist the understanding of the ISO/IEC Guide 98-3:2008 (GUM), its Supplement 1: 2008 (GUM S1), and other papers on uncertainty analysis. It cannot replace these papers nor can it provide the background and justification of the arguments leading to the concept of the ISO/IEC Guide 98-3:2008 (GUM) and its Supplement 1:2008 (GUM S1).

Finally, this Technical Report gives a very simple method to judge whether a measured result is significantly different from zero or not based on ISO 11929.

For better readability the correct terms are not always used throughout this technical report. For example, instead of “random variables of a quantity” only the “quantity” itself is stated.

### 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050 (all parts): *International Electrotechnical Vocabulary* (available at <http://www.electropedia.org>)

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO/IEC Guide 98-3, Supplement 1:2008, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995) – Propagation of distributions using a Monte Carlo method*

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**