

| | | |
|------------|--|--------------------------------------|
| STN | Zdravotnícke elektrické prístroje Časť 2-28: Osobitné požiadavky na základnú bezpečnosť a nevyhnutné prevádzkové vlastnosti röntgenových žiaričov na lekársku diagnostiku | STN EN IEC 60601-2-28 |
| | | 36 4800 |

Medical electrical equipment - Part 2-28: Particular requirements for the basic safety and essential performance of X-ray tube assemblies for medical diagnosis

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 01/20

Obsahuje: EN IEC 60601-2-28:2019, IEC 60601-2-28:2017

Oznámením tejto normy sa od 07.08.2022 ruší
STN EN 60601-2-28 (36 4800) z januára 2011

130209

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN IEC 60601-2-28

September 2019

ICS 11.040.55

Supersedes EN 60601-2-28:2010 and all of its
amendments and corrigenda (if any)

English Version

**Medical electrical equipment - Part 2-28: Particular requirements
for the basic safety and essential performance of X-ray tube
assemblies for medical diagnosis
(IEC 60601-2-28:2017)**

Appareils électromédicaux - Partie 2-28 : Exigences particulières pour la sécurité de base et les performances essentielles des gaines équipées pour diagnostic médical
(IEC 60601-2-28:2017)

Medizinische elektrische Geräte - Teil 2-28: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Röntgenstrahlern für die medizinische Diagnostik
(IEC 60601-2-28:2017)

This European Standard was approved by CENELEC on 2019-08-07. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
 Comité Européen de Normalisation Electrotechnique
 Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN IEC 60601-2-28:2019 (E)**European foreword**

The text of document 62B/1040/FDIS, future edition 3 of IEC 60601-2-28, prepared by SC 62B "Diagnostic imaging equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 60601-2-28:2019.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2020-05-07
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2022-08-07

This document supersedes EN 60601-2-28:2010.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 60601-2-28:2017 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

| | | |
|------------------|------|--|
| IEC 60601-1-9 | NOTE | Harmonized as EN 60601-1-9 |
| IEC 60601-1-10 | NOTE | Harmonized as EN 60601-1-10 |
| IEC 61010-1:2010 | NOTE | Harmonized as EN 61010-1:2010 (not modified) |
| ISO 13732-1 | NOTE | Harmonized as EN ISO 13732-1 |

Annex ZA
(normative)

**Normative references to international publications
with their corresponding European publications**

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

The annex ZA of EN 60601-1:2006 applies, except as follows:

| <u>Publication Addition</u> | <u>Year</u> | <u>Title</u> | <u>EN/HD</u> | <u>Year</u> |
|---------------------------------|-------------|--|--|-------------|
| IEC 60336 | - | Medical electrical equipment - X-ray tube assemblies for medical diagnosis - Characteristics of focal spots | EN 60336 | - |
| IEC 60522 | - | Determination of the permanent filtration of X-ray tube assemblies | EN 60522 | - |
| IEC TR 60788 | 2004 | Medical electrical equipment - Glossary of defined terms | | - |
| <i>Replacement</i> | | | | |
| IEC 60601-1-3 | 2008 | Medical electrical equipment - Part 1-3: General requirements for basic safety and essential performance - Collateral Standard: Radiation protection in diagnostic X-ray equipment | EN 60601-1-3:2008/corrigendum Mar. 2010 +A11 | 2008 |
| | | | +EN 60601-1-2010 3:2008/corrigendum Mar. 2010 +A11 | 2016 |



IEC 60601-2-28

Edition 3.0 2017-06

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –

**Part 2-28: Particular requirements for the basic safety and essential performance
of X-ray tube assemblies for medical diagnosis**

Appareils électromédicaux –

**Partie 2-28: Exigences particulières pour la sécurité de base et les performances
essentielles des gaines équipées pour diagnostic médical**





THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2017 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
 3, rue de Varembé
 CH-1211 Geneva 20
 Switzerland

Tel.: +41 22 919 02 11
 Fax: +41 22 919 03 00
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

IEC Catalogue - webstore.iec.ch/catalogue

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

IEC publications search - www.iec.ch/searchpub

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing 20 000 terms and definitions in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

65 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: csc@iec.ch.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Catalogue IEC - webstore.iec.ch/catalogue

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

Recherche de publications IEC - www.iec.ch/searchpub

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalelement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

65 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: csc@iec.ch.



INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Medical electrical equipment –
Part 2-28: Particular requirements for the basic safety and essential performance
of X-ray tube assemblies for medical diagnosis**

**Appareils électromédicaux –
Partie 2-28: Exigences particulières pour la sécurité de base et les performances
essentielles des gaines équipées pour diagnostic médical**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

| | |
|---|----|
| FOREWORD | 3 |
| 201. 1 Scope, object and related standards | 5 |
| 201. 2 Normative references | 6 |
| 201. 3 Terms and definitions | 7 |
| 201. 4 General requirements | 7 |
| 201. 5 General requirements for testing ME EQUIPMENT | 8 |
| 201. 6 Classification of ME EQUIPMENT and ME SYSTEMS | 8 |
| 201. 7 ME EQUIPMENT identification, marking and documents | 9 |
| 201. 8 Protection against electrical HAZARDS from ME EQUIPMENT | 13 |
| 201. 9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS | 14 |
| 201. 10 Protection against unwanted and excessive RADIATION HAZARDS | 15 |
| 201. 11 Protection against excessive temperatures and other HAZARDS | 15 |
| 201. 12 Accuracy of controls and instruments and protection against hazardous outputs | 16 |
| 201. 13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT | 16 |
| 201. 14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS) | 17 |
| 201. 15 Construction of ME EQUIPMENT | 17 |
| 201. 16 ME SYSTEMS | 17 |
| 201. 17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS | 17 |
| 203 RADIATION protection in diagnostic X-RAY EQUIPMENT | 18 |
| Annexes | 20 |
| Annex AA (informative) Test of X-RAY TUBE ASSEMBLIES for expelled parts-related and/or tube implosion-related RISKS | 20 |
| Bibliography | 22 |
| Index of defined terms used in this particular standard | 23 |

INTERNATIONAL ELECTROTECHNICAL COMMISSION**MEDICAL ELECTRICAL EQUIPMENT –****Part 2-28: Particular requirements for the basic safety
and essential performance of X-ray tube assemblies for medical diagnosis****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-28 has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 2010. This edition constitutes a technical revision.

The third edition of this particular standard has been prepared to fit IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 (the amended third edition of IEC 60601-1), which is referred to as the general standard. Apart from the changes related to the amendment of IEC 60601-1, changes related to technical improvements are also included.

The text of this standard is based on the following documents:

| FDIS | Report on voting |
|---------------|------------------|
| 62B/1040/FDIS | 62B/1051/RVD |

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- requirements and definitions: roman type;
- *test specifications*: italic type;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-28: Particular requirements for the basic safety and essential performance of X-ray tube assemblies for medical diagnosis

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 Scope

Replacement:

This part of IEC 60601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of X-RAY TUBE ASSEMBLIES and to components thereof, intended for medical diagnosis and imaging.

Where the general standard IEC 60601-1 and the collateral standard IEC 60601-1-3 refer to ME EQUIPMENT, this is interpreted as X-RAY TUBE ASSEMBLIES in this particular standard. If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

NOTE This document is also applicable to the X-RAY TUBE ASSEMBLY aspects of X-RAY SOURCE ASSEMBLIES and X-RAY TUBE HEADS.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for X-RAY TUBE ASSEMBLIES for medical diagnosis.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-3:2008 and IEC 60601-1-3:2008/AMD1:2013 applies as modified in Clause 203. IEC 60601-1-2, IEC 60601-1-6, IEC 60601-1-8, IEC 60601-1-9, IEC 60601-1-10, IEC 60601-1-11 and IEC 60601-1-12 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

NOTE 101 IEC 60601-1-2 does not apply because RISKS for the X-RAY TUBE ASSEMBLY outside the system may only be indicative of RISKS for the system due to the difference in electromagnetic environment.

NOTE 102 IEC 60601-1-6 and IEC 60601-1-8 do not apply because X-RAY TUBE ASSEMBLIES are not operated as a stand-alone device.

NOTE 103 X-RAY TUBE ASSEMBLIES are not in the scope of IEC 60601-1-10, IEC 60601-1-11 and IEC 60601-1-12.

¹ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 are referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix “201” (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix “20x” where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are SPECIFIED by the use of the following words:

“*Replacement*” means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

“*Addition*” means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

“*Amendment*” means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However due to the fact that definitions in the general standard are numbered 3.1 through 3.147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term “this document” is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the Bibliography.

Clause 2 of the general standard applies, except as follows:

Replacement:

IEC 60601-1-3:2008, *Medical electrical equipment – Part 1-3: General requirements for basic safety and essential performance – Collateral Standard: Radiation protection in diagnostic X-ray equipment*

IEC 60601-1-3:2008/AMD1:2013

Addition:

IEC 60336, *Medical electrical equipment – X-ray tube assemblies for medical diagnosis – Characteristics of focal spots*

IEC 60522, *Determination of the permanent filtration of X-ray tube assemblies*

IEC 60613:2010, *Electrical and loading characteristics of X-ray tube assemblies for medical diagnosis*

IEC TR 60788:2004, *Medical electrical equipment – Glossary of defined terms*

koniec náhľadu – text d'alej pokračuje v platenej verzii STN