

<b>STN</b>	<b>Zdravotnícke elektrické prístroje Časť 2-59: Osobitné požiadavky na základnú bezpečnosť a nevyhnutné prevádzkové vlastnosti skriningových termografov na skrining ľudí s horúčkou</b>	<b>STN EN IEC 80601-2-59</b>  36 4800
------------	--	---

Medical electrical equipment - Part 2-59: Particular requirements for the basic safety and essential performance of screening thermographs for human febrile temperature screening

Táto norma obsahuje anglickú verziu európskej normy.  
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 01/20

Obsahuje: EN IEC 80601-2-59:2019, IEC 80601-2-59:2017

Oznámením tejto normy sa od 11.10.2022 ruší  
STN EN 80601-2-59 (36 4800) z júna 2010

**130213**

EUROPEAN STANDARD

**EN IEC 80601-2-59**

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2019

ICS 11.040.55

Supersedes EN 80601-2-59:2009 and all of its amendments and corrigenda (if any)

English Version

**Medical electrical equipment - Part 2-59: Particular requirements  
for the basic safety and essential performance of screening  
thermographs for human febrile temperature screening  
(IEC 80601-2-59:2017)**

Appareils électromédicaux - Partie 2-59: Exigences  
particulières pour la sécurité de base et les performances  
essentiels des imageurs thermiques pour le dépistage des  
humains fébriles  
(IEC 80601-2-59:2017)

Medizinische elektrische Geräte - Teil 2-59: Besondere  
Anforderungen für die Sicherheit einschließlich der  
wesentlichen Leistungsmerkmale von Wärmebildkameras  
für Reihenuntersuchungen von Menschen auf Fieber  
(IEC 80601-2-59:2017)

This European Standard was approved by CENELEC on 2017-10-24. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

**EN IEC 80601-2-59:2019 (E)****European foreword**

The text of document 62D/1501/FDIS, future edition 2 of IEC 80601-2-59, prepared by SC 62D "Electromedical equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 80601-2-59:2019.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2020-04-11
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2022-10-11

This document supersedes EN 80601-2-59:2009.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

**Endorsement notice**

The text of the International Standard IEC 80601-2-59:2017 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

ISO 80601-2-56	NOTE	Harmonized as EN ISO 80601-2-56
IEC 60601-1-10	NOTE	Harmonized as EN 60601-1-10

## Annex ZA (normative)

### Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: [www.cenelec.eu](http://www.cenelec.eu).

*The Annex ZA of EN 60601-1:2006 applies, except as follows:*

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<i>Replacement</i>				
IEC 60601-1-2	2014	Medical electrical equipment - Part 1-2:EN 60601-1-2 General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests	EN 60601-1-2	2015
IEC 60601-1-6	2010	Medical electrical equipment - Part 1-6:EN 60601-1-6 General requirements for basic safety and essential performance - Collateral standard: Usability	EN 60601-1-6	2010
IEC 60601-1-8	2006	Medical electrical equipment - Part 1-8:- General requirements for basic safety and essential performance - Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems		-
<i>Addition</i>				
IEC 60601-1	2005	Medical electrical equipment - Part 1:EN 60601-1 General requirements for basic safety and essential performance	EN 60601-1	2006
			+A12	2014
			+EN 60601-2010 1:2006/corrigendum Mar. 2010	
			+AC	2014
ISO/TR 13154	-	Medical electrical equipment --- Deployment, implementation and operational guidelines for indentifying febrile humans using a screening thermograph		-



IEC 80601-2-59

Edition 2.0 2017-09

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE



**Medical electrical equipment –  
Part 2-59: Particular requirements for the basic safety and essential performance  
of screening thermographs for human febrile temperature screening**

**Appareils électromédicaux –  
Partie 2-59: Exigences particulières pour la sécurité de base et les performances  
essentielles des imageurs thermiques pour le dépistage des humains fébriles**



## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2017 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
Fax: +41 22 919 03 00  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

#### IEC Catalogue - [webstore.iec.ch/catalogue](http://webstore.iec.ch/catalogue)

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

#### IEC publications search - [www.iec.ch/searchpub](http://www.iec.ch/searchpub)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary of electronic and electrical terms containing 20 000 terms and definitions in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### IEC Glossary - [std.iec.ch/glossary](http://std.iec.ch/glossary)

65 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [csc@iec.ch](mailto:csc@iec.ch).

### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Catalogue IEC - [webstore.iec.ch/catalogue](http://webstore.iec.ch/catalogue)

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

#### Recherche de publications IEC - [www.iec.ch/searchpub](http://www.iec.ch/searchpub)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

#### Glossaire IEC - [std.iec.ch/glossary](http://std.iec.ch/glossary)

65 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [csc@iec.ch](mailto:csc@iec.ch).



IEC 80601-2-59

Edition 2.0 2017-09

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE



**Medical electrical equipment –**

**Part 2-59: Particular requirements for the basic safety and essential performance of screening thermographs for human febrile temperature screening**

**Appareils électromédicaux –**

**Partie 2-59: Exigences particulières pour la sécurité de base et les performances essentielles des imageurs thermiques pour le dépistage des humains fébriles**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 11.040.55

ISBN 978-2-8322-4815-7

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards .....	8
201.2 Normative references.....	9
201.3 Terms and definitions.....	10
201.4 General requirements .....	12
201.5 General requirements for testing ME EQUIPMENT .....	13
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	13
201.7 ME EQUIPMENT identification, marking and documents .....	13
201.8 Protection against electrical HAZARDS from ME EQUIPMENT .....	14
201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS .....	14
201.10 Protection against unwanted and excessive radiation HAZARDS .....	15
201.11 Protection against excessive temperatures and other HAZARDS .....	15
201.12 Accuracy of controls and instruments and protection against hazardous outputs .....	15
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT .....	16
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	16
201.15 Construction of ME EQUIPMENT .....	16
201.16 ME SYSTEMS .....	16
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS .....	16
201.101 LABORATORY ACCURACY of a SCREENING THERMOGRAPH.....	17
201.102 SCREENING THERMOGRAPH ALARM CONDITIONS .....	20
202 Electromagnetic disturbances – Requirements and tests.....	21
206 USABILITY .....	21
Annexes .....	23
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	23
Annex AA (informative) Particular guidance and rationale .....	25
Annex BB (normative) CALIBRATION SOURCE .....	32
Annex CC (informative) Reference to the essential principles .....	33
Bibliography.....	36
Index of defined terms used in this document .....	40
Figure AA.1 – Illustration of TARGET in the visible spectrum.....	27
Figure AA.2 – Illustration of TARGET in the infrared spectrum .....	27
Figure AA.3 – Relative drift of 4 DETECTORS as a function of time.....	30
Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements .....	13
Table 201.C.101 – Marking on the outside of a SCREENING THERMOGRAPH or its parts.....	23
Table 201.C.102 – ACCOMPANYING DOCUMENTS, general of a SCREENING THERMOGRAPH .....	23
Table 201.C.103 – ACCOMPANYING DOCUMENTS, instructions for use of a SCREENING THERMOGRAPH .....	24



Table 201.C.104 – ACCOMPANYING DOCUMENTS, technical description of a SCREENING THERMOGRAPH .....	24
Table AA.1 – Example of relevant uncertainty terms for a SCREENING THERMOGRAPH .....	28
Table CC.1 – Correspondence between this document and the essential principles .....	33

# INTERNATIONAL ELECTROTECHNICAL COMMISSION

## MEDICAL ELECTRICAL EQUIPMENT –

### **Part 2-59: Particular requirements for the basic safety and essential performance of screening thermographs for human febrile temperature screening**

#### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 80601-2-59 has been prepared by a Joint Working Group of IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice, and of ISO subcommittee SC3: Lung ventilators and related equipment, of ISO technical committee 121: Anaesthetic and respiratory equipment.

This second edition cancels and replaces the first edition published in 2008. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) updates of the normative references and the bibliography;
- b) expansion of the applicability to pandemic infectious diseases in general.

The text of this document is based on the following documents:

FDIS	Report on voting
62D/1501/FDIS	62D/1515/RVD

Full information on the voting for the approval of this document can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term:

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this document ;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document ;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of National Committees and Member Bodies is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC or ISO publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

**IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.**

## INTRODUCTION

The minimum safety requirements specified in this document are considered to provide for a practical degree of safety in the operation of ME EQUIPMENT for human febrile temperature screening.

This document describes ME EQUIPMENT that uses infrared technology to detect naturally emitted heat at the skin surface of the FACE. Such ME EQUIPMENT can be useful at ports-of-entry or ports-of-exit and the entrances to buildings under controlled environmental conditions to separate febrile from afebrile individuals to help prevent the spread of communicable diseases. Care can be needed when evaluating individuals under changing environmental conditions, but the region medially adjacent to the inner canthus of the eye has been demonstrated to be a robust measurement site and is supplied by the internal carotid artery. [1]<sup>1</sup>

A body core temperature of 38 °C or above was used as the criterion to restrict traveling during the SARS (severe acute respiratory syndrome) epidemic (April 2003). [2] The US Centers for Disease Control advises that SARS typically begins with a temperature above 38 °C, which is 1 °C higher than normal human body core temperature which averages around 37 °C. [3] It is hard to give an accurate assessment of how many people were checked by infrared temperature measurements in China during the SARS epidemic. There is official Chinese government data indicating that during a two-month period in the spring of 2003, 30 million travellers were screened in China. From this cohort, 9 292 travellers with elevated temperature were detected and 38 were suspected of being SARS carriers. SARS was diagnosed in 21 of these cases. All elevated temperatures were confirmed using traditional clinical temperature measurements of body temperature. Although it is hard to determine the human body's core temperature accurately by infrared measurement of SKIN TEMPERATURE, it can be used for screening for elevated temperature values. [2] [4] [5] Improved rates of detection may result from improved techniques.[6]

International travellers were screened during the 2009 H1N1 influenza outbreak. [7] [8] The pandemic potential of other influenzas such as H7N9 [9] is of concern to the World Health Organization (WHO). [10]

Middle East Respiratory Syndrome Coronavirus (MERS-CoV) was first reported in Saudi Arabia in 2012, and a total of 1 026 laboratory-confirmed cases resulting in at least 376 deaths (36,7%) have been confirmed by the World Health Organization (WHO) as of 25 February 2015. [11] Most identified cases have had fever, although some mild and/or asymptomatic cases have been reported. [11] [12] [13] [14] The possibility of widespread dissemination of MERS-CoV during religious pilgrimage [11] and other regional travel has been investigated, but appears to be under control [15], although WHO continues to express concern. [13] [14] Fever screening at airports has also been employed during outbreaks of Dengue in Taiwan. [16] [17]

The 2014 Ebola outbreak originating in West Africa has brought issues of the potential for global pandemic to the forefront. [18] [19] [20] [21] Controversy has arisen over the effectiveness of thermography for fever screening at airports and other checkpoints [22] [23], while empirical data has demonstrated the effectiveness of this technology when used in compliance with appropriate international standards [24] [25] [26] [27] and WHO guidance. [10] [20] [21]

This document is intended to be applicable for thermographic fever screening devices for the above-mentioned and any other fever-producing infectious diseases. [10] [15] [28] [29].

---

<sup>1</sup> Figures in square brackets refer to the Bibliography

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-59: Particular requirements for the basic safety and essential performance of screening thermographs for human febrile temperature screening

#### 201.1 Scope, object and related standards

Clause 1 of the general standard<sup>2</sup> applies, except as follows:

##### 201.1.1 \* Scope

*Replacement:*

This part of IEC 80601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of SCREENING THERMOGRAPHS intended to be used for the individual non-invasive febrile temperature screening of a human under controlled environmental conditions, hereafter referred to as ME EQUIPMENT. This document sets laboratory characterization test limits for the SCREENING THERMOGRAPH.

NOTE 101 A SCREENING THERMOGRAPH is intended for screening of a human subject and detection of SKIN TEMPERATURE elevated above normal. An elevated SKIN TEMPERATURE needs to be followed up by a subsequent temperature measurement using a clinical thermometer (see ISO 80601-2-56 [30]).

NOTE 102 The main part of such equipment is commonly referred to as an infrared camera.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

##### 201.1.2 Object

*Replacement:*

The object of this particular document is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for SCREENING THERMOGRAPHS as defined in 201.3.209.

##### 201.1.3 Collateral standards

*Addition:*

This particular document refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this document.

IEC 60601-1-2:2014, IEC 60601-1-6:2010 and IEC 60601-1-6:2010/AMD1:2013 apply as modified in Clauses 202 and 206 respectively. IEC 60601-1-3, IEC 60601-1-10, IEC 60601-1-11 and IEC 60601-1-12 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

---

<sup>2</sup> The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

#### 201.1.4 Particular standards

##### *Replacement:*

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 are referred to in this particular document as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this document corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x", where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this document addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this document addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"*Replacement*" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this document.

"*Addition*" means that the text of this document is additional to the requirements of the general standard or applicable collateral standard.

"*Amendment*" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this document.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular document taken together.

Where there is no corresponding clause or subclause in this document, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

#### 201.2 Normative references

NOTE Informative references are listed in the Bibliography.

Clause 2 of the general standard applies, except as follows:

**Replacement:**

IEC 60601-1-2:2014, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests*

IEC 60601-1-6:2010, *Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability*  
IEC 60601-1-6:2010/AMD1:2013

IEC 60601-1-8:2006, *Medical electrical equipment – Part 1-8: General requirements for basic safety and essential performance – Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*  
IEC 60601-1-8:2006/AMD1:2012

**Addition:**

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*  
IEC 60601-1:2005/AMD1:2012

ISO/TR 13154, *Medical electrical equipment – Deployment, implementation and operational guidelines for identifying febrile humans using a screening thermograph*

ASTM E1213-14<sup>3</sup>, *Standard Test Method for Minimum Resolvable Temperature Difference for Thermal Imaging Systems*

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**