

<b>STN</b>	<b>Kvalita vody Technécium 99 Časť 2: Skúšobná metóda s použitím hmotnostnej spektrometrie s indukčne viazanou plazmou (ICP-MS) (ISO 22125-2: 2019)</b>	<b>STN EN ISO 22125-2</b>  75 7624
------------	---	--

Water quality - Technetium-99 - Part 2: Test method using inductively coupled plasma mass spectrometry (ICP-MS) (ISO 22125-2:2019)

Táto norma obsahuje anglickú verziu európskej normy.  
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 02/20

Obsahuje: EN ISO 22125-2:2019, ISO 22125-2:2019

**130384**



EUROPEAN STANDARD

**EN ISO 22125-2**

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2019

ICS 13.060.60; 17.240

English Version

**Water quality - Technetium-99 - Part 2: Test method using  
inductively coupled plasma mass spectrometry (ICP-MS)  
(ISO 22125-2:2019)**

Qualité de l'eau - Technétium-99 - Partie 2: Méthode  
d'essai par spectrométrie de masse couplée à un  
plasma induit (ISO 22125-2:2019)

Wasserbeschaffenheit - Technetium 99 - Teil 2:  
Verfahrens mittels Massenspektromie und induktiv  
gekoppeltem Plasma (ICP-MS) (ISO 22125-2:2019)

This European Standard was approved by CEN on 8 September 2019.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

**EN ISO 22125-2:2019 (E)**

<b>Contents</b>	<b>Page</b>
<b>European foreword.....</b>	<b>3</b>

## **European foreword**

This document (EN ISO 22125-2:2019) has been prepared by Technical Committee ISO/TC 147 "Water quality" in collaboration with Technical Committee CEN/TC 230 "Water analysis" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2020, and conflicting national standards shall be withdrawn at the latest by May 2020.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## **Endorsement notice**

The text of ISO 22125-2:2019 has been approved by CEN as EN ISO 22125-2:2019 without any modification.



**INTERNATIONAL  
STANDARD**

**ISO  
22125-2**

First edition  
2019-11

---

---

**Water quality — Technetium-99 —**

**Part 2:**

**Test method using inductively coupled  
plasma mass spectrometry (ICP-MS)**

*Qualité de l'eau — Technétium-99 —*

*Partie 2: Méthode d'essai par spectrométrie de masse couplée à un  
plasma induit (ICP-MS)*



Reference number  
ISO 22125-2:2019(E)

© ISO 2019

**ISO 22125-2:2019(E)****COPYRIGHT PROTECTED DOCUMENT**

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland



# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms, definitions and symbols</b> .....	<b>2</b>
<b>4 Principle</b> .....	<b>4</b>
<b>5 Sampling, handling and storage</b> .....	<b>4</b>
<b>6 Procedure</b> .....	<b>4</b>
6.1 Sample preparation for measurement.....	4
6.2 Sample measurement.....	5
<b>7 Quality assurance and quality control program</b> .....	<b>5</b>
7.1 General.....	5
7.2 Variables that could influence the measurement.....	5
7.3 Instrument verification.....	5
7.4 Contamination.....	5
7.5 Interference control.....	6
7.6 Method verification.....	6
7.7 Demonstration of analyst capability.....	6
<b>8 Expression of results</b> .....	<b>6</b>
8.1 Using Re, <sup>97</sup> Tc, or <sup>98</sup> Tc as a recovery tracer.....	6
8.1.1 Calculation of mass of tracer and analyte added.....	6
8.1.2 Measurement bias.....	7
8.1.3 Sample mass concentration.....	7
8.1.4 Detection limit.....	8
8.1.5 Limit of quantification.....	8
8.2 Using <sup>95m</sup> Tc, <sup>97m</sup> Tc or <sup>99m</sup> Tc as a recovery tracer.....	8
8.2.1 Calculation of activity of tracer, mass of analyte and mass of internal standard added.....	8
8.2.2 Purification step recovery.....	9
8.2.3 Measurement bias.....	9
8.2.4 Sample mass concentration.....	9
8.2.5 Detection limit.....	9
8.2.6 Limit of quantification.....	10
8.2.7 Conversion of mass concentration to activity concentration.....	10
8.2.8 Conversion of mass concentration to volume unit.....	10
8.3 Correction for the presence of <sup>99</sup> Tc in the tracer.....	11
<b>9 Test report</b> .....	<b>11</b>
<b>Annex A (informative) Method 1 — TEVA resin</b> .....	<b>12</b>
<b>Annex B (informative) Method 2 — TRU resin</b> .....	<b>15</b>
<b>Annex C (informative) Method 3 — Anion exchange resin</b> .....	<b>18</b>
<b>Bibliography</b> .....	<b>21</b>

## ISO 22125-2:2019(E)

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 3, *Radioactivity measurements*.

A list of all the parts in the ISO 22125 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

Radioactivity from several naturally-occurring and anthropogenic sources is present throughout the environment. Thus, water bodies (such as surface waters, ground waters, sea waters) can contain radionuclides of natural, human-made, or both origin.

- Natural radionuclides, including  $^{40}\text{K}$ ,  $^3\text{H}$ ,  $^{14}\text{C}$ , and those originating from the thorium and uranium decay series, in particular  $^{226}\text{Ra}$ ,  $^{228}\text{Ra}$ ,  $^{234}\text{U}$ ,  $^{238}\text{U}$ ,  $^{210}\text{Po}$  and  $^{210}\text{Pb}$  can be found in water for natural reasons (such as desorption from the soil and washoff by rain water) or can be released from technological processes involving naturally occurring radioactive materials (such as the mining and processing of mineral sands or phosphate fertilizers production and use).
- Human-made radionuclides such as transuranium elements (americium, plutonium, neptunium, curium),  $^3\text{H}$ ,  $^{14}\text{C}$ ,  $^{90}\text{Sr}$ , and gamma emitting radionuclides can also be found in natural waters. Small quantities of these radionuclides are discharged from nuclear fuel cycle facilities into the environment as a result of authorized routine releases. Some of these radionuclides used for medical and industrial applications are also released into the environment after use. Anthropogenic radionuclides are also found in waters as a result of past fallout contaminations resulting from the explosion in the atmosphere of nuclear devices and accidents such as those that occurred in Chernobyl and Fukushima.

Radionuclide activity concentration in water bodies can vary according to local geological characteristics and climatic conditions and can be locally and temporally enhanced by releases from nuclear installation during planned, existing, and emergency exposure situations<sup>[1]</sup>. Drinking water may thus contain radionuclides at activity concentrations which could present a risk to human health.

The radionuclides present in liquid effluents are usually controlled before being discharged into the environment<sup>[2]</sup> and water bodies. Drinking waters are monitored for their radioactivity as recommended by the World Health Organization (WHO)<sup>[3]</sup> so that proper actions can be taken to ensure that there is no adverse health effect to the public. Following these international recommendations, national regulations usually specify radionuclide authorized concentration limits for liquid effluent discharged to the environment and radionuclide guidance levels for waterbodies and drinking waters for planned, existing, and emergency exposure situations. Compliance with these limits can be assessed using measurement results with their associated uncertainties as specified by ISO/IEC Guide 98-3 and ISO 5667-20<sup>[4]</sup>.

Depending on the exposure situation, there are different limits and guidance levels that would result in an action to reduce health risk. As an example, during planned or existing situation, the WHO guidelines for guidance level in drinking water is  $100 \text{ Bq}\cdot\text{l}^{-1}$  for  $^{99}\text{Tc}$ <sup>[3]</sup> activity concentration.

NOTE 1 The guidance level is the activity concentration with an intake of 2 l/d of drinking water for one year that results in an effective dose of 0,1 mSv/a for members of the public. This is an effective dose that represents a very low level of risk and which is not expected to give rise to any detectable adverse health effects<sup>[3]</sup>.

In the event of a nuclear emergency, the WHO Codex Guideline Levels<sup>[5]</sup> mentioned that the activity concentration in contaminated food might not be greater than  $10\,000 \text{ Bq}\cdot\text{kg}^{-1}$  for  $^{99}\text{Tc}$ .

NOTE 2 The Codex guidelines levels (GLs) apply to radionuclides contained in foods destined for human consumption and traded internationally, which have been contaminated following a nuclear or radiological emergency. These GLs apply to food after reconstitution or as prepared for consumption, i.e. not to dried or concentrated foods, and are based on an intervention exemption level of 1 mSv in a year for members of the public (infant and adult)<sup>[5]</sup>.

Thus, the test method can be adapted so that the characteristic limits, decision threshold, detection limit and uncertainties ensure that the radionuclide activity concentrations test results can be verified to be below the guidance levels required by a national authority for either planned/existing situations or for an emergency situation<sup>[5][6][7]</sup>.

Usually, the test methods can be adjusted to measure the activity concentration of the radionuclide(s) in either wastewaters before storage or in liquid effluents before being discharged to the environment.

**ISO 22125-2:2019(E)**

The test results will enable the plant/installation operator to verify that, before their discharge, wastewaters/liquid effluent radioactive activity concentrations do not exceed authorized limits.

The test method(s) described in this document may be used during planned, existing and emergency exposure situations as well as for wastewaters and liquid effluents with specific modifications that could increase the overall uncertainty, detection limit, and threshold.

The test method(s) may be used for water samples after proper sampling, sample handling, and test sample preparation (see the relevant part of the ISO 5667 series).

This document has been developed to answer the need of test laboratories carrying out these measurements, that are sometimes required by national authorities, as they may have to obtain a specific accreditation for radionuclide measurement in drinking water samples.

This document is one of a set of International Standards on test methods dealing with the measurement of the activity concentration of radionuclides in water samples.

# Water quality — Technetium-99 —

## Part 2:

# Test method using inductively coupled plasma mass spectrometry (ICP-MS)

**WARNING** — Persons using this document should be familiar with normal laboratory practices. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to determine the applicability of any other restrictions.

**IMPORTANT** — It is absolutely essential that tests conducted according to this test method be carried out by suitably trained staff.

## 1 Scope

This document specifies a method for the measurement of  $^{99}\text{Tc}$  in all types of water by inductively coupled plasma mass spectrometry (ICP-MS).

The method is applicable to test samples of supply/drinking water, rainwater, surface and ground water, as well as cooling water, industrial water, domestic, and industrial wastewater after proper sampling and handling and test sample preparation. A filtration of the test sample is necessary.

The detection limit depends on the sample volume and the instrument used. The method described in this document, using currently available ICP-MS, has a detection limit of approximately  $0,2 \text{ ng}\cdot\text{kg}^{-1}$  to  $0,5 \text{ ng}\cdot\text{kg}^{-1}$  ( $0,1 \text{ Bq}\cdot\text{kg}^{-1}$  to  $0,3 \text{ Bq}\cdot\text{kg}^{-1}$ ), which is much lower than the WHO criteria for safe consumption of drinking water ( $100 \text{ Bq}\cdot\text{l}^{-1}$ )<sup>[3]</sup>. The method presented in this document is not intended for the determination of ultra-trace amount of  $^{99}\text{Tc}$ .

The mass concentration values in this document are expressed by sample mass unit instead of sample volume unit as it is usually the case in similar standards. The reason is that  $^{99}\text{Tc}$  is measured in various matrix types such as fresh water or sea water, which have significant differences in density. The mass concentration values can be easily converted to sample volume unit by measuring the sample volume. However, it increases the uncertainty on the mass concentration result.

The method described in this document is applicable in the event of an emergency situation, but not if  $^{99\text{m}}\text{Tc}$  is present at quantities that could cause interference.

The analysis of Tc adsorbed to suspended matter is not covered by this method.

It is the user's responsibility to ensure the validity of this test method for the water samples tested.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO/IEC Guide 99, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*

**ISO 22125-2:2019(E)**

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes and sampling techniques*

ISO 5667-3, *Water quality — Sampling — Part 3: Preservation and handling of water samples*

ISO 5667-10, *Water quality — Sampling — Part 10: Guidance on sampling of waste waters*

ISO 10703, *Water quality — Determination of the activity concentration of radionuclides — Method by high resolution gamma-ray spectrometry*

ISO 11929 (all parts), *Determination of the characteristic limits (decision threshold, detection limit and limits of the confidence interval) for measurements of ionizing radiation — Fundamentals and application*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

ISO 17294-2, *Water quality — Application of inductively coupled plasma mass spectrometry (ICP-MS) — Part 2: Determination of selected elements including uranium isotopes*

ISO 20042, *Measurement of radioactivity — Gamma emitting radionuclides — Generic test method using gamma spectrometry*

ISO 80000-10, *Quantities and units — Part 10: Atomic and nuclear physics*

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**