

STN	Mlieko, mliečne výrobky, dojčenská výživa a nutričné doplnky pre dospelých Stanovenie minerálnych a stopových prvkov Metóda hmotnostnej spektrometrie s indukčne viazanou plazmou (ICP-MS) (ISO 21424: 2018)	STN EN ISO 21424
		57 0548

Milk, milk products, infant formula and adult nutritionals - Determination of minerals and trace elements - Inductively coupled plasma mass spectrometry (ICP-MS) method (ISO 21424:2018)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 11/20

Obsahuje: EN ISO 21424:2020, ISO 21424:2018

131571

EUROPEAN STANDARD

EN ISO 21424

NORME EUROPÉENNE

EUROPÄISCHE NORM

June 2020

ICS 67.100.10

English Version

Milk, milk products, infant formula and adult nutritionals -
Determination of minerals and trace elements -
Inductively coupled plasma mass spectrometry (ICP-MS)
method (ISO 21424:2018)

Lait, produits laitiers, formules infantiles et produits nutritionnels pour adultes - Détermination de la teneur en minéraux et en oligo-éléments - Méthode par spectrométrie de masse avec plasma à couplage inductif (ICP-SM) (ISO 21424:2018)

Milch, Milcherzeugnisse, Säuglingsnahrung und Nahrungsergänzungsmittel für Erwachsene - Bestimmung von Mineralien und Spurenelementen - Verfahren mit induktiv gekoppelter Plasma-Massenspektrometrie (ICP-MS) (ISO 21424:2018)

This European Standard was approved by CEN on 10 May 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 21424:2020 (E)

Contents	Page
European foreword.....	3

European foreword

The text of ISO 21424:2018 has been prepared by Technical Committee ISO/TC 34 "Food products" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 21424:2020 by Technical Committee CEN/TC 302 "Milk and milk products - Methods of sampling and analysis" the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2020, and conflicting national standards shall be withdrawn at the latest by December 2020.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 21424:2018 has been approved by CEN as EN ISO 21424:2020 without any modification.

**INTERNATIONAL
STANDARD**

**ISO
21424**

IDF 243

First edition
2018-11

**Milk, milk products, infant formula and
adult nutritionals — Determination
of minerals and trace elements —
Inductively coupled plasma mass
spectrometry (ICP-MS) method**

*Lait, produits laitiers, formules infantiles et produits nutritionnels
pour adultes — Détermination de la teneur en minéraux et en oligo-
éléments — Méthode par spectrométrie de masse avec plasma à
couplage inductif (ICP-SM)*



Reference numbers
ISO 21424:2018(E)
IDF 243:2018(E)

© ISO and IDF 2018

ISO 21424:2018(E)
IDF 243:2018(E)**COPYRIGHT PROTECTED DOCUMENT**

© ISO and IDF 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

International Dairy Federation
Silver Building • Bd Auguste Reyers 70/B
B-1030 Brussels
Phone: +32 2 325 67 40
Fax: +32 2 325 67 41
Email: info@fil-idf.org
Website: www.fil-idf.org

Published in Switzerland

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	2
5 Reagents	2
6 Preparation of standards and solutions	3
6.1 Surfactant solution (optional, approximately 5 %).....	3
6.2 Nitric acid rinse solution for autosampler rinse port, 2 %.....	3
6.3 P/A factor tuning working solution.....	3
6.4 Calibration blank (Cal Blk) and preparation blank (PB) solution.....	3
6.5 Calibration standard solution set.....	4
7 Apparatus	4
8 Preparation of test sample	5
9 Determination	6
10 Calculation	7
11 Precision data	8
11.1 General.....	8
11.2 Repeatability.....	8
11.3 Reproducibility.....	8
12 Test report	8
Annex A (informative) Precision data	9
Bibliography	31

ISO 21424:2018(E)
IDF 243:2018(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products* and the International Dairy Federation (IDF), in collaboration with AOAC INTERNATIONAL.

It is being published jointly by ISO and IDF and separately by AOAC INTERNATIONAL. The method described in this document is equivalent to the AOAC Official Method 2015.06: *Minerals and Trace Elements in Infant Formula*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

IDF (the International Dairy Federation) is a non-profit private sector organization representing the interests of various stakeholders in dairying at the global level. IDF members are organized in National Committees, which are national associations composed of representatives of dairy-related national interest groups including dairy farmers, dairy processing industry, dairy suppliers, academics and governments/food control authorities.

ISO and IDF collaborate closely on all matters of standardization relating to methods of analysis and sampling for milk and milk products. Since 2001, ISO and IDF jointly publish their International Standards using the logos and reference numbers of both organizations.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. IDF shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

This document was prepared by the IDF Standing Committee on Analytical Methods for Composition and ISO Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*, in collaboration with AOAC INTERNATIONAL.

It is being published jointly by ISO and IDF and separately by AOAC INTERNATIONAL. The method described in this document is equivalent to the AOAC Official Method 2015.06: *Minerals and Trace Elements in Infant Formula*. All work was carried out by the ISO/IDF Action Team on C40 of the Standing Committee on Analytical Methods for Composition under the aegis of its project leader, Mr. H. Crujisen (NL).

Milk, milk products, infant formula and adult nutritionals — Determination of minerals and trace elements — Inductively coupled plasma mass spectrometry (ICP-MS) method

1 Scope

This document specifies a method for the quantitative determination of calcium (Ca), copper (Cu), iron (Fe), magnesium (Mg), manganese (Mn), phosphorus (P), potassium (K), sodium (Na), zinc (Zn), chromium (Cr), molybdenum (Mo) and selenium (Se) using inductively coupled plasma and mass spectrometry (ICP-MS).

The method is applicable for the determination of all 12 elements in infant formula and adult nutritional products. The method is also applicable for milk, milk powder, whey powder, butter and cheese excluding the determination of Cr, because all Cr results were below the quantification limit and reproducibility could not be determined in these matrices^[1]. The present method is an extension of ISO 20649 | IDF 235 (AOAC 2011.19^[2]) which was validated only for Cr, Mo and Se in infant formula and adult nutritional products.

The ranges given in [Table 1](#) are in scope (see also [Table A.1](#)).

Table 1 — Analytical ranges

	Ca	Cu	Fe	Mg	Mn	P	K	Na	Zn	Cr	Se	Mo
Lower analytical range ^a , in mg/100 g	3	0,002	0,04	0,7	0,002	3	3	2	0,07	0,002	0,000 6	0,000 2
Upper analytical range ^a , in mg/100 g	1 280	1,2	20	110	1,0	800	2 000	850	18	0,16	0,05	0,10

^a Concentrations apply to

— milk and "ready-to-feed" liquid as-is, using a typical sample size of 1 g per final analytical solution volume of 50 ml, and

— reconstituted milk powder, reconstituted infant formula powder and reconstituted adult nutritional powder (25 g into 200 g of water), using a typical sample size of 1,8 g of the reconstituted slurry per final analytical solution volume of 50 ml.

Ranges for non-reconstituted dairy ingredients (butter, cheese, whey powder, whey protein concentrate) are adjusted proportionally upward from these values based upon the sample size used for the ingredient. For example, if 0,3 g of cheese is digested the ranges will be 1 g/ 0,3 g = 3,3 × higher.

2 Normative references

There are no normative references in this document.

koniec náhľadu – text ďalej pokračuje v platenej verzii STN