

<b>STN</b>	<b>Tuhé environmentálne matrice Stanovenie polychlórovaných bifenylov (PCB) plynovou chromatografiou – hmotnostným spektrometrom (GC-MS) alebo s detektorom elektrónového záchytu (GC-ECD)</b>	<b>STN EN 17322</b>  83 8265
------------	--	--

Environmental Solid Matrices - Determination of polychlorinated biphenyls (PCB) by gas chromatography - mass selective detection (GC-MS) or electron-capture detection (GC-ECD)

Táto norma obsahuje anglickú verziu európskej normy.  
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 11/20

Obsahuje: EN 17322:2020

Oznámením tejto normy sa ruší  
STN EN 15308 (83 8265) z apríla 2017

STN EN 16167+AC (83 8464) z júla 2019

**131837**

EUROPEAN STANDARD

EN 17322

NORME EUROPÉENNE

EUROPÄISCHE NORM

July 2020

ICS 13.030.01; 13.030.10; 13.030.20; 13.080.10

Supersedes EN 15308:2016, EN 16167:2018

English Version

Environmental Solid Matrices - Determination of  
polychlorinated biphenyls (PCB) by gas chromatography -  
mass selective detection (GC-MS) or electron-capture  
detection (GC-ECD)

Matrices solides environnementales - Dosage des  
polychlorobiphényles (PCB) par chromatographie en  
phase gazeuse-spectrométrie de masse (CG-SM) ou  
chromatographie en phase gazeuse avec détection par  
capture d'électrons (CG-ECD)

Feststoffe in der Umwelt - Bestimmung von  
polychlorierten Biphenylen (PCB) mittels  
Gaschromatographie und massenspektrometrischer  
Detektion (GC-MS) oder Elektronen-Einfang-Detektion  
(GC-ECD)

This European Standard was approved by CEN on 15 June 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

<b>Contents</b>	<b>Page</b>
European foreword.....	3
Introduction .....	4
1 Scope.....	5
2 Normative references.....	5
3 Terms and definitions .....	6
4 Principle .....	7
5 Interferences .....	8
5.1 Interference with sampling and extraction.....	8
5.2 Interference with GC.....	8
6 Safety remarks .....	8
7 Reagents .....	9
7.1 General.....	9
7.2 Reagents for extraction.....	9
7.3 Reagents for clean-up.....	9
7.4 Gas chromatographic analysis .....	12
7.5 Standards .....	12
7.6 Preparation of standard solutions.....	14
8 Apparatus.....	15
8.1 Extraction and clean-up procedures.....	15
8.2 Gas chromatograph .....	16
9 Sample storage and preservation .....	16
9.1 Sample storage.....	16
9.2 Sample pre-treatment .....	16
10 Procedure.....	17
10.1 Blank test .....	17
10.2 Extraction.....	17
10.3 Concentration.....	20
10.4 Clean-up of the extract .....	20
10.5 Addition of the injection standard.....	24
10.6 Gas chromatographic analysis (GC).....	24
10.7 Mass spectrometry (MS).....	24
10.8 Electron capture detection (ECD).....	28
11 Performance characteristics.....	30
12 Precision.....	30
13 Test report.....	30
Annex A (informative) Repeatability and reproducibility data .....	31
A.1 Materials used in the inter-laboratory comparison study.....	31
A.2 Inter-laboratory comparison results .....	32
Annex B (informative) Examples for gas chromatographic conditions and retention times of PCBs.....	36
Annex C (informative) Calculation method for the estimation of total PCB content .....	37
Bibliography.....	45

## **European foreword**

This document (EN 17322:2020) has been prepared by Technical Committee CEN/TC 444 “Environmental characterization of solid matrices”, the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2021, and conflicting national standards shall be withdrawn at the latest by January 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 16167:2018 and EN 15308:2016.

This document is the result of the merging of EN 16167:2018 and EN 15308:2016, with minor technical modifications.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

**EN 17322:2020 (E)****Introduction**

Polychlorinated biphenyls (PCB) have been widely used as additives in industrial applications where chemical stability has been required. This stability on the other hand creates environmental problems when PCB are eventually released into the environment. Since some of these PCB compounds are highly toxic, their presence in the environment (air, water, soil, sediment and waste) is regularly monitored and controlled. At present determination of PCB is carried out in these matrices in most of the routine laboratories following the preceding steps for sampling, pre-treatment, extraction and clean-up, by measurement of specific PCB by means of gas chromatography in combination with mass spectrometric detection (GC-MS) or gas chromatography with electron capture detector (GC-ECD).

This document was developed by merging of EN 16167:2018, initially elaborated as a CEN Technical Specification in the European project 'HORIZONTAL' and validated by CEN/TC 400 with the support of BAM, with EN 15308, published by CEN/TC 292.

Considering the different matrices and possible interfering compounds, this document does not contain one single possible way of working. Several choices are possible, in particular relating to clean-up. Detection with both MS-detection and ECD-detection is possible. Two different extraction procedures are described and 9 clean-up procedures. The use of internal and injection standards is described in order to have an internal check on choice of the extraction and clean-up procedure. The method is as far as possible in agreement with the method described for PAH (EN 16181:2018 and EN 15527:2008). It has been tested for ruggedness.

This document is applicable and validated for several types of matrices as indicated in Table 1 (see also Annex A for the results of the validation).

**Table 1 — Matrices for which this European Standard is applicable and validated**

<b>Matrix</b>	<b>Materials used for validation</b>
Soil	Sandy soil Mix of soil from the vicinity of Berlin, Germany and PCB-free German reference soil
Sludge	Mix of municipal waste water treatment plant sludge from North Rhine Westphalia, Germany
Biowaste	Mix of compost from the vicinity of Berlin, Germany and sludge from North Rhine Westphalia, Germany
Waste	Contaminated soil, building debris, waste wood, sealant waste, electronic waste, shredder light fraction, cable shredder waste

**WARNING** — Persons using this document should be familiar with usual laboratory practice. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

**IMPORTANT** — It is absolutely essential that tests conducted according to this document be carried out by suitably trained staff.

## 1 Scope

This document specifies methods for quantitative determination of seven selected polychlorinated biphenyls (PCB28, PCB52, PCB101, PCB118, PCB138, PCB153 and PCB180) in soil, sludge, sediment, treated biowaste, and waste using GC-MS and GC-ECD (see Table 2).

**Table 2 — Target analytes of this European Standard**

Target analyte		CAS-RN <sup>a</sup>
PCB28	2,4,4'-trichlorobiphenyl	7012-37-5
PCB52	2,2',5,5'-tetrachlorobiphenyl	35693-99-3
PCB101	2,2',4,5,5'-pentachlorobiphenyl	37680-73-2
PCB118	2,3',4,4',5-pentachlorobiphenyl	31508-00-6
PCB138	2,2',3,4,4',5'-hexachlorobiphenyl	35065-28-2
PCB153	2,2',4,4',5,5'-hexachlorobiphenyl	35065-27-1
PCB180	2,2',3,4,4',5,5'-heptachlorobiphenyl	35065-29-3
<sup>a</sup> CAS-RN Chemical Abstracts Service Registry Number.		

The limit of detection depends on the determinants, the equipment used, the quality of chemicals used for the extraction of the sample and the clean-up of the extract.

Under the conditions specified in this document, lower limit of application from 1 µg/kg (expressed as dry matter) for soils, sludge and biowaste to 10 µg/kg (expressed as dry matter) for solid waste can be achieved. For some specific samples the limit of 10 µg/kg cannot be reached.

Sludge, waste and treated biowaste may differ in properties, as well as in the expected contamination levels of PCB and presence of interfering substances. These differences make it impossible to describe one general procedure. This document contains decision tables based on the properties of the sample and the extraction and clean-up procedure to be used.

NOTE The analysis of PCB in insulating liquids, petroleum products, used oils and aqueous samples is referred to in EN 61619, EN 12766-1 and EN ISO 6468 respectively.

The method can be applied to the analysis of other PCB congeners not specified in the scope, provided suitability is proven by proper in-house validation experiments.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 5667-15, *Water quality — Sampling — Part 15: Guidance on the preservation and handling of sludge and sediment samples (ISO 5667-15)*

EN ISO 16720, *Soil quality — Pretreatment of samples by freeze-drying for subsequent analysis (ISO 16720)*

EN ISO 22892, *Soil quality — Guidelines for the identification of target compounds by gas chromatography and mass spectrometry (ISO 22892)*

ISO 8466-1, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function*

**EN 17322:2020 (E)**

ISO 18512, *Soil quality — Guidance on long and short-term storage of soil samples*

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**