

STN	Stomatológia Magnetické zásuvné spoje (ISO 13017: 2020)	STN EN ISO 13017 85 6365
------------	--	--

Dentistry - Magnetic attachments (ISO 13017:2020)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 10/20

Obsahuje: EN ISO 13017:2020, ISO 13017:2020

Oznámením tejto normy sa ruší
STN EN ISO 13017 (85 6365) z novembra 2012

131841

EUROPEAN STANDARD

EN ISO 13017

NORME EUROPÉENNE

EUROPÄISCHE NORM

July 2020

ICS 11.060.10

Supersedes EN ISO 13017:2012

English Version

Dentistry - Magnetic attachments (ISO 13017:2020)Médecine bucco-dentaire - Attaches magnétiques (ISO
13017:2020)Zahnheilkunde - Magnetische Retentionselemente (ISO
13017:2020)

This European Standard was approved by CEN on 27 June 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

EN ISO 13017:2020 (E)

Contents	Page
European foreword.....	3

European foreword

This document (EN ISO 13017:2020) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2021, and conflicting national standards shall be withdrawn at the latest by January 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 13017:2012.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 13017:2020 has been approved by CEN as EN ISO 13017:2020 without any modification.

INTERNATIONAL STANDARD

ISO
13017

Second edition
2020-07

Dentistry — Magnetic attachments

Médecine bucco-dentaire — Attaches magnétiques



Reference number
ISO 13017:2020(E)

© ISO 2020

ISO 13017:2020(E)**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	3
4.1 Material.....	3
4.1.1 Magnet core.....	3
4.1.2 Components other than the magnet core.....	3
4.1.3 Reported chemical composition.....	3
4.2 Hazardous elements.....	3
4.2.1 Recognized hazardous elements.....	3
4.2.2 Permitted limits for the hazardous elements cadmium, beryllium and lead.....	3
4.2.3 Manufacturer's reported nickel content and permitted deviation.....	4
4.3 Risk analysis.....	4
4.4 Magnetic flux leakage.....	4
4.5 Retentive force.....	4
4.6 Corrosion resistance.....	4
4.6.1 Released ions.....	4
4.6.2 Breakdown potential.....	4
5 Preparation of test specimens	4
5.1 Retentive force.....	4
5.2 Static immersion test.....	4
5.3 Anodic polarization.....	5
6 Test methods	5
6.1 Information, instructions and marking.....	5
6.2 Magnetic flux leakage.....	5
6.2.1 Apparatus.....	5
6.2.2 Test procedure.....	5
6.3 Retentive force.....	5
6.3.1 Apparatus.....	5
6.3.2 Materials.....	6
6.3.3 Fixing procedure.....	7
6.3.4 Test procedure.....	7
6.3.5 Analysis.....	7
6.4 Corrosion resistance.....	9
6.4.1 Static immersion test.....	9
6.4.2 Anodic polarization.....	10
7 Information and instructions for use	11
8 Marking and labelling	11
8.1 Marking.....	11
8.2 Labelling.....	11
Bibliography	12

ISO 13017:2020(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 2, *Prosthetic materials*. This second edition cancels and replaces the first edition (ISO 13017:2012), which has been technically revised. It also incorporates the Amendment ISO 13017:2012/Amd.1:2015. The main changes compared to the previous edition are as follows:

- addition of ISO 14233 to [Clause 2](#);
- addition of lead as a hazardous element;
- addition of the cleaning method of test specimens prepared for retentive force;
- change of the device for retentive force to a single shaft type;
- change of [Figure 3](#) to the single shaft type device;
- specification of the performance of the device with respect to moving friction force and modification of specimen tables;
- change of a cross-head speed in measuring retentive force from 5,0 mm min⁻¹ to 2,0 mm min⁻¹;
- addition of materials for fixing a specimen on the table such as cyanoacrylate adhesive and self-curing acrylic resin;
- deletion of the description of the adhesive double sided tape to fix a specimen on the table;
- specification of the procedures to fix a specimen on the table;
- addition of detailed method of measuring retentive force;
- addition of explaining the calculation method of retentive force;
- addition of a figure that shows a retentive force curve as [Figure 4](#);

- specification of quantitative analyses in the static immersion test using definition of determination limit and detection limit;
- addition of “quantity” to labelling.

Any feedback or questions on this document should be directed to the user’s national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

ISO 13017:2020(E)**Introduction**

The early practical uses of permanent magnets were as navigational compasses. Magnets have since become firmly integrated into today's modern electronic device technology. The development of magnetic technology has generated rare earth magnets. Their excellent magnetic character properties permit predictable clinical applications and use. Dental magnetic attachments are one of the products composed of rare earth magnets, providing retention, support and stabilization of dental and maxillofacial appliances.

Specific qualitative and quantitative test methods for demonstrating freedom from unacceptable biological hazard are not included in this document, but for the assessment of possible biological or toxicological hazards, reference can be made to ISO 10993-1 and ISO 7405.

Dentistry — Magnetic attachments

1 Scope

This document specifies requirements and test methods for assessing the applicability of dental magnetic attachments that provide retention, support and stabilization of removable prostheses (crowns and bridges, partial dentures and overdentures), superstructures of dental implants and orthodontic or maxillofacial prostheses including obturators.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 3585, *Borosilicate glass 3.3 — Properties*

ISO 5832-1, *Implants for surgery — Metallic materials — Part 1: Wrought stainless steel*

ISO 10271, *Dentistry — Corrosion test methods for metallic materials*

ISO 14233, *Dentistry — Polymer-based die materials*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO 22674, *Dentistry — Metallic materials for fixed and removable restorations and appliances*

IEC 60404-8-1, *Magnetic materials — Part 8-1: Specifications for individual materials — Magnetically hard materials*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN