

<b>STN</b>	<b>Vozidlá záchranej zdravotnej služby a ich vybavenie Cestné ambulancie</b>	<b>STN EN 1789</b>  84 7080
------------	--	---------------------------------------

Medical vehicles and their equipment - Road ambulances

Táto norma obsahuje anglickú verziu európskej normy.  
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 02/21

Obsahuje: EN 1789:2020

Oznámením tejto normy sa od 31.03.2022 ruší  
STN EN 1789+A2 (84 7080) z augusta 2015

**132439**



EUROPEAN STANDARD

EN 1789

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 2020

ICS 11.160; 43.160

Supersedes EN 1789:2007+A2:2014

English Version

## Medical vehicles and their equipment - Road ambulances

Véhicules de transport sanitaire et leurs équipements -  
Ambulances routières

Rettungsdienstfahrzeuge und deren Ausrüstung -  
Krankenkraftwagen

This European Standard was approved by CEN on 13 April 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

## Contents

Page

European foreword.....	4
Introduction .....	5
1 Scope .....	6
2 Normative references.....	6
3 Terms and definitions .....	9
4 Requirements .....	10
4.1 General requirements .....	10
4.2 Electrical requirements.....	10
4.2.1 General.....	10
4.2.2 Electromagnetic compatibility (EMC).....	11
4.2.3 Battery and alternator .....	11
4.2.4 Electrical installation.....	12
4.2.5 Visual warning system and audible warning system (siren) .....	13
4.2.6 Reversing systems .....	13
4.2.7 Exterior illumination lights.....	13
4.3 Vehicle body.....	14
4.3.1 Fire safety.....	14
4.3.2 Driver's seat configuration.....	14
4.3.3 Minimum passenger capacity .....	14
4.3.4 Bulkhead .....	15
4.3.5 Openings (doors, windows, emergency exits) .....	15
4.3.6 Loading area.....	16
4.4 Patient's compartment.....	18
4.4.1 General.....	18
4.4.2 Safety .....	18
4.4.3 Hygiene .....	18
4.4.4 Patient's compartment dimensions .....	19
4.4.5 Patient and crew seating .....	24
4.4.6 Ventilation and anaesthetic gas scavenging systems.....	25
4.4.7 Temperature control system .....	25
4.4.8 Interior lighting .....	26
4.4.9 Interior noise level .....	26
4.4.10 Holding system for infusion .....	26
4.4.11 Retention, fixation and restraint systems.....	27
4.4.12 Mass reserve .....	27
5 Testing.....	27
5.1 General.....	27
5.2 Testing of the interior noise level.....	27
5.2.1 Specific measurement conditions .....	27
5.2.2 Measurements.....	28
5.3 Testing of retention systems and fixation of the equipment in the patient's compartment .....	28
5.3.1 General.....	28
5.3.2 Testing of the stretcher fixation on the vehicle floor .....	30
5.3.3 Testing of the medical devices fixation .....	30
5.3.4 Testing of furniture .....	30

5.3.5	Test procedure .....	31
5.4	Testing of rounded edges and radius inside the patient's compartment.....	32
5.4.1	Testing of rounded edges.....	32
5.4.2	Testing of radius inside the patient's compartment.....	33
5.5	Procedure to verify the patient's compartment specifications .....	33
5.6	Procedure to verify the loading area specifications .....	33
5.6.1	General .....	33
5.6.2	Procedure to verify the loading angle of 16° .....	33
5.7	Procedure to verify the dimensions of the patient's compartment .....	35
5.7.1	Type A and B road ambulances.....	35
5.7.2	Type C road ambulances .....	35
5.8	Procedure to verify the seats dimensions of the patient's compartment .....	36
5.9	Testing of the ventilation system.....	36
5.10	Testing of the heating system.....	37
5.11	Testing of the cooling system .....	37
5.11.1	Test procedure .....	37
5.11.2	Testing of independent air conditioning system .....	38
5.12	Testing of interior lighting.....	38
5.13	Testing of infusion holding system .....	38
6	Equipment and medical devices.....	39
6.1	Provision of medical devices .....	39
6.2	Medical devices storage.....	39
6.3	Requirements for medical devices.....	39
6.3.1	General .....	39
6.3.2	Temperature .....	39
6.3.3	Humidity and ingress of liquids .....	39
6.3.4	Mechanical strength .....	39
6.3.5	Fixation of devices.....	40
6.3.6	Electrical safety .....	40
6.3.7	User interface.....	40
6.3.8	Gas installation .....	40
6.3.9	Marking and instructions .....	42
6.3.10	Maintenance .....	42
6.4	List of equipment.....	42
	Annex A (informative) Test summary .....	52
	Annex B (informative) Recognition.....	53
B.1	Recognition and visibility of ambulances.....	53
B.2	Recognition of crew .....	53
	Annex C (informative) Hygiene.....	54
	Annex D (informative) A-deviations .....	55
D.1	Deviation in Spain.....	55
	Annex ZA (informative) Relationship between this European Standard and the essential requirements of Directive 93/42/EEC [OJ L 169] aimed to be covered.....	56
	Bibliography .....	57

**EN 1789:2020 (E)****European foreword**

This document (EN 1789:2020) has been prepared by Technical Committee CEN/TC 239 “Rescue systems”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2021, and conflicting national standards shall be withdrawn at the latest by March 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 1789:2007+A2:2014.

This document has been prepared under a standardization request given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## Introduction

Road ambulances are subject to a higher risk in use. The exact circumstances of operation cannot always be planned or anticipated in detail.

Vehicles are designed so as to be safe. Design requirements can be derived from European and national occupational safety and health legislation.

Under EU law, employers are responsible for carrying out a risk assessment (89/391/EEC, OSH framework directive) and for provision of safe work equipment (89/655/EEC, use of work equipment directive) that allows employees to work without their health being at risk.

The document was first developed in the late 1990s to define a common approach to requirements to enhance patient and crew safety. The document has evolved and matured through several amendments and revisions.

This latest revision work of EN 1789 has had two key objectives:

- The first objective was to revise the technical side of the document with more manageable verification in mind, while maintaining the high quality and strict nature of the requirements.
- The second objective was to check all the references and regulations, paying special attention to EU regulations and updated standardization rules.

Testing of special purpose vehicle, such as an ambulance, is complex. Multiple functions (e.g. fixations, maintain systems, noise, illumination, heating, cooling etc.) may require numerous tests, which can be destructive. In this edition, carefully planned tests according to worst-case scenario strategies have reduced the number of destructive tests without sacrificing test qualities.

The previous edition of this standard (EN 1789:2007+A2:2014) contained a number of direct references to EU regulations. According to CEN Internal Regulations Part 3:2017 and to avoid duplication as well as outdated references and to enable use of this standard independently of the ECE rules, EU regulations and directives, these references have now been removed from the normative section of the standard.

This document is a reference document which can be used in support of regulations.

For the purpose of verification of an ambulance according to EU vehicle approval process, a section of EN 1789:2007+A1:2010+A2:2014 (i.e. patient's compartment) has been referenced directly in Regulation (EU) 2018/858.

CEN/TC 239 has agreed to a transition period of a maximum of 18 months in order to accommodate the different organisational structures that are necessary for the transport of patients are responsible for providing sufficient time for the technical implementation. At the date of publication of EN 1789, the presumption of conformity of the superseded standard has not yet been established in the Official Journal of the European Union. Users of the standard are invited to check the date in the Official Journal of the European Union against the transition period established by CEN/TC 239.

**EN 1789:2020 (E)****1 Scope**

This document specifies requirements for the design, testing, performance and equipping of road ambulances used for the transport, monitoring, treatment and care of patients. It contains requirements for the patient's compartment in terms of the working environment, ergonomic design and the safety of the crew and patients. This document does not cover the training of the crew, which is the responsibility of the authority/authorities in the country where the ambulance is to be registered.

This document is applicable to road ambulances capable of transporting at least one patient on a stretcher and excludes the transportation of hospital beds.

This document also specifies requirements for ambulances intended to carry transport incubator systems.

This document covers the specific requirements of each type of road ambulance, which are designated according to the patient condition.

This document gives general requirements for medical devices carried in road ambulances and used therein and outside hospitals and clinics in situations where the ambient conditions can differ from normal indoor conditions.

**2 Normative references**

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

CEN/TS 16165:2016, *Determination of slip resistance of pedestrian surfaces - Methods of evaluation*

DIN 51130:2014, *Testing of floor coverings - Determination of the anti-slip property - Workrooms and fields of activities with slip danger - Walking method - Ramp test*

EN 3-7:2004+A1:2007, *Portable fire extinguishers - Part 7: Characteristics, performance requirements and test methods*

EN 443:2008, *Helmets for fire fighting in buildings and other structures*

EN 455-1:2020, *Medical gloves for single use - Part 1: Requirements and testing for freedom from holes*

EN 455-2:2015, *Medical gloves for single use - Part 2: Requirements and testing for physical properties*

EN 794-3:1998+A2:2009, *Lung ventilators - Part 3: Particular requirements for emergency and transport ventilators*

EN 1041:2008+A1:2013, *Information supplied by the manufacturer of medical devices*

EN 1865-1:2010+A1:2015, *Patient handling equipment used in road ambulances - Part 1: General stretcher systems and patient handling equipment*

EN 1865-2:2010+A1:2015, *Patient handling equipment used in road ambulances - Part 2: Power assisted stretcher*

EN 1865-4:2012, *Patient handling equipment used in road ambulances - Part 4: Foldable patient transfer chair*

EN 1865-5:2012, *Patient handling equipment used in road ambulances - Part 5: Stretcher support*



EN 12470-1:2000+A1:2009, *Clinical thermometers - Part 1: Metallic liquid-in-glass thermometers with maximum device*

EN 13544-1:2019, *Respiratory therapy equipment - Part 1: Nebulizing systems and their components*

EN 13976-1:2018, *Rescue systems - Transportation of incubators - Part 1: Interface requirements*

EN 60601-1:2006+A1:2013, *Medical electrical equipment - Part 1: General requirements for basic safety and essential performance (IEC 60601-1:2005 + Cor.:2006 + Cor.:2007 + A1:2012)*

EN 60601-1-12:2015, *Medical electrical equipment - Part 1-12: General requirements for basic safety and essential performance - Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment*

EN 60601-2-4:2011, *Medical electrical equipment - Part 2-4: Particular requirements for the basic safety and essential performance of cardiac defibrillators*

EN 60601-2-27:2014, *Medical electrical equipment - Part 2-27: Particular requirements for the basic safety and essential performance of electrocardiographic monitoring equipment*

EN ISO 407:2004, *Small medical gas cylinders - Pin-index yoke-type valve connections (ISO 407:2004)*

EN ISO 5359:2014+A1:2017, *Anaesthetic and respiratory equipment - Low-pressure hose assemblies for use with medical gases (ISO 5359:2014 + Amd 1:2017)*

EN ISO 9170-1:—,<sup>1</sup> *Terminal units for medical gas pipeline systems — Part 1: Terminal units for use with compressed medical gases and vacuum (ISO/DIS 9170-1:2016)*

EN ISO 7396-1:2016+A1:2019, *Medical gas pipeline systems - Part 1: Pipeline systems for compressed medical gases and vacuum (ISO 7396-1:2016 + Amd 1:2017)*

EN ISO 10079-1:2015+A1:2019, *Medical suction equipment - Part 1: Electrically powered suction equipment (ISO 10079-1:2015 + Amd 1:2018)*

EN ISO 10079-2:2014, *Medical suction equipment - Part 2: Manually powered suction equipment (ISO 10079-2:2014)*

EN ISO 10079-3:2014, *Medical suction equipment - Part 3: Suction equipment powered from a vacuum or positive pressure gas source (ISO 10079-3:2014)*

EN ISO 10524-1:2019, *Pressure regulators for use with medical gases - Part 1: Pressure regulators and pressure regulators with flow-metering devices (ISO 10524-1:2018)*

EN ISO 10524-2:2019, *Pressure regulators for use with medical gases - Part 2: Manifold and line pressure regulators (ISO 10524-2:2018)*

EN ISO 10524-3:2019, *Pressure regulators for use with medical gases - Part 3: Pressure regulators integrated with cylinder valves (VIPRs) (ISO 10524-3:2019)*

EN ISO 11197:2018,<sup>2</sup> *Medical supply units (ISO/DIS 11197:2018)*

---

<sup>1</sup> Under preparation. Stage at time of publication: prEN ISO 9170-1:2017

**EN 1789:2020 (E)**

EN ISO 14971:2019, *Medical devices - Application of risk management to medical devices (ISO 14971:2019)*

EN ISO 15002:2008, *Flow-metering devices for connection to terminal units of medical gas pipeline systems (ISO 15002:2008)*

EN ISO 15223-1:2016, *Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements (ISO 15223-1:2016, Corrected version 2017-03)*

EN ISO 19054:2006+A1:2016, *Rail systems for supporting medical equipment (ISO 19054:2005+Amd1:2016)*

EN ISO 20471:2013+A1:2016, *High visibility clothing — Test methods and requirements (ISO 20471:2013, Corrected version 2013-06-01+Amd 1:2016)*

EN ISO 21420:2020-06, *Protective gloves - General requirements and test methods (ISO 21420:2020)*

EN ISO 80601-2-55:2018, *Medical electrical equipment - Part 2-55: Particular requirements for the basic safety and essential performance of respiratory gas monitors (ISO 80601-2-55:2018)*

EN ISO 80601-2-61:2019, *Medical electrical equipment - Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment (ISO 80601-2-61:2017, Corrected version 2018-02)*

IEC 60364-7-721:2017, *Low-voltage electrical installations — Part 7-721: Requirements for special installations or locations — Electrical installations in caravans and motor caravans*

ISO 3795:1989, *Road vehicles, and tractors and machinery for agriculture and forestry — Determination of burning behaviour of interior materials*

ISO 5128:1980, *Acoustics — Measurement of noise inside motor vehicles*

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**