

STN	Srdcovo-cievne implantáty Protézy srdcovej chlopne Časť 3: Náhrady srdcovej chlopne implantované minimálne invazívnymi technikami (ISO 5840-3: 2021)	STN EN ISO 5840-3 85 2922
------------	---	---

Cardiovascular implants - Cardiac valve prostheses - Part 3: Heart valve substitutes implanted by transcatheter techniques (ISO 5840-3:2021)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 07/21

Obsahuje: EN ISO 5840-3:2021, ISO 5840-3:2021

Oznámením tejto normy sa ruší
STN EN ISO 5840-3 (85 2926) z júla 2013

EUROPEAN STANDARD

EN ISO 5840-3

NORME EUROPÉENNE

EUROPÄISCHE NORM

February 2021

ICS 11.040.40

Supersedes EN ISO 5840-3:2013

English Version

Cardiovascular implants - Cardiac valve prostheses - Part 3: Heart valve substitutes implanted by transcatheter techniques (ISO 5840-3:2021)

Implants cardiovasculaires - Prothèses valvulaires -
Partie 3: Valves cardiaques de substitution implantées
par des techniques transcathéter (ISO 5840-3:2021)

Herz- und Gefäßimplantate - Herzklappenprothesen -
Teil 3: Durch minimal-invasive Methoden
implantierter Herzklappenersatz (ISO 5840-3:2021)

This European Standard was approved by CEN on 22 September 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 5840-3:2021 (E)**Contents**

Page

European foreword.....	3
-------------------------------	----------

European foreword

This document (EN ISO 5840-3:2021) has been prepared by Technical Committee ISO/TC 150 "Implants for surgery" in collaboration with Technical Committee CEN/TC 285 "Non-active surgical implants" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2021, and conflicting national standards shall be withdrawn at the latest by August 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 5840-3:2013.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 5840-3:2021 has been approved by CEN as EN ISO 5840-3:2021 without any modification.

INTERNATIONAL STANDARD

ISO
5840-3

Second edition
2021-01

Cardiovascular implants — Cardiac valve prostheses —

Part 3: Heart valve substitutes implanted by transcatheter techniques

Implants cardiovasculaires — Prothèses valvulaires —

*Partie 3: Valves cardiaques de substitution implantées par des
techniques transcathéter*



Reference number
ISO 5840-3:2021(E)

© ISO 2021

ISO 5840-3:2021(E)**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Abbreviations	4
5 Fundamental requirements	5
6 Device description	5
6.1 General	5
6.2 Intended use	6
6.3 Design inputs	6
6.3.1 Operational specifications	6
6.3.2 Performance specifications	6
6.3.3 Implant procedure	7
6.3.4 Packaging, labelling and sterilization	7
6.4 Design outputs	7
6.5 Design transfer (manufacturing verification/validation)	7
6.6 Risk management	7
7 Design verification and validation	8
7.1 General requirements	8
7.2 <i>In vitro</i> assessment	8
7.2.1 General	8
7.2.2 Test conditions, sample selection and reporting requirements	8
7.2.3 Material property assessment	8
7.2.4 Hydrodynamic performance assessment	8
7.2.5 Structural performance assessment	10
7.2.6 Design- or procedure-specific testing	10
7.2.7 Device MRI compatibility	12
7.2.8 Simulated use	12
7.2.9 Human factors and usability assessment	12
7.2.10 Implant thrombogenic and haemolytic potential assessment	12
7.3 Preclinical <i>in vivo</i> evaluation	12
7.3.1 General	12
7.3.2 Overall requirements	13
7.3.3 Methods	14
7.3.4 Test report	15
7.4 Clinical investigations	16
7.4.1 General	16
7.4.2 Study considerations	17
7.4.3 Study endpoints	18
7.4.4 Ethical considerations	19
7.4.5 Pivotal studies: Distribution of subjects and investigators	19
7.4.6 Statistical considerations including sample size and duration	20
7.4.7 Patient selection criteria	22
7.4.8 Valve thrombosis prevention	22
7.4.9 Clinical data requirements	23
Annex A (informative) Description of the transcatheter heart valve system	28
Annex B (informative) Transcatheter heart valve substitute hazard analysis example	30
Annex C (informative) Guidelines for verification of hydrodynamic performance — Pulsatile flow testing	32

ISO 5840-3:2021(E)

Annex D (normative) Requirements for delivery system design and evaluation	40
Annex E (informative) Examples of design specific testing	42
Annex F (informative) Preclinical <i>in vivo</i> evaluation.....	44
Annex G (normative) Adverse event classification during clinical investigation	47
Annex H (informative) Multimodality imaging of TAVI and TMVI pre, peri and post- implantation assessments — Examples.....	53
Bibliography	56

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 285, *Non-active surgical implants*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 5840-3:2013), which has been technically revised.

The main changes compared to the previous edition are as follows: the engineering and clinical requirements in the ISO 5840 series have been updated to current specifications and integrated and harmonized across all parts.

A list of all parts in the ISO 5840 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

ISO 5840-3:2021(E)

Introduction

This document has been prepared for transcatheter heart valve substitutes with emphasis on providing guidance for *in vitro* testing, preclinical *in vivo* and clinical evaluations, reporting of all *in vitro*, preclinical *in vivo*, and clinical evaluations and labelling and packaging of the device. This process is intended to clarify the required procedures prior to market release and to enable prompt identification and management of any subsequent issues.

This document is used in conjunction with ISO 5840-1 and ISO 5840-2.

Cardiovascular implants — Cardiac valve prostheses —

Part 3: Heart valve substitutes implanted by transcatheter techniques

1 Scope

This document is applicable to all devices intended for implantation as a transcatheter heart valve substitute.

This document is applicable to transcatheter heart valve substitutes and to the accessory devices, packaging and labelling required for their implantation and for determining the appropriate size of heart valve substitute to be implanted.

This document establishes an approach for verifying/validating the design and manufacture of a transcatheter heart valve substitute through risk management. The selection of appropriate verification/validation tests and methods are to be derived from the risk assessment. The tests can include those to assess the physical, chemical, biological and mechanical properties of heart valve substitutes and of their materials and components. The tests can also include those for preclinical *in vivo* evaluation and clinical evaluation of the finished heart valve substitute.

This document defines operational conditions and performance requirements for transcatheter heart valve substitutes where adequate scientific and/or clinical evidence exists for their justification.

This document includes considerations for implantation of a transcatheter heart valve substitute inside a pre-existing prosthetic device (e.g. valve-in-valve and valve-in-ring configurations).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5840-1:2021, *Cardiovascular implants — Cardiac valve prostheses — Part 1: General requirements*

ISO 10993-2, *Biological evaluation of medical devices — Part 2: Animal welfare requirements*

ISO 14155, *Clinical investigation of medical devices for human subjects — Good clinical practice*

ISO 14630, *Non-active surgical implants — General requirements*

ISO 14971, *Medical devices — Application of risk management to medical devices*

IEC 62366 (all parts), *Medical devices — Application of usability engineering to medical devices*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN