

STN	Rádiologická ochrana X a gama referenčné žiarenie na kalibráciu dozimetrov a dávkomerov a na stanovenie ich odzvy ako funkcie energie fotónov Časť 4: Kalibrácia priestorových a osobných dozimetrov v nízkoenergetických poliach referenčného žiarenia X (ISO 4037-4: 2019)	STN EN ISO 4037-4 40 1413
------------	---	---

Radiological protection - X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy - Part 4: Calibration of area and personal dosimeters in low energy X reference radiation fields (ISO 4037-4:2019)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 07/21

Obsahuje: EN ISO 4037-4:2021, ISO 4037-4:2019

133219

EUROPEAN STANDARD

EN ISO 4037-4

NORME EUROPÉENNE

EUROPÄISCHE NORM

February 2021

ICS 17.240

English Version

**Radiological protection - X and gamma reference radiation
for calibrating dosimeters and doserate meters and for
determining their response as a function of photon energy
- Part 4: Calibration of area and personal dosimeters in
low energy X reference radiation fields (ISO 4037-4:2019)**

Radioprotection - Rayonnements X et gamma de
référence pour l'étalonnage des dosimètres et des
débitmètres et pour la détermination de leur réponse
en fonction de l'énergie des photons - Partie 4:
Étalonnage des dosimètres de zone et individuels dans
des champs de référence X de faible énergie (ISO 4037-
4:2019)

Strahlenschutz - Röntgen- und Gamma-
Referenzstrahlungsfelder zur Kalibrierung von
Dosimetern und Dosisleistungsmessgeräten und zur
Bestimmung ihres Ansprechvermögens als Funktion
der Photonenenergie - Teil 4: Kalibrierung von Orts-
und Personendosimetern in niedrigerenergetischen
Röntgen-Referenzstrahlungsfeldern (ISO 4037-4:2019)

This European Standard was approved by CEN on 18 January 2021.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

European foreword

The text of ISO 4037-4:2019 has been prepared by Technical Committee ISO/TC 85 "Nuclear energy, nuclear technologies, and radiological protection" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 4037-4:2021 by Technical Committee CEN/TC 430 "Nuclear energy, nuclear technologies, and radiological protection" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2021, and conflicting national standards shall be withdrawn at the latest by August 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 4037-4:2019 has been approved by CEN as EN ISO 4037-4:2021 without any modification.

**INTERNATIONAL
STANDARD****ISO
4037-4**Second edition
2019-01

**Radiological protection — X and
gamma reference radiation for
calibrating dosimeters and doserate
meters and for determining their
response as a function of photon
energy —****Part 4:
Calibration of area and personal
dosimeters in low energy X reference
radiation fields***Radioprotection — Rayonnements X et gamma de référence
pour l'étalonnage des dosimètres et des débitmètres et pour la
détermination de leur réponse en fonction de l'énergie des photons —**Partie 4: Étalonnage des dosimètres de zone et individuels dans des
champs de référence X de faible énergie*Reference number
ISO 4037-4:2019(E)

© ISO 2019

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Symbols (and abbreviated terms).....	2
5 General procedures for calibrating and determining response.....	3
6 Characterisation and production of low energy X-ray reference radiations.....	3
6.1 General.....	3
6.2 Tube potential.....	4
6.3 Spectral fluence and conversion coefficients.....	4
7 Dosimetry of low energy reference radiations.....	4
7.1 General.....	4
7.2 Stability check facility.....	4
8 Calibration and determination of the response as a function of photon energy and angle of radiation incidence.....	4
8.1 General.....	4
8.2 Selection of calibration method.....	5
8.3 Calibration by using reference instruments for K_a	5
8.3.1 General.....	5
8.3.2 Conventional quantity value of the air kerma.....	5
8.3.3 Conventional quantity value of the dose equivalent quantities $H_p(0,07)$ and $H'(0,07)$	6
8.3.4 Conventional quantity value of the dose equivalent quantities $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$	6
8.3.5 Performing the calibration.....	8
8.4 Calibration by using reference instruments which measure the ICRU dose equivalent quantities.....	8
8.4.1 General.....	8
8.4.2 Conventional quantity value of the dose equivalent quantities $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$	9
8.4.3 Performing the calibration.....	10
8.5 Statement of uncertainty.....	11
Annex A (normative) Correction for air density.....	12
Bibliography.....	18

ISO 4037-4:2019(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies and radiological protection*, Subcommittee SC 2, *Radiological protection*.

This second edition cancels and replaces the first edition (ISO 4037-4:2004), which has been technically revised.

A list of all the parts in the ISO 4037 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

The maintenance release of this document adjusts this fourth part to the second edition of the first three parts. This includes the improvements on high voltage generators from 1996 to 2017 (e.g., the use of high frequency switching supplies providing nearly constant potential), and the spectral measurements at irradiation facilities equipped with such generators (e.g., the catalogue of X-ray spectra by Ankerhold[1]). It also incorporates all published information with the aim to adjust the requirements for the technical parameters of the reference fields to the targeted overall uncertainty of about 6 % to 10 % for the phantom related operational quantities of the International Commission on Radiation Units and Measurements (ICRU)[2]. It does not change the concept of ISO 4037.

ISO 4037, focusing on photon reference radiation fields, is divided into four parts. ISO 4037-1 gives the methods of production and characterization of reference radiation fields in terms of the quantities photon fluence and air kerma free-in-air. ISO 4037-2 describes the dosimetry of the reference radiation qualities in terms of air kerma and in terms of the phantom related operational quantities of the International Commission on Radiation Units and Measurements (ICRU)[2]. ISO 4037-3 describes the methods for calibrating and determining the response of dosimeters and doserate meters in terms of the operational quantities of the ICRU[2]. This document gives special considerations and additional requirements for calibration of area and personal dosimeters in low energy X reference radiation fields, which are reference fields with generating potential lower or equal to 30 kV.

The general procedures described in ISO 29661 including Amendment 1 are used as far as possible in this document. In addition, the symbols used are in line with ISO 29661.

NOTE For irradiation of the whole body, $H_p(10)$ and $H^*(10)$ are relevant for radiation protection, as long as they are closer to their limit than $H'(0,07)$ and $H_p(0,07)$. This is the case down to about 15 keV.

Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy —

Part 4: Calibration of area and personal dosimeters in low energy X reference radiation fields

1 Scope

This document gives guidelines on additional aspects of the characterization of low energy photon radiations and on the procedures for calibration and determination of the response of area and personal dose(rate)meters as a function of photon energy and angle of incidence. This document concentrates on the accurate determination of conversion coefficients from air kerma to $H_p(10)$, $H^*(10)$, $H_p(3)$ and $H'(3)$ and for the spectra of low energy photon radiations. As an alternative to the use of conversion coefficients the direct calibration in terms of these quantities by means of appropriate reference instruments is described.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 4037-1, *Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 1: Radiation characteristics and production methods*

ISO 4037-2:2019, *Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 2: Dosimetry for radiation protection over the energy ranges from 8 keV to 1,3 MeV and 4 MeV to 9 MeV*

ISO 4037-3:2019, *Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 3: Calibration of area and personal dosimeters and the measurement of their response as a function of energy and angle of incidence*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO 29661, *Reference radiation fields for radiation protection — Definitions and fundamental concepts*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN