

STN P	Molekulárno-diagnostické vyšetrenia <i>in vitro</i> Špecifikácie postupov pred vyšetrením aspiračnej biopsie tenkou ihlou (FNAs) Časť 1: Izolovaná bunková RNA	STN P CEN/TS 17688-1 85 6577
------------------	---	--

Molecular in vitro diagnostic examinations - Specifications for pre-examination processes for Fine Needle Aspirates (FNAs) - Part 1: Isolated cellular RNA

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 02/22

Táto predbežná STN je určená na overenie. Pripomienky zasielajte ÚNMS SR najneskôr do 15. 11. 2023.

Obsahuje: CEN/TS 17688-1:2021

134509



TECHNICAL SPECIFICATION
SPÉCIFICATION TECHNIQUE
TECHNISCHE SPEZIFIKATION

CEN/TS 17688-1

December 2021

ICS 11.100.10

English Version

**Molecular in vitro diagnostic examinations - Specifications
for pre-examination processes for Fine Needle Aspirates
(FNAs) - Part 1: Isolated cellular RNA**

Analyses moléculaires de diagnostic in vitro -
Spécifications pour les processus préanalytiques pour
les ponctions à l'aiguille fine - Partie 1 : ARN cellulaire
extrait

Molekularanalytische in-vitro-diagnostische Verfahren
- Spezifikationen für präanalytische Prozesse für
Feinnadelaspiration (FNA) - Teil 1: Isolierte zelluläre
RNA

This Technical Specification (CEN/TS) was approved by CEN on 15 November 2021 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

CEN/TS 17688-1:2021 (E)

Contents	Page
European foreword.....	4
Introduction	5
1 Scope.....	6
2 Normative references.....	6
3 Terms and definitions	6
4 General considerations.....	12
5 Outside the laboratory	13
5.1 Specimen collection.....	13
5.1.1 General.....	13
5.1.2 Information about the patient/specimen donor	13
5.1.3 Information about the specimen.....	14
5.1.4 Selection of the primary FNA collection device.....	14
5.1.5 FNA specimen collection and stabilization from the donor/patient.....	14
5.2 Specimen storage and transport	16
6 Inside the laboratory	16
6.1 Specimen reception.....	16
6.2 Specimen/sample storage after transport and reception.....	16
6.2.1 General.....	16
6.2.2 Storage of FNA specimen/samples using collection devices with stabilizer	17
6.2.3 Storage of FNA specimen/samples using collection devices without stabilizers	17
6.3 Specimen/sample processing for cytological examination prior to RNA isolation	17
6.3.1 General.....	17
6.3.2 Handling of cell suspension.....	18
6.3.3 Preparation of paraffin-embedded cell blocks.....	19
6.3.4 Preparation of cell suspension slides	20
6.4 Evaluation of the pathology of the specimen or sample(s).....	20
6.5 Processed sample storage, transport and reception.....	21
6.5.1 General.....	21
6.5.2 Storage and transport of cell suspension	21
6.5.3 Storage and transport of paraffin-embedded cell blocks.....	21
6.5.4 Storage and transport of cell suspension slides	22
6.6 Isolation of RNA	22
6.6.1 General.....	22
6.6.2 Using a commercial RNA isolation kit intended for diagnostic use	22
6.6.3 Using a laboratory developed RNA isolation procedure.....	23
6.6.4 Further specifications	23
6.7 Quantity and quality assessment of isolated RNA.....	24
6.7.1 General.....	24
6.7.2 Quantity assessment.....	25
6.7.3 Quality assessment.....	25
6.8 Storage of isolated RNA.....	25
6.8.1 General.....	25
6.8.2 RNA isolated using commercially available kits.....	26
6.8.3 RNA isolated using a laboratory developed procedure	26

Annex A (informative) Impact of pre-analytical variables on FNA sample quality, RNA quantity and quality	27
A.1 Introduction.....	27
A.2 Method – FNA model sample	27
A.3 Result - Impact of specimen stabilization method on FNA cell quality.....	28
A.3.1 General	28
A.3.2 Method.....	28
A.3.3 Result/conclusion.....	30
A.4 Result - Impact of FNA sample stabilization on isolated RNA quality.....	31
A.4.1 General	31
A.4.2 Method.....	31
A.4.3 Result/conclusion, part 1	31
A.4.4 Result/conclusion, part 2	33
A.4.5 Result/conclusion, part 3	34
A.5 Conclusions	36
Bibliography	37

CEN/TS 17688-1:2021 (E)**European foreword**

This document (CEN/TS 17688-1:2021) has been prepared by Technical Committee CEN/TC 140 “In vitro diagnostic medical devices”, the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users’ national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

Molecular *in vitro* diagnostics has enabled significant progress in medicine. Further progress is expected by new technologies analysing profiles of nucleic acids, proteins, and metabolites in human tissues and body fluids. However, the profiles of these molecules can change drastically during the pre-examination process, including the specimen collection, transport, storage and processing.

Examination of RNA is commonly used in clinical practice. This includes e.g. prognostic and predictive biomarker examinations. This is a fast growing field in molecular diagnostics.

Fine needle aspiration is a non-surgical procedure that uses a thin, hollow-bore needle and syringe to collect a specimen from patients for cytopathological and molecular investigation. As a minimally-invasive technique, fine needle aspirates (FNAs) are commonly used to diagnose and monitor for example a range of cancer types, e.g. breast, lung and thyroid cancer, and other diseases, such as inflammatory diseases. FNAs also provide the opportunity to sample metastatic sites (e.g. lymph nodes) and otherwise non-resectable tissues.

Besides cytological assessment, molecular biological analysis of FNAs is expected to become increasingly used for cancer and other disease diagnostics, including companion diagnostics.

One of the challenges facing molecular analysis of FNA samples is their small size and diversity in composition (cells, blood, body fluid). The low cellular content of FNAs means that the yield of isolated RNA is typically towards the lower end of detection for molecular examination. Therefore, the RNA isolation procedure should provide a sufficient amount of RNA as required by the specific examination.

RNA profiles and RNA integrities in FNAs can change drastically [15] during and after collection (due to, e.g. gene induction, gene down regulation and RNA degradation). RNA expression can change differently in different donors'/patients' FNAs.

Therefore, standardization of the entire process from specimen collection to RNA examination is needed to minimize RNA degradation and other RNA profile changes during and after FNA collection. This document describes special measures which need to be taken to obtain good quality FNA specimens/samples and isolated RNA therefrom for molecular examination.

In this document, the following verbal forms are used:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” indicates a possibility or a capability.

CEN/TS 17688-1:2021 (E)

1 Scope

This document gives guidelines on the handling, documentation, storage and processing of fine needle aspirates (FNAs) intended for RNA examination during the pre-examination phase before a molecular examination is performed.

This document is applicable to molecular *in vitro* diagnostic examination including laboratory developed tests performed by medical laboratories and molecular pathology laboratories that examine RNA isolated from FNAs. It is also intended to be used by laboratory customers, *in vitro* diagnostics developers and manufacturers, biobanks, institutions and commercial organisations performing biomedical research, and regulatory authorities.

Different dedicated measures are taken for collecting, stabilizing, transporting and storing of core needle biopsies (FNA Biopsy or FNA B) and are not covered in this document, but in EN ISO 20184-1, *Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for frozen tissue — Part 1: Isolated RNA* and EN ISO 20166-1, *Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for formalin fixed and paraffin-embedded (FFPE) tissue — Part 1: Isolated RNA*.

This document is not applicable for RNA examination by *in situ* detection.

NOTE International, national or regional regulations or requirements can also apply to specific topics covered in this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 15189, *Medical laboratories — Requirements for quality and competence (ISO 15189)*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN