

<b>TNI</b>	<b>Aditívna výroba pre zdravotníctvo Údaje Optimalizované zdravotnícke obrazové údaje (ISO/ASTM TR 52916: 2022)</b>	<b>TNI CEN ISO/ASTM/TR 52916</b>  18 0057
------------	---	---

Additive manufacturing for medical - Data - Optimized medical image data (ISO/ASTM TR 52916:2022)

Táto technická normalizačná informácia obsahuje anglickú verziu CEN ISO/ASTM/TR 52916:2022, ISO/ASTM TR 52916:2022.

This Technical standard information includes the English version of CEN ISO/ASTM/TR 52916:2022, ISO/ASTM TR 52916:2022.

Táto technická normalizačná informácia bola oznámená vo Vestníku ÚNMS SR č. 04/22

**134755**

TECHNICAL REPORT  
RAPPORT TECHNIQUE  
TECHNISCHER BERICHT

**CEN ISO/ASTM/TR**  
**52916**

February 2022

---

ICS 25.030

English Version

**Additive manufacturing for medical - Data - Optimized  
medical image data (ISO/ASTM TR 52916:2022)**

Fabrication additive dans le secteur médical - Données  
- Données d'images médicales optimisées (ISO/ASTM  
TR 52916:2022)

Additive Fertigung - Datenformate - Normspezifikation  
für optimierte medizinische Bilddaten (ISO/ASTM TR  
52916:2022)

This Technical Report was approved by CEN on 18 January 2022. It has been drawn up by the Technical Committee CEN/TC 438.

This European Standard was corrected and reissued by the CEN-CENELEC Management Centre on 09 March 2022.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

---

**CEN ISO/ASTM/TR 52916:2022 (E)**

<b>Contents</b>	<b>Page</b>
<b>European foreword.....</b>	<b>3</b>

## **European foreword**

This document (CEN ISO/ASTM/TR 52916:2022) has been prepared by Technical Committee ISO/TC 261 "Additive manufacturing" in collaboration with Technical Committee CEN/TC 438 "Additive Manufacturing" the secretariat of which is held by AFNOR.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

## **Endorsement notice**

The text of ISO/ASTM TR 52916:2022 has been approved by CEN as CEN ISO/ASTM/TR 52916:2022 without any modification.

**TECHNICAL  
REPORT**

**ISO/ASTM TR  
52916**

First edition  
2022-01

---

---

**Additive manufacturing for medical —  
Data — Optimized medical image data**

*Fabrication additive dans le secteur médical — Données — Données  
d'images médicales optimisées*



Reference number  
ISO/ASTM TR 52916:2022(E)

© ISO/ASTM International 2022

**ISO/ASTM TR 52916:2022(E)****COPYRIGHT PROTECTED DOCUMENT**

© ISO/ASTM International 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester. In the United States, such requests should be sent to ASTM International.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11

Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

ASTM International  
100 Barr Harbor Drive, PO Box C700  
West Conshohocken, PA 19428-2959, USA  
Phone: +610 832 9634  
Fax: +610 832 9635  
Email: [khooper@astm.org](mailto:khooper@astm.org)  
Website: [www.astm.org](http://www.astm.org)

# Contents

Page

<b>Foreword</b> .....	<b>v</b>
<b>Introduction</b> .....	<b>vi</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Medical images generation for AM</b> .....	<b>3</b>
4.1 General medical image data generation.....	3
4.2 General error occurrence steps in medical images generation.....	3
4.3 Medical image extraction.....	4
4.3.1 Introduction of medical image extraction.....	4
4.3.2 CT image error generation factors.....	4
4.3.3 MRI Image error generation factors.....	5
<b>5 Image segmentation</b> .....	<b>6</b>
5.1 Introduction of segmentation.....	6
5.2 Segmentation techniques.....	6
5.2.1 Thresholding algorithm.....	6
5.2.2 Region growing algorithm.....	6
5.2.3 Morphological image algorithm.....	7
5.2.4 Level-set algorithm.....	7
5.2.5 Other partial segmentation algorithm.....	7
<b>6 Reconstruction</b> .....	<b>7</b>
6.1 Introduction of reconstruction.....	7
6.2 Reconstruction process.....	7
<b>7 Smoothing</b> .....	<b>8</b>
7.1 Marching cubes.....	8
7.2 Mesh smoothing.....	8
<b>8 3D visualization method</b> .....	<b>8</b>
8.1 Surface rendering.....	8
8.1.1 Introduction of surface shaded rendering.....	8
8.1.2 Surface shaded rendering feature.....	9
8.2 Volume rendering.....	9
8.2.1 Introduction of volume rendering.....	9
8.2.2 Volume rendering feature.....	9
8.2.3 Ray casting techniques.....	9
8.2.4 3D texture mapping techniques.....	9
<b>9 Additional processing for additive manufacturing</b> .....	<b>10</b>
<b>10 Methods</b> .....	<b>10</b>
10.1 Image isotropic conversion.....	10
10.2 Image enhancement.....	11
10.3 Image segmentation.....	12
<b>11 Minimizing error of software and equipment</b> .....	<b>14</b>
11.1 Introduction of software and equipment error.....	14
11.2 Software error.....	14
11.2.1 Background.....	14
11.2.2 Verification method using main inflection.....	14
11.2.3 Improving accuracy and precision.....	14
11.3 Equipment error.....	15
11.3.1 Background.....	15
11.3.2 Standard computational mesh model data creation for an evaluation method.....	15

**ISO/ASTM TR 52916:2022(E)**

11.4 Tolerance error situations.....	15
<b>Annex A (informative) Medical CAD for additive manufacturing tolerance.....</b>	<b>16</b>
<b>Bibliography.....</b>	<b>24</b>



## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by ISO/TC 261, *Additive manufacturing*, in cooperation with ASTM Committee F42, *Additive Manufacturing Technologies*, on the basis of a partnership agreement between ISO and ASTM International with the aim to create a common set of ISO/ASTM standards on additive manufacturing, and in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 438, *Additive manufacturing*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## ISO/ASTM TR 52916:2022(E)

### Introduction

This document has been developed in close cooperation of ISO/TC 261 and ASTM F 42 on basis of a partnership agreement between ISO and ASTM international with the aim to create a common set of ISO/ASTM standards on additive manufacturing.

Digital imaging and communications in medicine (DICOM) image files cannot be used directly for 3D printing; further steps are necessary to make them readable by additive manufacturing system. In particular, as the thickness of the computed tomography slice increases, there is a problem that the error in 3D reconstruction of the anatomical structure increases. Therefore, the focus of this technical report is to automatically reconfigure the slice interval through the application of isotropic conversion technology to utilize the existing dicom file and visualization and editing software as it is. In addition, in order to present a method for optimized medical image data for additive manufacturing, tomography metadata without compression is used by editing and processing the output format file without loss in the AM equipment system, or tomography within the maximum allowable range of radiation. Consider reducing the spacing of slices as much as possible and increasing the resolution per image as much as possible.

This document benefits from the direction of development and high quality additive manufacturing output through the technical optimization of medical imaging for additive manufacturing: medical academics, clinic and industry fields for AM like as anatomical measurements, 3D analysis, finite element analysis and surgical planning or simulation, patient-specific implant and device design. There are many affected stakeholder like as medical AM system manufacturer, AM feedstock manufacturer, AM feedstock supplier and vendor, medical AM hardware manufacturer, medical AM software manufacturer, medical AM system manufacturer, medical AM platform manufacturer, AM based medical device manufacturer, medical 3D scanning and digitizing device manufacturer, surgical simulation AM model manufacturer, AM surgical implant manufacturer, AM surgical guide manufacturer, AM physical model for clinical education and diagnostic treatment, disposable medical AM consumable devices.

# Additive manufacturing for medical — Data — Optimized medical image data

## 1 Scope

This document includes the creation of optimized data for medical additive manufacturing (MAM). These data are generated from static modalities, such as magnetic resonance imaging (MRI), computed tomography (CT). This document addresses improved medical image data, and medical image data acquisition processing and optimization approaches for accurate solid medical models, based on real human and animal data.

Solid medical models are generally created from stacked 2D images output from medical imaging systems. The accuracy of the final model depends on the resolution and accuracy of the original image data. The main factors influencing accuracy are the resolution of the image, the amount of image noise, the contrast between the tissues of interest and artefacts inherent in the imaging system.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/ASTM 52900, *Additive manufacturing — General principles — Fundamentals and vocabulary*

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**