

STN	Anestetické a dýchacie prístroje Prístroje na reguláciu koncentrácie vzduchu (ISO 23372: 2022)	STN EN ISO 23372 85 2118
------------	---	--

Anaesthetic and respiratory equipment - Air entrainment devices (ISO 23372:2022)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 07/22

Obsahuje: EN ISO 23372:2022, ISO 23372:2022

Oznámením tejto normy sa ruší
STN EN 13544-3+A1 (85 2125) z januára 2010

135328

EUROPEAN STANDARD

EN ISO 23372

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2022

ICS 11.040.10

Supersedes EN 13544-3:2001+A1:2009

English Version

Anaesthetic and respiratory equipment - Air entrainment devices (ISO 23372:2022)Matériel d'anesthésie et de réanimation respiratoire -
Dispositifs d'entraînement d'air(ISO 23372:2022)Atemtherapiegeräte - Luftbeimischgeräte (ISO
23372:2022)

This European Standard was approved by CEN on 7 February 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

EN ISO 23372:2022 (E)

Contents	Page
European foreword.....	3

European foreword

This document (EN ISO 23372:2022) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2022, and conflicting national standards shall be withdrawn at the latest by November 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 13544-3:2001+A1:2009.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 23372:2022 has been approved by CEN as EN ISO 23372:2022 without any modification.

INTERNATIONAL
STANDARD

ISO
23372

First edition
2022-05

**Anaesthetic and respiratory
equipment — Air entrainment devices**

*Matériel d'anesthésie et de réanimation respiratoire — Dispositifs
d'entraînement d'air*



Reference number
ISO 23372:2022(E)

© ISO 2022

ISO 23372:2022(E)**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General requirements	2
5 Materials	2
5.1 General.....	2
5.2 Biocompatibility of breathing gas pathways.....	2
6 Design requirements	2
6.1 General.....	2
6.2 Oxygen inlet connectors.....	3
6.3 Outlet connectors.....	3
6.4 Air inlet attachments.....	3
7 Information to be provided by the manufacturer	3
7.1 General.....	3
7.2 Marking.....	3
7.3 Instructions for use.....	4
Annex A (normative) Test method for delivered oxygen concentration	5
Bibliography	8

ISO 23372:2022(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 2, *Airways and related equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Air entrainment devices, commonly known as venturi masks, are used to provide a known concentration of oxygen to a patient at a known set flow. This is achieved by driving the oxygen through a controlled diameter orifice and entraining room air through side openings. These devices are available in various concentrations and can ensure continuity over a long period of time within relatively close limits of accuracy.

However, the use of these devices does not guarantee that the patient receives the designated oxygen concentration as there are physiological factors such as the patient's ventilatory pattern, lung compliance and airway resistance, and physical factors such as the fit of the mask, movement by the patient, etc^[2].

Anaesthetic and respiratory equipment — Air entrainment devices

1 Scope

This document specifies minimum performance and safety requirements for *air entrainment devices* used for delivery of designated oxygen concentrations to patients. It provides a test method to check the accuracy of the oxygen concentration in the air/oxygen mixture generated by the *air entrainment devices*. *Air entrainment devices* can be fixed to deliver a single oxygen concentration or adjustable, to deliver a range of oxygen concentration outputs.

This document also specifies marking requirements and recommends an optional system of colour coding to assist the user in identifying the designated oxygen concentration.

This document does not cover *air entrainment devices* which are integral with medical devices specified in other standards (e.g. emergency lung ventilators, humidifiers, nebulizers).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5356-1, *Anaesthetic and respiratory equipment — Conical connectors — Part 1: Cones and sockets*

ISO 15002, *Flow-metering devices for connection to terminal units of medical gas pipeline systems*

ISO 18190:2016, *Anaesthetic and respiratory equipment — General requirements for airways and related equipment*

ISO 18562-1, *Biocompatibility evaluation of breathing gas pathways in healthcare applications — Part 1: Evaluation and testing within a risk management process*

ISO 20417, *Medical devices — Information to be supplied by the manufacturer*

ISO 80369-2¹⁾, *Small-bore connectors for liquids and gases in healthcare applications — Part 2: Connectors for respiratory applications*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN

1) Under preparation. Stage at the time of publication: ISO/DIS 80369-2:2022.