

|            |   |   |
|------------|---|---|
| <b>STN</b> | <b>Plasty</b><br><b>Stanovenie hmotnostného indexu toku taveniny (MFR) a objemového indexu toku taveniny (MVR) termoplastov</b><br><b>Časť 1: Normalizovaná metóda (ISO 1133-1: 2022)</b> | <b>STN</b><br><b>EN ISO 1133-1</b><br><br>64 0861 |
|------------|---|---|

Plastics - Determination of the melt mass-flow rate (MFR) and melt volume-flow rate (MVR) of thermoplastics - Part 1: Standard method (ISO 1133-1:2022)

Táto norma obsahuje anglickú verziu európskej normy.  
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 09/22

Obsahuje: EN ISO 1133-1:2022, ISO 1133-1:2022

Oznámením tejto normy sa ruší  
STN EN ISO 1133-1 (64 0861) z októbra 2017

**135540**

EUROPEAN STANDARD

EN ISO 1133-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

July 2022

ICS 83.080.20

Supersedes EN ISO 1133-1:2011

English Version

## Plastics - Determination of the melt mass-flow rate (MFR) and melt volume-flow rate (MVR) of thermoplastics - Part 1: Standard method (ISO 1133-1:2022)

Plastiques - Détermination de l'indice de fluidité à chaud des thermoplastiques, en masse (MFR) et en volume (MVR) - Partie 1: Méthode normale (ISO 1133-1:2022)

Kunststoffe - Bestimmung der Schmelze-Massefließrate (MFR) und der Schmelze-Volumenfließrate (MVR) von Thermoplasten - Teil 1: Allgemeines Prüfverfahren (ISO 1133-1:2022)

This European Standard was approved by CEN on 25 June 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

**EN ISO 1133-1:2022 (E)**

| <b>Contents</b>               | <b>Page</b> |
|-------------------------------|-------------|
| <b>European foreword.....</b> | <b>3</b>    |

## **European foreword**

This document (EN ISO 1133-1:2022) has been prepared by Technical Committee ISO/TC 61 "Plastics" in collaboration with Technical Committee CEN/TC 249 "Plastics" the secretariat of which is held by NBN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2023, and conflicting national standards shall be withdrawn at the latest by January 2023.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 1133-1:2011.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## **Endorsement notice**

The text of ISO 1133-1 has been approved by CEN as EN ISO 1133-1:2022 without any modification.

# INTERNATIONAL STANDARD

# ISO 1133-1

Second edition  
2022-06

---

---

## Plastics — Determination of the melt mass-flow rate (MFR) and melt volume-flow rate (MVR) of thermoplastics —

### Part 1: Standard method

*Plastiques — Détermination de l'indice de fluidité à chaud des  
thermoplastiques, en masse (MFR) et en volume (MVR) —*

*Partie 1: Méthode normale*



Reference number  
ISO 1133-1:2022(E)

© ISO 2022

**ISO 1133-1:2022(E)****COPYRIGHT PROTECTED DOCUMENT**

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

|  | Page      |
|--|-----------|
| Foreword.....  | v         |
| Introduction.....  | vi        |
| <b>1 Scope.....</b>  | <b>1</b>  |
| <b>2 Normative references.....</b>   | <b>1</b>  |
| <b>3 Terms and definitions.....</b>  | <b>1</b>  |
| <b>4 Principle.....</b>  | <b>2</b>  |
| <b>5 Apparatus.....</b>  | <b>3</b>  |
| 5.1 Extrusion plastometer.....   | 3         |
| 5.2 Accessory equipment.....   | 7         |
| 5.2.1 General.....   | 7         |
| 5.2.2 Equipment for procedure A (see <a href="#">Clause 8</a> ).....   | 8         |
| 5.2.3 Equipment for procedure B (see <a href="#">Clause 9</a> ): Piston displacement transducer/<br>timer.....   | 8         |
| <b>6 Test sample.....</b>  | <b>8</b>  |
| 6.1 Sample form.....   | 8         |
| 6.2 Conditioning.....  | 9         |
| <b>7 Temperature verification, cleaning and maintenance of the apparatus.....</b>  | <b>9</b>  |
| 7.1 Verification of the temperature control system.....  | 9         |
| 7.1.1 Verification procedure.....  | 9         |
| 7.1.2 Material used during temperature verification.....   | 10        |
| 7.2 Cleaning the apparatus.....  | 10        |
| 7.3 Vertical alignment of the instrument.....  | 10        |
| <b>8 Procedure A: mass-measurement method.....</b>   | <b>10</b> |
| 8.1 Selection of temperature and load.....   | 10        |
| 8.2 Cleaning.....  | 11        |
| 8.3 Selection of sample mass and charging the cylinder.....  | 11        |
| 8.4 Measurements.....  | 12        |
| 8.5 Expression of results.....   | 13        |
| 8.5.1 General.....   | 13        |
| 8.5.2 Expression of results: standard die.....   | 13        |
| 8.5.3 Expression of results: half size die.....  | 13        |
| <b>9 Procedure B: displacement-measurement method.....</b>   | <b>14</b> |
| 9.1 Selection of temperature and load.....   | 14        |
| 9.2 Cleaning.....  | 14        |
| 9.3 Minimum piston displacement distance.....  | 14        |
| 9.4 Selection of sample mass and charging the cylinder.....  | 14        |
| 9.5 Measurements.....  | 14        |
| 9.6 Expression of results.....   | 15        |
| 9.6.1 General.....   | 15        |
| 9.6.2 Expression of results: standard die.....   | 15        |
| 9.6.3 Expression of results: half size die.....  | 16        |
| <b>10 Flow rate ratio.....</b>   | <b>16</b> |
| <b>11 Precision.....</b>   | <b>17</b> |
| <b>12 Test report.....</b>   | <b>17</b> |
| <b>Annex A (normative) Test conditions for MFR and MVR determinations.....</b>   | <b>19</b> |
| <b>Annex B (informative) Conditions specified in International Standards for the<br/>determination of the melt flow rate of thermoplastic materials.....</b> | <b>21</b> |

**ISO 1133-1:2022(E)**

|   |           |
|---|-----------|
| <b>Annex C (informative) Device and procedure for performing a compacted charge of material by compression</b> .....        | <b>22</b> |
| <b>Annex D (informative) Precision data for polypropylene obtained from an intercomparison of MFR and MVR testing</b> ..... | <b>25</b> |
| <b>Bibliography</b> .....   | <b>26</b> |

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 61, *Plastics*, Subcommittee SC 5, *Physical-chemical properties*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 249, *Plastics*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 1133-1:2011), of which it constitutes a minor revision. The changes are as follows:

- references to withdrawn standards in [Annex B](#) (informative), [Annex D](#) (informative) and Bibliography have been updated;
- editorial corrections.

A list of all parts in the ISO 1133 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## ISO 1133-1:2022(E)

### Introduction

For stable materials that are not rheologically sensitive to the time-temperature history experienced during melt flow rate testing, this document is recommended.

For materials whose rheological behaviour is sensitive to the test's time-temperature history, e.g. materials which degrade during the test, ISO 1133-2 is recommended. Also, ISO 1133-2 is considered to be particularly relevant for moisture-sensitive plastics.

**NOTE** At the time of publication, there is no evidence to suggest that the use of ISO 1133-2 for stable materials results in better precision in comparison with the use of this document.

# Plastics — Determination of the melt mass-flow rate (MFR) and melt volume-flow rate (MVR) of thermoplastics —

## Part 1: Standard method

**WARNING** — Persons using this document should be familiar with normal laboratory practice, if applicable. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any regulatory requirements.

### 1 Scope

This document specifies two procedures for the determination of the melt mass-flow rate (MFR) and the melt volume-flow rate (MVR) of thermoplastic materials under specified conditions of temperature and load. Procedure A is a mass-measurement method. Procedure B is a displacement-measurement method. Normally, the test conditions for measurement of melt flow rate are specified in the material standard with a reference to this document. The test conditions normally used for thermoplastics are listed in [Annex A](#).

The MVR is particularly useful when comparing materials of different filler content and when comparing filled with unfilled thermoplastics. The MFR can be determined from MVR measurements, or vice versa, provided the melt density at the test temperature is known.

This document is also possibly applicable to thermoplastics for which the rheological behaviour is affected during the measurement by phenomena such as hydrolysis (chain scission), condensation and cross-linking, but only if the effect is limited in extent and only if the repeatability and reproducibility are within an acceptable range. For materials which show significantly affected rheological behaviour during testing, this document is not appropriate. In such cases, ISO 1133-2 applies.

**NOTE** The rates of shear in these methods are much smaller than those used under normal conditions of processing, and therefore it is possible that data obtained by these methods for various thermoplastics will not always correlate with their behaviour during processing. Both methods are used primarily in quality control.

### 2 Normative references

There are no normative references in this document.

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**