

| | | |
|------------|--|---|
| STN | Meranie prietoku tekutiny pomocou zariadení so snímačmi diferenčného tlaku vložených do úplne vyplneného potrubia kruhového prierezu Časť 2: Clony (ISO 5167-2: 2022) | STN EN ISO 5167-2 25 7710 |
|------------|--|---|

Measurement of fluid flow by means of pressure differential devices inserted in circular cross-section conduits running full - Part 2: Orifice plates (ISO 5167-2:2022)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 09/22

Obsahuje: EN ISO 5167-2:2022, ISO 5167-2:2022

Oznámením tejto normy sa ruší
STN EN ISO 5167-2 (25 7710) z júla 2005

135586

EUROPEAN STANDARD

EN ISO 5167-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

June 2022

ICS 17.120.10

Supersedes EN ISO 5167-2:2003

English Version

Measurement of fluid flow by means of pressure
differential devices inserted in circular cross-section
conduits running full - Part 2: Orifice plates (ISO 5167-
2:2022)

Mesurage de débit des fluides au moyen d'appareils
déprimogènes insérés dans des conduites en charge de
section circulaire - Partie 2: Diaphragmes (ISO 5167-
2:2022)

Durchflussmessung von Fluiden mit Drosselgeräten in
voll durchströmten Leitungen mit Kreisquerschnitt -
Teil 2: Blenden (ISO 5167-2:2022)

This European Standard was approved by CEN on 17 June 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 5167-2:2022 (E)

| Contents | Page |
|-------------------------------|-------------|
| European foreword..... | 3 |

European foreword

This document (EN ISO 5167-2:2022) has been prepared by Technical Committee ISO/TC 30 "Measurement of fluid flow in closed conduits" in collaboration with CCMC.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2022, and conflicting national standards shall be withdrawn at the latest by December 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 5167-2:2003.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 5167-2:2022 has been approved by CEN as EN ISO 5167-2:2022 without any modification.

INTERNATIONAL STANDARD

ISO 5167-2

Second edition
2022-06

Measurement of fluid flow by means of pressure differential devices inserted in circular cross-section conduits running full —

Part 2: Orifice plates

*Mesurage de débit des fluides au moyen d'appareils déprimogènes
insérés dans des conduites en charge de section circulaire —*

Partie 2: Diaphragmes



Reference number
ISO 5167-2:2022(E)

© ISO 2022

ISO 5167-2:2022(E)**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

| | Page |
|---|-----------|
| Foreword | v |
| Introduction | vi |
| 1 Scope | 1 |
| 2 Normative references | 1 |
| 3 Terms, definitions and symbols | 1 |
| 4 Principles of the method of measurement and computation | 2 |
| 5 Orifice plates | 2 |
| 5.1 Description..... | 2 |
| 5.1.1 General..... | 2 |
| 5.1.2 General shape..... | 2 |
| 5.1.3 Upstream face A..... | 3 |
| 5.1.4 Downstream face B..... | 4 |
| 5.1.5 Thicknesses E and e | 4 |
| 5.1.6 Angle of bevel, α | 5 |
| 5.1.7 Edges G, H and I..... | 5 |
| 5.1.8 Diameter of orifice, d | 5 |
| 5.1.9 Bidirectional plates..... | 6 |
| 5.1.10 Material and manufacture..... | 6 |
| 5.2 Pressure tappings..... | 6 |
| 5.2.1 General..... | 6 |
| 5.2.2 Orifice plate with D and $D/2$ tappings or flange tappings..... | 6 |
| 5.2.3 Orifice plate with corner tappings..... | 8 |
| 5.3 Coefficients and corresponding uncertainties of orifice plates..... | 11 |
| 5.3.1 Limits of use..... | 11 |
| 5.3.2 Coefficients..... | 12 |
| 5.3.3 Uncertainties..... | 13 |
| 5.4 Pressure loss, $\Delta\varpi$ | 14 |
| 6 Installation requirements | 15 |
| 6.1 General..... | 15 |
| 6.2 Minimum upstream and downstream straight lengths for installation between various fittings and the orifice plate..... | 16 |
| 6.3 Flow conditioners..... | 21 |
| 6.3.1 General..... | 21 |
| 6.3.2 19-tube bundle flow straightener (1998)..... | 21 |
| 6.3.3 The Zanker flow conditioner plate..... | 27 |
| 6.4 Circularity and cylindricity of the pipe..... | 28 |
| 6.5 Location of orifice plate and carrier rings..... | 29 |
| 6.6 Method of fixing and gaskets..... | 30 |
| 7 Flow calibration of orifice meters | 30 |
| 7.1 General..... | 30 |
| 7.2 Test facility..... | 31 |
| 7.3 Meter installation..... | 31 |
| 7.4 Design of the test programme..... | 31 |
| 7.5 Reporting the calibration results..... | 31 |
| 7.6 Uncertainty analysis of the calibration..... | 32 |
| 7.6.1 General..... | 32 |
| 7.6.2 Uncertainty of the test facility..... | 32 |
| 7.6.3 Uncertainty of the orifice meter..... | 32 |
| Annex A (informative) Tables of discharge coefficients and expansibility [expansion] factors | 33 |
| Annex B (informative) Flow conditioners | 47 |

ISO 5167-2:2022(E)

| | |
|---------------------------|-----------|
| Bibliography | 52 |
|---------------------------|-----------|

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

ISO 5167-2 was prepared by Technical Committee ISO/TC 30, *Measurement of fluid flow in closed conduits*, Subcommittee SC 2, *Pressure differential devices*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/SS F05, *Measuring instruments*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition of ISO 5167-2 cancels and replaces the first edition (ISO 5167-2:2003), which has been technically revised.

The main changes are as follows:

- a revised maximum orifice edge thickness is given for $\beta < 0,2$;
- a correction has been made to the required spacing between two 45° bends for which the straight length upstream of an orifice plate is stated;
- a clearer specification has been given for the tee for which the straight length upstream of an orifice plate is stated;
- flow calibration of orifice plates is included;
- there is improved wording of the rules for spacing of multiple fittings but no change in actual requirements.

A list of all parts in the ISO 5167 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

ISO 5167-2:2022(E)

Introduction

ISO 5167, consisting of six parts, covers the geometry and method of use (installation and operating conditions) of orifice plates, nozzles, Venturi tubes, cone meters and wedge meters when they are inserted in a conduit running full to determine the flow rate of the fluid flowing in the conduit. It also gives necessary information for calculating the flow rate and its associated uncertainty.

ISO 5167 (all parts) is applicable only to pressure differential devices in which the flow remains subsonic throughout the measuring section and where the fluid can be considered as single-phase, but is not applicable to the measurement of pulsating flow. Furthermore, each of these devices can only be used uncalibrated in accordance with this standard within specified limits of pipe size and Reynolds number, or alternatively they can be used across their calibrated range.

ISO 5167 (all parts) deals with devices for which direct calibration experiments have been made, sufficient in number, spread and quality to enable coherent systems of application to be based on their results and coefficients to be given with certain predictable limits of uncertainty. ISO 5167 also provides methodology for bespoke calibration of differential pressure meters.

The devices introduced into the pipe are called primary devices. The term primary device also includes the pressure tappings. All other instruments or devices required to facilitate the instrument readings are known as secondary devices, and the flow computer that receives these readings and performs the algorithms is known as a tertiary device. ISO 5167 (all parts) covers primary devices; secondary devices (see ISO 2186) and tertiary devices will be mentioned only occasionally.

Aspects of safety are not dealt with in ISO 5167-1 to ISO 5167-6. It is the responsibility of the user to ensure that the system meets applicable safety regulations.

Measurement of fluid flow by means of pressure differential devices inserted in circular cross-section conduits running full —

Part 2: Orifice plates

1 Scope

This document specifies the geometry and method of use (installation and operating conditions) of orifice plates when they are inserted in a conduit running full to determine the flow rate of the fluid flowing in the conduit.

This document also provides background information for calculating the flow rate and is applicable in conjunction with the requirements given in ISO 5167-1.

This document is applicable to primary devices having an orifice plate used with flange pressure tapplings, or with corner pressure tapplings, or with D and $D/2$ pressure tapplings. Other pressure tapplings such as “vena contracta” and pipe tapplings are not covered by this document. This document is applicable only to a flow which remains subsonic throughout the measuring section and where the fluid can be considered as single phase. It is not applicable to the measurement of pulsating flow^[1]. It does not cover the use of orifice plates in pipe sizes less than 50 mm or more than 1 000 mm, or where the pipe Reynolds numbers are below 5 000.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 4006, *Measurement of fluid flow in closed conduits — Vocabulary and symbols*

ISO 5167-1, *Measurement of fluid flow by means of pressure differential devices inserted in circular-cross section conduits running full — Part 1: General principles and requirements*

ISO 5168, *Measurement of fluid flow — Procedures for the evaluation of uncertainties*

ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN