

STN	Chirurgické implantáty Aktívne implantovateľné zdravotnícke pomôcky Časť 5: Prístroje na podporu obehu (ISO 14708-5: 2020)	STN EN ISO 14708-5 85 3001
------------	---	--

Implants for surgery - Active implantable medical devices - Part 5: Circulatory support devices (ISO 14708-5:2020)

Táto norma obsahuje anglickú verziu európskej normy.

This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 09/22

Obsahuje: EN ISO 14708-5:2022, ISO 14708-5:2020

135671

EUROPEAN STANDARD

EN ISO 14708-5

NORME EUROPÉENNE

EUROPÄISCHE NORM

July 2022

ICS 11.040.40

English version

Implants for surgery - Active implantable medical devices - Part 5: Circulatory support devices (ISO 14708-5:2020)

Implants chirurgicaux - Dispositifs médicaux
implantables actifs - Partie 5: Dispositifs d'assistance
circulatoire (ISO 14708-5:2020)

Chirurgische Implantate - Aktive implantierbare
medizinische Geräte - Teil 5: Besondere
Anforderungen an Kreislaufunterstützungssysteme
(ISO 14708-5:2020)

This European Standard was approved by CEN on 6 July 2022.

CEN and CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN and CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN and CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN and CENELEC members are the national standards bodies and national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



**CEN-CENELEC Management Centre:
Rue de la Science 23, B-1040 Brussels**

EN ISO 14708-5:2022 (E)

Contents	Page
European foreword.....	3

European foreword

This document (EN ISO 14708-5:2022) has been prepared by Technical Committee ISO/TC 150 "Implants for surgery" in collaboration with Technical Committee CEN-CENELEC/ JTC 16 "Active Implantable Medical Devices" the secretariat of which is held by DKE.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2023, and conflicting national standards shall be withdrawn at the latest by January 2023.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN-CENELEC shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a Standardization Request given to CEN and CENELEC by the European Commission and the European Free Trade Association.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN and CENELEC websites.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 14708-5:2020 has been approved by CEN-CENELEC as EN ISO 14708-5:2022 without any modification.

INTERNATIONAL STANDARD

ISO
14708-5

Second edition
2020-05

Implants for surgery — Active implantable medical devices —

Part 5: Circulatory support devices

*Implants chirurgicaux — Dispositifs médicaux implantables actifs —
Partie 5: Dispositifs d'assistance circulatoire*



Reference number
ISO 14708-5:2020(E)

© ISO 2020

ISO 14708-5:2020(E)**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	vi
Introduction	vii
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Symbols and abbreviations	5
5 General requirements for active implantable medical devices	5
5.1 General requirements for non-implantable parts	5
5.2 General requirements for software	5
5.3 Usability of non-implantable parts	5
5.4 Data security and protection from harm caused by unauthorized information tampering	6
5.5 General requirements for risk management	6
5.6 Misconnection of parts of the active implantable medical device	6
5.7 Wireless coexistence and wireless quality of service	6
6 Requirements for particular active implantable medical devices	6
6.1 Intended clinical use/indications	6
6.2 System description	6
6.2.1 General	6
6.2.2 System configuration	8
6.2.3 System performance and operating limits	8
6.3 Design analysis	8
6.3.1 General	8
6.3.2 Human factors analysis	8
6.4 Risk analysis	9
6.5 Human factors	10
6.6 In vitro design evaluation and system performance testing	10
6.6.1 Objective	10
6.6.2 System characterization	10
6.6.3 Subsystem component testing	13
6.7 Electromagnetic compatibility	17
6.8 Materials qualification	17
6.9 Biocompatibility	18
6.10 Dynamic haemolysis	18
6.11 Environmental testing	18
6.12 <i>In vivo</i> evaluation	18
6.12.1 Objective	18
6.12.2 Definition of success or failure	19
6.12.3 Test articles	19
6.12.4 Test system	19
6.12.5 Control	20
6.12.6 Test equipment	20
6.12.7 Preoperative animal care	20
6.12.8 Implant procedure	20
6.12.9 Special instructions for early termination	20
6.12.10 Postoperative care	21
6.12.11 Anticoagulation	21
6.12.12 Adverse events	21
6.12.13 System performance	21
6.12.14 Measurement of physiological parameters	21
6.12.15 Clinical pathology	21
6.12.16 Necropsy and device retrieval	21
6.12.17 Macroscopic examination	22

ISO 14708-5:2020(E)

6.12.18	Histological examination	22
6.12.19	Explanted device analysis	22
6.12.20	Data analysis	22
6.13	Reliability	22
6.14	Clinical evaluation	23
7	General arrangement of the packaging	23
8	General <i>markings</i> for active implantable medical devices	24
9	Markings on the sales packaging	24
10	Construction of the <i>sales packaging</i>	24
11	Markings on the sterile pack	24
12	Construction of the non-reusable pack	25
13	Markings on the active implantable medical device	25
14	Protection from unintentional biological effects being caused by the active implantable medical device	25
15	Protection from harm to the patient or user caused by external physical features of the active implantable medical device	26
16	Protection from harm to the patient caused by electricity	26
17	Protection from harm to the patient caused by heat	26
17.1	Protection from harm to the patient caused by heat	26
17.2	Active implantable medical device intended to supply heat	26
18	Protection from ionizing radiation released or emitted from the active implantable medical device	26
19	Protection from unintended effects caused by the active implantable medical device	26
20	Protection of the active implantable medical device from damage caused by external defibrillators	27
21	Protection of the active implantable medical device from changes caused by electrical fields applied directly to the patient	27
22	Protection of the active implantable medical device from changes caused by miscellaneous medical treatments	27
23	Protection of the active implantable medical device from mechanical forces	28
24	Protection of the active implantable medical device from damage caused by electrostatic discharge	28
25	Protection of the active implantable medical device from damage caused by atmospheric pressure changes	28
26	Protection of the active implantable medical device from damage caused by temperature changes	28
27	Protection of the active implantable medical device from electromagnetic non-ionizing radiation	29
27.1	General	29
27.2	Test conditions	29
27.2.1	Acceptance criteria	29
27.2.2	Test configuration and setup	29
27.2.3	Operating functions, modes, and settings	30
27.2.4	Patient physiological simulation	30
27.2.5	Immunity test levels	30
27.3	Risk management file and test report file documentation	30
27.4	Protection from static magnetic fields of flux density up to 50 mT	31
27.5	Protection from AC magnetic fields in the range of 1 kHz to 140 kHz	31

27.6	Protection from proximity fields due to RF wireless communications equipment.....	32
28	Accompanying documentation.....	32
Annex A	(informative) Relationship between the fundamental principles in ISO/TR 14283 and the clauses of this document.....	36
Annex B	(informative) Rationale.....	55
Annex C	(informative) Pre-clinical <i>in vitro</i>/<i>in silico</i> evaluation.....	61
Annex D	(informative) Active implantable medical device hazards, associated failure modes, and evaluation methods.....	65
Bibliography	67

ISO 14708-5:2020(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 6, *Active implants*.

This second edition cancels and replaces the first edition (ISO 14708-5:2010), which has been technically revised. The main change compared to the previous edition is as follows:

— alignment to the revised ISO 14708-1:2014.

A list of all parts in the ISO 14708 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document specifies requirements for safety and performance of active implantable circulatory support devices. It amends and supplements ISO 14708-1:2014, hereinafter referred to as ISO 14708-1. The requirements of this document take priority over those of ISO 14708-1.

Heart failure is a major public health problem. It is estimated that worldwide more than 5 million people die per year due to heart failure. In addition, it accounts for a large portion of health care expenditure and rehospitalisation (see Reference [35]). Circulatory support devices are needed for promoting myocardial recovery following acute heart failure as well as long-term support until eventual transplantation or permanent therapy. Circulatory support devices may be fully implanted, partially implanted, or delivered by percutaneous approach. The growth of heart failure is expected to increase with the aging population (see Reference [30]).

The requirements of this document supplement or modify those of ISO 14708-1.

In this document, terms printed in italics are used as defined in [Clause 3](#). Where a defined term is used as a qualifier in another term, it is not printed in italics unless the concept thus qualified is also defined.

Information is also provided in [Annex A](#) that explains the relationship between ISO/TR 14283, ISO 14708-1 and this document.

Notes on this document are provided in [Annex B](#) for information.

[Annex C](#) provides guidance on pre-clinical in vitro and in silico evaluation. [Annex D](#) provides information device hazards, associated failure modes, and evaluation methods. All annexes are informative.

Implants for surgery — Active implantable medical devices —

Part 5: Circulatory support devices

1 Scope

This document specifies requirements for safety and performance of active implantable circulatory support devices, including type tests, animal studies and clinical evaluation requirements.

NOTE The device that is commonly referred to as an active implantable medical device can in fact be a single device, a combination of devices, or a combination of a device or devices and one or more accessories. Not all of these parts are required to be either partially or totally implantable, but there is a need to specify main requirements of non-implantable parts and accessories if they could affect the safety or performance of the implantable device.

The tests that are specified in this document are type tests and are to be carried out on a sample of a device to assess device behavioural responses and are not intended to be used for the routine testing of manufactured products.

Included in the scope of this document are:

- ventricular assist devices (VAD), left or right heart support;
- total artificial hearts (TAH);
- biventricular assist devices (biVAD);
- percutaneous assist devices;
- paediatric assist devices.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1:2018, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 14708-1:2014, *Implants for surgery — Active implantable medical devices — Part 1: General requirements for safety, marking and for information to be provided by the manufacturer*

ISO 14971:2019, *Medical devices — Application of risk management to medical devices*

IEC 60068-1:2013, *Environmental testing — Part 1: General and guidance*

IEC 60068-2-27:2008, *Environmental testing — Part 2-27: Tests — Test Ea and guidance: shock*

IEC 60068-2-31:2008, *Environmental testing — Part 2-31: Tests — Test Ec: rough handling shocks, primarily for equipment-type specimens*

ISO 14708-5:2020(E)

IEC 60068-2-64:2008, *Environmental testing — Part 2-64: Tests — Test Fh: vibration, broadband random and guidance*

IEC 60601-1:2018, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*

IEC 60601-1-2:2014, *Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral standard: Electromagnetic compatibility — Requirements and tests*

IEC 60601-1-6:2010, *Medical electrical equipment — Part 1-6: General requirements for basic safety and essential performance — Collateral standard: Usability*

IEC 60601-1-10:2007, *Medical electrical equipment — Part 1-10: General requirements for basic safety and essential performance — Collateral standard: Requirements for the development of physiologic closed-loop controllers*

IEC 60601-1-11:2015, *Medical electrical equipment — Part 1-11: General requirements for basic safety and essential performance — Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*

IEC 62304:2006, *Medical device software — Software life cycle processes*

IEC 62366-1:2015, *Medical devices — Part 1: Application of usability engineering to medical devices*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN