

STN	Pracovné ovzdušie Čerpadlá na osobný odber chemických a biologických látok Požiadavky a skúšobné metódy (ISO 13137: 2022)	STN EN ISO 13137 83 8701
------------	--	--

Workplace atmospheres - Pumps for personal sampling of chemical and biological agents - Requirements and test methods (ISO 13137:2022)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 12/22

Obsahuje: EN ISO 13137:2022, ISO 13137:2022

Oznámením tejto normy sa ruší
STN EN ISO 13137 (83 3701) z apríla 2014

136023

EUROPEAN STANDARD

EN ISO 13137

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2022

ICS 13.040.30

Supersedes EN ISO 13137:2013

English Version

Workplace atmospheres - Pumps for personal sampling of chemical and biological agents - Requirements and test methods (ISO 13137:2022)

Air des lieux de travail - Pompes pour le prélèvement individuel des agents chimiques et biologiques - Exigences et méthodes d'essai (ISO 13137:2022)

Arbeitsplatzatmosphäre - Pumpen für die personenbezogene Probenahme von chemischen und biologischen Arbeitsstoffen - Anforderungen und Prüfverfahren (ISO 13137:2022)

This European Standard was approved by CEN on 2 September 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 13137:2022 (E)

Contents	Page
European foreword.....	3

European foreword

This document (EN ISO 13137:2022) has been prepared by Technical Committee ISO/TC 146 "Air quality" in collaboration with Technical Committee CEN/TC 137 "Assessment of workplace exposure to chemical and biological agents" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2023, and conflicting national standards shall be withdrawn at the latest by April 2023.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 13137:2013.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 13137:2022 has been approved by CEN as EN ISO 13137:2022 without any modification.

INTERNATIONAL STANDARD

ISO 13137

Second edition
2022-08

Workplace atmospheres — Pumps for personal sampling of chemical and biological agents — Requirements and test methods

*Air des lieux de travail — Pompes pour le prélèvement individuel des
agents chimiques et biologiques — Exigences et méthodes d'essai*



Reference number
ISO 13137:2022(E)

© ISO 2022

ISO 13137:2022(E)**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	v
Introduction.....	vi
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Type of pump.....	2
5 Requirements.....	2
5.1 Features.....	2
5.2 Mass.....	2
5.3 Design safety.....	2
5.4 Operating time.....	2
5.5 Start-up and long-term performance.....	3
5.6 Short-term interruption of air flow.....	3
5.7 Temperature dependence.....	3
5.8 Mechanical strength.....	3
5.9 Pulsation of flow rate (for type P pumps only).....	4
5.10 Flow rate stability under increasing pressure drop.....	4
5.10.1 Pumps with a nominal flow rate range less or equal than 5 000 ml · min ⁻¹	4
5.10.2 Pumps with a nominal flow rate range above 5 000 ml · min ⁻¹	5
5.11 Timer accuracy.....	5
5.12 Electromagnetic compatibility.....	5
5.13 Explosion hazard.....	5
6 Test conditions.....	5
6.1 Number of test objects.....	5
6.2 Test instruments.....	5
6.3 Preconditioning and sequence of tests.....	6
6.4 Adjustment of volumetric flow rate and pressure drop.....	6
6.5 Test set-up and performance.....	6
7 Test methods.....	7
7.1 General.....	7
7.2 Features.....	7
7.3 Mass.....	7
7.4 Design safety.....	8
7.5 Operating time.....	8
7.6 Start-up and long-term performance.....	8
7.6.1 Test set-up.....	8
7.6.2 Flow rate and pressure drop adjustment.....	8
7.6.3 Procedure.....	8
7.7 Short-term interruption of air flow.....	9
7.7.1 Test set-up.....	9
7.7.2 Flow rate and pressure drop adjustment.....	9
7.7.3 Procedure.....	9
7.8 Temperature dependence.....	9
7.8.1 Test set-up.....	9
7.8.2 Flow rate and pressure drop adjustment.....	9
7.8.3 Procedure.....	10
7.9 Mechanical strength.....	10
7.9.1 Test set-up.....	10
7.9.2 Flow rate and pressure drop adjustment.....	11
7.9.3 Procedure.....	11
7.10 Pulsation of flow rate (for type P pumps only).....	12
7.10.1 Test set-up.....	12

ISO 13137:2022(E)

7.10.2	Flow rate and pressure drop adjustment.....	12
7.10.3	Procedure.....	12
7.11	Flow rate stability under increasing pressure drop.....	14
7.11.1	Test set-up.....	14
7.11.2	Flow rate adjustment.....	14
7.11.3	Procedure.....	15
7.12	Timer accuracy.....	15
7.13	Electromagnetic compatibility.....	15
7.14	Explosion hazard.....	15
8	Test report.....	15
9	Instructions for use.....	15
10	Charger.....	16
10.1	Requirements.....	16
10.2	Testing.....	16
11	Marking.....	16
Annex A (informative) Types of pump mechanisms and control systems.....		17
Annex B (informative) Internal sensors of sampling pumps.....		20
Annex C (informative) User tests for pumps and flow meters.....		22
Annex D (informative) Pressure drop due to collection substrates.....		25
Annex E (informative) Test instruments.....		28
Bibliography.....		29

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 2, *Workplace atmospheres*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 137, *Assessment of workplace exposure to chemical and biological agents*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 13137:2013), which has been technically revised.

The main changes are as follows:

- definitions that appear in ISO 18158 have been removed from this document, with ISO 18158 being added as a reference (replacing references to EN 1540);
- references to EN 482 have been replaced with ISO 20581;
- the text has been editorially updated.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

ISO 13137:2022(E)

Introduction

Many different methods are used to determine the concentration of chemical and biological agents in workplace air. Many of these methods involve the use of a pump and sampler connected by a flexible tube. Air is drawn through the sampler and chemical and biological agents are trapped, e.g. on a filter, sorbent tube or long-term detector tube, or in a gas washing bottle. In personal sampling, the pump and sampler are attached to the worker to collect chemical and biological agents in the breathing zone.

The volume of air drawn by the pump during the sampling period is one of the quantities in the calculation of the concentration of the chemical and biological agents in air. Therefore, the volume of air sampled should be determined accurately and, in order to facilitate this, the flow rate should be maintained within acceptable limits throughout the sampling period. For particle size selective sampling, the short-term fluctuation of the flow rate should also be maintained within acceptable limits in order to ensure that the sampler exhibits the required collection characteristics.

ISO 20581^[1] specifies general performance criteria for methods for measuring the concentration of chemical and biological agents in workplace air. These performance criteria include maximum values of expanded uncertainty that are not to be exceeded under prescribed laboratory conditions. In addition, the performance criteria should also be met under a wider variety of environmental influences, representative of workplace conditions. The contribution of the sampling pump to measurement uncertainty should be kept to a minimum.

This document is intended to enable manufacturers and users of personal sampling pumps to adopt a consistent approach to, and provide a framework for, the assessment of the specified performance criteria. Manufacturers are urged to ensure that pumps meet the requirements laid down in this document, including environmental influences which can be expected to affect performance.

Workplace atmospheres — Pumps for personal sampling of chemical and biological agents — Requirements and test methods

1 Scope

This document specifies performance requirements for battery powered pumps used for personal sampling of chemical and biological agents in workplace air. It also specifies test methods in order to determine the performance characteristics of such pumps under prescribed laboratory conditions.

This document is applicable to battery powered pumps having a nominal volumetric flow rate above $10 \text{ ml} \cdot \text{min}^{-1}$, as used with combinations of sampler and collection substrate for sampling of gases, vapours, dusts, fumes, mists and fibres.

This document is primarily intended for flow-controlled pumps.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 18158, *Workplace air — Terminology*

IEC 60079-0, *Explosive atmospheres — Part 0: Equipment — General requirements*

IEC 61000-6-1, *Electromagnetic compatibility (EMC) — Part 6-1: Generic standards — Immunity standard for residential, commercial and light-industrial environments*

IEC 61000-6-3, *Electromagnetic compatibility (EMC) — Part 6-3: Generic standards — Emission standard for residential, commercial and light-industrial environments*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN