

<b>STN</b>	<b>Rádiologická ochrana</b> <b>Minimálne kritériá na elektrónovú paramagnetickú</b> <b>rezonančnú (EPR) spektroskopiu na</b> <b>retrospektívne meranie dávky ionizujúceho žiarenia</b> <b>Časť 1: Všeobecné zásady (ISO 13304-1: 2020)</b>	<b>STN</b> <b>EN ISO 13304-1</b>  40 1415
------------	--	--

Radiological protection - Minimum criteria for electron paramagnetic resonance (EPR) spectroscopy for retrospective dosimetry of ionizing radiation - Part 1: General principles (ISO 13304-1:2020)

Táto norma obsahuje anglickú verziu európskej normy.  
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 03/23

Obsahuje: EN ISO 13304-1:2022, ISO 13304-1:2020

**136486**



EUROPEAN STANDARD

**EN ISO 13304-1**

NORME EUROPÉENNE

EUROPÄISCHE NORM

December 2022

ICS 13.280; 17.240

English Version

**Radiological protection - Minimum criteria for electron  
paramagnetic resonance (EPR) spectroscopy for  
retrospective dosimetry of ionizing radiation - Part 1:  
General principles (ISO 13304-1:2020)**

Radioprotection - Critères minimaux pour la  
spectroscopie par résonance paramagnétique  
électronique (RPE) pour la dosimétrie rétrospective  
des rayonnements ionisants - Partie 1: Principes  
généraux (ISO 13304-1:2020)

Strahlenschutz - Mindestanforderungen an die  
Elektronenspinresonanz (EPR-Spektroskopie) für die  
retrospektive Dosimetrie ionisierender Strahlung - Teil  
1: Allgemeine Grundsätze (ISO 13304-1:2020)

This European Standard was approved by CEN on 18 December 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

**EN ISO 13304-1:2022 (E)**

<b>Contents</b>	<b>Page</b>
<b>European foreword.....</b>	<b>3</b>

## **European foreword**

The text of ISO 13304-1:2020 has been prepared by Technical Committee ISO/TC 85 "Nuclear energy, nuclear technologies, and radiological protection" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 13304-1:2022 by Technical Committee CEN/TC 430 "Nuclear energy, nuclear technologies, and radiological protection" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2023, and conflicting national standards shall be withdrawn at the latest by June 2023.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

## **Endorsement notice**

The text of ISO 13304-1:2020 has been approved by CEN as EN ISO 13304-1:2022 without any modification.

# INTERNATIONAL STANDARD

# ISO 13304-1

Second edition  
2020-07

---

---

## **Radiological protection — Minimum criteria for electron paramagnetic resonance (EPR) spectroscopy for retrospective dosimetry of ionizing radiation —**

### **Part 1: General principles**

*Radioprotection — Critères minimaux pour la spectroscopie par  
résonance paramagnétique électronique (RPE) pour la dosimétrie  
rétrospective des rayonnements ionisants —*

*Partie 1: Principes généraux*



Reference number  
ISO 13304-1:2020(E)

© ISO 2020

**ISO 13304-1:2020(E)****COPYRIGHT PROTECTED DOCUMENT**

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Confidentiality and ethical considerations</b> .....	<b>2</b>
<b>5 Laboratory safety requirements</b> .....	<b>3</b>
5.1 Magnetic field.....	3
5.2 Electromagnetic frequency.....	3
5.2.1 in vitro measurement.....	3
5.2.2 in vivo measurement.....	3
5.3 Biohazards from samples.....	3
<b>6 Collection/selection and identification of samples</b> .....	<b>3</b>
<b>7 Transportation and storage of samples</b> .....	<b>4</b>
<b>8 Preparation of samples</b> .....	<b>4</b>
<b>9 Apparatus</b> .....	<b>5</b>
9.1 Principles of EPR spectroscopy.....	5
9.2 Requirements for EPR spectrometers.....	6
9.3 Requirements for the resonator.....	6
9.4 Measurements of the background signals.....	6
9.5 Spectrometer stability and monitoring/control of environmental conditions.....	6
9.6 Baseline drift.....	7
<b>10 Measurements of the samples</b> .....	<b>7</b>
10.1 General principles.....	7
10.2 Choice and optimization of the measurement parameters.....	7
10.2.1 General.....	7
10.2.2 Microwave-related parameters.....	8
10.2.3 Magnetic field parameters.....	8
10.2.4 Signal channel parameters.....	8
10.3 Sample positioning and loading.....	9
10.4 Microwave bridge tuning.....	10
10.5 Use of standard samples as field markers and amplitude monitors.....	10
10.6 Monitoring reproducibility.....	10
10.7 Procedure to measure anisotropic samples.....	10
10.8 Coding of spectra and samples.....	11
<b>11 Determination of the absorbed dose in the samples</b> .....	<b>11</b>
11.1 Determination of the radiation-induced signal intensity.....	11
11.2 Conversion of the EPR signal into an estimate of absorbed dose.....	11
11.2.1 Conversion of the EPR signal into an estimate of absorbed dose for in vitro dosimetry.....	11
11.2.2 Conversion of the EPR signal into an estimate of absorbed dose for in vivo tooth dosimetry.....	12
<b>12 Measurement uncertainty</b> .....	<b>12</b>
<b>13 Investigation of dose that has been questioned</b> .....	<b>12</b>
<b>14 Quality assurance (QA) and quality control (QC)</b> .....	<b>13</b>
<b>15 Minimum documentation requirements</b> .....	<b>14</b>
<b>Bibliography</b> .....	<b>16</b>



## ISO 13304-1:2020(E)

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*.

A list of all parts in the ISO 13304 series can be found on the ISO website.

This second edition cancels and replaces the first edition (ISO 13304-1:2013), of which it constitutes a minor revision. The changes compared to the previous edition are as follows:

- inclusion of bibliographic references in the text;
- informative reference to ISO 13304-2 added;
- update of Bibliography.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

Electron paramagnetic resonance (EPR) has become an important approach for retrospective dosimetry in any situation where dosimetric information is potentially incomplete or unknown for an individual. It is now applied widely for retrospective evaluation of doses that were delivered at considerable times in the past (e.g. EPR dosimetry is one of the methods of choice for retrospective evaluation of doses to the involved populations from the atomic weapon exposures in Japan and after the Chernobyl accident) and has received attention for use for triage after an incident in which large numbers of people have potentially been exposed to clinically significant levels of radiation [1] to [12]. Various materials may be analysed by EPR to provide information on dose [13] to [41]. Thus, EPR is a versatile tool for retrospective dosimetry, pertinent as well for acute exposures (past or recent, whole or partial body) and prolonged exposures. Doses estimated with EPR were mainly used to correlate the biological effect of ionizing radiation to received dose, to validate other techniques or methodologies, to manage casualties, or for forensic expertise for judicial authorities [42]. It uses mainly biological tissues of the person as the dosimeter and also can use materials from personal objects as well as those located in the immediate environment. EPR dosimetry is based on the fundamental properties of ionizing radiation: the generation of unpaired electron species (often but not exclusively free radicals) proportional to absorbed dose. The technique of EPR specifically and sensitively detects the amount of unpaired electrons that have sufficient stability to be observed after their generation; while the amount of the detectable unpaired electrons is usually directly proportional to the amount that was generated, these species can react, and therefore, the relationship between the intensity of the EPR signal and the radiation dose needs to be established for each type of use. The most extensive use of the technique has been with calcified tissue, especially with enamel from teeth [43] to [50]. An IAEA technical report was published on the use for tooth enamel [51]. To extend the possibility of EPR retrospective dosimetry, new materials possibly suitable for EPR dosimetry are regularly studied and associated protocols established. This document is aimed to make this technique more widely available, more easily applicable and useful for dosimetry. Specifically, this document proposes a methodological frame and recommendations to set up, validate, and apply protocols from sample collection to dose reporting. The application of this document to *ex vivo* human tooth enamel dosimetry is described in ISO 13304-2 [52].



# Radiological protection — Minimum criteria for electron paramagnetic resonance (EPR) spectroscopy for retrospective dosimetry of ionizing radiation —

## Part 1: General principles

### 1 Scope

The primary purpose of this document is to provide minimum acceptable criteria required to establish a procedure for retrospective dosimetry by electron paramagnetic resonance spectroscopy and to report the results.

The second purpose is to facilitate the comparison of measurements related to absorbed dose estimation obtained in different laboratories.

This document covers the determination of absorbed dose in the measured material. It does not cover the calculation of dose to organs or to the body. It covers measurements in both biological and inanimate samples, and specifically:

- a) based on inanimate environmental materials like glass, plastics, clothing fabrics, saccharides, etc., usually made at X-band microwave frequencies (8 GHz to 12 GHz);
- b) in vitro tooth enamel using concentrated enamel in a sample tube, usually employing X-band frequency, but higher frequencies are also being considered;
- c) in vivo tooth dosimetry, currently using L-band (1 GHz to 2 GHz), but higher frequencies are also being considered;
- d) in vitro nail dosimetry using nail clippings measured principally at X-band, but higher frequencies are also being considered;
- e) in vivo nail dosimetry with the measurements made at X-band on the intact finger or toe;
- f) in vitro measurements of bone, usually employing X-band frequency, but higher frequencies are also being considered.

For biological samples, in vitro measurements are carried out in samples after their removal from the person or animal and under laboratory conditions, whereas the measurements in vivo are carried out without sample removal and may take place under field conditions.

NOTE The dose referred to in this document is the absorbed dose of ionizing radiation in the measured materials.

### 2 Normative references

There are no normative references in this document.

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**