

STN	Zdravotnícke elektrické prístroje Prístroje používané na neinvazívne meranie napätia na röntgenke v diagnostickej rádiológii	STN EN IEC 61676 36 4768
------------	---	--

Medical electrical equipment - Dosimetric instruments used for non-invasive measurement of X-ray tube voltage in diagnostic radiology

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 08/23

Obsahuje: EN IEC 61676:2023, IEC 61676:2023

Oznámením tejto normy sa od 25.04.2026 ruší
STN EN 61676 (36 4768) z júna 2004

137289

Úrad pre normalizáciu, metrológiu a skúšobníctvo Slovenskej republiky, 2023
Slovenská technická norma a technická normalizačná informácia je chránená zákonom č. 60/2018 Z. z. o technickej normalizácii
v znení neskorších predpisov.

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN IEC 61676

May 2023

ICS 11.040.50; 11.040.55

Supersedes EN 61676:2002; EN 61676:2002/A1:2009

English Version

**Medical electrical equipment - Dosimetric instruments used for
non-invasive measurement of X-ray tube voltage in diagnostic
radiology
(IEC 61676:2023)**

Appareils électromédicaux - Appareils de dosimétrie pour le
mesurage non invasif de la tension du tube radiogène dans
la radiologie de diagnostic
(IEC 61676:2023)

Medizinische elektrische Geräte - Geräte für die nicht-
invasive Messung der Röntgenröhrenspannung in der
diagnostischen Radiologie
(IEC 61676:2023)

This European Standard was approved by CENELEC on 2023-04-25. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN IEC 61676:2023 (E)**European foreword**

The text of document 62C/830/CDV, future edition 2 of IEC 61676, prepared by SC 62C "Equipment for radiotherapy, nuclear medicine and radiation dosimetry" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 61676:2023.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2024-01-25
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2026-04-25

This document supersedes EN 61676:2002 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

Endorsement notice

The text of the International Standard IEC 61676:2023 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standard indicated:

- | | | |
|----------------|------|--|
| IEC 60580:2019 | NOTE | Approved as EN IEC 60580:2020 (not modified) |
| IEC 60731:2011 | NOTE | Approved as EN 60731:2012 (not modified) |

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60417	-	Graphical symbols for use on equipment. Index, survey and compilation of the single sheets.	-	-
IEC 60601-1	2005	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	EN 60601-1	2006
-	-		+ corrigendum Mar.	2010
+ A1	2012		+ A1	2013
-	-		+ A12	2014
+ A2	2020		+ A2	2021
IEC/TR 60788	2004	Medical electrical equipment - Glossary of defined terms	-	-
IEC 61000-4-2	-	Electromagnetic compatibility (EMC) - Part 4-2: Testing and measurement techniques - Electrostatic discharge immunity test	EN 61000-4-2	-
IEC 61000-4-3	-	Electromagnetic compatibility (EMC) - Part 4-3: Testing and measurement techniques - Radiated, radio-frequency, electromagnetic field immunity test	EN IEC 61000-4-3	-
IEC 61000-4-4	-	Electromagnetic compatibility (EMC) - Part 4-4: Testing and measurement techniques - Electrical fast transient/burst immunity test	EN 61000-4-4	-
IEC 61000-4-5	-	Electromagnetic compatibility (EMC) - Part 4-5: Testing and measurement techniques - Surge immunity test	EN 61000-4-5	-
IEC 61000-4-6	-	Electromagnetic compatibility (EMC) - Part 4-6: Testing and measurement techniques - Immunity to conducted disturbances, induced by radio-frequency fields	EN 61000-4-6	-

EN IEC 61676:2023 (E)

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 61000-4-11	-	Electromagnetic compatibility (EMC) - Part 4-11: Testing and measurement techniques - Voltage dips, short interruptions and voltage variations immunity tests for equipment with input current up to 16 A per phase	EN IEC 61000-4-11	-
IEC 61010-1	-	Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements	-	-
IEC 61187	-	Electrical and electronic measuring equipment - Documentation	EN 61187	-
ISO 7000	2019	Graphical symbols for use on equipment - Registered symbols	-	-



IEC 61676

Edition 2.0 2023-03

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment – Dosimetric instruments used for non-invasive measurement of X-ray tube voltage in diagnostic radiology

Appareils électromédicaux – Appareils de dosimétrie pour le mesurage non invasif de la tension du tube radiogène dans la radiologie de diagnostic

**THIS PUBLICATION IS COPYRIGHT PROTECTED****Copyright © 2023 IEC, Geneva, Switzerland**

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications previews. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 300 terminological entries in English and French, with equivalent terms in 19 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -**webstore.iec.ch/advsearchform**

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 300 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 19 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.



IEC 61676

Edition 2.0 2023-03

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment – Dosimetric instruments used for non-invasive measurement of X-ray tube voltage in diagnostic radiology

Appareils électromédicaux – Appareils de dosimétrie pour le mesurage non invasif de la tension du tube radiogène dans la radiologie de diagnostic

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.50; 11.040.55

ISBN 978-2-8322-6698-4

Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	7
2 Normative references	7
3 Terms and definitions	8
4 General performance requirements for measurement of PRACTICAL PEAK VOLTAGE measurements.....	11
4.1 Quantity to be measured	11
4.2 Limits of PERFORMANCE CHARACTERISTICS	11
4.2.1 Limits	11
4.2.2 Maximum error	11
4.2.3 Over and under range indications	12
4.2.4 Repeatability	13
4.2.5 Long term stability	13
4.3 LIMITS OF VARIATION for effects of INFLUENCE QUANTITIES	13
4.3.1 INFLUENCE QUANTITIES	13
4.3.2 MINIMUM RATED RANGE of use	13
4.3.3 REFERENCE CONDITIONS	14
4.3.4 STANDARD TEST CONDITIONS	14
4.3.5 LIMITS OF VARIATION.....	14
4.4 Performance test procedures	15
4.4.1 General remarks.....	15
4.4.2 Dependence of instrument RESPONSE on voltage waveform and frequency	16
4.4.3 Dependence of instrument RESPONSE on ANODE ANGLE.....	17
4.4.4 Dependence of instrument RESPONSE on FILTRATION.....	17
4.4.5 Dependence of instrument RESPONSE on dose rate	17
4.4.6 Dependence of instrument RESPONSE on IRRADIATION TIME	18
4.4.7 Dependence of instrument RESPONSE on field size	18
4.4.8 Dependence of instrument RESPONSE on focus-to-detector distance.....	18
4.4.9 Dependence of instrument RESPONSE on angle of incidence of RADIATION.....	19
4.4.10 Dependence of instrument RESPONSE on angle of detector rotation with respect to the X-RAY TUBE axis.....	19
4.4.11 Dependence of instrument RESPONSE on temperature and humidity	19
4.4.12 Dependence of instrument RESPONSE on operating voltage	20
4.4.13 Dependence of instrument RESPONSE on electromagnetic compatibility	21
4.4.14 Additional tungsten filtration (tube aging).....	22
5 Special instrumental requirements and marking.....	23
5.1 Requirements for the complete instruments	23
5.2 General.....	23
5.3 Display	23
5.4 Range of measurement.....	23
5.5 Connectors and cables	23
6 ACCOMPANYING DOCUMENTS	24
6.1 General.....	24
6.2 Information provided	24
6.3 Instrument description.....	24

6.4	Detector	24
6.5	Delay time	24
6.6	Measurement window	24
6.7	Data outlet	24
6.8	Transport and storage	24
Annex A (informative) COMBINED STANDARD UNCERTAINTY		25
Annex B (informative) Additional information on PRACTICAL PEAK VOLTAGE		26
B.1	Overview	26
B.2	Simplified formalism for the determination of the PRACTICAL PEAK VOLTAGE $\hat{U}$	26
Bibliography		32
Index of defined terms		33
Figure B.1 – Example of a waveform of a two-pulse generator		28
Figure B.2 – Example of a waveform of a constant-voltage generator		28
Figure B.3 – Example of falling load waveform		29
Table 1 – Minimum effective ranges		11
Table 2 – Minimum RATED RANGE OF USE, REFERENCE CONDITIONS, STANDARD TEST CONDITIONS, LIMITS OF VARIATION ($\pm L$) and INTRINSIC ERROR (E) over the EFFECTIVE RANGE of use, for the pertaining INFLUENCE QUANTITY		14
Table 3 – Minimum test points and test values of PRACTICAL PEAK VOLTAGE for INFLUENCE QUANTITIES		16
Table 4 – Maximum HALF-VALUE LAYER (HVL) depending on anode angle		23
Table A.1 – Example for assessment of the COMBINED STANDARD UNCERTAINTY – Instruments used for NON-INVASIVE MEASUREMENT of X-RAY TUBE VOLTAGE		25
Table B.1 – Values of 20 samples of the falling load waveform in Figure B.3		29
Table B.2 – Voltage bins, probability and weighting factors for the 20 samples of the falling load waveform in Figure B.3		30
Table B.3 – Weighting factors for the 20 equally spaced samples of the falling load waveform in Figure B.3		31

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEDICAL ELECTRICAL EQUIPMENT – DOSIMETRIC INSTRUMENTS
USED FOR NON-INVASIVE MEASUREMENT OF X-RAY TUBE VOLTAGE
IN DIAGNOSTIC RADIOLOGY****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 61676 has been prepared by subcommittee 62C: Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC technical committee 62: Medical equipment, software, and systems. It is an International Standard.

This second edition of IEC 61676 cancels and replaces first edition published in 2002, Amendment 1:2008. This edition constitutes a technical revision.

It includes an assessment of the COMBINED STANDARD UNCERTAINTY for the performance of a hypothetical instrument for the non-invasive measurement of the tube high voltage (in Annex A) which replaces Annex A of the edition 1.1 titled "Recommended performance criteria for the invasive divider".

The text of this document is based on the following documents:

Draft	Report on voting
62C/830/CDV	62C/866/RVC

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/standardsdev/publications.

In this document the following print types are used:

- requirements, compliance with which can be tested, and definitions: in roman type;
- notes, explanations, advice, general statements and exceptions: in small roman type;
- *test specifications: in italic type*;
- TERMS USED THROUGHOUT THIS DOCUMENT THAT HAVE BEEN DEFINED IN CLAUSE 3 OR IN IEC 60601-1 AND ITS COLLATERAL STANDARDS: IN SMALL CAPITALS.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The committee knows this second edition of the document does still not address all problems associated with non-invasive high voltage measurements. For mammography only molybdenum filtration is considered in conjunction with a molybdenum anode although in addition tungsten and rhodium anodes with other filtrations are in use like rhodium, aluminium, copper, silver or titanium. At the time when this document was drafted there were not enough data available in the literature to define realistic limits of variation for these types of INFLUENCE QUANTITIES. On the other hand, the committee was informed that several international projects were started to examine the general behaviour of non-invasive X-ray multimeters of the main MANUFACTURERS. Results from these studies were to be expected within about 5 years. Therefore, the committee decided to set a short stability time for the second edition and update the document as soon as the results from these new examinations will be available.

INTRODUCTION

The result of a measurement of the X-RAY TUBE VOLTAGE by means of invasive or non-invasive instruments is normally expressed in the form of one single number for the value of the tube voltage, irrespective of whether the tube voltage is constant potential or shows a time dependent waveform. Non-invasive instruments for the measurement of the X-RAY TUBE VOLTAGE on the market usually indicate the "MEAN PEAK VOLTAGE". But the quantity "MEAN PEAK VOLTAGE" is not unambiguously defined and can be any mean of all voltage peaks. It is impossible to establish test procedures for the performance requirements of non-invasive instruments for the measurement of the X-RAY TUBE VOLTAGE without the definition of the quantity under consideration. Therefore, this document is based on a quantity called "PRACTICAL PEAK VOLTAGE". The PRACTICAL PEAK VOLTAGE is unambiguously defined and applicable to any waveform. This quantity is related to the spectral distribution of the emitted X-RADIATION and the image properties. X-RAY GENERATORS operating at the same value of the PRACTICAL PEAK VOLTAGE produce the same low-level contrast in the RADIOGRAMS, even when the waveforms of the tube voltages are different. Detailed information on this concept is provided in Annex B. An example for the calculation of the PRACTICAL PEAK VOLTAGE in the case of a "falling load" waveform is also given in Annex B.

The CALIBRATION and adjustment of the X-RAY TUBE VOLTAGE of an X-RAY GENERATOR is generally performed by the MANUFACTURER using a direct INVASIVE MEASUREMENT. Instruments utilising NON-INVASIVE MEASUREMENTS can also be used to check the CALIBRATION or to adjust the X-RAY TUBE VOLTAGE. These instruments are used to have uncertainties of the voltage measurement comparable with the INVASIVE MEASUREMENT. One of the most important parameters of diagnostic X-RAY EQUIPMENT is the voltage applied to the X-RAY TUBE, because both the image quality in diagnostic radiology and the DOSE received by the PATIENT undergoing radiological examinations are dependent on the X-RAY TUBE VOLTAGE. An overall uncertainty below $\pm 5\%$ is applicable, and this value serves as a guide for the LIMITS OF VARIATION for the effects of INFLUENCE QUANTITIES.

MEDICAL ELECTRICAL EQUIPMENT – DOSIMETRIC INSTRUMENTS USED FOR NON-INVASIVE MEASUREMENT OF X-RAY TUBE VOLTAGE IN DIAGNOSTIC RADIOLOGY

1 Scope

This document specifies the performance requirements of instruments as used in the NON-INVASIVE MEASUREMENT of X-RAY TUBE VOLTAGE up to 150 kV and the relevant compliance tests. This document also describes the method for CALIBRATION and gives guidance for estimating the uncertainty in measurements performed under conditions different from those during CALIBRATION.

Applications for such measurement are found in diagnostic RADIOLOGY including mammography, COMPUTED TOMOGRAPHY (CT), dental radiology and RADIOSCOPY. This document is not concerned with the safety aspect of such instruments. The requirements for electrical safety applying to them are contained in IEC 61010-1.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60417, *Graphical symbols for use on equipment*, available at <http://www.graphical-symbols.info/equipment>

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1:2005/AMD1:2012

IEC 60601-1:2005/AMD2:2020

IEC TR 60788:2004, *Medical electrical equipment – Glossary of defined terms*

IEC 61000-4-2, *Electromagnetic compatibility (EMC) – Part 4-2: Testing and measurement techniques – Electrostatic discharge immunity test*

IEC 61000-4-3, *Electromagnetic compatibility (EMC) – Part 4-3: Testing and measurement techniques – Radiated, radio-frequency, electromagnetic field immunity test*

IEC 61000-4-4, *Electromagnetic compatibility (EMC) – Part 4-4: Testing and measurement techniques – Electrical fast transient/burst immunity test*

IEC 61000-4-5, *Electromagnetic compatibility (EMC) – Part 4-5: Testing and measurement techniques – Surge immunity test*

IEC 61000-4-6, *Electromagnetic compatibility (EMC) – Part 4-6: Testing and measurement techniques – Immunity to conducted disturbances, induced by radio-frequency fields*

IEC 61000-4-11, *Electromagnetic compatibility (EMC) – Part 4-11: Testing and measurement techniques – Voltage dips, short interruptions and voltage variations immunity tests for equipment with input current up to 16 A per phase*

IEC 61010-1, *Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements*

IEC 61187, *Electrical and electronic measuring equipment – Documentation*

ISO 7000:2019, *Graphical symbols for use on equipment – Registered symbol*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN