

STN	Pasívne neutrónové dozimetrické systémy Časť 2: Metodika a kritériá na kvalifikáciu osobných dozimetrických systémov na pracoviskách (ISO 21909-2: 2021)	STN EN ISO 21909-2 40 1427
------------	---	--

Passive neutron dosimetry systems - Part 2: Methodology and criteria for the qualification of personal dosimetry systems in workplaces (ISO 21909-2:2021)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 09/23

Obsahuje: EN ISO 21909-2:2023, ISO 21909-2:2021

137462

Úrad pre normalizáciu, metrológiu a skúšobníctvo Slovenskej republiky, 2023
Slovenská technická norma a technická normalizačná informácia je chránená zákonom č. 60/2018 Z. z. o technickej normalizácii v znení neskorších predpisov.

EUROPEAN STANDARD

EN ISO 21909-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

July 2023

ICS 13.280

English Version

Passive neutron dosimetry systems - Part 2: Methodology and criteria for the qualification of personal dosimetry systems in workplaces (ISO 21909-2:2021)

Systèmes dosimétriques passifs pour les neutrons -
Partie 2: Méthodologie et critères de qualification des
systèmes dosimétriques individuels aux postes de
travail (ISO 21909-2:2021)

This European Standard was approved by CEN on 16 July 2023.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 21909-2:2023 (E)

Contents	Page
European foreword.....	3

European foreword

The text of ISO 21909-2:2021 has been prepared by Technical Committee ISO/TC 85 "Nuclear energy, nuclear technologies, and radiological protection" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 21909-2:2023 by Technical Committee CEN/TC 430 "Nuclear energy, nuclear technologies, and radiological protection" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2024, and conflicting national standards shall be withdrawn at the latest by January 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 21909-2:2021 has been approved by CEN as EN ISO 21909-2:2023 without any modification.

INTERNATIONAL STANDARD

ISO 21909-2

First edition
2021-12

Passive neutron dosimetry systems — Part 2: Methodology and criteria for the qualification of personal dosimetry systems in workplaces

Systèmes dosimétriques passifs pour les neutrons —

*Partie 2: Méthodologie et critères de qualification des systèmes
dosimétriques individuels aux postes de travail*



Reference number
ISO 21909-2:2021(E)

© ISO 2021

ISO 21909-2:2021(E)**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms, definitions and symbols	2
3.1 General terms and definitions.....	2
3.2 Quantities.....	3
3.3 Calibration and evaluation.....	4
3.4 Symbols.....	6
4 Documentation and communication to the users	7
5 Recommendations about the workplace to consider	7
6 Methodologies and criteria to qualify the personal dosimetry system at a certain workplace	8
6.1 Choice of the methods to be used for the qualification at workplaces.....	8
6.2 Quantification of the impact of the non-correct behaviour in terms of energy and angle responses of the dosimetry system.....	9
6.2.1 General.....	9
6.2.2 Computational approach.....	10
6.2.3 Experimental approach.....	10
6.3 Qualification based on experimental tests of the dosimetry systems at the workplace.....	10
6.3.1 General method.....	10
6.3.2 Performance test criteria.....	11
6.3.3 First solution: tests at three levels of dose at the workplace.....	12
6.3.4 Second solution: tests at one level of dose at the workplace.....	13
6.3.5 Complementary tests based on ISO 21909-1.....	14
6.3.6 Unique correction for several workplaces.....	15
Annex A (normative) Methodologies to characterize the workplace field	16
Annex B (normative) Determination of the neutron personal dose equivalent $H_p(10)$ - Practical methods	19
Annex C (informative) Example of a complete characterization of the workplace field	26
Annex D (informative) Determination of field-specific correction factors or functions - Practical example: use of information from literature	28
Annex E (informative) Links between ISO 21909-1 and ISO 21909-2	29
Bibliography	31

ISO 21909-2:2021(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*.

A list of all the parts in the ISO 21909 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

ISO 21909-1 provides laboratory-based type tests, and performance requirements for passive neutron dosimetry systems to be used for measurement of personal dose equivalent, $H_p(10)$, for neutrons ranging from thermal energy to approximately 20 MeV^[1]. No distinction between the different techniques available in the marketplace is made in the description of the tests. ISO 21909 (series) aims at covering all passive neutron detectors that can be used as a personal dosimeter in parts of, or in the complete above-mentioned neutron energy range.

The main objective of ISO 21909 series is to achieve correspondence between performance tests and conditions of use at the workplaces. Dosimetry systems complying totally with ISO 21909-1 should give consistent dosimetry results in workplace environments without the requirement of precise information on the characteristics of the radiation fields (neutron energy and direction distributions).

For the case that a dosimetry system does not comply with the full range of requirements in ISO 21909-1 with regard to the dependence of the response on the energy and direction distributions of the neutron fluence, it remains necessary to evaluate the performance of the dosimetry system for the conditions of the workplace. That means that this document is systematically used to qualify at workplaces a dosimetry system that does not fulfil the criteria of ISO 21909-1 on the dependence of the response on neutron energy and direction of incidence.

This document aims to address dosimetry systems with responses that show energy and directional dependencies that do not comply with the test requirements in ISO 21909-1, but that are able to give consistent and reliable dosimetry results at selected workplaces. In this case, a specific study of the workplace where the dosimetry systems are used is necessary to demonstrate that the dosimetry systems are suited for the workplace of application and, if needed, to determine the appropriate corrections to be applied. This document gives requirements for the qualification of the dosimetry system as well as methods for evaluating its performance and qualifying it for use in the workplace.

In cases where the dosimetry system meets the requirements of ISO 21909-1, it may still be desirable to perform a similar study at the workplace to improve the performance of the neutron dosimeters. It is also recommended that this document may be implemented, not only for passive dosimetry systems, but for active dosimetry systems as well.

No qualification or correction of the dosimetry system at a workplace is required if the dosimetry system fulfils the criteria of ISO 21909-1.

All the estimations of the uncertainties in this document have to be considered in accordance with the GUM^[2]. Uncertainties quoted in this document are provided using a coverage factor $k=2$.

Passive neutron dosimetry systems —

Part 2:

Methodology and criteria for the qualification of personal dosimetry systems in workplaces

1 Scope

This document provides methodology and criteria to qualify the dosimetry system at workplaces where it is used. The criteria in this document apply to dosimetry systems which do not meet the criteria with regard to energy and direction dependent responses described in ISO 21909-1.

The qualification of the dosimetry system at workplace aims to demonstrate that:

- either, the non-conformity of the dosimetry system to some of the requirements on the energy or direction dependent responses defined in ISO 21909-1 does not lead to significant discrepancies in the dose determination for a certain workplace field;
- or, that the correction factor or function used for this specific studied workplace enables the dosimetry system to accurately determine the conventional dose value with uncertainties similar to the ones given in ISO 21909-1.

NOTE This document is directed at all stakeholders who are involved: IMSs, accreditation or regulatory bodies, and users of the particular dosimetry (the user is meant as the entity which assigns the dosimetry system to the radiation worker and records the assigned dose.)

The methodologies to characterize the work place field in order to perform the qualification of the dosimetry system are given in [Annex A](#). [Annex B](#) is complementary as it gives the practical methods to follow, once one methodology is chosen.

The provider of the dosimetry system shall provide the type test results corresponding to ISO 21909-1. However, when the dosimetry system to be qualified does not comply with all the criteria of ISO 21909-1 dealing with the energy and angle dependence of the response, some tests of the ISO 21909-1 can be not performed.

The links between ISO 21909-1 and ISO 21909-2 are described in [Annex E](#).

This document only addresses neutron personal monitoring and not criticality accident conditions.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitute requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 21909-1:2021, *Passive neutron dosimetry systems — Part 1: Performance and test requirements for personal dosimetry*

ISO 8529-2:2000, *Reference neutron radiations — Part 2: Calibration fundamentals of radiation protection devices related to the basic quantities characterizing the radiation field*

ISO 8529-3:1998, *Reference neutron radiations — Part 3: Calibration of area and personal dosimeters and determination of response as a function of energy and angle of incidence*