

STN	Injekčné systémy na báze ihiel na zdravotnícke použitie Požiadavky a skúšobné metódy Časť 3: Zásobníky a integrované dráhy tekutín (ISO 11608-3: 2022, opravená verzia 2023-01)	STN EN ISO 11608-3 85 5930
------------	--	--

Needle-based injection systems for medical use - Requirements and test methods - Part 3: Containers and integrated fluid paths (ISO 11608-3:2022, Corrected version 2023-01)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 02/24

Obsahuje: EN ISO 11608-3:2022, ISO 11608-3 Corrected version 2023-01:2022

Oznámením tejto normy sa ruší
STN EN ISO 11608-3 (85 5930) z augusta 2022

138068

Úrad pre normalizáciu, metrológiu a skúšobníctvo Slovenskej republiky, 2024
Slovenská technická norma a technická normalizačná informácia je chránená zákonom č. 60/2018 Z. z. o technickej normalizácii v znení neskorších predpisov.

EUROPEAN STANDARD

EN ISO 11608-3

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2022

ICS 11.040.25

Supersedes EN ISO 11608-3:2012

English Version

Needle-based injection systems for medical use - Requirements and test methods - Part 3: Containers and integrated fluid paths (ISO 11608-3:2022)

Systèmes d'injection à aiguille pour usage médical -
Exigences et méthodes d'essai - Partie 3: Conteneurs et
chemins de fluide intégrés (ISO 11608-3:2022)

Kanülenbasierte Injektionssysteme zur medizinischen
Verwendung - Anforderungen und Prüfverfahren - Teil
3: NIS-Behälter und integrierte Flüssigkeitspfade (ISO
11608-3:2022)

This European Standard was approved by CEN on 2 January 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 11608-3:2022 (E)

Contents	Page
European foreword.....	3

European foreword

This document (EN ISO 11608-3:2022) has been prepared by Technical Committee ISO/TC 84 "Devices for administration of medicinal products and catheters" in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2022, and conflicting national standards shall be withdrawn at the latest by November 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11608-3:2012.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 11608-3:2022 has been approved by CEN as EN ISO 11608-3:2022 without any modification.

INTERNATIONAL STANDARD

ISO 11608-3

Third edition
2022-04

Corrected version
2023-01

Needle-based injection systems for medical use — Requirements and test methods —

Part 3: Containers and integrated fluid paths

*Systèmes d'injection à aiguille pour usage médical — Exigences et
méthodes d'essai —*

Partie 3: Conteneurs et chemins de fluide intégrés



Reference number
ISO 11608-3:2022(E)

© ISO 2022

ISO 11608-3:2022(E)**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Requirements	3
4.1 General.....	3
4.2 Container integrity.....	4
4.2.1 Container Closure Integrity (CCI).....	4
4.2.2 Resealability — All multi-dose cartridges or reservoirs with discs.....	4
4.2.3 Fragmentation (disc coring) – cartridges or reservoirs with discs.....	4
4.3 Cannula requirements (as part of the fluid path).....	5
4.3.1 Rigid needles.....	5
4.3.2 Soft cannulas.....	5
4.4 Fluid line connections.....	5
4.5 Medicinal product compatibility.....	5
4.5.1 General.....	5
4.5.2 Medicinal product compatibility with reservoir and integrated fluid path materials.....	6
4.5.3 Reservoir and integrated fluid path particulate matter.....	6
4.5.4 Reservoir and fluid path pyrogenicity.....	6
4.5.5 Reservoir and integrated fluid path leachables.....	7
4.5.6 Sterilization of the reservoir and/or integrated fluid path.....	7
4.6 Medicinal product leakage.....	8
5 Test methods	8
5.1 Resealability for multi-dose cartridges or reservoirs.....	8
5.2 Fragmentation (disc coring) – cartridges or reservoirs.....	9
5.3 Sub-visible particulates.....	10
5.4 Visible particulates.....	10
6 Information supplied with the container	10
6.1 General.....	10
6.2 Marking on the unit packaging.....	10
Annex A (informative) Medicinal product compatibility references – Requirements, guidance, standards or compendia material	11
Annex B (informative) Historical references to previous editions	14
Annex C (informative) Theoretical support for resealability requirements	17
Annex D (informative) Reservoir and integrated fluid path leachables	20
Annex E (informative) Medicinal product compatibility	22
Annex F (informative) Primary container closure as compared to reservoir and fluid path	24
Bibliography	27

ISO 11608-3:2022(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 84, *Devices for administration of medicinal products and catheters*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 205, *Non-active medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 11608-3:2012), which has been technically revised.

The main changes are as follows:

- test methods and dimensions specific to traditional pen-injector “Type A” cartridges have been removed.

A list of all parts in the ISO 11608 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This corrected version of ISO 11608-3:2022 incorporates the following corrections:

- "Unless otherwise justified" has been added to the beginning of [4.5.3.3](#).

Introduction

Developers and manufacturers of NIS are encouraged to investigate and determine if there are any other requirements relevant to the safety of their products.

Previous editions of this document focused on multi-dose pen-injector cartridges, important dimensions (e.g. inner diameter) and related attributes (e.g., disc seal eccentricity, meniscus) deemed critical for pen-injector form, fit, and function. The previous edition (i.e. ISO 11608-3:2012) included a more general discussion of "other containers" like syringes given their role in single dose NIS with automated functions (commonly referred to as auto-injectors).

Needle-based injection systems for medical use — Requirements and test methods —

Part 3: Containers and integrated fluid paths

1 Scope

This document specifies requirements and test methods for design verification of containers and integrated fluid paths used with Needle-Based Injection Systems (NISs) according to ISO 11608-1.

It is applicable to single and multi-dose containers either filled by the manufacturer (primary container closure) or by the end-user (reservoir) (e.g. cartridges, syringes) and fluid paths that are integrated with the NIS at the point of manufacture.

This document is also applicable to prefilled syringes (see ISO 11040-8) when used with a NIS (see also scope of ISO 11608-1:2022).

This document is not applicable to the following products:

- sterile hypodermic needles;
- sterile hypodermic syringes;
- sterile single-use syringes, with or without needle, for insulin;
- containers that can be refilled multiple times;
- containers intended for dental use;
- catheters or infusion sets that are attached or assembled separately by the user.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7864:2016, *Sterile hypodermic needles for single use — Requirements and test methods*

ISO 8872, *Aluminium caps for transfusion, infusion and injection bottles — General requirements and test methods*

ISO 9626:2016, *Stainless steel needle tubing for the manufacture of medical devices — Requirements and test methods*

ISO 10555-1:2013, *Intravascular catheters — Sterile and single-use catheters — Part 1: General requirements*

ISO 10555-5:2013, *Intravascular catheters — Sterile and single-use catheters — Part 5: Over-needle peripheral catheters*

ISO 10993-11, *Biological evaluation of medical devices — Part 11: Tests for systemic toxicity*

ISO 11608-3:2022(E)

ISO 10993-17, *Biological evaluation of medical devices — Part 17: Establishment of allowable limits for leachable substances*

ISO 11040-4, *Prefilled syringes — Part 4: Glass barrels for injectables and sterilized subassembled syringes ready for filling*

ISO 11040-5, *Prefilled syringes — Part 5: Plunger stoppers for injectables*

ISO 11040-6, *Prefilled syringes — Part 6: Plastic barrels for injectables and sterilized subassembled syringes ready for filling*

ISO 11040-8, *Prefilled syringes — Part 8: Requirements and test methods for finished prefilled syringes*

ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*

ISO 11608-1:2022, *Needle-based injection systems for medical use — Requirements and test methods — Part 1: Needle-based injection systems*

ISO 13926-1, *Pen systems — Part 1: Glass cylinders for pen-injectors for medical use*

ISO 13926-2, *Pen systems — Part 2: Plunger stoppers for pen-injectors for medical use*

ISO 13926-3, *Pen systems — Part 3: Seals for pen-injectors for medical use*

ISO 21881, *Sterile packaged ready for filling glass cartridges*

ISO 80369-7, *Small-bore connectors for liquids and gases in healthcare applications — Part 7: Connectors for intravascular or hypodermic applications*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN