

<b>STN</b>	<b>Mikrobiológia potravinárskeho reťazca Všeobecné požiadavky a pokyny na mikrobiologické skúšanie (ISO 7218: 2024)</b>	<b>STN EN ISO 7218</b>  56 0104
------------	---	---

Microbiology of the food chain - General requirements and guidance for microbiological examinations (ISO 7218:2024)

Táto norma obsahuje anglickú verziu európskej normy.  
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 10/24

Rozpracovaná prekladom.

Obsahuje: EN ISO 7218:2024, ISO 7218:2024

Oznámením tejto normy sa ruší  
STN EN ISO 7218 (56 0104) z januára 2008

**139135**

---

Úrad pre normalizáciu, metrológiu a skúšobníctvo Slovenskej republiky, 2024  
Slovenská technická norma a technická normalizačná informácia je chránená zákonom č. 60/2018 Z. z. o technickej normalizácii  
v znení neskorších predpisov.

EUROPEAN STANDARD

EN ISO 7218

NORME EUROPÉENNE

EUROPÄISCHE NORM

July 2024

ICS 07.100.30

Supersedes EN ISO 7218:2007, EN ISO  
7218:2007/A1:2013

English Version

## Microbiology of the food chain - General requirements and guidance for microbiological examinations (ISO 7218:2024)

Microbiologie de la chaîne alimentaire - Exigences  
générales et recommandations pour les examens  
microbiologiques (ISO 7218:2024)

Mikrobiologie der Lebensmittelkette - Allgemeine  
Anforderungen und Leitlinien für mikrobiologische  
Untersuchungen (ISO 7218:2024)

This European Standard was approved by CEN on 28 April 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

**EN ISO 7218:2024 (E)**

<b>Contents</b>	<b>Page</b>
<b>European foreword.....</b>	<b>3</b>

## **European foreword**

This document (EN ISO 7218:2024) has been prepared by Technical Committee ISO/TC 34 "Food products" in collaboration with Technical Committee CEN/TC 463 "Microbiology of the food chain" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2025, and conflicting national standards shall be withdrawn at the latest by January 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 7218:2007.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

## **Endorsement notice**

The text of ISO 7218:2024 has been approved by CEN as EN ISO 7218:2024 without any modification.



# International Standard

**ISO 7218**

## **Microbiology of the food chain — General requirements and guidance for microbiological examinations**

*Microbiologie de la chaîne alimentaire — Exigences générales et  
recommandations pour les examens microbiologiques*

**Fourth edition  
2024-06**

**ISO 7218:2024(en)****COPYRIGHT PROTECTED DOCUMENT**

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

## ISO 7218:2024(en)

## Contents

	Page
<b>Foreword</b> .....	<b>vii</b>
<b>Introduction</b> .....	<b>viii</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Premises</b> .....	<b>5</b>
4.1 General.....	5
4.2 Biosafety considerations.....	5
4.3 Laboratory design.....	5
4.4 Laboratory areas.....	5
4.4.1 General.....	5
4.4.2 Areas associated with samples and testing.....	6
4.4.3 General areas.....	6
4.5 Layout and fittings of the premises.....	6
4.5.1 Objectives.....	6
4.5.2 Fittings.....	7
4.5.3 Other arrangements for laboratory premises.....	7
4.5.4 Cleaning and disinfection.....	8
<b>5 Personnel</b> .....	<b>8</b>
5.1 General.....	8
5.2 Competence.....	8
5.3 Verification of ongoing staff competence.....	9
5.4 Hygiene.....	9
<b>6 Equipment and consumables</b> .....	<b>9</b>
6.1 General.....	9
6.2 Sterilization and other heating equipment.....	10
6.2.1 General.....	10
6.2.2 Autoclave.....	10
6.2.3 Culture media preparator.....	11
6.2.4 Steamers, including boiling-water baths.....	12
6.2.5 Sterilizing oven.....	12
6.2.6 Microwave oven.....	13
6.2.7 Hotplate, induction cooker and heating mantle.....	14
6.2.8 Gas burner or wire incinerator.....	14
6.3 Temperature controlled equipment and monitoring devices.....	15
6.3.1 General.....	15
6.3.2 Incubator.....	15
6.3.3 Thermostatically controlled bath.....	16
6.3.4 Heating blocks.....	17
6.3.5 Refrigerators and cold-storage rooms.....	18
6.3.6 Freezer and deep freezer/ultra-low temperature freezer.....	19
6.3.7 Temperature-monitoring devices, including automatic recorders.....	19
6.3.8 Balances and gravimetric diluters.....	20
6.4 Defined volume inoculation equipment.....	21
6.4.1 Pipettes and pipettors.....	21
6.4.2 Dispensers.....	22
6.4.3 Spiral platers.....	23
6.4.4 Serial diluters.....	24
6.5 Protective cabinets.....	24
6.5.1 Description.....	24
6.5.2 Use.....	25
6.5.3 Cleaning and disinfection.....	25
6.5.4 Maintenance and inspection.....	26

**ISO 7218:2024(en)**

6.6	Homogenizers, blenders, mixers and shakers .....	26
6.6.1	Homogenizers and blenders .....	26
6.6.2	Vortex mixers .....	27
6.7	Stills, deionizers and reverse-osmosis units .....	28
6.7.1	Description .....	28
6.7.2	Use .....	28
6.7.3	Maintenance .....	28
6.7.4	Verification .....	28
6.8	Separation and concentration equipment .....	28
6.8.1	Immunomagnetic separator (IMS) .....	28
6.8.2	Centrifuge .....	29
6.8.3	Filtration systems .....	29
6.9	Modified atmosphere equipment .....	29
6.9.1	Description .....	29
6.9.2	Use .....	29
6.9.3	Maintenance .....	30
6.9.4	Verification .....	30
6.10	Other equipment .....	30
6.10.1	pH meter .....	30
6.10.2	Colony-counting device .....	31
6.10.3	Timers and timing devices .....	31
6.10.4	Optical microscope .....	32
6.10.5	Glass washers, glassware and other laboratory ware .....	32
6.10.6	Disposable equipment and consumables .....	33
6.10.7	Other equipment and software .....	34
<b>7</b>	<b>Sterilization/decontamination and disposal of laboratory materials .....</b>	<b>34</b>
7.1	Sterilization .....	34
7.1.1	General .....	34
7.1.2	Sterilization by dry heat .....	34
7.1.3	Sterilization by moist heat (steam) .....	34
7.2	Decontamination and disinfection .....	34
7.2.1	Decontamination of glassware and materials before use .....	34
7.2.2	Decontamination of glassware and materials after use .....	34
7.3	Waste management .....	35
7.4	Washing .....	35
<b>8</b>	<b>Preparation and use of culture media and reagents .....</b>	<b>35</b>
<b>9</b>	<b>Laboratory samples .....</b>	<b>36</b>
9.1	Sampling techniques and sampling plans .....	36
9.1.1	General .....	36
9.1.2	Sampling .....	36
9.2	Sample transport .....	36
9.3	Sample receipt .....	37
9.4	Sample handling .....	37
9.4.1	General .....	37
9.4.2	Storage before examination .....	38
9.4.3	Test portions .....	38
9.4.4	Storage of laboratory samples after examination .....	38
9.5	Pre-testing of samples .....	38
<b>10</b>	<b>Examination .....</b>	<b>39</b>
10.1	Hygienic precautions during sample preparation and examination .....	39
10.1.1	General .....	39
10.1.2	Basic precautions .....	39
10.1.3	Sample handling .....	39
10.1.4	Sample handling tools and implements .....	40
10.1.5	Spillages .....	40
10.1.6	Process controls .....	40
10.1.7	Aerosols .....	40

**ISO 7218:2024(en)**

10.1.8	Molecular methods.....	41
10.2	Preparation of initial suspension and dilutions.....	41
10.2.1	General.....	41
10.2.2	Concentration.....	41
<b>11</b>	<b>Enumeration (quantitative) methods.....</b>	<b>41</b>
11.1	General.....	41
11.2	Enumeration using a solid medium.....	42
11.2.1	General.....	42
11.2.2	Pour plate technique.....	42
11.2.3	Surface plating techniques.....	43
11.2.4	Enumeration of yeasts and moulds.....	44
11.2.5	Incubation.....	45
11.2.6	Calculation and expression of results obtained with solid culture media.....	45
11.2.7	Calculations for enumeration methods.....	47
11.3	Enumeration using liquid media.....	54
11.3.1	Principle.....	54
11.3.2	General MPN procedure.....	54
11.3.3	Limitations of MPN.....	54
11.3.4	Inoculation procedure.....	55
11.3.5	Choice of MPN configuration.....	55
11.3.6	Incubation.....	56
11.3.7	Interpretation and expression of results.....	56
11.3.8	Determination of MPN values using MPN calculators.....	56
11.3.9	Rarity categories.....	57
11.4	Estimates of uncertainty of test results.....	57
<b>12</b>	<b>Detection (qualitative) methods.....</b>	<b>58</b>
12.1	General.....	58
12.2	Principle.....	58
<b>13</b>	<b>Confirmation and identification methods.....</b>	<b>58</b>
13.1	General.....	58
13.2	Preparation of a pure culture.....	59
13.3	Confirmation methods.....	59
13.3.1	Latex agglutination test.....	59
13.3.2	Nucleic acid hybridization or molecular amplification methods.....	59
13.3.3	Slide agglutination tests.....	60
13.4	Identification methods.....	60
13.4.1	Biochemical galleries.....	60
13.4.2	DNA sequencing.....	60
13.4.3	Mass spectrometry.....	61
<b>14</b>	<b>Selection and characterization of control microorganisms.....</b>	<b>61</b>
14.1	General.....	61
14.2	Characterization of microorganisms.....	62
14.2.1	General.....	62
14.2.2	Phenotypic characterization.....	62
14.2.3	Molecular characterization.....	62
14.3	Selection of control microorganisms.....	62
<b>15</b>	<b>Test report.....</b>	<b>63</b>
<b>16</b>	<b>Laboratory quality control in microbiology.....</b>	<b>64</b>
16.1	General.....	64
16.2	Internal quality control.....	65
16.2.1	General.....	65
16.2.2	Process controls.....	65
16.2.3	Replicate testing.....	66
16.2.4	Spiked samples.....	66
16.2.5	IQC assessment using control charts.....	66
16.3	External quality assessment.....	66

**ISO 7218:2024(en)**

<b>17</b>	<b>Validation and verification of microbiological methods</b> .....	<b>67</b>
17.1	General.....	67
17.2	Performance characteristics.....	67
17.3	Validation.....	67
17.4	Verification.....	68
<b>Annex A</b>	<b>(informative) Properties of disinfectants</b> .....	<b>69</b>
<b>Annex B</b>	<b>(informative) Confidence intervals for colony count technique</b> .....	<b>70</b>
<b>Annex C</b>	<b>(normative) General confirmation tests</b> .....	<b>73</b>
<b>Bibliography</b>	.....	<b>78</b>

## ISO 7218:2024(en)

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at [www.iso.org/patents](http://www.iso.org/patents). ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 9, *Microbiology*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 463, *Microbiology of the food chain*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This fourth edition cancels and replaces the third edition (ISO 7218:2007), which has been technically revised. It also incorporates the Amendment ISO 7218:2007/Amd 1:2013.

The main changes are as follows:

- the calculations section has been simplified and two further calculators have been added;
- the equipment section has been reorganized into groups with similar purposes and requirements;
- cross-references have been added to other general microbiology standards such as those for media, validation and verification, and uncertainty to reduce repetition;
- information on laboratory quality control and characterization of control microorganisms has been added.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## ISO 7218:2024(en)

### Introduction

When conducting microbiological examinations, it is especially important that:

- only those microorganisms present in the samples are detected and/or enumerated;
- these microorganisms do not contaminate the environment.

To achieve this, good laboratory practices are essential, including personal hygiene and aseptic working techniques which exclude extraneous contamination as far as possible.

Only limited information on the precautions to be taken during microbiological examinations is given in this document, so a thorough knowledge of the microbiological techniques and microorganisms involved is essential. It is important that examinations are conducted safely, correctly and as carefully as possible, including monitoring and recording aspects that can affect results, calculating numbers of microorganisms and assessing the uncertainty of test results.

The most common risks and their control in the microbiological laboratory are given in this document. However, work processes in each laboratory can differ and appropriate risk analysis should be considered to ensure good laboratory practices. Periodic evaluation and control of critical points not only maintains safe and hygienic practices but can also improve reliability of test results.

The purpose of this document is to help to ensure the validity of microbiology examinations in the food chain. In particular, to ensure that general techniques for conducting examinations are the same in all laboratories, to achieve consistent results in different laboratories and to contribute to safety of laboratory personnel by preventing risks of infection.

This document includes the main measures necessary for conducting the wide range of microbiological examinations. Additional information is available from the literature listed in the Bibliography (see References [43] to [47]).

In this document, the following verbal forms are used:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” indicates a possibility or a capability.

In addition, the imperative mood is used to give instructions or where actions are required.

# Microbiology of the food chain — General requirements and guidance for microbiological examinations

## 1 Scope

This document specifies general requirements and gives guidance on microbiological examinations.

It is applicable to:

- the implementation of specific horizontal or vertical International Standards developed by ISO/TC 34/SC 9 or ISO/TC 34/SC 5 for detection or enumeration of microorganisms, named hereafter “specific standards”;
- good laboratory practices for microbiology laboratories testing samples from the food chain;
- guidance for microbiological laboratories testing samples from the food chain on the technical requirements for conforming to ISO/IEC 17025.

The requirements of this general standard supersede corresponding ones in existing specific standards.

Additional instructions for examinations using the polymerase chain reaction (PCR) are specified in ISO 22174.

This document is applicable to examinations for bacteria, yeasts and moulds and can be used, if supplemented with specific guidance, for parasites and viruses. It does not apply to examinations for toxins or other metabolites (e.g. amines) from microorganisms.

This document is applicable to microbiology of the food chain, from primary production stage to food and animal feed products, including the premises where the food or feed production and handling takes place. It is also applicable to the microbiological examination of water where water is used in food production or is regarded as a food in national legislation.

## 2 Normative references

There are no normative references in this document.

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**