

STN P	Nanotechnológie Charakterizácia nanoobjektov použitím hmotnostnej spektrometrie jednej častice s indukčne viazanou plazmou (ISO/TS 19590: 2024)	STN P CEN ISO/TS 19590 60 3035
------------------	--	--

Nanotechnologies - Characterization of nano-objects using single particle inductively coupled plasma mass spectrometry (ISO/TS 19590:2024)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 11/24

Táto predbežná slovenská technická norma je určená na overenie. Prípadné pripomienky pošlite do augusta 2026 Úradu pre normalizáciu, metrológiu a skúšobníctvo Slovenskej republiky.

Obsahuje: CEN ISO/TS 19590:2024, ISO/TS 19590:2024

Oznámením tejto normy sa ruší
STN P CEN ISO/TS 19590 (60 3035) z augusta 2019

139409



Úrad pre normalizáciu, metrológiu a skúšobníctvo Slovenskej republiky, 2024
Slovenská technická norma a technická normalizačná informácia je chránená zákonom č. 60/2018 Z. z. o technickej normalizácii v znení neskorších predpisov.

TECHNICAL SPECIFICATION

CEN ISO/TS 19590

SPÉCIFICATION TECHNIQUE

TECHNISCHE SPEZIFIKATION

August 2024

ICS 07.120

Supersedes CEN ISO/TS 19590:2019

English Version

Nanotechnologies - Characterization of nano-objects using single particle inductively coupled plasma mass spectrometry (ISO/TS 19590:2024)

Nanotechnologies - Caractérisation des nano-objets par
spectrométrie de masse à plasma induit en mode
particule unique (ISO/TS 19590:2024)

Nanotechnologien - Charakterisierung von
Nanoobjekten mit Hilfe der Massenspektrometrie mit
induktiv gekoppeltem Einzelpartikelplasma (ISO/TS
19590:2024)

This Technical Specification (CEN/TS) was approved by CEN on 12 August 2024 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

CEN ISO/TS 19590:2024 (E)

Contents	Page
European foreword.....	3

European foreword

This document (CEN ISO/TS 19590:2024) has been prepared by Technical Committee ISO/TC 229 "Nanotechnologies" in collaboration with Technical Committee CEN/TC 352 "Nanotechnologies" the secretariat of which is held by AFNOR.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN ISO/TS 19590:2019.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO/TS 19590:2024 has been approved by CEN as CEN ISO/TS 19590:2024 without any modification.



Technical Specification

ISO/TS 19590

Nanotechnologies — Characterization of nano-objects using single particle inductively coupled plasma mass spectrometry

*Nanotechnologies — Caractérisation des nano-objets par
spectrométrie de masse à plasma induit en mode particule unique*

**Second edition
2024-08**

ISO/TS 19590:2024(en)**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

ISO/TS 19590:2024(en)

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Abbreviated terms	3
5 Principles of operation	4
5.1 Introduction to spICP-MS.....	4
5.2 Reference material dependent calibration methods.....	6
5.2.1 Particle frequency method.....	6
5.2.2 Particle size method.....	8
5.3 Reference material free calibration methods.....	9
5.3.1 Dynamic mass flow method.....	9
5.3.2 Microdroplet calibration method.....	11
5.4 Particle number concentration determination.....	13
5.5 Particle mass and corresponding spherical equivalent diameter determination.....	14
5.6 Dissolved element fraction.....	17
5.7 Multi-isotope and multi-elemental analysis.....	17
5.8 Data treatment.....	18
6 Method development	19
6.1 Sample specification.....	19
6.2 Sample preparation.....	19
6.2.1 Aqueous suspensions and paste.....	20
6.2.2 Non-aqueous suspensions and creams.....	20
6.2.3 Powders.....	21
6.2.4 Larger pieces of solids.....	21
6.3 Selection of reference materials, quality control materials and representative test materials.....	21
6.4 Optimization of ICP-MS operating conditions.....	22
7 Qualification, performance criteria and measurement uncertainty	23
7.1 Applicability of spICP-MS.....	23
7.2 System qualification and quality control.....	23
7.3 Method performance criteria.....	24
7.3.1 Particle number concentration.....	24
7.3.2 Particle mass and equivalent spherical diameter.....	24
7.4 Method precision and measurement uncertainty.....	25
8 General measurement procedure	25
9 Test report	26
9.1 Apparatus and measurement parameters.....	26
9.2 Reporting test results.....	26
Bibliography	27

ISO/TS 19590:2024(en)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 229, *Nanotechnologies*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 352, *Nanotechnologies*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO/TS 19590:2017), which has been technically revised.

The main changes are as follows:

- general restructuring;
- expansion of text on the test method;
- inclusion of considerations regarding method precision and measurement uncertainty;
- updates to normative and bibliographical references.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

ISO/TS 19590:2024(en)

Introduction

Following the introduction of single particle inductively coupled plasma mass spectrometry (spICP-MS) by Degueldre in 2003,^[1] the technique has increasingly been used for nano-object characterization due to its high sensitivity, elemental specificity, the fact that often minimal sample preparation is needed and the development of much improved instrumentation, along with user-friendly data analysis software.

In spICP-MS, a very diluted suspension containing nano-objects is introduced continuously into an ICP-MS system with the intent that the ion cloud from one particle at a time arrives at the detector, set to acquire data with a high time resolution (i.e. dwell time). Following the nebulization, a fraction of the nano-objects enter the plasma where they are atomized, and the individual atoms ionized. Every atomized particle results in a cloud of ions which is then sampled by the mass spectrometer. The mass spectrometer can be tuned to measure any specific element. Typically, only one mass-to-charge value per single particle will be monitored with a quadrupole-based MS instrumentation. However, the technique can also be used with time-of-flight (TOF) mass spectrometers, allowing simultaneous multi-element and multi-isotope detection.

The number of events detected in each run (time scan) is directly proportional to the number of nano-objects in the suspension introduced but necessitates calibration of the sample transport efficiency to calculate the particle number concentration. Several available approaches to measure the transport efficiency are described in detail in this document. The intensity of the measured signal is directly proportional to the mass of the measured element in the nano-object, which can be derived following appropriate calibration of the instrument's response factor, also described in this document. For particles of known geometry, composition and density, the mass can be related to particle size. Most of the currently available, commercial data analysis software assumes spherical geometry; particle diameter is proportional to the cubic root of the mass of element(s) in a spherical nano-object. In addition to nano-object characterization with spICP-MS, mass concentrations of dissolved element present in the same sample can also be determined from the same data, if a good separation between the dissolved and particulate fraction is achieved. This represents one of the key advantages of the technique.

spICP-MS was once predominantly the domain of specialist laboratories, but with recent developments in commercially available hardware and software, the technique is now more commonly used and increasingly popular for high-throughput analysis as well as high accuracy reference measurements.

Further information on spICP-MS can be found in ISO/TS 24672, and References [1], [2], [3], [4] and [5].

Nanotechnologies — Characterization of nano-objects using single particle inductively coupled plasma mass spectrometry

1 Scope

This document specifies parameters, conditions and considerations for the reliable detection, characterization and quantification of nano-objects in aqueous suspension by spICP-MS.

Particle number concentration, particle mass, particle mass concentration, particle spherical equivalent diameter, and number-based size distribution are considered the main measurands, but the technique also allows for determination of the dissolved element mass fraction in the sample. This document provides general guidelines and procedures related to spICP-MS application, and specifies minimal reporting requirements.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 18115-1, *Surface chemical analysis — Vocabulary — Part 1: General terms and terms used in spectroscopy*

ISO/TS 80004-6, *Nanotechnologies — Vocabulary — Part 6: Nano-object characterization*

ISO/TS 80004-8, *Nanotechnologies — Vocabulary — Part 8: Nanomanufacturing processes*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN