

<b>STN</b>	<b>Oftalmologické prístroje Základné požiadavky a skúšobné metódy Časť 2: Ochrana pred svetelným nebezpečenstvom (ISO 15004-2: 2024)</b>	<b>STN EN ISO 15004-2</b>  19 5115
------------	--	--

Ophthalmic instruments - Fundamental requirements and test methods - Part 2: Light hazard protection (ISO 15004-2:2024)

Táto norma obsahuje anglickú verziu európskej normy.

This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 02/25

Obsahuje: EN ISO 15004-2:2024, ISO 15004-2:2024

Oznámením tejto normy sa ruší

STN EN ISO 15004-2 (19 5115) z júla 2007

**140081**



Úrad pre normalizáciu, metrológiu a skúšobníctvo Slovenskej republiky, 2025

Slovenská technická norma a technická normalizačná informácia je chránená zákonom č. 60/2018 Z. z. o technickej normalizácii v znení neskorších predpisov.

EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**EN ISO 15004-2**

December 2024

ICS 11.040.70

Supersedes EN ISO 15004-2:2007

English Version

**Ophthalmic instruments - Fundamental requirements and  
test methods - Part 2: Light hazard protection (ISO 15004-  
2:2024)**

Instruments ophtalmiques - Exigences fondamentales  
et méthodes d'essai - Partie 2: Protection contre les  
dangers de la lumière (ISO 15004-2:2024)

Ophthalmische Instrumente - Grundlegende  
Anforderungen und Prüfverfahren - Teil 2: Schutz  
gegen Gefährdung durch Licht (ISO 15004-2:2024)

This European Standard was approved by CEN on 18 November 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

Contents	Page
European foreword.....	3

## **European foreword**

This document (EN ISO 15004-2:2024) has been prepared by Technical Committee ISO/TC 172 "Optics and photonics" in collaboration with Technical Committee CEN/TC 170 "Ophthalmic optics" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2025, and conflicting national standards shall be withdrawn at the latest by June 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 15004-2:2007.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

## **Endorsement notice**

The text of ISO 15004-2:2024 has been approved by CEN as EN ISO 15004-2:2024 without any modification.



# International Standard

## ISO 15004-2

### Ophthalmic instruments — Fundamental requirements and test methods —

#### Part 2: Light hazard protection

*Instruments ophtalmiques — Exigences fondamentales et  
méthodes d'essai —*

*Partie 2: Protection contre les dangers de la lumière*

### Second edition 2024-12

## ISO 15004-2:2024(en)



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

**ISO 15004-2:2024(en)****Contents**

Page

<b>Foreword</b>	<b>v</b>
<b>Introduction</b>	<b>vi</b>
<b>1 Scope</b>	<b>1</b>
<b>2 Normative references</b>	<b>1</b>
<b>3 Terms and definitions</b>	<b>1</b>
<b>4 Classification principles</b>	<b>7</b>
4.1 General	7
4.2 Group 1 instruments	9
4.3 Group 2 instruments	9
<b>5 Requirements for classification</b>	<b>9</b>
5.1 General	9
5.2 Requirements for classification as a Group 1 instrument	9
5.3 Requirements for classification as a Group 2 instrument	10
5.4 Exposure limits for Group 1 classification	10
5.4.1 Continuous wave instruments	10
5.4.2 Pulsed instruments	10
5.4.3 Time-limited instruments	11
5.4.4 Dose limited instruments	11
5.4.5 Scanning instruments	11
5.4.6 Instruments with multiple sources	12
5.4.7 Instruments for long-term repetitive daily use	12
5.5 Recommended maximum exposure (RME) values for Group 2 instruments	12
5.5.1 Continuous wave instruments	12
5.5.2 Pulsed instruments	12
5.5.3 Time-limited instruments	13
5.5.4 Scanning instruments	13
5.5.5 Instruments with multiple sources	13
5.5.6 Instruments for long-term repetitive daily use	14
5.6 Exposure limits and maximum recommended exposure values	14
<b>6 Test methods</b>	<b>19</b>
6.1 General	19
6.2 Classification of instruments into Group 1 or Group 2	19
6.3 Spectral measurements	20
6.4 Group 2 instruments: determination of time and number of pulses to reach recommended maximum exposure	20
6.4.1 Determination of time $t_{\max}$ to reach the recommended maximum exposure for weighted corneal and lenticular ultraviolet radiation radiant exposure, $H_{S-CL}$	20
6.4.2 Determination of time $t_{\max}$ to reach the recommended maximum exposure for photochemical aphakic retinal radiant exposure	20
6.4.3 Determination of the number of pulses necessary to reach the recommended maximum exposure for photochemical aphakic retinal exposure, $n_{\max}$ (for pulsed instruments)	21
<b>7 Information supplied by the manufacturer</b>	<b>21</b>
7.1 Information provided on request	21
7.2 Accompanying documents	21
7.2.1 General	21
7.2.2 Dose-limited instruments	21
7.2.3 Cautionary statements	21
7.3 Marking	25
7.3.1 Radiation output	25
7.3.2 Consult accompanying documents	25
7.3.3 Safety information for optical radiation	25

ISO 15004-2:2024(en)

**Annex A** (normative) **Spectral weighting functions**.....26

**Annex B** (informative) **Measurement instruments**.....33

**Annex C** (informative) **Measurement methods for radiance/irradiance** .....37

**Annex D** (informative) **Guidance on the direct measurement of irradiance**.....42

**Annex E** (informative) **Examples of measurement methods for specific ophthalmic instruments**.....44

**Annex F** (informative) **Safety signs**.....55

**Bibliography** .....56



**ISO 15004-2:2024(en)****Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at [www.iso.org/patents](http://www.iso.org/patents). ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 170, *Ophthalmic optics*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 15004-2:2007), which has been technically revised.

The main changes are as follows:

- The terms and definitions include dose limited instruments and time limited instruments;
- The safe exposure limits have been reorganized into 4 tables ([Tables 2 to 5](#)), and the associated measurement conditions have been reorganized into a companion table ([Table 6](#));
- A number of Group 1 exposure limits and Group 2 recommended maximum exposures (RMEs) have been updated to conform to recent research and relevant standards;
- The language has been clarified and simplified throughout the document, and a flowchart has been added as a guide to make the standard more accessible to first-time users;
- Clauses and associated exposure limits have been added for long-term repetitive exposures, such as may apply to extensive use of head-mounted displays by people with visual impairments;
- Provisions have been added to ensure that the exposure limits and RMEs applicable to specific devices are easily accessible to end users.

A list of all parts in the ISO 15004 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

**ISO 15004-2:2024(en)**

## Introduction

Ophthalmic instruments are classified into two groups to distinguish instruments that can present a potential hazard from those that cannot. The two groups are named Group 1 and Group 2. They are defined as follows:

Group 1 instruments: ophthalmic instruments for which no potential light hazard exists when used as intended (see [Clause 4](#)).

Group 2 instruments: ophthalmic instruments for which a potential light hazard exists (see [Clause 4](#)).

Limits and guidelines for optical radiation exposure of the eye during ophthalmic examination can differ from those of non-ophthalmic applications. They can be more restrictive because of pupils dilation or retinal image stabilization, or less restrictive based on benefit/risk ratios. Furthermore, interruptions of exposure during surgical procedures mitigate the risk.

All group 2 instruments pose a potential risk of injury at the upper emission values of the instrument. This is true for both photochemical (where time is critical) and thermal, where transmission and absorption values can vary. Clinical judgement of individual susceptibility is also required.

NOTE 1 The basic limits and guidelines in this document are based on the International Commission on Non-Ionizing Radiation Protection (ICNIRP) guidelines for human exposure to optical radiation. They also cover cases where pupils are dilated, or the image is stabilized on the retina during ophthalmic examinations.

NOTE 2 This document provides exposure limits for ocular tissues. The exposures can be calculated based upon the measured instrument emissions.

The flow chart in [Figure 1](#) at the beginning of [Clause 4](#) provides guidance on how to apply this document to any device to be tested or designed for light hazards conformity.

# Ophthalmic instruments — Fundamental requirements and test methods —

## Part 2: Light hazard protection

### 1 Scope

This document specifies fundamental requirements for optical radiation safety for ophthalmic instruments and is applicable to all ophthalmic instruments that direct optical radiation into or at the eye. It is also applicable to all new and emerging ophthalmic instruments that direct optical radiation into or at the eye, as well as to those portions of therapeutic or surgical systems that direct optical radiation into or at the eye for diagnostic, illumination, measurement, imaging or alignment purposes.

**NOTE** For the purpose of this document, optical radiation relates to the wavelength range of 250 nm to 2 500 nm.

This document does not apply to therapeutic radiation. However, in the case of the treatment beams of therapeutic devices, when conducting risk assessments for non-target tissues, the limits given in this document may be applied to those parts of the treatment beam that strike non-target tissue.

Where vertical (instrument-specific) International Standards contain specific light hazard requirements different from those given in ISO 15004-2, then those in the vertical International Standard take precedence.

This document classifies ophthalmic instruments into either Group 1 or Group 2 to distinguish instruments that are non-hazardous from those that are potentially hazardous.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 15004-1, *Ophthalmic instruments — Fundamental requirements and test methods — Part 1: General requirements applicable to all ophthalmic instruments*

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**