

STN P	Molekulárno-diagnostické vyšetrenia <i>in vitro</i> Špecifikácie predbežných vyšetrení cirkulujúcich nádorových buniek (CTC) v žilovej plnej krvi Časť 2: Izolovaná DNA (ISO/TS 7552-2: 2024)	STN P CEN ISO/TS 7552-2 85 1027
------------------	--	---

Molecular in vitro diagnostic examinations - Specifications for pre-examination processes for circulating tumour cells (CTCs) in venous whole blood - Part 2: Isolated DNA (ISO/TS 7552-2:2024)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 02/25

Táto predbežná slovenská technická norma je určená na overenie. Prípadné pripomienky pošlite do novembra 2026 Úradu pre normalizáciu, metrológiu a skúšobníctvo.

Obsahuje: CEN ISO/TS 7552-2:2024, ISO/TS 7552-2:2024

Oznámením tejto normy sa ruší
STN P CEN/TS 17390-2 (85 1027) z apríla 2020

140105

Úrad pre normalizáciu, metrológiu a skúšobníctvo Slovenskej republiky, 2025
Slovenská technická norma a technická normalizačná informácia je chránená zákonom č. 60/2018 Z. z. o technickej normalizácii v znení neskorších predpisov.

TECHNICAL SPECIFICATION
SPÉCIFICATION TECHNIQUE
TECHNISCHE SPEZIFIKATION

CEN ISO/TS 7552-2

November 2024

ICS 11.100.10

Supersedes CEN/TS 17390-2:2020

English Version

**Molecular in vitro diagnostic examinations - Specifications
for pre-examination processes for circulating tumour cells
(CTCs) in venous whole blood - Part 2: Isolated DNA
(ISO/TS 7552-2:2024)**

Analyses de diagnostic moléculaire in vitro -
Spécifications relatives aux processus préanalytiques
pour les cellules tumorales circulantes (CTC) dans le
sang total veineux - Partie 2: ADN isolé (ISO/TS 7552-
2:2024)

Spezifikationen für präanalytische Prozesse für
zirkulierende Tumorzellen (CTC) in venösen
Vollblutproben - Teil 2: Isolierte DNA (ISO/TS 7552-
2:2024)

This Technical Specification (CEN/TS) was approved by CEN on 10 November 2024 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

CEN ISO/TS 7552-2:2024 (E)

Contents	Page
European foreword.....	3

European foreword

This document (CEN ISO/TS 7552-2:2024) has been prepared by Technical Committee ISO/TC 212 "Medical laboratories and in vitro diagnostic systems" in collaboration with Technical Committee CEN/TC 140 "In vitro diagnostic medical devices" the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN/TS 17390-2:2020.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO/TS 7552-2:2024 has been approved by CEN as CEN ISO/TS 7552-2:2024 without any modification.



Technical Specification

ISO/TS 7552-2

Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for circulating tumour cells (CTCs) in venous whole blood —

Part 2: Isolated DNA

*Analyses de diagnostic moléculaire in vitro — Spécifications
relatives aux processus préanalytiques pour les cellules tumorales
circulantes (CTC) dans le sang total veineux —*

Partie 2: ADN isolé

**First edition
2024-11**

ISO/TS 7552-2:2024(en)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

ISO/TS 7552-2:2024(en)**Contents**

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General considerations	5
5 Activities outside the laboratory	6
5.1 Specimen collection.....	6
5.1.1 General.....	6
5.1.2 Information about the patient/specimen donor.....	6
5.1.3 Selection of the venous whole blood collection tube by the laboratory.....	6
5.1.4 Venous whole blood specimen collection from the patient/donor.....	7
5.2 Specimen storage and transport.....	7
5.2.1 General.....	7
5.2.2 Storage and transport using blood collection tubes with stabilizers.....	8
5.2.3 Storage and transport using blood collection tubes without stabilizers.....	8
6 Activities inside the laboratory	8
6.1 Specimen reception.....	8
6.2 Specimen storage after transport and reception.....	9
6.3 Enrichment of CTCs.....	9
6.3.1 General.....	9
6.3.2 Using a commercial CTC enrichment system intended for diagnostic use.....	9
6.3.3 Using the laboratory developed CTC enrichment procedure.....	10
6.4 Quality of enriched CTCs.....	10
6.5 Storage of enriched CTCs.....	10
6.6 Isolation of CTCs.....	11
6.6.1 General.....	11
6.6.2 Using a commercial CTC isolation system intended for diagnostic use.....	11
6.6.3 Using the laboratory-developed CTC isolation procedure.....	11
6.7 Isolation of DNA from an enriched CTC sample.....	11
6.7.1 General.....	11
6.7.2 Using a commercial DNA isolation kit intended for diagnostic use.....	12
6.7.3 Using a laboratory-developed CTC DNA isolation procedure.....	12
6.8 Quantity and quality assessment of isolated DNA from enriched or isolated CTCs.....	13
6.8.1 General.....	13
6.8.2 Quantity assessment of CTC DNA.....	13
6.8.3 Quality assessment of CTC DNA.....	13
6.9 Storage of isolated DNA from enriched CTCs.....	13
6.9.1 General.....	13
6.9.2 Storage of DNA isolated with a commercially available kit intended for diagnostic use.....	14
6.9.3 Storage of DNA isolated with the laboratory developed procedure.....	14
Annex A (informative) Decision guideline for critical steps of the CTC pre-analytical workflow	15
Bibliography	17

ISO/TS 7552-2:2024(en)**Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 212, *Medical laboratories and in vitro diagnostic systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 140, *In vitro diagnostic medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 7552 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

ISO/TS 7552-2:2024(en)

Introduction

Solid tumours release cells and bioanalytes into blood and other body fluids. This has opened the option of utilizing such body fluids (liquid biopsies) for a minimally-invasive procedure for tumour detection, diagnosis and characterization. Liquid biopsies can enable earlier detection and diagnosis of cancers and advance personalized patient treatment.^[20,22]

These applications have become one of the fastest growing segments of the entire diagnostic market.

Circulating tumour cells (CTCs) in venous whole blood can reflect the disease complexity that evolves during tumour progression, with distinct genetic, epigenetic and gene expression biomarkers.^[22]

Beside the prognostic role of CTC identification and enumeration in cancer progression, CTC molecular characterization can improve disease outcome prediction, therapeutic guidance and post-treatment monitoring of the patient.^[20]

CTCs are now considered as a surrogate of tumour tissue in cancer early development, progression, and metastatic phase.^[23]

Molecular characterization of CTCs can provide a strategy for monitoring cancer genotypes during systemic therapies,^[24] identifying mechanisms of disease progression, identifying novel targets for biological treatment^[25] and selecting targeted therapies.^[20] Moreover, CTC single-cell sequencing is an important tool for tumour genomic heterogeneity analysis.^[26-28] Molecular examination techniques such as qPCR, dPCR and sequencing methods including next generation sequencing (NGS) enable characterization of the CTC specific DNA features.

CTCs are fragile and tend to degrade within a few hours when collected in conventional blood collection tubes, e.g. EDTA containing tubes, without dedicated CTC stabilizers. CTCs are extremely rare, especially in early disease, e.g. less than 10 cells per 10 ml of blood, representing a ratio of approximately 1:10⁷ CTCs to white blood cells (WBCs). This low ratio represents a significant challenge to CTC enrichment required for examination. Furthermore, co-enrichment of normal blood cells causes a dilution of CTCs. The challenge is to minimize the amount of co-enriched WBCs for subsequent accurate and sensitive detection of CTC specific genetic and epigenetic alterations, especially when dealing with minor tumour cell clones.

Special measures to remove the WBCs are necessary in order to obtain good quality DNA samples characterized by high purity and thus representative of the mutational pattern within the tumour.

Standardization includes all steps of the pre-examination process, including blood collection and stabilization, transport, storage, CTC enrichment, CTC isolation (if included), and DNA isolation. This pre-examination standardization is crucial to ensure reliable examination results in current clinical use and is also critical to develop new CTC based diagnostic examinations and to establish these in clinical healthcare.^[29]

An illustration of critical steps of the CTC pre-analytical workflow is provided in [Annex A](#).

This document describes special measures to obtain appropriate quality and quantity of DNA from CTC-containing blood specimens for subsequent examination.

Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for circulating tumour cells (CTCs) in venous whole blood —

Part 2: Isolated DNA

1 Scope

This document specifies requirements and gives recommendations on the handling, storage, CTC enrichment and isolation, DNA isolation and storage, and documentation of venous whole blood specimens intended for the examination of DNA isolated from circulating tumour cells (CTCs) during the pre-examination phase before a molecular examination is performed.

This document is applicable to molecular in vitro diagnostic examinations including laboratory developed tests performed by medical laboratories. It is also intended to be used by laboratory customers, in vitro diagnostics developers and manufacturers, biobanks, institutions, and commercial organizations performing biomedical research, and regulatory authorities.

This document does not cover the isolation of genomic DNA directly from venous whole blood containing CTCs. This is covered in ISO 20186-2.

This document does not cover the isolation of specific white blood cells and subsequent isolation of genomic DNA therefrom or the pre-analytical workflow requirements for viable CTC cryopreservation and culturing.

NOTE 1 The requirements given in this document can also be applied to other circulating rare cells (e.g. foetal cells).

NOTE 2 International, national, or regional regulations or requirements can also apply to specific topics covered in this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 15189, *Medical laboratories — Requirements for quality and competence*

ISO 15190, *Medical laboratories — Requirements for safety*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN