

<b>STN</b>	<b>Meranie rádioaktivity Stanovenie aktivity beta žiaričov Skúšobná metóda využívajúca počítanie kvapalnej scintilácie (ISO 19361: 2025)</b>	<b>STN EN ISO 19361</b>  40 1018
------------	--	--

Measurement of radioactivity - Determination of beta emitters activities - Test method using liquid scintillation counting (ISO 19361:2025)

Táto norma obsahuje anglickú verziu európskej normy.  
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 09/25

Obsahuje: EN ISO 19361:2025, ISO 19361:2025

Oznámením tejto normy sa ruší  
STN EN ISO 19361 (40 1018) z augusta 2020

**141127**

---

Úrad pre normalizáciu, metrológiu a skúšobníctvo Slovenskej republiky, 2025  
Slovenská technická norma a technická normalizačná informácia je chránená zákonom č. 60/2018 Z. z. o technickej normalizácii  
v znení neskorších predpisov.



EUROPEAN STANDARD

**EN ISO 19361**

NORME EUROPÉENNE

EUROPÄISCHE NORM

July 2025

ICS 17.240

Supersedes EN ISO 19361:2020

English Version

## Measurement of radioactivity - Determination of beta emitters activities - Test method using liquid scintillation counting (ISO 19361:2025)

Mesurage de la radioactivité - Détermination de l'activité des radionucléides émetteurs bêta - Méthode d'essai par comptage des scintillations en milieu liquide (ISO 19361:2025)

This European Standard was approved by CEN on 21 July 2025.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

**EN ISO 19361:2025 (E)**

<b>Contents</b>	<b>Page</b>
<b>European foreword.....</b>	<b>3</b>

## **European foreword**

This document (EN ISO 19361:2025) has been prepared by Technical Committee ISO/TC 85 "Nuclear energy, nuclear technologies, and radiological protection " in collaboration with Technical Committee CEN/TC 430 "Nuclear energy, nuclear technologies, and radiological protection" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2026, and conflicting national standards shall be withdrawn at the latest by January 2026.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 19361:2020.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

## **Endorsement notice**

The text of ISO 19361:2025 has been approved by CEN as EN ISO 19361:2025 without any modification.



# International Standard

**ISO 19361**

## **Measurement of radioactivity — Determination of beta emitters activities — Test method using liquid scintillation counting**

*Mesurage de la radioactivité — Détermination de l'activité des radionucléides émetteurs bêta — Méthode d'essai par comptage des scintillations en milieu liquide*

**Second edition  
2025-07**

**ISO 19361:2025(en)****COPYRIGHT PROTECTED DOCUMENT**

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

**ISO 19361:2025(en)****Contents**

Page

<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms, definitions</b> .....	<b>2</b>
<b>4 Symbols</b> .....	<b>2</b>
<b>5 Principle</b> .....	<b>3</b>
<b>6 Reagents and equipment</b> .....	<b>3</b>
6.1 Reagents.....	3
6.1.1 Blank material.....	3
6.1.2 Calibration source solutions.....	4
6.1.3 Scintillation cocktail.....	4
6.1.4 Quenching agent.....	4
6.2 Equipment.....	5
6.2.1 General.....	5
6.2.2 Liquid scintillation counter.....	5
6.2.3 Counting vials.....	5
<b>7 Sampling and samples</b> .....	<b>5</b>
7.1 Sampling.....	5
7.2 Sample storage.....	6
<b>8 Procedure</b> .....	<b>6</b>
8.1 Determination of background.....	6
8.2 Determination of counting efficiency.....	6
8.3 Quench correction.....	6
8.4 Sample preparation.....	7
8.5 Preparation of the scintillation sources to be measured.....	7
8.6 Counting procedure.....	8
8.6.1 Control and calibration.....	8
8.6.2 Measurement conditions.....	8
8.6.3 Interference control.....	8
<b>9 Expression of results</b> .....	<b>9</b>
9.1 General.....	9
9.2 Calculation of activity concentration, without sample treatment prior to measurement.....	9
9.3 Decision threshold, without sample treatment prior to measurement.....	10
9.4 Detection limit, without sample treatment prior to measurement.....	10
9.5 Limit of coverage interval.....	11
9.5.1 Limits of the probabilistically symmetric coverage interval.....	11
9.5.2 Limits of the shortest coverage interval.....	11
9.6 Calculations using the activity per unit of mass, without sample treatment prior to measurement.....	12
<b>10 Test report</b> .....	<b>12</b>
<b>Annex A (informative) Internal standard method</b> .....	<b>13</b>
<b>Annex B (informative) TDCR liquid scintillation counting</b> .....	<b>15</b>
<b>Annex C (informative) Cerenkov measurement with liquid scintillation and TDCR counter</b> .....	<b>18</b>
<b>Bibliography</b> .....	<b>21</b>

**ISO 19361:2025(en)****Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at [www.iso.org/patents](http://www.iso.org/patents). ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 430, *Nuclear energy, nuclear technologies, and radiological protection*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 19361:2017), which has been technically revised.

The main changes are as follows:

- those driven by the ISO 11929 series evolution for expressing results.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## ISO 19361:2025(en)

### Introduction

Everyone is exposed to natural radiation. The natural sources of radiation are cosmic rays and naturally occurring radioactive substances which exist in the earth and within the human body. Human activities involving the use of radiation and radioactive substances add to the radiation exposure from this natural exposure. Some of those activities, such as the mining and use of ores containing naturally-occurring radioactive materials (NORM) and the production of energy by burning coal that contains such substances, simply enhance the exposure from natural radiation sources. Nuclear power plants and other nuclear installations use radioactive materials and produce radioactive effluent and waste during operation and on their decommissioning. The use of radioactive materials in industry, agriculture and research is expanding around the globe.

All these human activities give rise to an average exposure to radiation that represents only a small fraction of the global average level of exposure, taking into account all contributions (natural, medical, radon, etc.). The medical use of radiation is the largest and a growing human-made source of radiation exposure in developed countries. It includes diagnostic radiology, radiotherapy, nuclear medicine and interventional radiology.

Radiation exposure also occurs as a result of occupational activities. It is incurred by workers in industry, medicine and research using radiation or radioactive substances, as well as by passengers and crew during air travel and by astronauts. The average level of occupational exposures is generally similar to the global average level of natural radiation exposure<sup>[1]</sup>.

As uses of radiation increase, so do the potential health risk and the public's concerns. Thus, all these exposures are regularly assessed in order to

- a) improve the understanding of global levels and temporal trends of public and worker exposure,
- b) evaluate the components of exposure so as to provide a measure of their relative importance, and
- c) identify emerging issues that may warrant more attention and study.

While doses to workers are mostly directly assessed, doses to the public are usually assessed by indirect methods using radioactivity measurements performed on waste, effluent and/or environmental samples.

To ensure that the data obtained from radioactivity monitoring programs support their intended use, it is essential that the stakeholders (for example, nuclear site operators, regulatory and local authorities) agree on appropriate methods and procedures for obtaining representative samples and then handling, storing, preparing and measuring the test samples. An assessment of the overall measurement uncertainty needs also to be carried out systematically. As reliable, comparable and 'fit for purpose' data are an essential requirement for any public health decision based on radioactivity measurements, international standards of tested and validated radionuclide test methods are an important tool for the production of such measurement results. The application of standards serves also to guarantee comparability over time of the test results and between different testing laboratories. Laboratories apply them to demonstrate their technical qualifications and to successfully complete proficiency tests during interlaboratory comparison, two prerequisites for obtaining national accreditation. Today, over a hundred international standards, prepared by Technical Committees of the International Standardization Organization (ISO), including those produced by ISO/TC 85, and the International Electrotechnical Commission (IEC), are available for application by testing laboratories to measure the main radionuclides.

Generic standards help testing laboratories to manage the measurement process by setting out the general requirements and methods to calibrate and validate techniques. These standards underpin specific standards which describe the test methods to be performed by staff, for example, for different types of samples. The specific standards cover test methods for:

- Naturally-occurring radionuclides (including  $^{40}\text{K}$ ,  $^3\text{H}$ ,  $^{14}\text{C}$  and those originating from the thorium and uranium decay series, in particular  $^{226}\text{Ra}$ ,  $^{228}\text{Ra}$ ,  $^{234}\text{U}$ ,  $^{238}\text{U}$ ,  $^{210}\text{Pb}$ ) which can be found in materials from natural sources or can be released from technological processes involving naturally occurring radioactive materials (e.g. the mining and processing of mineral sands or phosphate fertilizer production and use);

**ISO 19361:2025(en)**

- Anthropogenic radionuclides, such as transuranium elements (americium, plutonium, neptunium, and curium),  $^3\text{H}$ ,  $^{14}\text{C}$ ,  $^{90}\text{Sr}$  and gamma emitting radionuclides found in waste, liquid and gaseous effluent, in environmental matrices (water, air, soil, biota) and food and feed as a result of authorized releases into the environment and of fallout resulting from the explosion in the atmosphere of nuclear devices and accidents, such as those that occurred in Chernobyl and Fukushima.

Many of these radionuclides are beta emitters that can be measured by liquid scintillation counting, following appropriate sample preparation.

A generic international standard on liquid scintillation counting is justified for test laboratories carrying out beta emitter measurements in fulfilment of national authority requirements. For example, testing laboratories need to obtain a specific accreditation for radionuclide measurement for the monitoring of drinking water, food, the environment or the discharges, as well as for biological samples for medical purpose.

This document describes (after appropriate sampling, sample handling and test sample preparation) the generic requirements to quantify the activity concentration of beta emitters using liquid scintillation counting. In the absence of a specific pre-treatment of the test sample (such as distillation for  $^3\text{H}$  measurement, or after benzene synthesis for  $^{14}\text{C}$  measurement), this document is to be used as a screening method unless the interference of beta emitters, others than those to be quantified, is considered negligible in the test portion.

This document is one of a set of generic International Standards on measurement of radioactivity.

# Measurement of radioactivity — Determination of beta emitters activities — Test method using liquid scintillation counting

## 1 Scope

This document applies to the determination of beta emitters activity concentration using liquid scintillation counting. The method requires the preparation of a scintillation source, which is obtained by mixing the test sample and a scintillation cocktail. The test sample can be liquid (aqueous or organic), or solid (particles or filter or planchet).

NOTE Planchet are samples, described in [8.5](#), out of solid material e.g. small metal, plastic or glass pans or support material made of these materials

This document describes the conditions for measuring the activity concentration of beta emitter radionuclides by liquid scintillation counting<sup>[2]</sup>.

The choice of the test method using liquid scintillation counting involves the consideration of the potential presence of other beta-, alpha- and gamma emitter radionuclides in the test sample. In this case, a specific sample treatment by separation or extraction is implemented to isolate the radionuclide of interest in order to avoid any interference with other beta-, alpha- and gamma-emitting radionuclides during the counting phase.

This document is applicable to all types of liquid samples having an activity concentration ranging from about  $1 \text{ Bq}\cdot\text{l}^{-1}$  to  $10^6 \text{ Bq}\cdot\text{l}^{-1}$ . For a liquid test sample, it is possible to dilute liquid test samples in order to obtain a solution having an activity compatible with the measuring instrument. For solid samples, the activity of the prepared scintillation source shall be compatible with the measuring instrument.

The measurement range is related to the test method used: nature of test portion, preparation of the scintillator - test portion mixture, measuring assembly as well as to the presence of the co-existing activities due to interfering radionuclides.

Test portion preparations (such as distillation for  $^3\text{H}$  measurement, or benzene synthesis for  $^{14}\text{C}$  measurement, etc.) are outside the scope of this document and are described in specific test methods using liquid scintillation<sup>[3][4][5][6][7][8][9][10]</sup>.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes and sampling techniques*

ISO 5667-3, *Water quality — Sampling — Part 3: Preservation and handling of water samples*

ISO 80000-10, *Quantities and units — Part 10: Atomic and nuclear physics*

ISO 18589-2, *Measurement of radioactivity in the environment — Soil — Part 2: Guidance for the selection of the sampling strategy, sampling and pre-treatment of samples*

ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

**ISO 19361:2025(en)**

ISO/IEC Guide 99, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**