

STN	Stomatológia Fluoridové laky (ISO 17730: 2025)	STN EN ISO 17730 85 6375
------------	---------------------------------------------------------------	--------------------------------------------

Dentistry - Fluoride varnishes (ISO 17730:2025)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 12/25

Obsahuje: EN ISO 17730:2025, ISO 17730:2025

Oznámením tejto normy sa ruší
STN EN ISO 17730 (85 6375) z januára 2021

141408



EUROPEAN STANDARD

EN ISO 17730

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 2025

ICS 71.100.70

Supersedes EN ISO 17730:2020

English Version

Dentistry - Fluoride varnishes (ISO 17730:2025)Médecine bucco-dentaire - Vernis fluorés (ISO
17730:2025)

Zahnheilkunde - Fluoridlacke (ISO 17730:2025)

This European Standard was approved by CEN on 18 August 2025.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

EN ISO 17730:2025 (E)

Contents	Page
European foreword.....	3

European foreword

This document (EN ISO 17730:2025) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2026, and conflicting national standards shall be withdrawn at the latest by March 2026.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 17730:2020.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 17730:2025 has been approved by CEN as EN ISO 17730:2025 without any modification.



International Standard

ISO 17730

Dentistry — Fluoride varnishes

Médecine bucco-dentaire — Vernis fluorés

**Third edition
2025-09**

ISO 17730:2025(en)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

ISO 17730:2025(en)**Contents**

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	2
4.1 Total fluoride content.....	2
4.2 Fluoride release potential.....	2
5 Sampling	2
6 Measurement and test methods	2
6.1 Test conditions.....	2
7 Test report	5
8 Instructions for use	5
9 Packaging and marking/labelling	5
9.1 Packaging.....	5
9.2 Marking/labelling.....	6
Annex A (informative) Test for fluoride release potential	7
Bibliography	11

ISO 17730:2025(en)**Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 7, *Oral care products*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 55, *Dentistry*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 17730:2020), which has been technically revised.

The main changes are as follows:

- Abbreviated terms and jargon have been removed and edited for clarity.
- [A.2.3](#): The minimum size range of the plastic rod was changed from 30 mm² to 60 mm² to be more precise at 28,3 mm² to 60 mm².

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

ISO 17730:2025(en)**Introduction**

Fluoride varnishes are used in dentistry primarily for caries prevention and reduction of dentine hypersensitivity. Fluoride varnishes are applied in the oral cavity directly on the outer surfaces of teeth and fillings and are intended to remain for several hours.

Specific qualitative and quantitative test methods for demonstrating freedom from unacceptable biological hazards are not included in this document, but, for the assessment of possible biological or toxicological hazards, reference should be made to ISO 10993-1 and ISO 7405.

Dentistry — Fluoride varnishes

1 Scope

This document specifies requirements and test methods for total digestible fluoride content and a minimum soluble fluoride release potential in dental varnishes containing fluoride, intended for use in the oral cavity directly on the outer surfaces of teeth and fillings. This also specifies packaging and labelling requirements, including the instructions for use. This document covers fluoride varnishes to be applied by dental health care workers.

This document does not apply to fast acting topical fluoride products such as fluoride oral rinses.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 648:2008, *Laboratory glassware — Single-volume pipettes*

ISO 835:2007, *Laboratory glassware — Graduated pipettes*

ISO 1942, *Dentistry — Vocabulary*

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ISO 8601-1, *Date and time — Representations for information interchange — Part 1: Basic rules*

ISO 8601-2, *Date and time — Representations for information interchange — Part 2: Extensions*

ISO 19448, *Dentistry — Analysis of fluoride concentration in aqueous solutions by use of fluoride ion-selective electrode*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN