

STN	Chirurgické implantáty Kovové materiály Časť 2: Nelegovaný titán (ISO 5832-2: 2025)	STN EN ISO 5832-2 85 6364
------------	--	---

Implants for surgery - Metallic materials - Part 2: Unalloyed titanium (ISO 5832-2:2025)

Táto norma obsahuje anglickú verziu európskej normy.
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 12/25

Obsahuje: EN ISO 5832-2:2025, ISO 5832-2:2025

Oznámením tejto normy sa ruší
STN EN ISO 5832-2 (85 6364) z novembra 2018

141437

EUROPEAN STANDARD

EN ISO 5832-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 2025

ICS 11.040.40

Supersedes EN ISO 5832-2:2018

English Version

**Implants for surgery - Metallic materials - Part 2:
Unalloyed titanium (ISO 5832-2:2025)**Implants chirurgicaux - Matériaux métalliques - Partie
2: Titane non allié (ISO 5832-2:2025)Chirurgische Implantate - Metallische Werkstoffe - Teil
2: Unlegiertes Titan (ISO 5832-2:2025)

This European Standard was approved by CEN on 12 September 2025.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

EN ISO 5832-2:2025 (E)

Contents	Page
European foreword.....	3

European foreword

This document (EN ISO 5832-2:2025) has been prepared by Technical Committee ISO/TC 150 "Implants for surgery " in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2026, and conflicting national standards shall be withdrawn at the latest by March 2026.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 5832-2:2018.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 5832-2:2025 has been approved by CEN as EN ISO 5832-2:2025 without any modification.



International Standard

ISO 5832-2

Implants for surgery — Metallic materials —

Part 2: Unalloyed titanium

Implants chirurgicaux — Matériaux métalliques —

Partie 2: Titane non allié

**Fifth edition
2025-09**

ISO 5832-2:2025(en)**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

ISO 5832-2:2025(en)**Contents**

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Chemical composition	1
5 Microstructure	2
6 Mechanical properties	2
6.1 Tensile properties.....	2
6.2 Bending properties.....	2
7 Test methods	3
Bibliography	4

ISO 5832-2:2025(en)**Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, SC 1, *Materials* in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 55, *Dentistry*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This fifth edition cancels and replaces the fourth edition (ISO 5832-2:2018), which has been technically revised.

The main changes are as follows:

- the introduction has been updated;
- the requirement for cobalt in [Table 1](#) has been added;
- the wording for mechanical properties in [Table 2](#) has been updated;
- this document has been harmonized with the other parts of the ISO 5832 series.

A list of all parts in the ISO 5832 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

ISO 5832-2:2025(en)**Introduction**

While no known surgical implant material has ever been shown to cause absolutely no adverse reactions in the human body, long-term clinical experience with the material referred to in this document has shown that an acceptable level of biological response can be expected when the material is used in appropriate applications. However, this document covers the raw material and not finished medical devices, where the design and fabrication of the device can impact biological response.

Implants for surgery — Metallic materials —

Part 2: Unalloyed titanium

1 Scope

This document specifies the characteristics of, and corresponding test methods for, unalloyed titanium for use in the manufacture of surgical implants.

Six grades of titanium based on tensile strength are listed in [Table 2](#).

NOTE The mechanical properties of a sample obtained from a finished product made of this metal do not necessarily conform with those specified in this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 643, *Steels — Micrographic determination of the apparent grain size*

ISO 6892-1, *Metallic materials — Tensile testing — Part 1: Method of test at room temperature*

ISO 7438, *Metallic materials — Bend test*

ASTM E112, *Standard test methods for determining average grain size*

koniec náhľadu – text ďalej pokračuje v platenej verzii STN