

|            |  |   |
|------------|--|---|
| <b>STN</b> | <b>Zdravotnícke elektrické prístroje<br/>Časť 2-89: Osobitné požiadavky na základnú<br/>bezpečnosť a nevyhnutné prevádzkové<br/>vlastnosti zdravotníckych lôžok pre deti</b> | <b>STN<br/>EN IEC<br/>80601-2-89</b><br><br>36 4800 |
|------------|--|---|

Medical electrical equipment - Part 2-89: Particular requirements for the basic safety and essential performance of medical beds for children

Táto norma obsahuje anglickú verziu európskej normy.  
This standard includes the English version of the European Standard.

Táto norma bola oznámená vo Vestníku ÚNMS SR č. 04/26

Obsahuje: EN IEC 80601-2-89:2026, IEC 80601-2-89:2025

Oznámením tejto normy sa od 31.01.2029 ruší  
STN EN 50637 (36 4800) z mája 2018

**142396**



EUROPEAN STANDARD

**EN IEC 80601-2-89**

NORME EUROPÉENNE

EUROPÄISCHE NORM

January 2026

ICS 11.140

Supersedes EN 50637:2017

English Version

**Medical electrical equipment - Part 2-89: Particular requirements  
for the basic safety and essential performance of medical beds  
for children  
(IEC 80601-2-89:2025)**

Appareils électromédicaux - Partie 2-89: Exigences  
particulières pour la sécurité de base et les performances  
essentielles des lits médicaux pour enfants  
(IEC 80601-2-89:2025)

Medizinische elektrische Geräte - Teil 2-89: Besondere  
Festlegungen für die grundlegende Sicherheit und die  
wesentlichen Leistungsmerkmale von medizinischen Betten  
für Kinder  
(IEC 80601-2-89:2025)

This European Standard was approved by CENELEC on 2026-01-08. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.



European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

**EN IEC 80601-2-89:2026 (E)****European foreword**

The text of document 62D/2239/FDIS, future edition 1 of IEC 80601-2-89, prepared by SC 62D "Particular medical equipment, software, and systems" of IEC/TC 62 "Medical equipment, software, and systems" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 80601-2-89:2026.

The following dates are fixed:

- latest date by which the document has to be implemented at national (dop) 2027-01-31 level by publication of an identical national standard or by endorsement
- latest date by which the national standards conflicting with the (dow) 2029-01-31 document have to be withdrawn

This document supersedes EN 50637:2017 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

This document is read in conjunction with EN 60601-1:2006 and all of its amendments and corrigenda.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

**Endorsement notice**

The text of the International Standard IEC 80601-2-89:2025 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standard indicated:

|                       |      |   |
|-----------------------|------|---|
| ISO 20342 (series)    | NOTE | Approved as EN ISO 20342 (series)               |
| IEC 61032:1997        | NOTE | Approved as EN 61032:1998 (not modified)        |
| IEC 60601-2-19        | NOTE | Approved as EN IEC 60601-2-19                   |
| IEC 60601-1 (series)  | NOTE | Approved as EN 60601-1-9:2008/A2 (series)       |
| ISO 9999:2022         | NOTE | Approved as EN ISO 9999:2022 (not modified)     |
| ISO 9614-1            | NOTE | Approved as EN ISO 9614-1                       |
| ISO 13857:2019        | NOTE | Approved as EN ISO 13857:2019 (not modified)    |
| IEC 60601-1-12        | NOTE | Approved as EN 60601-1-12                       |
| ISO 7010:2019         | NOTE | Approved as EN ISO 7010:2020 (not modified)     |
| IEC 60601-1-10        | NOTE | Approved as EN 60601-1-10                       |
| ISO/IEC Guide 71:2014 | NOTE | Approved as CEN/CLC Guide 6:2014 (not modified) |
| IEC 60068-2-31        | NOTE | Approved as EN 60068-2-31                       |

## Annex ZA (normative)

### Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: [www.cencenelec.eu](http://www.cencenelec.eu).

Annex ZA of EN 60601-1:2006 and all of its amendments and corrigenda applies, except as follows:

*Addition:*

| <u>Publication</u> | <u>Year</u> | <u>Title</u>  | <u>EN/HD</u>       | <u>Year</u> |
|--------------------|-------------|---|--------------------|-------------|
| IEC 60529          | -           | Degrees of protection provided by enclosures (IP Code)  | -                  | -           |
| IEC 60601-1        | 2005        | Medical electrical equipment - Part 1: General requirements for basic safety and essential performance  | EN 60601-1         | 2006        |
| -                  | -           |   | + corrigendum Mar. | 2010        |
| + A1               | 2012        |   | + A1               | 2013        |
| -                  | -           |   | + A12              | 2014        |
| + A2               | 2020        |   | + A2               | 2021        |
| -                  | -           |   | + A13              | 2024        |
| ISO 48-5           | 2018        | Rubber, vulcanized or thermoplastic - Determination of hardness – Part 5: Indentation hardness by IRHD pocket meter method  | -                  | -           |
| ISO 3746           | -           | Acoustics - Determination of sound power levels and sound energy levels of noise sources using sound pressure - Survey method using an enveloping measurement surface over a reflecting plane | EN ISO 3746        | -           |
|                    |             | Safety of toys - Part 3: Migration of certain elements  | EN 71-3            | -           |
|                    |             | Furniture - Assessment of the ignitability of mattresses and upholstered bed bases - Part 1: Ignition source: Smouldering cigarette   | EN 597-1           | -           |
|                    |             | Furniture - Assessment of the ignitability of mattresses and upholstered bed bases -Part 2: Ignition source: Match flame equivalent   | EN 597-2           | -           |
|                    |             | Furniture - Children's cots and folding cots for domestic use - Part 2: Test methods  | EN 716-2           | -           |



IEC 80601-2-89

Edition 1.0 2025-12

# INTERNATIONAL STANDARD

---

**Medical electrical equipment -  
Part 2-89: Particular requirements for the basic safety and essential performance  
of medical beds for children**



**THIS PUBLICATION IS COPYRIGHT PROTECTED**  
**Copyright © 2025 IEC, Geneva, Switzerland**

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Secretariat  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

**About the IEC**

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

**About IEC publications**

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

**IEC publications search -**  
[webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

**IEC Just Published -** [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

**IEC Customer Service Centre -** [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

**IEC Products & Services Portal -** [products.iec.ch](http://products.iec.ch)

Discover our powerful search engine and read freely all the publications previews, graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

**Electropedia -** [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

**Warning! Make sure that you obtained this publication from an authorized distributor.**

## IEC 80601-2-89:2025 © IEC 2025

## CONTENTS

|  |     |
|--|-----|
| FOREWORD .....   | 4   |
| INTRODUCTION .....   | 7   |
| 201.1 Scope, object and related standards .....  | 8   |
| 201.2 Normative references .....   | 10  |
| 201.3 Terms and definitions .....  | 11  |
| 201.4 General requirements .....   | 17  |
| 201.5 *General requirements for testing of ME EQUIPMENT .....  | 18  |
| 201.6 Classification of ME EQUIPMENT and ME SYSTEMS .....  | 23  |
| 201.7 ME EQUIPMENT identification, marking and documents .....   | 23  |
| 201.8 Protection against electrical HAZARDS from ME EQUIPMENT .....                                    | 31  |
| 201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS .....                       | 32  |
| 201.10 Protection against unwanted and excessive radiation HAZARDS .....                               | 64  |
| 201.11 Protection against excessive temperatures and other HAZARDS .....                               | 64  |
| 201.12 Accuracy of controls and instruments and protection against hazardous outputs .....             | 68  |
| 201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT .....                                | 68  |
| 201.14 Programmable electrical medical systems (PEMS) .....  | 69  |
| 201.15 Construction of ME EQUIPMENT .....  | 69  |
| 201.16 ME SYSTEMS .....  | 73  |
| 201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS .....                              | 73  |
| Annexes .....  | 74  |
| Annex AA (informative) Particular guidance and rationale .....   | 75  |
| Annex BB (normative) Additional design requirements and recommendations for MEDICAL BEDS .....         | 90  |
| Annex CC (informative) Particular guidance for assessing RISK of entrapment in V-shaped openings ..... | 94  |
| Annex DD (informative) Guidance and recommendations for periodic inspection .....                      | 100 |
| Bibliography .....   | 102 |
| Index of defined terms .....   | 104 |
| Figure 201.101 – COT, general arrangement (example, schematic presentation only) .....                 | 13  |
| Figure 201.102 – CRIB, general arrangement (example, schematic presentation only) .....                | 13  |
| Figure 201.103 – MEDICAL BED, general arrangement (example, schematic presentation only) .....         | 15  |
| Figure 201.104 – HEAD DOWN TILT, Example .....   | 16  |
| Figure 201.105 – FOOT DOWN TILT, Example .....   | 17  |
| Figure 201.106 – Small finger probe $\varnothing$ 5,6 .....  | 18  |
| Figure 201.107 – Entrapment test TOOLS .....   | 20  |
| Figure 201.108 – Loading pad .....   | 21  |
| Figure 201.109 – Impactor .....  | 22  |
| Figure 201.110 – Side impactor TOOL .....  | 23  |
| Figure 201.111 – Graphic symbol for maximum PATIENT weight and SAFE WORKING LOAD .....                 | 24  |
| Figure 201.112 – Graphic symbol for mass; weight .....   | 24  |

## IEC 80601-2-89:2025 © IEC 2025

|  |    |
|--|----|
| Figure 201.113 – Graphic symbol for machine washable MEDICAL BED .....   | 24 |
| Figure 201.114 – Graphic Symbol for jet stream washable MEDICAL BEDS.....  | 25 |
| Figure 201.115 – Graphic symbol for manual cleaning only.....  | 25 |
| Figure 201.116 – Description of allowed length of PATIENT .....  | 26 |
| Figure 201.117 – Example of marking for positioning PATIENT in MEDICAL BED.....  | 27 |
| Figure 201.118 – MEDICAL BED function controls and actuators: guidelines for creating graphic symbols .....  | 28 |
| Figure 201.119 – Examples of marking on the MEDICAL BED of storage location for wired and wireless PENDANT CONTROLS .....  | 29 |
| Figure 201.120 – Example of MEDICAL BED with segmented or split SIDE RAIL .....  | 33 |
| Figure 201.121 – Example of MEDICAL BED with single piece SIDE RAIL and PROTECTION PANEL .....   | 34 |
| Figure 201.122 – Dimension of handle for LIFTING POLE .....  | 39 |
| Figure 201.123 – Allowable spacing for fingers in areas of normal reach around the perimeter of the MATTRESS SUPPORT PLATFORM.....                                 | 41 |
| Figure 201.124 – Example using barriers for clearance measurement around the perimeter of the MATTRESS SUPPORT PLATFORM to mitigate PATIENT-finger entrapment..... | 42 |
| Figure 201.125 – Clearance areas .....   | 43 |
| Figure 201.126 – Required minimum radii of edges and corners.....  | 44 |
| Figure 201.127 – Retention of loop and mass.....   | 45 |
| Figure 201.128 – Lateral stability test along the side of the MEDICAL BED .....  | 48 |
| Figure 201.129 – Longitudinal stability test with removable FOOT BOARD .....   | 48 |
| Figure 201.130 – Longitudinal stability test with fixed HEAD BOARDS and FOOT BOARDS.....   | 49 |
| Figure 201.131 – Distribution of SAFE WORKING LOAD per maximum PATIENT weight for tests.....   | 53 |
| Figure 201.132 – Position of loading pad and impactor (see Figure 201.108 for loading pad and Figure 201.109 for impactor).....                                    | 56 |
| Figure 201.133 – Impact to slats and solid elements of MEDICAL BEDS .....  | 58 |
| Figure 201.134 – Application of forces for test of SIDE RAIL.....  | 60 |
| Figure 201.135 – Height of PROTECTION PANELS, SIDE RAIL and PROTECTION PERIMETER.....  | 63 |
| Figure 201.136 – Direction of movement for rough handling test.....  | 70 |
| Figure 201.137 – Configurations of the MATTRESS SUPPORT PLATFORM .....   | 71 |
| Figure 201.138 – Small parts cylinder.....   | 73 |
| Figure AA.1 – Example of marking for compatible mattresses specified by the MANUFACTURER.....  | 77 |
| Figure AA.2 – Example of marking for detachable SIDE RAILS specified by the MANUFACTURER.....  | 77 |
| Figure AA.3 – Resultant forces without mattress .....  | 80 |
| Figure AA.4 – Resultant forces with mattress .....   | 80 |
| Figure AA.5 – 40 mm gap measurement of B.....  | 80 |
| Figure AA.6 – Angle measurement example of B .....   | 80 |
| Figure AA.7 – Placement of a measurement TOOL for the measurement of $D_X$ .....   | 81 |
| Figure AA.8 – Example of area $D_X$ measurement that passes .....  | 82 |
| Figure AA.9 – Example of area $D_X$ measurement that fails (on limit).....   | 82 |
| Figure AA.10 – Example of area $D_X$ measurement that fails.....   | 82 |

## IEC 80601-2-89:2025 © IEC 2025

|   |    |
|---|----|
| Figure AA.11 – Example of potential PATIENT entrapment in area A within the SIDE RAIL.....                  | 83 |
| Figure AA.12 – Example of potential PATIENT entrapment in area A below the SIDE RAIL .....                  | 83 |
| Figure AA.13 – Example of potential PATIENT entrapment in area B.....                                       | 83 |
| Figure AA.14 – Example of potential PATIENT entrapment in area C between segmented or split SIDE RAIL ..... | 83 |
| Figure AA.15 – Example of potential PATIENT entrapment in area C between SIDE RAIL and HEAD BOARD .....     | 83 |
| Figure AA.16 – Example of potential PATIENT entrapment in area D.....                                       | 83 |
| Figure AA.17 – Example of potential PATIENT entrapment in area A below a single piece SIDE RAIL .....       | 84 |
| Figure BB.1 – Schematic presentation of under MEDICAL BED clearance .....                                   | 92 |
| Figure BB.2 – Angles for different sections of the MATTRESS SUPPORT PLATFORM.....                           | 93 |
| Figure CC.1 – Wedge TOOL.....   | 95 |
| Figure CC.2 – V-shaped opening in relation to B.....  | 96 |
| Figure CC.3 – Acceptance criteria in relation to area B.....  | 96 |
| Figure CC.4 – Positioning of wedge TOOL.....  | 97 |
| Figure CC.5 – Acceptance criteria in relation to area C between HEAD BOARD and FOOT BOARD .....             | 98 |
| Figure CC.6 – Acceptance criteria in relation to area C between segmented or split SIDE RAILS .....         | 99 |
| Table 201.101 – Protection against PATIENT entrapment .....   | 35 |
| Table 201.102 – Minimum SAFE WORKING LOADS .....  | 52 |
| Table 201.103 – Protection against inadvertent PATIENT falls and climbing out of MEDICAL BED .....          | 63 |
| Table 201.104 – Allowable maximum temperatures for skin contact with MEDICAL BED APPLIED PARTS .....        | 64 |
| Table 201.105 – Machine washable compliance PROCEDURE .....   | 66 |
| Table AA.1 – Protection against PATIENT entrapment in non-moving parts.....                                 | 79 |
| Table AA.2 – Height of protective barriers according to age .....   | 88 |

IEC 80601-2-89:2025 © IEC 2025

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

---

**Medical electrical equipment -  
Part 2-89: Particular requirements for the basic safety  
and essential performance of medical beds for children**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 80601-2-89 has been prepared by a Joint Working Group of IEC subcommittee 62D: Particular medical equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems, and ISO technical committee 173: Assistive products. It is an International Standard.

This publication is published as a double logo standard.

## IEC 80601-2-89:2025 © IEC 2025

The text of this International Standard is based on the following documents of IEC:

|               |                  |
|---------------|------------------|
| Draft         | Report on voting |
| 62D/2239/FDIS | 62D/2272/RVD     |

Full information on the voting for its approval can be found in the report on voting indicated in the above table. In ISO, the document was approved by XXX P members out of YYY having cast a vote.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at [www.iec.ch/members\\_experts/refdocs](http://www.iec.ch/members_experts/refdocs). The main document types developed by IEC are described in greater detail at [www.iec.ch/publications](http://www.iec.ch/publications).

In this standard, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- terms defined in clause 3 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 AND IEC 60601-1:2005/AMD2:2020, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 80601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

## IEC 80601-2-89:2025 © IEC 2025

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under [webstore.iec.ch](http://webstore.iec.ch) in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

## IEC 80601-2-89:2025 © IEC 2025

## INTRODUCTION

IEC 80601-2-52[1]<sup>1</sup> applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL BEDS for ADULTS, hence not covering requirement for beds for CHILDREN and ADULTS with atypical anatomy. This particular standard is based on EN 50637[2], which was created pursuant to Mandate M/467 Medical beds issued by the European Commission with the following background information:

It appears, from a first analysis undertaken by EU Competent Authorities, that the current set of standards is not adapted to the needs of CHILDREN or ADULTS with an atypical anatomy. IEC 80601-2-52 does not foresee a maximum distance for the bars that is small enough to prevent accidents.

According to the EU Competent Authorities' representatives, a part of the safety problem is due to the fact that MEDICAL BEDS for ADULTS are not appropriately labelled as being designed only for ADULTS with a normal anatomy. Users are therefore not always aware of the risk of MEDICAL BEDS for young PATIENTS or for ADULTS with an atypical anatomy. Hospital administrations do not always see a need to buy MEDICAL BEDS which are appropriate for CHILDREN or for ADULTS with an atypical anatomy. Therefore, clear labelling of the targeted PATIENT groups for MEDICAL BEDS complying with IEC 80601-2-52 could reduce the risk of inappropriate use of this kind of MEDICAL BEDS for CHILDREN or for ADULTS with an atypical anatomy.

EU Competent Authorities' representatives also stated that there is a need for the development of requirements for MEDICAL BEDS and COTS for CHILDREN and ADULTS with an atypical anatomy.

In order to prevent IEC 80601-2-52 from being extraordinarily complex to use, TC 62 decided to develop this particular standard rather than further amending IEC 80601-2-52 in relation to use for CHILDREN and ADULTS with an atypical anatomy.

This standard is based on EN 50637 and IEC 80601-2-52 with input from the following standards and reports:

- EN 716-1, *Furniture – V Children's cots and folding cots for domestic use – Part 1: Safety requirements*
- EN 716-2, *Furniture – Children's cots and folding cots for domestic use – Part 2: Test methods*
- EN 1130, *Furniture – Children's furniture – Cribs – Safety requirements and test methods*
- EN 747-1, *Furniture – Bunk beds and high beds – Part 1: Safety, strength and durability requirements*
- EN 747-2, *Furniture – Bunk beds and high beds – Part 2: Test methods*
- CEN/TR 13387 (all parts), *Child use and care articles – General safety guidelines*
- DIN 32623, *Hospital children's cots made from metal and plastic – Safety requirements and testing*
- *Nordic Requirements specification for Adjustable beds for disabled children*

---

<sup>1</sup> Numbers in square brackets refer to the Bibliography.

## IEC 80601-2-89:2025 © IEC 2025

**201.1 Scope, object and related standards**

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, Clause 1, applies, except as follows:

**201.1.1 \* Scope***Replacement:*

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL BEDS, hereafter referred to as MEDICAL BEDS as defined in 201.3.219, intended for CHILDREN as defined in 201.3.207, and ADULTS with atypical anatomy (ADULTS ranging outside the definition for ADULTS in 201.3.201).

This document applies to both electrical and non-electrical (manual) MEDICAL BEDS with or without adjustable functions. This document applies to MEDICAL BEDS with an INTERNAL LENGTH of up to 180 cm suitable to a body length of 155 cm.

NOTE 1 The limitation of 180 cm is in order to minimize the foreseeable misuse, of a parent sharing the bed with the CHILD or that the bed will be used by an ADULT.

If a MANUFACTURER wishes to make a MEDICAL BED that can be used by both a CHILD and an ADULT, e.g. INTERNAL LENGTH of 180 cm or more, then IEC 80601-2-52 and this document apply.

This document does not apply to:

- ADULT *only* MEDICAL BEDS covered by IEC 80601-2-52;
- SPECIALITY MATTRESS covered by the ISO 20342 series[5];
- incubators covered by IEC 60601-2-19;
- devices for which the INTENDED USE is mainly for examination or transportation under medical supervision (e.g. stretcher, examination table).

If a clause or subclause is specifically intended to be applicable to a MEDICAL BED only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to MEDICAL BEDS and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of MEDICAL BEDS or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, 7.2.13 and 8.4.1.

NOTE 2 See also IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, 4.2.

NOTE 3 Whenever the term MEDICAL ELECTRICAL EQUIPMENT (MEE, ME Equipment) is used within the series of IEC 60601 standards, it refers to MEDICAL BEDS, both electrical and non-electrical.

**201.1.2 Object***Replacement:*

The object of this document is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements and test methods for MEDICAL BEDS as defined in 201.3.219 intended for CHILDREN as defined in 201.3.207 and ADULTS with atypical anatomy, i.e. ADULTS ranging outside the definition for ADULTS in 202.3.201.

## IEC 80601-2-89:2025 © IEC 2025

**201.1.3 Collateral standards***Addition:*

This document refers to those applicable collateral standards that are listed in Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and Clause 201.2 of this document.

IEC 60601-1-3 and IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

NOTE Some IEC 60601-1-8 requirements can be excluded if they do not affect PATIENT safety, could lead to user confusion, or are inappropriate to MEDICAL BED usage.

**201.1.4 Particular standards***Replacement:*

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, including the collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

Requirements of this document takes priority over IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020.

The numbering of clauses and subclauses of this particular standard corresponds to that of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this document addresses the content of Clause 4 of the IEC 60601-1-2:2015 collateral standard, 203.4 in this document addresses the content of Clause 4 of the IEC 60601-1-3:2008 collateral standard, etc.). The changes to the text of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are specified by the use of the following words:

*"Replacement"* means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard is replaced completely by the text of this document.

*"Addition"* means that the text of this document is additional to the requirements of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard.

*"Amendment"* means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard is amended as indicated by the text of this document.

Subclauses, figures or tables which are additional to those of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered starting from 201.101. However, due to the fact that definitions in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered 3.1 through 3.154, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

## IEC 80601-2-89:2025 © IEC 2025

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

## 201.2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE Informative references are listed in the Bibliography.

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, Clause 2, applies except as follows:

*Addition:*

IEC 60529, *Degrees of protection provided by enclosures (IP Code)*

IEC 60601-1:2005, *Medical electrical equipment - Part 1: General requirements for basic safety and essential performance*  
IEC 60601-1:2005/AMD1:2012  
IEC 60601-1:2005/AMD2:2020

ISO 48-5:2018, *Rubber, vulcanized or thermoplastic - Determination of hardness - Part 5: Indentation hardness by IRHD pocket meter method*

ISO 3746, *Acoustics - Determination of sound power levels of noise sources using sound pressure - Survey method using an enveloping measurement surface over a reflecting plane*

EN 71-3, *Safety of toys - Part 3: Migration of certain elements*

EN 597-1, *Furniture - Assessment of the ignitability of mattresses and upholstered bed bases - Part 1: Ignition source : Smouldering cigarette*

EN 597-2, *Furniture - Assessment of the ignitability of mattresses and upholstered bed bases - Part 2: Ignition source: Match flame equivalent*

EN 716-2, *Furniture - Children's cots and folding cots for domestic use - Part 2: Test methods*

**koniec náhľadu – text ďalej pokračuje v platenej verzii STN**